

Merge DICOM Toolkit™

5.16.0 C/C++ REFERENCE MANUAL

© Copyright Merge Healthcare Solutions Inc. 2023.

Licensed materials – Property of Merge Healthcare Solutions Inc.

The content of this document is confidential information of Merge Healthcare Solutions Inc. and its use and disclosure is subject to the terms of the agreement pursuant to which you obtained the software that accompanies the documentation.

Merge Healthcare and the Merge Healthcare logo are trademarks of Merge Healthcare Inc.

Microsoft, Windows, and the Windows logo are trademarks of Microsoft Corporation in the United States, other countries, or both. All other names are trademarks or registered trademarks of their respective companies.

U.S. GOVERNMENT RESTRICTED RIGHTS:

This product is a "Commercial Item" offered with "Restricted Rights." The Government's rights to use, modify, reproduce, release, perform, display or disclose this documentation are subject to the restrictions set forth in Federal Acquisition Regulation ("FAR") 12.211 and 12.212 for civilian agencies and in DFARS 227.7202-3 for military agencies. Contractor is Merge Healthcare Solutions Inc.



Merge Healthcare Incorporated 900 Walnut Ridge Drive Hartland, WI 53029

Symbols Glossary:

Symbol	Title
REF	Part or Catalog number
	Manufacturer

The full symbols glossary can be viewed at https://www.merative.com/content/dam/merative/documents/brief/ Merge_Healthcare_Symbols_Glossary.pdf.

INSTRUCTIONS FOR USE:

Instructions for use are available electronically in PDF format. Paper format can be requested for supported languages by contacting Customer Support and will be provided free of charge within seven calendar days of receiving the request. The paper format can also be provided at the time of delivery if so requested at the time of order.

For application support or to report issues with user documentation, contact Customer Support:

Substantiation States Stat

+31.20.514.5073 (Europe, the Middle East and Africa) 1800 316 746 (Australia)

MC3Support@merative.com

Part	Date	Revision	Description
COM-5278	January 2023	1.0	Updated bi-annually

The latest version of this document can be found at https://mergecustomer.force.com/mergeusercommunity/login.

Contents

OVERVIEW	13
Structure of Documentation	13
Conventions	14
CONFIGURATION	15
Initialization File Integration	16
APPLICATION PROGRAMMING INTERFACE	18
Overview of API	18
API Function Reference	
MC_Abort_Association	29
MC_Accept_Association MC_Accept_Association_With_Identity	
MC_Add_Nonstandard_Attribute	
MC_Add_Private_Attribute	
MC_Add_Private_Block	
MC_Add_Standard_Attribute	
MC_Byte_To_Unicode	
MC_Byte_Swap_OBOW	
MC_Cleanup_Memory	
MC_Clear_Negotiation_Info	40
MC_Close_Association	41
MC_Close_Encapsulated_Value	42
MC_Close_Listen_Port MC_Close_Listen_Port_On_Address	43
MC_Continue_Read_Message	44
MC_Continue_Read_Message_To_Stream	45
MC_Continue_Read_Message_To_Tag	47
MC_Create_Empty_File	49
MC_Create_File	50
MC_Delete_Attribute	51
MC_Delete_Current_Value	52
MC_Delete_Private_Attribute	54
MC_Delete_Private_Block	55
MC_Delete_Range	56
MC_Dir_Add_Entity	57
MC_Dir_Add_Record	58
MC_Dir_Delete_Record	59

MC_Dir_Entity_Count. 61 MC_Dir_First_Record. 62 MC_Dir_Item_Count. 64 MC_Dir_Next_Entity. 65 MC_Dir_Open_MRDR. 66 MC_Dir_Reference_MRDR. 67 MC_Dir_Root_Count. 70 MC_Dir_Root_Count. 70 MC_Dir_Root_Entity. 71 MC_Dir_Sort. 72 MC_Dir_Stream_Directory_Record 75 MC_Duplicate_Message 76 MC_Empty_File 80 MC_Empty_Item 81 MC_Empty_Message 81
MC_Dir_Item_Count. 64 MC_Dir_Next_Entity. 65 MC_Dir_Next_Record 66 MC_Dir_Open_MRDR. 67 MC_Dir_Reference_MRDR 68 MC_Dir_Remove_Ref_MRDR 69 MC_Dir_Root_Count. 70 MC_Dir_Root_Entity. 71 MC_Dir_Sort. 72 MC_Dir_Stream_Directory_Record 75 MC_Duplicate_Message 76 MC_Empty_File 80 MC_Empty_Item. 81 MC_Empty_Message 81
MC_Dir_Next_Entity 65 MC_Dir_Next_Record 66 MC_Dir_Open_MRDR 67 MC_Dir_Reference_MRDR 68 MC_Dir_Remove_Ref_MRDR 69 MC_Dir_Root_Count 70 MC_Dir_Root_Entity 71 MC_Dir_Sort 72 MC_Dir_Stream_Directory_Record 75 MC_Empty_File 80 MC_Empty_Item 81
MC_Dir_Next_Record 66 MC_Dir_Open_MRDR. 67 MC_Dir_Reference_MRDR. 68 MC_Dir_Remove_Ref_MRDR 69 MC_Dir_Root_Count. 70 MC_Dir_Root_Entity. 71 MC_Dir_Sort. 72 MC_Dir_Stream_Directory_Record 75 MC_Empty_File. 80 MC_Empty_Item. 81 MC_Empty_Message 81
MC_Dir_Open_MRDR67MC_Dir_Reference_MRDR.68MC_Dir_Remove_Ref_MRDR.69MC_Dir_Root_Count.70MC_Dir_Root_Entity71MC_Dir_Sort72MC_Dir_Stream_Directory_Record.75MC_Duplicate_Message.76MC_Empty_File.80MC_Empty_Item.81MC_Empty_Message.81
MC_Dir_Reference_MRDR68MC_Dir_Remove_Ref_MRDR69MC_Dir_Root_Count70MC_Dir_Root_Entity71MC_Dir_Sort72MC_Dir_Stream_Directory_Record75MC_Louplicate_Message76MC_Empty_File80MC_Empty_Item81MC_Empty_Message81
MC_Dir_Remove_Ref_MRDR 69 MC_Dir_Root_Count. 70 MC_Dir_Root_Entity. 71 MC_Dir_Sort. 72 MC_Dir_Stream_Directory_Record 75 MC_Duplicate_Message 76 MC_Empty_File 80 MC_Empty_Item. 81 MC_Empty_Message 81
MC_Dir_Root_Count .70 MC_Dir_Root_Entity. .71 MC_Dir_Sort. .72 MC_Dir_Stream_Directory_Record .75 MC_Duplicate_Message .76 MC_Empty_File .80 MC_Empty_Item .81 MC_Empty_Message .81
MC_Dir_Root_Entity. 71 MC_Dir_Sort. 72 MC_Dir_Stream_Directory_Record 75 MC_Duplicate_Message 76 MC_Empty_File 80 MC_Empty_Item. 81 MC_Empty_Message 81
MC_Dir_Sort. 72 MC_Dir_Stream_Directory_Record 75 MC_Duplicate_Message 76 MC_Empty_File 80 MC_Empty_Item 81 MC_Empty_Message 81
MC_Dir_Stream_Directory_Record
MC_Duplicate_Message
MC_Empty_File
MC_Empty_Item
MC_Empty_Message81
No Frankla Unicada Conversion
MC_Enable_Unicode_Conversion82
MC_Error_Message
MC_File_To_Message
MC_Free_Item
MC_Free_File
MC_Free_Message
MC_FreeService
MC_FreeServiceList
MC_FreeSyntaxList
MC_Get_Association_Info
MC_Get_Attribute_Info91
MC_Get_Bool_Config_Value
MC_Get_Encapsulated_Value_To_Function94
MC_Get_Enum_From_Transfer_Syntax97
MC_Get_File_Length
MC_Get_File_Preamble
MC_Get_Filename
MC_Get_First_Acceptable_Service
MC_Get_First_Attribute
MC_Get_Frame_To_Function104
MC_Get_Int_Config_Value

MC_Get_Listen_Socket	108
MC_Get_Listen_Socket_For_Port	109
MC_Get_Log_Destination	110
MC_Get_Long_Config_Value	111
MC_Get_MergeCOM_Service	112
MC_Get_MergeCOM_Service_From_UID	113
MC_Get_Message_Service	113
MC_Get_Message_Transfer_Syntax	115
MC_Get_Meta_ServiceName	116
MC_Get_Negotiation_Info	117
MC_Get_Next_Acceptable_Service	119
MC_Get_Next_Attribute	120
MC_Get_Next_Encapsulated_Value_To_Function	121
MC_Get_Next_pValue Functions	124
MC_Get_Next_Validate_Error	128
MC_Get_Next_Value Functions	130
MC_Get_Offset_Table_To_Function	134
MC_Get_pAttribute_Info	137
MC_Get_pTag_Info	138
MC_Get_pTags_Dict_Info	139
MC_Get_pValue Functions	140
MC_Get_pValue_Count	145
MC_Get_pValue_Length	147
MC_Get_pValue_To_Function	148
MC_Get_Stream_Length	151
MC_Get_String_Config_Value	153
MC_Get_Tag_Info	156
MC_Get_Tags_Dict_Info	157
MC_Get_Tag_Keyword	158
MC_Get_Transfer_Syntax_From_Enum	158
MC_Get_UID_From_MergeCOM_Service	160
MC_Get_User_Identity_Info	161
MC_Get_User_Identity_Length	163
MC_Get_Value Functions	164
MC_Get_Value_Count	168
MC_Get_Value_Length	169
MC_Get_Value_To_Function	171
MC_Get_Version_String	173

MC_Json_To_Message	173
MC_bibrary Initialization	
MC_Library_Release	
MC_Library_Reset	
MC_List_File (All toolkits except Windows versions)	
MC List File (Windows toolkit versions)	
MC_List_Item (All toolkits except Windows versions)	
MC_List_Item (Windows toolkit versions)	
MC_List_Message (All toolkits except Windows versions)	
MC_List_Message (Windows toolkit versions)	
MC_MemoryLog_To_Function	
MC_Message_To_File	
MC_Message_To_Ison	
MC_Message_To_SR	
MC_Message_To_Stream	
MC_Message_To_XML	
MC_Message_To_XML_Native	
MC_NewProposedServiceList MC_NewProposedServiceListAsync	
MC_NewSyntaxList	
MC NewService Functions	
MC_Open_Association	
MC_Open_Association_With_Connect_Timeout	
MC_Open_Association_With_Identity	
MC_Open_Association_With_Identity_With_Reject_Info	
MC Open Association With Identity With Reject Info With Connect Timeout	
MC_Open_Association_With_Reject_Info	
MC_Open_Association_With_Callback	
MC Open Association With All Optional Parameters	
MC_Open_Empty_Item	
MC_Open_Empty_Message	
MC_Open_File MC_Open_File_Bypass_OBOW MC_Open_File_Upto_Tag MC_Open_File_Upto_Tag_Bypass_Value	
MC_Open_Item	
MC_Open_Message	
MC_Open_Secure_Association MC_Open_Secure_Association_With_Reject_Info	
MC_Process_Association_Request MC_Process_Secure_Association_Request	
MC_Read_Message	
MC_Read_Message_To_Tag	
·······	

MC_Read_To_Stream	231
MC_Register_Application	234
MC_Register_Callback_Function MC_Register_pCallback_Function	235
MC_Register_Compression_Callbacks	243
MC_Register_Enhanced_MemoryLog_Function	246
MC_Register_MemoryLog_Function	248
MC_Register_Network_Capture_Callbacks (deprecated)	250
MC_Reject_Association MC_Reject_Association_With_Reason_Codes	253
MC_Release_Application	255
MC_Release_Callback_Function MC_Release_pCallback_Function	255
MC_Release_Library_Exception_Handler	256
MC_Release_Parent_Association (UNIX Only)	257
MC_Release_Parent_Connection (UNIX Only)	258
MC_Report_Memory	258
MC_Reset_Filename	259
MC_Reset_Message_Transfer_Syntax	260
MC_Send_Request_Message	261
MC_Send_Request_Message_For_Service	263
MC_Send_Request	265
MC_Send_Request_For_Service	267
MC_Send_Response_Message	269
MC_Send_Response	271
MC_Set_Bool_Config_Value	273
MC_Set_Encapsulated_Value_From_Function	275
MC_Set_File_Preamble	278
MC_Set_Int_Config_Value	278
MC_Set_Library_Exception_Handler	280
MC_Set_Log_Destination	281
MC_Set_Log_Prefix	282
MC_Set_Long_Config_Value	282
MC_Set_MergeINI MC_Set_MergeINI_Unicode	283
MC_Set_Message_Callbacks	284
MC_Set_Message_Transfer_Syntax	285
MC_Set_Negotiation_Info	286
MC_Set_Negotiation_Info_For_Association	287
MC_Set_Next_Encapsulated_Value_From_Function	289
MC_Set_Next_pValue Functions	291
MC_Set_Next_pValue_To_NULL	296

MC_Set_Next_Value Functions	
MC_Set_Next_Value_To_NULL	
MC_Set_pValue Functions	
MC_Set_pValue_From_Function	
MC_Set_pValue_Representation	311
MC_Set_pValue_To_Empty	312
MC_Set_pValue_To_NULL	314
MC_Set_Service_Command	315
MC_Set_String_Config_Value	317
MC_Set_Value Functions	
MC_Set_Value_From_Function	325
MC_Set_Value_Representation	327
MC_Set_Value_To_Empty	
MC_Set_Value_To_NULL	
MC_SR_Add_Child	
MC_SR_Add_Root	331
MC_SR_Delete_Child	
MC_SR_Get_Child_Count	
MC_SR_Get_First_Child	
MC_SR_Get_Location	
MC_SR_Get_Next_Child	
MC_SR_Get_Root	
MC_SR_To_Message	
MC_Stream_To_Message MC_Stream_To_Message_With_Offset	
MC_Thread_Release	343
MC_Unicode_Get_Substitution_Characters	
MC_Unicode_To_Byte	345
MC_Validate_Attribute	
MC_Validate_File	
MC_Validate_Message	352
MC_Wait_For_Association MC_Wait_For_Association_On_Port MC_Wait_For_Association	_On_Address 355
MC_Wait_For_Connection MC_Wait_For_Connection_On_Port MC_Wait_For_Connection	n_On_Address 358
MC_Wait_For_Secure_Association MC_Wait_For_Secure_Association_On_Port MC_Wait_For_Secure_Association_On_Address	
MC Write File MC Write File By Callback	
MC_XML_To_Message	
MC_XML_Native_To_Message	
High Level API Reference	

MC_DDH_Create	371
MC_DDH_Open	373
MC_DDH_Update	374
MC_DDH_Traverse_Records	376
MC_DDH_Get_Record_Type	377
MC_DDH_Get_Parent_Record	378
MC_DDH_Get_Next_Record	379
MC_DDH_Get_First_Lower_Record	381
MC_DDH_Copy_Values	
MC_DDH_Add_Record	
MC_DDH_Delete_Record	
MC_DDH_Release_Record	
MC_SRH_Create_SR	
MC_SRH_Free_SR	
MC_SRH_Add_Child	
MC_SRH_Add_TEXT_Child	
MC_SRH_Create_TEXT_Node	
MC_SRH_Set_Concept_Name	391
MC_SRH_Add_CODE_Child	
MC_SRH_Create_CODE_Node	
MC_SRH_Add_NUM_Child	
MC_SRH_Create_NUM_Node	395
MC_SRH_Set_NUM_Qualifier	
MC_SRH_Set_NUM_Next_Data	
MC_SRH_Add_DATETIME_Child	
MC_SRH_Create_DATETIME_Node	
MC_SRH_Add_DATE_Child	
MC_SRH_Create_DATE_Node	400
MC_SRH_Add_TIME_Child	401
MC_SRH_Create_TIME_Node	402
MC_SRH_Add_UIDREF_Child	403
MC_SRH_Create_UIDREF_Node	404
MC_SRH_Add_PNAME_Child	405
MC_SRH_Create_PNAME_Node	406
MC_SRH_Add_SCOORD_Child	407
MC_SRH_Set_SCOORD_Next_Data	408
MC_SRH_Create_SCOORD_Node	409
MC_SRH_Add_TCOORD_R_Child	411

MC_SRH_Create_TCOORD_R_Node	412
MC_SRH_Set_TCOORD_R_Next_Data	413
MC_SRH_Add_TCOORD_O_Child	414
MC_SRH_Create_TCOORD_O_Node	415
MC_SRH_Set_TCOORD_O_Next_Data	416
MC_SRH_Add_TCOORD_D_Child	417
MC_SRH_Create_TCOORD_D_Node	418
MC_SRH_Set_TCOORD_D_Next_Data	419
MC_SRH_Add_COMPOSITE_Child	420
MC_SRH_Create_COMPOSITE_Node	421
MC_SRH_Add_IMAGE_Child	422
MC_SRH_Create_IMAGE_Node	422
MC_SRH_Set_IMAGE_Frames	423
MC_SRH_Add_WAVEFORM_Child	424
MC_SRH_Create_WAVEFORM_Node	425
MC_SRH_Set_WAVEFORM_Channels	425
MC_SRH_Add_CONTAINER_Child	426
MC_SRH_Create_CONTAINER_Node	427
MC_SRH_Add_TABLE_Child	428
MC_SRH_Create_TABLE_Node	429
MC_SRH_Set_TABLE_Next_Row_Definition	430
MC_SRH_Set_TABLE_Next_Column_Definition	431
MC_SRH_Set_TABLE_Next_Cell_Values	432
MC_SRH_Add_Reference	432
MC_SRH_Get_Reference	433
MC_SRH_Get_NodeType	433
MC_SRH_Get_First_Child	435
MC_SRH_Get_Next_Child	436
MC_SRH_Get_Concept_Name	437
MC_SRH_Get_TEXT_Data	438
MC_SRH_Get_CODE_Data	439
MC_SRH_Get_NUM_Data	440
MC_SRH_Get_NUM_Next_Data	441
MC_SRH_Get_NUM_Qualifier	442
MC_SRH_Get_DATETIME_Data	443
MC_SRH_Get_DATE_Data	444
MC_SRH_Get_TIME_Data	444
MC_SRH_Get_UIDREF_Data	445

MC_SRH_Get_PNAME_Data	446
MC_SRH_Get_SCOORD_First_Data	446
MC_SRH_Get_SCOORD_Next_Data	447
MC_SRH_Get_TCOORD_D_First_Data	448
MC_SRH_Get_TCOORD_D_Next_Data	450
MC_SRH_Get_TCOORD_O_First_Data	451
MC_SRH_Get_TCOORD_O_Next_Data	452
MC_SRH_Get_TCOORD_R_First_Data	453
MC_SRH_Get_TCOORD_R_Next_Data	454
MC_SRH_Get_COMPOSITE_Data	455
MC_SRH_Get_IMAGE_Data	455
MC_SRH_Get_IMAGE_Frames	456
MC_SRH_Get_WAVEFORM_Data	457
MC_SRH_Get_WAVEFORM_Channels	458
MC_SRH_Get_TABLE_First_Cell_Values_Counts	459
MC_SRH_Get_TABLE_First_Column_Definition	
MC_SRH_Get_TABLE_First_Row_Definition	461
MC_SRH_Get_TABLE_Next_Cell_Values_Counts	462
MC_SRH_Get_TABLE_Next_Column_Definition	463
MC_SRH_Get_TABLE_Next_Row_Definition	464
MC_SRH_Get_TABLE_Cell_Values_Item	464
MC_SRH_Get_CONTAINER_Data	465
APPENDIX A: WRITING A DICOM CONFORMANCE STATEMENT	467
Conformance Statement Sections	467
Implementation Model	467
Application Data Flow	467
Functional Definition of Application Entities (AE)	468
AE Specifications	468
APPENDIX B: CONFIGURATION PARAMETERS	473
Initialization File	473
Application Profile	477
Service List	479
Transfer Syntax Lists	492
Role Negotiation	496
DICOM Asynchronous Communication	497
Extended Negotiation	
Related General SOP Classes and Service Classes	498
System Profile	

Service Profile	

Overview

This reference manual contains a detailed description of the functionality of the Merge DICOM Toolkit Library. This includes library configuration, application programmer's interface (API) specification, and a DICOM conformance statement for the toolkit.

Structure of Documentation

The Merge DICOM Toolkit documentation is structured as shown in Figure 1.




Figure 1: Merge DICOM Toolkit Documentation Roadmap

Read Me FIRST! The User's Manual is the foundation for all other documentation because it explains the concepts of DICOM and the toolkit. Before plunging into the Reference Manual or Sample Application Guide you should be comfortable with the material in the User's Manual.

The Reference Manual is where you go for detailed information on the DICOM Toolkit. This includes the Application Programming Interface (API), toolkit configuration, the runtime object database, and status logging. The Reference Manual also includes a DICOM conformance statement for the toolkit.

The DICOM Extended Toolkit Manual is an optional extension that describes the organization of the Merge DICOM Toolkit DICOM Database and how to use it to extend standard services and

define your own private services. Tools are supplied for converting the contents of the database into the binary runtime object database.

Sample applications

The Sample Application Guide describes approaches to developing specific classes of DICOM applications. It presents the pertinent information from Parts 3 or 4 of the DICOM Standard in a more readable way and in the context of the DICOM Toolkit. The Application Guide also details the DICOM messages that can be passed between applications on the network. Also, a sample application is described, and the application supplied in source form for your platform.

Platform specific information required to use the DICOM Toolkit on your target platform are specified in Platform Notes. This includes supported compilers, compiler options, link options, configuration, and run-time related issues.

Conventions

This manual follows a few formatting conventions.

Terms that are being defined are presented in **boldface**.

Sample Margin
NoteMargin notes (in the left margin) are used to highlight important points or sections of the
document.

Performance Portions of the document that can relate directly to the performance of your application are marked with the special margin note Performance Tuning.

Sample commands appear in **bold courier** font, while sample output, source code, and function calls appear in standard Courier font.

Hexadecimal numbers are written with a trailing H. For example, 16 decimal is equivalent to 10H hexadecimal.

Configuration

Toolkit configuration is accomplished through the use of initialization or configuration files. Initialization files, also called "ini" files, contain configuration information the toolkit will use to initialize its internal settings.

Each of the four toolkit initialization files use the same format. The format of the initialization files is the same format that is used by others in the industry, namely Microsoft. Configuration files are broken down into sections for easier organization and grouping of parameters. Each section has a section heading enclosed in square brackets. Next, parameters are defined by putting the parameter name to the left of an equal sign and its initial value to the right. Figure 2 illustrates the format of an "ini" file.




Figure 2: Format of a configuration file.

Notice that parameter names are relative to their header sections. For example, PARAMETER_1 and PARAMETER_2 are defined twice in the above example "ini" file. But, since each is defined in a different section, they are considered different entities.

Since the Merge DICOM Toolkit is very versatile and configurable, it uses initialization files extensively. The toolkit makes use of four configuration files: the Merge Initialization File, the Library Profile, the Application Profile and the Service Profile. The *Appendix B: Configuration Parameters* discusses each of these files in detail.

Any application using the Merge DICOM Toolkit library is required to initialize the library using the MC_Library_Initialization() call. In most cases the first parameter passed to this function will be NULL and initialization of the library will occur from the configuration files described in this section.

The genconf utility (described later in this section) can be used to convert these configuration files into a function that can be compiled into an object file and linked directly into your application. This generated function can also be specified as the first parameter to MC_Library_Initialization(). This approach should only be used in environments where the configuration is not likely to be updated in the field, where performance at initialization time is of great concern, or you running in an embedded environment (without a file system) with the embedded version of the library.

Initialization File Integration

The Merge DICOM Toolkit allows initialization files to be integrated into an application at "link" time. The toolkit provides utilities to "compile" your initialization files into ANSI-C source code files that can then be compiled and linked into an object file. This object file is linked into the application thus allowing your application to access the initial values at run-time. This is a great advantage to systems that require limited disk access and the only option for embedded systems.

As with all "compiled programs", the disadvantage is that when initial values change, the files must be "recompiled" and re-linked into your application.

Performance Generating a Static Configuration Function (genconf)

The genconf utility supplied with Merge DICOM Toolkit allows you to convert the four configuration files (the Initialization file, the Application profile, the System Profile and the Service Profile) into an ANSI-C function that can be compiled into an object file and statically linked to your application along with the toolkit library.

In most cases this approach should not be used, since it forces you to re-run genconf and relink your application whenever you wish to change the initial configuration values. It can be used in environments where high performance at initialization time is critical, or in embedded environments. It may also be of use in end user applications to reduce installation confusion.

To use genconf, follow these steps:

- 1. Make your current directory (i.e., cd to) the directory that contains the Initialization File (i.e. merge.ini).
- 2. Make certain that the MERGE INI environment variable is pointing to the Initialization File.
- 3. Make certain that the parameters dealing with locations of other configuration files within the Initialization File are correct. For example, make sure MERGECOM 3 PROFILE contains the path list of the System Profile.
- 4. Make sure that the initial values in ALL configuration files are set to the initial values you desire. Remember, if the initial values need to be changed, you must rerun genconf, recompile and re-link your application.
- 5. Run genconf.

Tuning

When genconf starts, it will display a banner similar to the following:

```
% genconf
```

```
MergeCOM (tm)
               Configuration Source Generator Version 3.9.0
(c) 2008 Merge OEM. All rights reserved.
```

An ANSI-C source file named mc3cfg.c is generated that defines several structures and a configuration function called MC Config Values (). Compile and link this source file into your application. You can then use this configuration function by specifying the MC Config Values () function as the first parameter of the MC Library Initialization () call.

Never modify the mc3cfg.cto change configuration parameters. Always modify the configuration files and rerun genconf. Note that the genconf application MUST ONLY be used with the version of the toolkit that was shipped to you. Using mixed version of the library and genconf will cause your applications to fail.

Generating a Static Dictionary Function (gendict)

Performance Tuning The gendict utility supplied with Merge DICOM Toolkit allows you to convert the data dictionary into an ANSI-C function that can be compiled into an object file and statically linked to your application along with the toolkit library.

In most cases this approach should not be used, since it forces you to re-run gendict and re-link your application whenever you wish to change the initial configuration values. It can be used in environments where high performance at initialization time is critical, or in embedded environments. It may also be of use in end user applications to reduce installation confusion.

To use gendict, follow these steps:

- 1. Make your current directory (i.e., cd to) the directory that contains the data dictionary (i.e. mrgcom3.dct).
- 2. Make certain that the MERGE INI environment variable is pointing to the Initialization File.
- 3. Make certain that the DICTIONARY_FILE parameter in the System Profile (i.e., mergecom.pro) is set correctly.
- 4. Run gendict.

When gendict starts, it will display a banner similar to the following:

% gendict

MergeCOM (tm) DICOM Dictionary Source Generator Version 3.9.0 (c) 2008 Merge OEM. All rights reserved.

An ANSI-C source file named mc3dict.c is generated that defines several structures and a configuration function called MC_Dictionary_Values. Compile and link this source file into your application. You can then use this configuration function by specifying the MC_Dictionary_Values function as the second parameter of the MC_Library_Initialization() call.

Never modify the mc3dict.c to change configuration parameters. Always modify the configuration files and rerun genconf.

Application Programming Interface

Overview of API

Merge DICOM Toolkit provides functions which are used to construct and manipulate "message objects" and "file objects" as well as functions to establish a communication session with other DICOM systems to exchange DICOM messages.

"Message objects" and "file objects" are defined in the DICOM standard and in the conformance statements of DICOM applications.

The functions of the base API may be divided into eleven major groups:

- 1. **Library initialization and reset functions** Must be called initially and when wishing to reset the library to its initially configured state.
- 2. Run time configuration

Allow your application to change configurable parameters at runtime.

3. Application Registration

Register your DICOM application entity with the library.

4. Association functions

Deal with opening DICOM associations and querying the characteristics of a proposed or open association.

5. Message object functions

Deal with the creation and manipulation of message objects.

6. Item object functions

Deal with the creation and freeing of the item objects that are contained in attributes of message objects that are of value representation SQ.

7. File object functions

Deal with the creation, manipulation, and freeing of file objects.

8. Message, file and item object functions

Used in the encoding, decoding, and manipulation of the attributes of messages, files and items. This includes the handling of private attributes.

9. Message Transfer

These are used in sending and receiving DICOM messages over an open association.

10. DICOMDIR Object Functions

These deal with the upkeep and manipulation of DICOMDIR objects.

11. Structured Reporting Object Functions

These deal with the upkeep and manipulation of Structured Reporting (SR) objects.

12. Miscellaneous Functions.

 $[\]textcircled{\mbox{\scriptsize C}}$ Copyright Merge Healthcare Solutions Inc. 2023

Additional to the base API the toolkit provides higher level functionality that can facilitate the manipulation of specific DICOM objects. The current version of the toolkit contains the following categories of high level functions:

- 1. **DICOMDIR Manipulation Functions (MC_DDH_ functions)** These functions provide enhanced access to DICOMDIR files.
- Structured Report Manipulation Functions (MC_SRH_ functions) These functions provide enhanced access to Structured Reporting (SR) objects.

Library Initialization and Reset

The Merge DICOM Toolkit library initialization functions follow:

MC_Library_Initialization MC_Library_Release MC_Library_Reset MC_Set_MergeINI MC_Get_Version_String

Run Time Configuration Functions

The Merge DICOM Toolkit run-time configuration functions follow:

MC_Release_Library_Exception_Handler MC_Set_Bool_Config_Value MC_Set_Int_Config_Value MC_Set_Library_Exception_Handler MC_Set_Log_Destination MC_Set_Long_Config_Value MC_Set_String_Config_Value MC_Get_Bool_Config_Value MC_Get_Int_Config_Value MC_Get_Log_Destination MC_Get_Long_Config_Value MC_Get_Long_Config_Value

Application Registration

The Merge DICOM Toolkit application registration functions may be grouped as follows:

MC_Register_Application MC_Release_Application MC_Register_Callback_Function MC_Register_pCallback_Function MC_Release_Callback_Function MC_Set_Message_Callbacks

Association Functions

The Merge DICOM Toolkit association functions may be grouped as follows:

[©] Copyright Merge Healthcare Solutions Inc. 2023

1. Association object creation and release.

MC_Open_Association MC_Open_Association_With_Connect_Timeout MC_Open_Association_With_Identity MC_Open_Association_With_Identity_With_Reject_Info MC_Open_Association_With_Identity_With_Reject_Info_With_Connect_Timeout MC_Open_Association_With_Reject_Info MC_Open_Association_With_Callback MC_Open_Association_With_All_Optional_Parameters MC_Open_Secure_Association MC_Open_Secure_Association_With_Reject_Info MC_Close_Association MC_Abort_Association MC_Release_Parent_Association MC_Wait_For_Association MC_Wait_For_Association_On_Port MC_Wait_For_Connection MC_Wait_For_Connection _On_Port MC_Wait_For_Secure_Association MC_Wait_For_Secure_Association_On_Port MC_Close_Listen_Port

MC_Process_Association_Request

MC_Process_Secure_Association_Request

MC_Accept_Association

MC_Accept_Association_With_Identity

MC_Reject_Association

MC_Reject_Association_With_Reason_Codes

MC_Set_Negotiation_Info (depracated) MC_Set_Negotiation_Info_For_Association

MC_Get_Negotiation_Info

MC_Clear_Negotiation_Info (depracated)

MC_NewProposedServiceList MC_NewProposedServiceListAsync MC_NewServiceFrom... Functions MC_NewSyntaxList MC_FreeServiceList MC_FreeService MC_FreeSyntaxList 2. Query of association characteristics.

MC_Get_Association_Info MC_Get_Listen_Socket MC_Get_Listen_Socket_For_Port MC_Get_First_Acceptable_Service MC_Get_Next_Acceptable_Service MC_Get_User_Identity_Info MC_Get_User_Identity_Length

Message Object Functions

The Merge DICOM Toolkit message object functions may be grouped as follows:

1. Message object creation and release.

MC_Open_Message MC_Free_Message MC_Open_Empty_Message MC_Set_Service_Command

2. Message duplication

MC_Duplicate_Message

3. Clearing all message attribute values.

MC_Empty_Message

4. Message validation functions: those dealing with ensuring that a message does not violate DICOM or conformance rules.

MC_Validate_Message MC_Validate_Attribute MC_Get_Next_Validate_Error

5. Message streaming functions: those dealing with retrieving message values from a DICOM stream or creating a DICOM stream from message values.

MC_Message_To_Stream MC_Stream_To_Message MC_Stream_To_Message_With_Offset MC_Get_Stream_Length

6. Message attribute information functions: those functions which report which attributes are in a message and information about message attributes.

MC_List_Message

 Message XML conversion functions: those dealing with converting a message to an XML string or reading attributes from an XML string into a message using either Merge DICOM Model or Native DICOM Model.

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Message_To_XML MC_XML_To_Message

MC_Message_To_XML_Native MC_XML_Native_To_Message

8. Message JSON conversion functions: those dealing with converting a message to an JSON string or reading attributes from a JSON string into a message using DICOM JSON Model.

MC_Message_To_Json MC_Json_To_Message

Item Object Functions

The Merge DICOM Toolkit message object functions may be grouped as follows:

1. Functions supporting attributes with a value representation of SQ (Sequence of Items):

MC_Open_Item MC_Free_Item MC_Empty_Item MC_List_Item

File Object Functions

The Merge DICOM Toolkit file object functions may be grouped as follows:

1. File object creation and release.

MC_Create_File MC_Create_Empty_File MC_Free_File

2. Clearing all file attribute values.

MC_Empty_File

3. Message validation functions: those dealing with ensuring that a message does not violate DICOM or conformance rules.

MC_Validate_File MC_Get_Next_Validate_Error

4. File opening, closing and information functions: those dealing with retrieving and writing DICOM file objects from or to media as well as getting information about the file object.

MC_Open_File MC_Open_File_Bypass_OBOW MC_Open_File_Upto_Tag MC_Open_File_Upto_Tag_Bypass_Value MC_Write_File MC_Write_File_By_Callback MC_Get_File_Length

[©] Copyright Merge Healthcare Solutions Inc. 2023

5. File manipulation functions: those functions that manipulate the elements that are specific to a file.

MC_Reset_Filename MC_Get_Filename MC_Set_File_Preamble MC_Get_File_Preamble

6. File transformation functions: those dealing with changing "file objects" to "message objects".

MC_File_To_Message MC_Message_To_File

7. Message attribute information functions: those functions which report which attributes are in a message and information about message attributes.

MC_List_File

Message, File and Item Object Functions

The Merge DICOM Toolkit message and item object functions can be grouped as follows:

1. Message attribute information functions: those functions which report which attributes are in a message and information about message attributes.

MC_Get_First_Attribute MC_Get_Next_Attribute MC_Get_Attribute_Info MC_Get_pAttribute_Info

2. Message attribute manipulation.

MC_Add_Standard_Attribute MC_Add_Nonstandard_Attribute MC_Add_Private_Attribute MC_Add_Private_Block MC_Byte_Swap_OBOW MC_Delete_Attribute MC_Delete_Current_Value MC_Delete_Private_Attribute MC_Delete_Private_Block MC_Delete_Range

3. Functions to change the value representation of a message attribute from Unknown_VR to a valid value representation code:

MC_Set_Value_Representation MC_Set_pValue_Representation

4. Setting values for standard attributes.

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Set_Value... Functions MC_Set_Value_From_Function MC_Set_Next_Value... Functions MC_Set_Value_To_NULL MC_Set_Next_Value_To_NULL MC_Set_Value_To_Empty MC_Set_Encapsulated_Value_From_Function MC_Set_Next_Encapsulated_Value_From_Function MC_Close_Encapsulated_Value

5. Setting values for <u>private</u> attributes.

MC_Set_pValue... Functions MC_Set_pValue_From_Function MC_Set_Next_pValue... Functions MC_Set_pValue_To_NULL MC_Set_Next_pValue_To_NULL MC_Set_pValue_To_Empty

6. Getting values of <u>standard</u> attributes.

MC_Get_Value... Functions MC_Get_Value_To_Function MC_Get_Value_Count MC_Get_Value_Length MC_Get_Next_Value...Functions MC_Get_Encapsulated_Value_To_Function MC_Get_Next_Encapsulated_Value_To_Function MC_Get_Frame_To_Function MC_Get_Offset_Table_To_Function

7. Getting values of private attributes.

MC_Get_pValue... Functions MC_Get_pValue_To_Function MC_Get_pValue_Count MC_Get_pValue_Length MC_Get_Next_pValue... Functions

Message Transfer

The Merge DICOM Toolkit message transfer functions follow:

MC_Send_Request_Message MC_Send_Request_Message_For_Service MC_Send_Request MC_Send_Request_For_Service MC_Send_Response_Message MC_Send_Response MC_Read_Message

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

MC_Read_Message_To_Tag MC_Continue_Read_Message MC_Continue_Read_Message_To_Tag MC_Read_To_Stream MC_Reset_Message_Transfer_Syntax MC_Set_Message_Transfer_Syntax MC_Get_Message_Transfer_Syntax

DICOMDIR Object Functions

The DICOMDIR object functions may be grouped as follows:

1. DICOMDIR entity functions: those functions that deal with the manipulation of DICOMDIR entities.

MC_Dir_Root_Entity MC_Dir_Next_Entity MC_Dir_Add_Entity MC_Dir_Delete_Referenced_Entity

2. DICOMDIR directory record functions: those functions that deal with the manipulation of DICOMDIR directory records.

MC_Dir_First_Record MC_Dir_Next_Record MC_Dir_Add_Record MC_Dir_Delete_Record

3. DICOMDIR MRDR directory record functions: those functions which allow users to create MRDR (Multi-referenced file directory record) records and set up references to them.

MC_Dir_Open_MRDR MC_Dir_Reference_MRDR MC_Dir_Remove_Ref_MRDR

4. DICOMDIR statistical functions: those functions which allow users to obtain counts of the over all items and entities contained within a DICOMDIR object.

MC_Dir_Root_Count MC_Dir_Entity_Count MC_Dir_Item_Count

5. DICOMDIR sorting functions: this function which allow users to sort the record types in each DICOMDIR entity, and the record entries within each record type in order within a DICOMDIR object.

MC_Dir_Sort

6. DICOMDIR storage functions: this function which allows users to set the storage for specific directory records in a DICOMDIR. This routine reduces memory usage for DICOMDIRs.

MC_Dir_Stream_Directory_Record

[©] Copyright Merge Healthcare Solutions Inc. 2023

Structured Reporting Object Functions

The Structured Reporting object functions are grouped as follows:

1. Getting values of SR objects:

MC_SR_Get_First_Child MC_SR_Get_Next_Child MC_SR_Get_Child_Count MC_SR_Get_Root

2. Adding values to SR objects:

MC_SR_Add_Child MC_SR_Add_Root MC_SR_Get_Location

3. Converting messages to or from SR objects:

MC_Message_To_SR MC_SR_To_Message

4. Deleting child SR objects:

MC_SR_Delete_Child

Miscellaneous Functions

The following miscellaneous functions are available:

MC_Error_Message MC_Get_Message_Service MC_Get_MergeCOM_Service MC_Get_UID_From_MergeCOM_Service MC_Get_Tag_Info MC_Get_Tag_Info MC_Get_Tag_Keyword MC_Get_Enum_From_Transfer_Syntax MC_Get_Transfer_Syntax_From_Enum MC_Set_Log_Prefix MC_Enable_Unicode_Conversion MC_Byte_To_Unicode MC_Unicode_To_Byte MC_Register_Network_Capture_Callbacks

Enhanced DICOMDIR Manipulation Functions

The enhanced DICOMDIR manipulation functions may be grouped as follows:

1. DICOMDIR file functions: functions that deal with reading and creating DICOMDIR files.

MC_DDH_Create MC_DDH_Open

[©] Copyright Merge Healthcare Solutions Inc. 2023

2. Directory record navigation functions: functions that deal with the traversal of the directory record hierarchy.

MC_DDH_Get_First_Lower_Record MC_DDH_Get_Next_Record MC_DDH_Get_Parent_Record MC_DDH_Traverse_Records MC_DDH_Release_Record MC_DDH_Get_Record_Type

3. DICODIR modification function: functions that deal with modifying the directory record hierarchy:

MC_DDH_Add_Record MC_DDH_Delete_Record MC_DDH_Update MC_DDH_Copy_Values

High Level Structured Reporting Manipulation Functions

The High Level Structured Reporting (SR) manipulation functions are grouped as follows:

1. SR creation and setup function:

MC_SRH_Create_SR MC_SRH_Free_SR MC_SRH_Add_TEXT_Child MC_SRH_Add_CODE_Child MC_SRH_Set_NUM_Qualifier MC_SRH_Set_NUM_Next_Data MC_SRH_Add_NUM_Child MC_SRH_Add_DATETIME_Child MC_SRH_Add_DATE_Child MC_SRH_Add_TIME_Child MC_SRH_Add_UIDREF_Child MC_SRH_Add_PNAME_Child MC_SRH_Add_SCOORD_Child MC_SRH_Set_SCOORD_Next_Data MC_SRH_Add_TCOORD_R_Child MC_SRH_Set_TCOORD_R_Next_Data MC_SRH_Add_TCOORD_O_Child MC_SRH_Set_TCOORD_O_Next_Data MC_SRH_Add_TCOORD_D_Child MC_SRH_Set_TCOORD_D_Next_Data MC_SRH_Add_COMPOSITE_Child MC_SRH_Add_IMAGE_Child MC_SRH_Set_IMAGE_Frames MC_SRH_Add_WAVEFORM_Child MC_SRH_Set_WAVEFORM_Channels MC_SRH_Add_CONTAINER_Child

MC_SRH_Add_TABLE_Child MC_SRH_Set_TABLE_Next_Row_Definition MC_SRH_Set_TABLE_Next_Column_Definition MC_SRH_Set_TABLE_Next_Cell_Values MC_SRH_Set_Concept_Name

2. SR parsing and reading functions:

MC_SRH_Get_NodeType MC_SRH_Get_First_Child MC_SRH_Get_Next_Child MC_SRH_Get_Concept_Name MC_SRH_Get_TEXT_Data MC_SRH_Get_CODE_Data MC_SRH_Get_NUM_Data MC_SRH_Get_NUM_Next_Data MC_SRH_Get_NUM_Qualifier MC_SRH_Get_DATETIME_Data MC_SRH_Get_DATE_Data MC_SRH_Get_TIME_Data MC_SRH_Get_UIDREF_Data MC_SRH_Get_PNAME_Data MC_SRH_Get_SCOORD_First_Data MC_SRH_Get_SCOORD_Next_Data MC_SRH_Get_TCOORD_D_First_Data MC_SRH_Get_TCOORD_D_Next_Data MC_SRH_Get_TCOORD_O_First_Data MC_SRH_Get_TCOORD_O_Next_Data MC_SRH_Get_TCOORD_R_First_Data MC_SRH_Get_TCOORD_R_Next_Data MC_SRH_Get_COMPOSITE_Data MC_SRH_Get_IMAGE_Data MC_SRH_Get_IMAGE_Frames MC_SRH_Get_WAVEFORM_Data MC_SRH_Get_WAVEFORM_Channels MC_SRH_Get_TABLE_First_Cell_Values_Counts MC_SRH_Get_TABLE_First_Column_Definition MC_SRH_Get_TABLE_First_Row_Definition MC_SRH_Get_TABLE_Next_Cell_Values_Counts MC_SRH_Get_TABLE_Next_Column_Definition MC_SRH_Get_TABLE_Next_Row_Definition MC_SRH_Get_TABLE_Cell_Values_Item MC_SRH_Get_CONTAINER_Data MC_SRH_Get_Reference

3. SR utility functions:

MC_SRH_Add_Child MC_SRH_Add_Reference MC_SRH_Create_TEXT_Node MC_SRH_Create_CODE_Node MC_SRH_Create_NUM_Node MC_SRH_Create_DATETIME_Node MC_SRH_Create_DATE_Node MC_SRH_Create_TIME_Node MC_SRH_Create_UIDREF_Node MC_SRH_Create_PNAME_Node MC_SRH_Create_SCOORD_Node MC_SRH_Create_TCOORD_R_Node MC_SRH_Create_TCOORD_O_Node MC_SRH_Create_TCOORD_D_Node MC_SRH_Create_COMPOSITE_Node MC_SRH_Create_IMAGE_Node MC_SRH_Create_WAVEFORM_Node MC_SRH_Create_TABLE_Node MC_SRH_Create_CONTAINER_Node

API Function Reference

The next section of this manual contains an alphabetically arranged function reference section. Each function reference provides:

- A brief description of the API function
- A synopsis of the API function which contains a list of include files required to use the function, the function prototype and a description of each function parameter
- Remarks outlining the use of the function
- A list of status codes returned by the function
 - A "See also" cross reference to other functions.

MC_Abort_Association

Terminates an open association immediately.

Synopsis

```
#include ``mergecom.h"
```

```
MC_STATUS MC_Abort_Association (
    int *AssociationID
```

)

AssociationID Address of the association object's identification number

[©] Copyright Merge Healthcare Solutions Inc. 2023

Remarks

MC_Abort_Association abruptly terminates the DICOM association. In effect, this function is used to tell the remote system that the association is no longer valid.

It is the responsibility of the requester of an association (client) to close a DICOM association. If a service provider (server) cannot proceed with the association it should use **MC_Abort_Association** to terminate the association. The service user (client) uses **MC_Close_Association** for normal association completion, or **MC_Abort_Association** if an error situation occurs.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The AssociationID parameter was NULL.
MC_INVALID_ASSOC_ID	*AssociationID is not a valid association object ID.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC Close Association	

MC_Close_Association MC_Open_Association MC_Open_Secure_Association MC_Wait_For_Association MC_Wait_For_Secure_Association

MC_Accept_Association MC_Accept_Association_With_Identity

Accept a remote application's request for a DICOM association.

Synopsis

AssociationID	The association object's identification number
ServerResponse	A server response to a User Identity request from the association request as defined in DICOM Supplement 99
ServerResponseLength	The length of the buffer passed in the <i>ServerResponse</i> parameter. If <i>ServerResponse</i> is NULL, <i>ServerResponseLength</i> should be set to 0.

Remarks

If a **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** function completes normally, one of two functions must be called. **MC_Accept_Association** informs the remote application that the association can proceed. Use **MC_Reject_Association** to reject the association request.

MC_Accept_Association_With_Identity can be used to send a user identity response to the association request when requested for the association as defined in DICOM Supplement 99. User Identity information within the association request can be retrieved through the use of the MC_Get_Association_Info, MC_Get_User_Identity_Info, and MC_Get_User_Identity_Length functions.

If the association request does not request a positive response or if no user identity information was contained in the association request, the **MC_Accept_Association** function should be used. When a positive response is requested, the **MC_Accept_Association_With_Identity** function can be used. If no server response information is required, the *ServerResponse* field can be set to NULL and the *ServerResponseLength* field can be set to 0.

If **MC_Accept_Association** returns **MC_ASSOCIATION_ABORTED** or **MC_SYSTEM_ERROR** no further calls may be made for the association.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_NO_REQUEST_PENDING	There are no pending association request for this AssociationID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_ASSOCIATION_ABORTED	The association has been aborted.
See Also	
MC_Wait_For_Association MC_Get_User_Identity_Length MC_Wait_For_Secure_Association	MC_Get_Association_Info MC_Get_User_Identity_Info MC_Open_Association_With_Identity MC_Reject_Association

MC_Add_Nonstandard_Attribute

Adds a new non-DICOM standard attribute to a message object.

Synopsis

<pre>#include ``mc3msg.h"</pre>		
<pre>MC_STATUS MC_Add_Nonstandard_Attribute (int MsgFileItemID, unsigned long Tag, MC_VR ValueRep)</pre>		
MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.	
Tag	DICOM tag which identifies this attribute.	
ValueRep	A code identifying the value representation of the attribute. The MC_VR enumerated type is defined in "mc3msg.h". The valid codes are:	
	AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR, UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL, OD, OF, SQ	

Remarks

MC_Add_Nonstandard_Attribute adds a nonstandard (and non-private) attribute that does not exist in the DICOM Data Dictionary to an existing message. A message can not have nonstandard attributes and still conform to DICOM. This implies **MC_Validate_Message** will return an error when validating a message with nonstandard attributes. This function should only be used in extreme cases: when communicating with non-conformant implementations of DICOM 3.0. Private attributes should be used to extend the standard, not nonstandard attributes.

Because nonstandard attributes are not defined in the Data Dictionary, you must supply the Value Representation code: *ValueRep*. If, however, an attempt is made to add an attribute which is in the Data Dictionary, an error status will be returned.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TAG	The attribute with an ID of <i>Tag</i> is a standard attribute found in the Data Dictionary. Instead, use MC_Add_Standard_Attribute .
MC_TAG_ALREADY_EXISTS MC_INVALID_VR_CODE	The message already contains an attribute with an ID of <i>Tag</i> . <i>ValueRep</i> is invalid.

MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID, or item object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Add_Standard_Attribute MC_Add_Private_Attribute MC_Add_Private_Block	MC_Delete_Attribute MC_Delete_Private_Attribute MC_Delete_Private_Block

MC_Delete_Range

MC_Add_Private_Attribute

Adds a private DICOM attribute to a message object.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Add_Private_Attribute (
    int MsgFileItemID,
    char *PrivateCode,
    unsigned short Group,
```

unsigned char ElementByte, MC_VR ValueRep

)

MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
PrivateCode	The code string which identifies which block in the private Group is to "own" the new attribute.
Group	The number identifying the private group. It must be an odd number.
ElementByte	The number identifying this attribute within the private block
ValueRep	A code identifying the value representation of the attribute. The MC_VR enumerated type is defined in "mc3msg.h".
	The valid codes are: AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR, UT, UI, SS, US, SV, UV, AT, SL, UL, FL, FD, OB, OW, OV, OL, OD. OF. SO

Remarks

MC_Add_Private_Attribute adds an attribute that does not exist in the Data Dictionary (a private attribute) to an existing message. A message can have private attributes and still conform to DICOM, as long as the attributes are not equivalent to and do not replace standard attributes. Private attributes are stored in an odd-number *Group* (the high-order portion of a DICOM tag). An identifying *PrivateCode* must be supplied to differentiate one owner's block of private attributes from another in any given *Group*. Since the attribute is not in the Data Dictionary, it is necessary to supply the value representation of the attribute's values: *ValueRep*.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>PrivateCode</i> was NULL.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID, or item object ID.
MC_NOT_FOUND	There is no private block identified by <i>PrivateCode</i> .
MC_INVALID_GROUP	<i>Group</i> is not an odd number.
MC_INVALID_VR_CODE	<i>ValueRep</i> is invalid.
MC_TAG_ALREADY_EXISTS	The private attribute already exists in the message object.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Add_Standard_Attribute MC_Add_Private_Attribute	MC_Delete_Attribute MC_Delete_Private_Attribute

MC_Delete_Private_Block

MC_Delete_Range

MC_Add_Nonstandard_Attribute

MC_Add_Private_Block

Creates a new DICOM private attribute block descriptor in a message object.

Synopsis

<pre>#include ``mc3msg.h"</pre>		
<pre>MC_STATUS MC_Add_Private_Block (int MsgFileItemID, char *PrivateCode, unsigned short Group)</pre>		
MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.	
PrivateCode	The code string which identifies the block which is being added to Group.	
Group	The number identifying the private group. It must be an odd number.	

Remarks

MC_Add_Private_Block adds an attribute to a given *Group* to identify a block of private attributes in the group. This must be done before **MC_Add_Private_Attribute** can be used to add attributes for the private block. The first unused tag in the range gggg0010 through gggg00FF will be assigned to this block and given *PrivateCode* as its value (gggg is *Group*).

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	PrivateCode was NULL.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_GROUP	<i>Group</i> is not an odd number.
MC_INVALID_VR_CODE	<i>ValueRep</i> is invalid.
MC_TOO_MANY_BLOCKS	There already exists the maximum 240 private blocks in the group.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Add_Standard_Attribute MC_Add_Private_Attribute MC_Add_Nonstandard_Attribute MC_Delete_Range MC_Delete_Attribute MC_Delete_Private_Attribute MC_Delete_Private_Block

MC_Add_Standard_Attribute

Adds a new DICOM standard attribute to a message object.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Add_Standard_Attribute (
    int MsgFileItemID,
    unsigned long Tag
)
MsgFileItem/ The identifier assigned to this object by the
D MC_Open_Message, MC_Open_Empty_Message,
    MC_Create_File, MC_Create_Empty_File, or
    MC_Open_Item functions.
Tag DICOM tag which identifies the attribute.
```

Remarks

MC_Add_Standard_Attribute adds an attribute that exists in the DICOM Data Dictionary (a standard attribute) to an existing message. Adding this attribute may make the message non-conformant to DICOM since all necessary attributes should be determined by the service and command specified in the **MC_Open_Message** function. This implies **MC_Validate_Message** would report an error when used to validate such a message. This function might be used to construct a message to communicate with another non-conformant application.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_TAG_ALREADY_EXISTS	The message already contains an attribute with an ID of <i>Tag</i> .
MC_INVALID_TAG	The Data Dictionary does not contain an attribute with an ID of <i>Tag</i> .
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_Add_Nonstandard_Attribute MC_Add_Private_Attribute MC_Add_Private_Block MC_Delete_Range MC_Delete_Attribute MC_Delete_Private_Attribute MC_Delete_Private_Block

MC_Byte_To_Unicode

A utility function to convert DICOM character set to Unicode.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Byte To Unicode (
       char *DefaultCharset,
       const char *Val,
       int InLen,
       int BufferSize,
       int *OutLen,
       MC Uhar *OutVal
)
 DefaultCharset
                  The DICOM specific character set name for the default
                  character set.
 Val
                  Input bytes array
 InLen
                  Input bytes count (excluding NULL terminator)
 BufferSize
                  Output buffer size
 OutLen
                  Output length (returned by this call)
 OutVal
                  Output buffer to hold the Unicode
```

Remarks

MC_Byte_To_Unicode requires MC_Enable_Unicode_Conversion being called first.

DefaultCharset should be set to the first character set of (0008,0005). If set to NULL or empty string, the default ISO-IR 6 (ASCII) will be used.

InLen can be set to -1 and the toolkit will calculate length if the input buffer is NULL terminated.

When the return status is MC_NORMAL_COMPLETION, *OutVal* contains the output Unicode string with U+0000 terminator. *OutLen* contains the number of Unicode characters present in the *OutVal* buffer (excluding the U+0000 terminator).

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.

MC_CANNOT_COMPLY	Fail to process the input. Check the Merge DICOM Toolkit log file for detail of failure.
MC_BUFFER_TOO_SMALL	The output buffer doesn't have enough space to hold the output.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_Unicode_To_Byte MC_Enable_Unicode_Conversion MC_Get_Value_To_UnicodeString MC_Get_Next_Value_To_UnicodeString MC_Set_Value_From_UnicodeString MC_Set_Next_Value_From_UnicodeString

MC_Byte_Swap_OBOW

Byte swaps an attribute of value representation OB, OW, or OF.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Byte_Swap_OBOW (
    int MsgFileItemID,
    unsigned long Tag
```

)

MsgFileIteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions
_	MC_Open_Item functions.
Tag	DICOM tag which identifies the attribute to byte swap.

Remarks

MC_Byte_Swap_OBOW byte swaps attributes of value representation OB, OW, or OF.

MC_Byte_Swap_OBOW can be of use when encoding or decoding 8-bit pixel data. A discussion of 8bit pixel data is contained in the *Merge DICOM Toolkit User's Manual*.

When data is OB, no operation is performed, and **MC_NORMAL_COMPLETION** is returned.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID, or item object ID.

MC_Set_Value_From_Function	MC_Get_Value_To_Function
See Also	
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CALLBACK_REGISTERED	This function is not allowed when the MC_Register_Callback_Function function was issued for the same attribute.
MC_INCOMPATIBLE_VR	The VR of <i>Tag</i> was not OB, OW, OD or OF.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_INVALID_TAG	<i>Tag</i> is not a valid tag in <i>MsgFileItemID</i> .

MC_Cleanup_Memory

Releases memory allocated but not in use by the Merge DICOM Toolkit library.

Synopsis		
#include "mergecom.h"		
MC_STATUS MC_Cleanup_Memory (int Timeout		
, Timeout	Maximum time, in seconds, that this function should attempt	
inneedt	to cleanup memory.	

Remarks

MC_Cleanup_Memory allows an application to reduce the amount of memory consumed by the toolkit. The Merge toolkit uses an internal memory management system to increase performance. This increase in performance is at the expense of keeping allocated memory block available by not freeing them to the operating system. This call allows memory blocks not in use to be returned to the operating system.

This call can take a significant amount of time if a large amount of memory is to be freed. The timeout can be used to make incremental improvements in the memory usage. Calls to this function pick up where the previous call left off. If the timeout expires before the function is complete, MC_TIMEOUT is returned

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_LIBRARY_NOT_INITIALIZED	Call made prior to library initialization
MC_TIMEOUT	The timeout expired before the entire cleanup was finished.
MC_VALUE_OUT_OF_RANGE	The timeout was less then 0.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_Clear_Negotiation_Info

De-registers extended negotiation information.

NOTE: Use of this call is deprecated.

MC_Set_Negotiation_Info_For_Association and setting of extended negotiation information through service lists should be used in place of this call and **MC_Set_Negotiation_Info**.

Synopsis

```
#include "mergecom.h"
MC_STATUS MC_Clear_Negotiation_Info (
    int ApplicationID,
    char *ServiceName
)
ApplicationID The identification number for the registered application.
ServiceName The name given to a valid DICOM service.
```

Remarks

The DICOM standard allows application entities to exchange "extended negotiation information" when establishing an association. The contents of the negotiation information must be known to both the association requestor application and the association acceptor application. Such extended negotiation is not often used for DICOM services, but some services may require it.

MC_Set_Negotiation_Info allows the caller to supply Merge DICOM Toolkit with extended negotiation information which it will use when establishing associations.

When a MC_Open_Association, MC_Open_Association_With_Identity or

MC_Open_Secure_Association call is made, Merge DICOM Toolkit sends any registered extended negotiation information to the association acceptor. The acceptor normally returns an updated version of the negotiation information which can be accessed using the **MC_Get_Negotiation_Info** call.

When a **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** call is made, Merge DICOM Toolkit stores any received extended negotiation information. This information may be accessed using the **MC_Get_Negotiation_Info** call. **MC_Set_Negotiation_Info** may be used to "update" the extended negotiation information before calling **MC_Accept_Association** to accept the association. Merge DICOM Toolkit will return any registered negotiation information to the remote application.

MC_Clear_Negotiation_Info is used to remove extended information previously registered for a service using the **MC_Set_Negotiation_Info** call.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_APPLICATION_ID	ApplicationID is not a valid application identifier.
MC_NULL_POINTER_PARM	ServiceName was NULL.
MC_UNKNOWN_SERVICE	<i>ServiceName</i> was not registered in the Merge DICOM Toolkit configuration files.
MC_NO_INFO_REGISTERED	There was no extended negotiation information registered for <i>ServiceName</i> for this application.
San Alan	

See Also

MC_Get_Negotiation_Info MC_Set_Negotiation_Info_For_Association MC_Set_Negotiation_Info

MC_Close_Association

Gracefully shuts down an open association with a remote DICOM application.

Synopsis

```
AssociationID The identifier assigned to this object by the
MC_Open_Association, MC_Open_Secure_Association,
MC_Process_Association_Request,
MC_Process_Secure_Association_Request,
MC_Wait_For_Association or
MC_Wait_For_Secure_Association function.
```

Remarks

MC_Close_Association gracefully shuts down the association connection and releases system resources used by the association. This function is used to end an association which has proceeded with no errors.

It is the responsibility of the requester of an association (client) to close a DICOM association. If a service provider (server) cannot proceed with the association it should use **MC_Abort_Association** to terminate the association. The service user (client) uses **MC_Close_Association** for normal association completion, or **MC_Abort_Association** if an error situation occurs.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The AssociationID parameter was NULL.
MC_INVALID_ASSOC_ID	*AssociationID is not a valid association object ID.
MC_ASSOCIATION_ABORTED	There was an error while closing the association. Error message logged.

See Also

MC_Abort_Association MC_Open_Association MC_Open_Secure_Association MC_Wait_For_Association MC_Wait_For_Secure_Association

MC_Close_Encapsulated_Value

Terminates an encapsulated value by placing the end delimiter in the stream.

Synopsis

#include ``mc3msg.h"

```
      MC_STATUS MC_Close_Encapsulated_Value (

      int MsgFileItemID

      unsigned long Tag

      )

      MsgFileItemID

      The identifier assigned to this object by the

      MC_Open_Message, MC_Open_Empty_Message,

      MC_Create_File, MC_Create_Empty_File, or

      MC_Open_Item functions.

      Tag

      The attribute which contains encapsulated data, but no end delimiter.
```

Remarks

MC_Close_Encapsulated_Value places an end delimiter for encapsulated data. This should only be done when at least MC_Set_Encapsulated_Value_From_Function has been used to store encapsulated data in the message for the attribute, and can also be used after MC_Set_Next_Encapsulated_Value_From_Function.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_INVALID_TRANSFER_SYNTAX	The message's transfer syntax is non-encapsulated.
MC_INVALID_MESSAGE_ID	<i>MessageID</i> is not a valid message object ID.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .

MC_Set_Encapsulated_Value_From_Function MC_Set_Next_Encapsulated_Value_From_Function

MC_Close_Listen_Port MC_Close_Listen_Port_On_Address

Stops listening for incoming DICOM connections on a port

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Close Listen Port (
       int Port
)
MC STATUS MC Close Listen Port On Address (
       int Port
       const char *Address
)
 Port
                   Listen port opened by a call to
                   MC_Wait_For_Association,
                   MC_Wait_For_Association_On_Port,
                   MC_Wait_For_Association_On_Address,
                   MC_Wait_For_Connection,
                   MC_Wait_For_Connection_On_Port,
                   MC_Wait_For_Connection_On_Address
                   MC_Wait_For_Secure_Association,
                   MC_Wait_For_Secure_Association_On_Port or
                   MC_Wait_For_Secure_Association_On_Address
                   function.
                   The network name or IP address of the network interface
 Address
                   on which the listener was started by a call to
                   MC_Wait_For_Association_On_Address,
                   MC_Wait_For_Connection_On_Address or
                   MC_Wait_For_Secure_Association_On_Address
                   function, exactly as passed to those functions.
```

Remarks

MC_Close_Listen_Port causes Merge DICOM Toolkit to stop listening for incoming DICOM associations on a TCP/IP listen port. The port may be the default listen port as configured by the TCPIP_LISTEN_PORT option from **MC_Wait_For_Association**, **MC_Wait_For_Connection** or **MC_Wait_For_Secure_Association**. It may also be a listen port specified in a call to the

MC_Wait_For_Association_On_Port, MC_Wait_For_Connection_On_Port, or MC_Wait_For_Secure_Association_On_Port functions. If a specific network interface was specified for the listener via a call to MC_Wait_For_Association_On_Address, MC_Wait_For_Connection_On_Address or MC_Wait_For_Secure_Association_On_Address, MC_Close_Listen_Port_On_Address should be used to stop the listener.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_INVALID_PORT_NUMBER	<i>Port</i> is not a valid listen port.
See Also	
MC_Wait_For_Association MC_Wait_For_Connection MC_Wait_For_Secure_Association MC_Wait_For_Connection_On_Addr MC_Wait_For_Secure_Association_(MC_Wait_For_Secure_Association_(Dn_Port

MC_Continue_Read_Message

Continues the reading of the current message that arrived from the remote application up until the end of the message. Continuing requires that a previous **MC_Read_Message_To_Tag** or **MC_Continue_Read_Message_To_Tag** function was called.

Synopsis

```
#include ``mergecom.h"
MC_STATUS MC_Continue_Read_Message (
    int AssociationID,
    int *MessageID,
)
AssociationID The association object's identification number.
MessageID The ID of the message to be continued up until StopTag.
```

Remarks

MC_Continue_Read_Message reads from after previous stop tag, up to the end of the message.

The message passed in must have been previously read with **MC_Read_Message_To_Tag**, or **MC_Continue_Read_Message_To_Tag** with a stop tag less then (FFFF,FFFF).

MC_Read_Message_To_Tag, or **MC_Read_Message** may be used immediately following this function to receive a new message from the remote application.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MessageID</i> is not a valid message object ID.
MC_NULL_POINTER_PARM	The MessageID parameter was NULL.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of the	e following are returned:
MC_NETWORK_SHUT_DOWN	The network connect unexpectedly dropped.
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_ASSOCIATION_CLOSED	The association has been closed.
MC_INACTIVITY_TIMEOUT	A timeout occurred in the middle of receiving a message.
MC_CONFIG_INFO_ERROR	The message information file describing the message's service/command pair could not be accessed.
MC_INVALID_MESSAGE_RECEIVEI	An improperly formatted DICOM message was received. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
-	_Continue_Read_Message_To_Tag _Continue_Read_Message_To_Stream

MC_Continue_Read_Message_To_Stream

Continues the reading of the current message that arrived from the remote application to a callback function. Continuing requires that a previous **MC_Read_Message_To_Tag** or **MC_Continue_Read_Message_To_Tag** function was called.

Synopsis

[©] Copyright Merge Healthcare Solutions Inc. 2023

MessagelD	The ID of the message to be continued.
UserInfo	Address of data which will be passed on to <i>YourReceiveStreamFunction</i> each time it is called. This may be NULL.
YourReceiveStreamFunctio n	Name of a function which will be called repeatedly to provide blocks of streamed DICOM message data.

The function must be prototyped as follows:

```
MC_STATUS YourReceiveStreamFunction (
    int CBmessageID,
    void* CBuserInfo,
    int CBdataSize,
    void* CBdataBuffer,
    int CBisFirst,
    int CBisLast
)
```

```
CBmessageID
                The identifier assigned to the message object by the
                MC_Open_Message function.
CBuserInfo
                Address of data which is being passed from the
                MC_Continue_Read_Message_To_Stream function. This
                may be NULL.
CBdataSize
                The number of bytes of stream data being provided to you in
                 CbdataBuffer.
CBdataBuffer
                The buffer containing stream data from the message object.
CBisFirst
                Is TRUE (not zero) when Merge DICOM Toolkit is providing
                the first block of stream data.
CBisLast
                Is TRUE (not zero) when Merge DICOM Toolkit is providing
                the last block of stream data.
```

Remarks

MC_Continue_Read_Message_To_Stream reads from after previous stop tag until the end of the message. The message passed in must have been read with MC_Read_Message_To_Tag, or MC_Continue_Read_Message_To_Tag.

MC_Continue_Read_Message_To_Stream can be used to increase performance by not having Merge DICOM Toolkit process a message and pass it directly to disk as it is being read off of the network. The **MC_Read_Message_To_Tag** can be used to get basic information about the message such as the SOP Instance UID of the message from the Affected SOP Instance UID tag or the SOP Class UID of the message from the Affected SOP Class UID tag. The transfer syntax that the message is encoded in can be determined from the **MC_Get_Message_Transfer_Syntax** function after the message has been initially read with **MC_Read_Message_To_Tag**.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MessageID</i> is not a valid message object ID or MC_Read_Message_To_Tag was not called for <i>MessageID</i> .
MC_NULL_POINTER_PARM	The MessageID parameter was NULL.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_STATE_VIOLATION	<i>AssociationID</i> is not a valid association that has been accepted.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of the	e following are returned:
MC_NETWORK_SHUT_DOWN	The network connect unexpectedly dropped.
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_ASSOCIATION_CLOSED	The association has been closed.
MC_INACTIVITY_TIMEOUT	A timeout occurred in the middle of receiving a message.
MC_CONFIG_INFO_ERROR	The message information file describing the message's service/command pair could not be accessed.
MC_INVALID_MESSAGE_RECEIVED	An improperly formatted DICOM message was received. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CALLBACK_CANNOT_COMPLY	YourReceiveStreamFunction returned an error.
See Also	
MC_Read_Message MC_Continue_Read_Message	MC_Read_Message_To_Tag MC_Get_Message_Transfer_Syntax

MC_Continue_Read_Message_To_Tag

Continues the reading of the current message that arrived from the remote application up to and including the specified tag. Continuing requires that a previous **MC_Read_Message_To_Tag** or **MC_Continue_Read_Message_To_Tag** function was called.

Synopsis

```
#include "mergecom.h"
MC_STATUS MC_Continue_Read_Message_To_Tag (
    int AssociationID,
    unsigned long StopTag,
    int *MessageID,
)
```

AssociationID The association object's identification number.

[©] Copyright Merge Healthcare Solutions Inc. 2023

StopTag	The tag to stop a read from the remote application.
MessageID	The ID of the message to be continued up until <i>StopTag.</i>

MC_Continue_Read_Message_To_Tag reads from after previous stop tag, up to and including *StopTag*.

The message passed in must have been read with **MC_Read_Message_To_Tag**, or **MC_Continue_Read_Message_To_Tag** with a stop tag less then *StopTag*.

The message returned must be read until the last tag (FFFF,FFFF) either explicitly with **MC_Continue_Read_Message_To_Tag**, or by calling **MC_Continue_Read_Message** before using **MC_Read_Message_To_Tag**, or **MC_Read_Message** again.

MC_Continue_Read_Message_To_Tag may be called multiple times with increasing stop tags.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	MessageID is not a valid message object ID.
MC_INVALID_TAG	<i>StopTag</i> was within the command set (less then (0001,0000)), or the tag requested has already been read in.
MC_MUST_CONTINUE_BEFORE_RE	ADING A previous message was not finished before this call
MC_NULL_POINTER_PARM	The MessageID parameter was NULL.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of the	e following are returned:
MC_NETWORK_SHUT_DOWN	The network connect unexpectedly dropped.
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_ASSOCIATION_CLOSED	The association has been closed.
MC_INACTIVITY_TIMEOUT	A timeout occurred in the middle of receiving a message.
MC_CONFIG_INFO_ERROR	The message information file describing the message's service/command pair could not be accessed.
MC_INVALID_MESSAGE_RECEIVED	An improperly formatted DICOM message was received. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_Read_Message MC_Read_Message_To_Tag MC_Continue_Read_Message MC_Continue_Read_Message_To_Stream

MC_Create_Empty_File

Creates a new empty file object

Synopsis		
<pre>#include ``mc3media.h"</pre>		
<pre>MC_STATUS MC_Create_Empty_File (int *FileIDPtr, char *Filename</pre>		
FileIDPtr	Upon successful completion, the file object identifier is returned here.	
Filename	String filename to be associated with this file object.	

Remarks

The **MC_Create_Empty_File** function creates a "file object" which contains no attributes. The resulting file object is given an identification number which is returned in **FileIDPtr*. All functions dealing with this file must provide this file ID number.

The opened file object is not associated with any particular DICOM service or command. If the file object is going to be converted to a message and sent to a network partner, or if **MC_Validate_File** is to be called for the file object, **MC_Set_Service_Command** must be called first to associate the file object with a given DICOM service and command.

If a file is opened using **MC_Create_Empty_File**, it is not necessary to add attributes to the file object before setting attribute values. If one of the set value functions (e.g. **MC_Set_Value_From_String**) is used for an attribute, the attribute will automatically be added to the file object before the value is set.

NOTE: This is <u>not the case</u> if a file object is opened with **MC_Create_File** or if **MC_Set_Service_Command** is called for the file. In these cases, the file <u>is</u> associated with a given service/command pair and attributes other than those associated with that service and command must be explicitly added to the file before setting values for the added attributes

If a file object generated by this function is to be used as a DICOMDIR, the **MC_Set_Service_Command** function must be called for the object in order for the **MC_Dir_...** functions to work.

NOTE:

MC_Set_Service_Command will not parse attributes contained in a file object. If directory records are contained in a file created with
 MC_Create_Empty_File and MC_Set_Service_Command is called on the object, the MC_Dir_... functions will not recognize the directory records.

The DICOM file prefix for the new file object is set to "DICM". All bytes in the file preamble are set to 00H.

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about commands associated with messages and files.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The <i>FileIDPtr</i> or <i>Filename</i> parameter was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment.
See Also	
MC_Create_File MC_Open_Item	MC_Open_Message MC_Reset_Filename

MC_Create_File

Creates a new file object

Synopsis

```
#include ``mc3media.h"
MC_STATUS MC_Create_File (
    int *FileIDPtr,
    char *Filename,
    char *ServiceName,
    MC_COMMAND Command
)
```

```
)FileIDPtrUpon successful completion, the file object identifier is<br/>returned here.FilenameFilename associated with file object.
```

- ServiceName String name of a DICOM service to be associated with this message object. Name are found in the file mergecom.srv.
- *Command* The **MC_COMMAND** enumerated values are defined in "mc3msg.h".

The **MC_Create_File** function creates a "file object" which contains all the attributes of a DICOM file which will be used for the given *ServiceName* and *Command*. The resulting file object is given an identification number which is returned in **FileIDPtr*. All functions dealing with this file must provide this file ID number. The *ServiceName* and *Command* are used to access configuration information which describes the attributes of the file. If such configuration information is not available, an empty file object is created and a warning message is logged.

This function can also be used to create a DICOMDIR object. When the service is set for a DICOMDIR, Merge DICOM Toolkit creates the internal data structures needed to manipulate a DICOMDIR. An empty root directory entity is created in the new DICOMDIR object and the root directory record offset attributes are initialized.

MC_Create_File generates in each file object created the DICOM File Meta Information attributes used by DICOM media services. The user is responsible for filling in these attributes. The DICOM prefix for the file is set to "DICM". All bytes in the file preamble are set to 00H.

MC_Create_Empty_File should be used if the service and command are not yet known, or if there is no need to validate that values will be set only for attributes assigned to a given service/command pair.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The <i>FileIDPtr</i> , <i>Filename</i> , or <i>ServiceName</i> parameter was NULL.
MC_INVALID_COMMAND	Command is not a supported command.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Create_Empty_File MC_Open_Item	MC_Open_Message

MC_Delete_Attribute

Removes an attribute from a message object.

Synopsis

#include ``mc3msg.h"

[©] Copyright Merge Healthcare Solutions Inc. 2023

NOTE:

MC_STATUS MC_Delete_Attribute (int MsgFileItemID, unsigned long Tag		
) MsgFileItemI	The identifier assigned to this object by the	
D	MC_Open_Message, MC_Open_Empty_Message,	
	MC_Create_File, MC_Create_Empty_File, or	
Tag	MC_Open_Item functions. DICOM tag which identifies the attribute.	
	Dicolaritag which identifies the attribute.	

Remarks

MC_Delete_Attribute removes an attribute (standard or nonstandard) from an existing message. Removing this attribute may make the message non-conformant to DICOM since all necessary attributes are determined by the service and command provided by **MC_Open_Message**. This implies **MC_Validate_Message** might report an error when used to validate such a message. This function could be used to clean up a message from a non-conformant application.

If the attribute has a value representation of SQ, the item objects identified by the attribute's values are automatically freed (by calling **MC_Free_Item**). If the freed item object is in use as a value in any other sequence of items, the value in those sequences is cleared.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
See Also	
MC_Add_Standard_Attribute MC_Add_Private_Attribute MC_Add_Private_Block MC_Add_Nonstandard_Attribute	MC_Delete_Range MC_Delete_Private_Attribute MC_Delete_Private_Block

MC_Delete_Current_Value

Deletes the current value within an attribute's value list.

Synopsis

#include ``mc3msg.h"

MC_STATUS MC_Delete_Current_Value (
 int MsgFileItemID,

[©] Copyright Merge Healthcare Solutions Inc. 2023

unsigned long Tag	
)	
MsgFileIteml	The identifier assigned to this object by the
D	MC_Open_Message, MC_Open_Empty_Message,
	MC_Create_File, MC_Create_Empty_File, or
	MC_Open_Item functions.
Tag	DICOM tag which identifies the attribute.

The **MC_Delete_Current_Value** function deletes the current value within an attribute's value list. The current value is defined as the last value retrieved with a call to the **MC_Get_Next_Value...** or **MC_Get_Value...** functions.

MC_Delete_Current_Value will not work with attributes whose value representation is OB, OW, or OF.

Performance
TuningCalling this function for an attribute that has a large enough value to be stored in a temporary file is not
recommended for performance reasons. For this condition, all of the attribute's values are read into
memory, the specific value within the attribute is deleted, and the remaining values are re-written to
the temporary file.

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these commands.

One of the enumerated MC_STATUS codes defined in "mcstatus.h ":	
Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_NO_MORE_VALUES	There are no more values for this attribute.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INCOMPATIBLE_VR	The function was called to delete an attribute value whose value representation (VR) was OB, OW, or OF.
MC_NULL_VALUE	The attribute's value was set to NULL.
MC_EMPTY_VALUE	The attribute is empty.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

NOTE: If the attribute has a value representation of SQ, the item objects identified by the attribute's value are automatically freed (by calling **MC_Free_Item**). If the freed item object is in use as a value in other sequence of items, the value in those sequences is cleared.

MC_Get_Value... Functions

MC_Get_Next_Value... Functions

MC_Delete_Private_Attribute

Removes a private attribute from a message object.

Synopsis

```
#include ``mc3msg.h"
```

```
MC_STATUS MC_Delete_Private_Attribute (
    int MsgFileItemID,
    char *PrivateCode,
    unsigned short Group,
    unsigned short ElementByte
```

)

MsgFileItemI D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group. It must be an odd number.
ElementByte	The number identifying this attribute within the private block.

Remarks

MC_Delete_Private_Attribute deletes an attribute that does not exist in the Data Dictionary (a private attribute) to an existing message. A message can have private attributes and still conform to DICOM, as long as the attributes are not equivalent to and do not replace standard attributes. Private attributes are stored in an odd-number *Group* (the high-order portion of a DICOM tag). An identifying *PrivateCode* must be supplied to differentiate one owner's block of private attributes from another in any given *Group*. The block number assigned to *PrivateCode* along with the *ElementByte* form the effective tag of the private attribute

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>PrivateCode</i> was NULL.
MC_INVALID_TAG	The private attribute does not exist in the message object.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_GROUP	<i>Group</i> was not an odd number.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_NOT_FOUND

See Also

The private block identified by *PrivateCode* was not in *Group*.

MC_Add_Standard_Attribute MC_Add_Private_Attribute MC_Add_Private_Block MC_Add_Nonstandard_Attribute

MC_Delete_Attribute MC_Delete_Range MC_Delete_Private_Block

MC_Delete_Private_Block

Removes a private block descriptor and its attributes from a message object.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Delete Private Block (
       int MsqFileItemID,
       char *PrivateCode,
       unsigned short Group
)
 MsgFileItemI
                The identifier assigned to this object by the
                MC_Open_Message, MC_Open_Empty_Message,
 D
                MC_Create_File, MC_Create_Empty_File, or
                MC_Open_Item functions.
 PrivateCode
                The code string which identifies which block in the private
                Group is to be deleted.
                The number identifying the private group. It must be an odd
 Group
                number.
```

Remarks

MC_Delete_Private_Block removes a block of private attributes from an existing message. A block of private attributes are all attributes in the private group having the same specified *PrivateCode*. The private block's identification attribute is also removed from the message.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>PrivateCode</i> was NULL.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_NOT_FOUND	The private block identified by <i>PrivateCode</i> was not in <i>Group</i>

MC_Add_Standard_Attribute MC_Add_Private_Attribute MC_Add_Private_Block MC_Add_Nonstandard_Attribute MC_Delete_Attribute MC_Delete_Private_Attribute MC_Delete_Range

MC_Delete_Range

Removes a range of attributes from a message object.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Delete Range (
       int MsgFileItemID,
       unsigned long FirstTag,
       unsigned long LastTag
)
 MsgFileIteml
                The identifier assigned to this object by the
 D
                MC_Open_Message, MC_Open_Empty_Message,
                MC_Create_File, MC_Create_Empty_File, or
                MC_Open_Item functions.
                Identifier of the first attribute which is to be removed from the
 FirstTag
                message object.
                Identifier of the last attribute which is to be removed from the
 LastTag
                message object.
```

Remarks

MC_Delete_Range removes a range of attributes (standard or nonstandard) from an existing message. Removing these attributes may make the message non-conformant to DICOM since all necessary attributes are determined by the service and command specified by **MC_Open_Message**. All attributes with tags equal to or greater than *FirstTag* and equal to or less than *LastTag* will be removed. *FirstTag* and *LastTag* need not exist in the message.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
See Also	
MC_Add_Standard_Attribute MC_Add_Private_Attribute	MC_Delete_Attribute MC_Delete_Private_Attribute

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Add_Private_Block MC_Add_Nonstandard_Attribute

MC_Delete_Private_Block

MC_Dir_Add_Entity

Adds a lower level directory entity to a specified DICOMDIR directory record and creates the first directory record in the new directory entity

Synopsis

#include "mc3media.h"
MC_STATUS MC_Dir_Add_Entity (
 int DirID,
 int ItemID,
 char *NewItemName,
 int *NewEntityID,
 int *NewItemID

)

DirlD	The identifier assigned to this DICOMDIR object by the MC_Create_File or MC_Create_Empty_File functions.
ItemID	The identifier assigned to the record to which the lower level entity is to be added.
NewltemNam e	String name of the item to be associated with this item object.
NewEntityID	Upon successful completion, the entity object identifier is returned here.
NewItemID	Upon successful completion, the item object identifier of the first directory record in the new entity object is returned here.

Remarks

MC_Dir_Add_Entity creates a new lower level directory entity referenced by the item object *Item/D*, and places it in the DICOMDIR object identified by *Dir/D*. **MC_Dir_Add_Entity** creates the first directory record, of type *NewItemName*, within the new directory entity.

The new directory record is placed in *DirlD*'s directory record sequence and all internal links to the new entity are adjusted by Merge DICOM Toolkit.

NewItemName is used to access configuration information which describes the attributes of the new directory record. If such configuration information is not available, an empty item object is created, and a warning message is logged.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.

O Copyright Merge Healthcare Solutions Inc. 2023

MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid directory record object ID.
MC_NULL_POINTER_PARM	The <i>NewEntityID</i> , <i>NewItemID</i> , or <i>NewItemName</i> parameter was NULL.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file for the directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_LOWER_DIR_RECORD	The directory record type specified by <i>NewItemName</i> is an invalid lower level type for the directory record <i>ItemID</i> .
MC_BAD_DIR_RECORD_TYPE	The directory record type specified in the directory record <i>ItemID</i> is invalid.
See Also	
MC_Dir_Add_Record	MC_Dir_Delete_Referenced_Entity MC_Dir_First_Record MC_Dir_Next_Entity MC_Dir_Next_Record MC_Dir_Root_Entity

MC_Dir_Add_Record

Adds a DICOMDIR directory record to a specified directory entity

Synopsis

```
#include ``mc3media.h"
MC_STATUS MC_Dir_Add_Record (
       int DirID,
       int EntityID,
       char *NewItemName,
       int *NewItemID
)
 DirlD
                 The identifier assigned to this DICOMDIR object by the
                 MC_Create_File or MC_Create_Empty_File functions.
                 The identifier of the directory entity object in which the
 EntityID
                 directory record is being added.
                 String name of the item to be associated with this item object.
 NewItemNam
                 See mergecom.srv item table for list of names.
 е
 NewItemID
                 Upon successful completion, the item object identifier of the
                 new directory record is returned here.
```

Remarks

MC_Dir_Add_Record creates a new directory record of type *NewItemName* and appends it to the end of the entity object specified by *EntityID*. A pointer to the new directory record is returned in *NewItemID*.

The *NewItemName* parameter is used to access configuration information which describes the attributes of the new directory record. If such configuration information is not available, an empty item object is created, and a warning message is logged.

The new directory record is placed in *DirID*'s directory record sequence attribute and all internal links are adjusted by Merge DICOM Toolkit to reflect the new record.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ENTITY_ID	The <i>EntityID</i> value is not a valid entity object ID.
MC_NULL_POINTER_PARM	The <i>NewItemID</i> or <i>NewItemName</i> parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	Merge DICOM Toolkit found a corrupted record ID in a directory record offset attribute while trying to traverse the DICOMDIR object pointed to by <i>DirID</i> .
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file for the directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_LOWER_DIR_RECORD	The directory record type specified by <i>NewItemName</i> is an invalid lower level type for the parent directory record of <i>EntityID</i> .
MC_BAD_DIR_RECORD_TYPE	The directory record type specified in the parent directory record of <i>EntityID</i> is invalid.
See Also	
MC_Dir_Add_Entity	MC_Dir_Delete_Record MC_Dir_Delete_Referenced_Entity MC_Dir_First_Record MC_Dir_Next_Entity MC_Dir_Next_Record MC_Dir_Root_Entity

MC_Dir_Delete_Record

Deletes a DICOMDIR directory record and any directory entities it references

Synopsis

#include ``mc3media.h"

```
MC_STATUS MC_Dir_Delete_Record (
    int DirID,
    int ItemID
)
```

DirlD	The identifier assigned to this DICOMDIR object by the
	MC_Create_File or MC_Create_Empty_File functions.
ItemID	The identifier assigned to the record to be deleted.

MC_Dir_Delete_Record deletes the directory record pointed to by *Item/D*. The resources allocated to the item are automatically freed. All lower level directory entities referenced by that item are also freed.

If the record is the only entry in a directory entity, the directory entity object will also be deleted. If the deleted record is the last one to point to an MRDR the MRDR will be deleted. The resources allocated to the directory entities and records are automatically freed.

All internal links will be adjusted within the affected directory records by Merge DICOM Toolkit.

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid record object ID.
MC_NULL_POINTER_PARM	The <i>DirlD</i> parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	Merge DICOM Toolkit found a corrupted record ID in a directory record offset attribute while trying to traverse the DICOMDIR object pointed to by <i>DirID</i> .
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Dir_Add_Entity	MC_Dir_Add_Record MC_Dir_Delete_Referenced_Entity MC_Dir_First_Record MC_Dir_Next_Entity MC_Dir_Next_Record

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

MC_Dir_Delete_Referenced_Entity

Deletes the lower level directory entity of a specified DICOMDIR directory record

Synopsis

DirlD	The identifier assigned to this DICOMDIR object by the MC_Create_File or MC_Create_Empty_File functions.
ltemID	The identifier assigned to the directory record whose next lower level directory entity is being deleted.

MC_Dir_Delete_Referenced_Entity deletes the directory record which is referenced by ItemID. If the deleted record is the last one to point to a multiple reference directory record (MRDR) the MRDR will be deleted.

If there was no referenced entity for *ItemID*, **MC_Dir_Delete_Referenced_Entity** will return **MC_NORMAL_COMPLETION**.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid record object ID.
MC_INVALID_ENTITY_ID	The <i>EntityID</i> value is not a valid entity object ID.
MC_NULL_POINTER_PARM	The <i>DirID</i> parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	Merge DICOM Toolkit found a corrupted record ID in a directory record offset attribute while trying to traverse the DICOMDIR object pointed to by <i>DirID</i> .
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Dir_Add_Entity	MC_Dir_Add_Record MC_Dir_Delete_Record MC_Dir_First_Record MC_Dir_Next_Entity MC_Dir_Next_Record MC_Dir_Root_Entity

MC_Dir_Entity_Count

Returns a count of the number of items contained within a particular DICOMDIR entity.

Synopsis

#include ``mc3media.h"

```
MC_STATUS MC_Dir_Entity_Count (
    int DirID,
    int EntityID,
    int *Count
)
```

```
© Copyright Merge Healthcare Solutions Inc. 2023
```

DirlD	The identifier assigned to this DICOMDIR object by the
	MC_Create_File or MC_Create_Empty_File functions.
EntityID	The identifier assigned to the DICOMDIR entity which is going to be counted.
Count	Upon successful completion, the number of items contained within a particular entity is returned here.

The count returned by **MC_Dir_Entity_Count** is maintained internally whenever items and entities are added or deleted from a DICOMDIR. Because this count is maintained internally, there is no performance penalty associated with this function call. The entity referenced by *EntityID* can be obtained by a call to **MC_Dir_Add_Entity**, **MC_Dir_Next_Entity**, **MC_Dir_First_Record**, **MC_Dir_Next_Record**, or **MC_Dir_Root_Entity**.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirlD</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid directory record object ID.
MC_INVALID_ENTITY_ID	The <i>EntityID</i> value is not a valid DICOMDIR entity object ID.
MC_UNABLE_TO_GET_ITEM_ID	The <i>DirlD</i> value does not contain a reference to a valid root item.
MC_NULL_POINTER_PARM	The <i>Count</i> parameter was NULL.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file for the directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Dir_Root_Count	MC_Dir_Item_Count

MC_Dir_First_Record

Retrieve a pointer to the first directory record in a DICOMDIR directory entity

Synopsis

#include ``mc3media.h"

MC_STATUS MC_Dir_First_Record (
 int DirID,
 int EntityID,
 int *ItemID,

[©] Copyright Merge Healthcare Solutions Inc. 2023

char **ItemName,	
int *Is	Last
)	
DirlD	The identifier assigned to this DICOMDIR object by the
	MC_Create_File or MC_Create_Empty_File functions.
EntityID	The identifier assigned to this directory entity object.
ltemID	Upon successful completion, the item object identifier of the first directory record in the directory entity <i>EntityID</i> is returned here.
ltemName	Upon successful completion, a pointer to the string name of the item associated with this item object is returned here.
lsLast	Upon successful completion, this parameter is set to true (non-zero) if the first record is also the last record in the directory entity (i.e., it is the only element)

MC_Dir_First_Record retrieves the identifier of the first directory record in the directory entity *EntityID*.

The item type for *Item/D* is also returned in the parameter *Item/Name*. This item type can be used to create another record with **MC_Dir_Add_Record** or a new record in a new entity with **MC_Dir_Add_Entity**.

If *EntityID* is an empty root directory entity, **IsLast* is set to non-zero, *ItemID* is set to 0 and *ItemName* is set to NULL.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ENTITY_ID	The <i>EntityID</i> value is not a valid entity object ID.
MC_NULL_POINTER_PARM	The <i>ItemID</i> , <i>ItemName</i> , or <i>IsLast</i> parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	Merge DICOM Toolkit found a corrupted record ID in a directory record offset attribute while trying to traverse the DICOMDIR object pointed to by <i>DirID</i> .
MC_EMPTY_ROOT_ENTITY	The entity requested is empty.
See Also	
MC_Dir_Next_Entity	MC_Dir_Next_Record MC_Dir_Root_Entity

MC_Dir_Item_Count

Returns a count of the number of items contained "below" a particular DICOMDIR item, including the starting item.

Synopsis

```
#include ``mc3media.h"
MC_STATUS MC_Dir_Item_Count (
    int DirID,
    int ItemID,
    int *Count
)
DirID The identifier assigned to this DICOMDIR object by the
    MC_Create_File or MC_Create_Empty_File functions.
ItemID The identifier assigned to the DICOMDIR item the counting
    process is starting from.
Count Upon successful completion, the number of items contained
    "below" a particular item is returned here.
```

Remarks

The count returned by **MC_Dir_Item_Count** is computed during this function call. Because this count is generated on the fly, there is a performance penalty associated with this function call. During this function call, each entity referenced by the given item is traversed, and the total number of items is generated.

The item referenced by *Item/D* can be obtained by a call to **MC_Dir_Add_Record**, **MC_Dir_First_Record** or **MC_Dir_Next_Record**.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirlD</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid directory record object ID.
MC_INVALID_ENTITY_ID	The <i>EntityID</i> value is not a valid DICOMDIR entity object ID.
MC_UNABLE_TO_GET_ITEM_ID	The <i>DirlD</i> value does not contain a reference to a valid root item.
MC_NULL_POINTER_PARM	The <i>Count</i> parameter was NULL.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file for the directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

```
MC_Dir_Root_Count
```

MC_Dir_Entity_Count

MC_Dir_Next_Entity

Retrieves a pointer to the next lower directory entity linked to a specified DICOMDIR directory record

Synopsis

```
#include ``mc3media.h"
MC STATUS MC Dir Next Entity (
        int DirID,
       int ItemID,
       int *NextEntityID,
       int *NextItemID,
       char **NextItemName,
       int *IsLast
)
                  The identifier assigned to this DICOMDIR object by the
 DirlD
                  MC_Create_File or MC_Create_Empty_File functions.
 ItemID
                  The item object identifier of the directory record to retrieve
                  the next lower directory entity from.
 NextEntityID
                  Upon successful completion, the entity object identifier of
                  the next lower directory entity referenced by ItemID is
                  returned here.
 NextItemID
                  Upon successful completion, the item object identifier of the
                  first directory record in the directory entity NextEntityID is
                  returned here.
                  Upon successful completion, a pointer to the string name of
 NextItemName
                  the item associated with the item object NextItemID is
                  returned here.
 lsLast
                  Upon successful completion, this parameter is set to true
                  (non-zero) if the first record in the directory entity
                   NextEntityID is also the last record (i.e., it is the only element)
```

Remarks

MC_Dir_Next_Entity returns *NextEntity/D*, the entity object identifier of the next lower directory entity linked to *Item/D*, and the identifier *NextItem/D* and type *NextItemName* of the first directory record of *NextEntity/D*.

The item type for *NextItemID* is also returned in *NextItemName*. This item type can be used to create another record with **MC_Dir_Add_Record** or a new record in a new entity with **MC_Dir_Add_Entity**

If *ItemID* does not contain a referenced entity, *NextEntityID* and *NextItemID* are set to zero, and ItemName is set to NULL.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid item object ID.
MC_NULL_POINTER_PARM	The <i>NextEntityID</i> , <i>NextItemID</i> , or <i>NextItemName</i> parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	Merge DICOM Toolkit found a corrupted record ID in a directory record offset attribute while trying to traverse the DICOMDIR object pointed to by <i>DirID</i> .
See Also	
MC_Dir_First_Record	MC_Dir_Next_Record MC_Dir_Root_Entity

MC_Dir_Next_Record

Retrieves the next directory record in a directory entity

Synopsis

```
#include ``mc3media.h"
MC STATUS MC Dir Next Record (
       int DirID,
       int EntityID,
       int *ItemID,
       char **ItemName,
       int *IsLast
)
 DirlD
                 The identifier assigned to this DICOMDIR object by the
                 MC_Create_File or MC_Create_Empty_File functions.
 EntityID
                 The identifier assigned to this directory entity object.
                 Upon successful completion, the item object identifier of the
 ItemID
                 directory record is returned here.
 ItemName
                 Upon successful completion, a pointer to the string name of
                 the item associated with the item object Item/D is returned
                 here.
 lsLast
                 Upon successful completion, this parameter is set to true
                 (non-zero) if the retrieved directory record is the last record
                 in EntityID.
```

MC_Dir_Next_Record retrieves the next directory record in the directory entity specified by *EntityID*. If no records have been fetched from this entity, **MC_Dir_Next_Record** behaves the same as **MC_Dir_First_Record**. If one or more records have been fetched from this entity, the next record will be returned.

IsLast is nonzero if the retrieved directory record is the last record in *Entity/D*. If **MC_Dir_First_Record** is called after the last record was retrieved, **IsLast* will be non-zero, **ItemID* will be set to 0, and **ItemName* will be NULL. If *Entity/D* is an empty root directory entity, **IsLast* will be non-zero, **ItemID* will be set to 0, and **ItemName* will be NULL.

The first call to **MC_Dir_Next_Record** after **MC_Dir_Root_Entity** or **MC_Dir_Next_Entity** was called for *Entity/D* will return the second record in *Entity/D*.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid item or record object ID.
MC_NULL_POINTER_PARM	The ItemID, ItemName, or isLast parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	Merge DICOM Toolkit found a corrupted record ID in a directory record offset attribute while trying to traverse the DICOMDIR object pointed to by <i>DirID</i> .
See Also	
MC_Dir_First_Record	MC_Dir_Next_Entity MC_Dir_Root_Entity

MC_Dir_Open_MRDR

Creates a DICOMDIR multiple reference directory record (MRDR) and references it with a directory record

Synopsis

```
#include ``mc3media.h"
```

```
MC_STATUS MC_Dir_Open_MRDR (
    int DirID,
    int ItemID,
    char *MrdrName,
    int *MrdrID
)
```

```
DirIDThe identifier assigned to this DICOMDIR object by theMC_Create_File or MC_Create_Empty_File functions.
```

ltemID	The identifier of the directory record that will reference the MRDR.
MrdrName	The string name of the item to be associated with the MRDR.
MrdrID	Upon successful completion, the item object identifier of the
	MRDR created will be returned here.

MC_Dir_Open_MRDR creates a new MRDR and associates it with the directory record specified by *ItemID*. The use count attribute of the MRDR is initialized and the internal links are adjusted by Merge DICOM Toolkit. To remain DICOM conformant, the parameter *MrdrName* should always be the default data dictionary name for an MRDR item.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirlD</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid item or record object ID.
MC_NULL_POINTER_PARM	The MrdrID or MrdrName parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	Merge DICOM Toolkit found a corrupted record ID in a directory record offset attribute while trying to traverse the DICOMDIR object pointed to by <i>DirID</i> .
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Dir_Reference_MRDR	MC_Dir_Remove_Ref_MRDR

MC_Dir_Reference_MRDR

Causes a DICOMDIR directory record to reference a multiple reference directory record (MRDR)

Synopsis

```
#include ``mc3media.h"
```

```
MC_STATUS MC_Dir_Reference_MRDR (
    int DirID,
    int ItemID,
    int MrdrID
)
```

DirIDThe identifier assigned to this DICOMDIR object by theMC_Create_File or MC_Create_Empty_File functions.

ItemID	The item object identifier of the directory record that will
	reference the MRDR.
MrdrID	The identifier of the MRDR that <i>ItemID</i> will reference.

MC_Dir_Reference_MRDR references an MRDR with a standard directory record. Merge DICOM Toolkit adjusts the use count attribute of the MRDR and the internal links within the affected directory records. The user is responsible for filling the other attributes within the MRDR and the directory record *ItemID*.

If *ItemID* already references another MRDR, **MC_Dir_Reference_MRDR** deletes that reference.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid item object ID.
MC_INVALID_MRDR_ID	The <i>MrdrID</i> value is not a valid MRDR object ID.
MC_UNABLE_TO_GET_ITEM_ID	While trying to traverse the DICOMDIR object pointed to by <i>DirID</i> , Merge DICOM Toolkit found an invalid item ID contained in a directory record offset attribute.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Dir_Open_MRDR	MC_Dir_Remove_Ref_MRDR

MC_Dir_Remove_Ref_MRDR

Removes a reference from a DICOMDIR directory record to a multiple reference directory record (MRDR)

Synopsis

```
#include ``mc3media.h"
MC_STATUS MC_Dir_Remove_Ref_MRDR (
    int DirID,
    int ItemID
)
DirID The identifier assigned to this DICOMDIR object by the
    MC_Create_File or MC_Create_Empty_File functions.
ItemID The item object identifier of the directory record that the
    reference to an MRDR will be removed from.
```

MC_Dir_Remove_Ref_MRDR removes the reference from a directory record to an MRDR. Merge DICOM Toolkit decrements the use count attribute of the MRDR. If the use count equals zero, the MRDR will be automatically freed.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirlD</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid item or record object ID.
MC_UNABLE_TO_GET_ITEM_ID	While trying to traverse the DICOMDIR object pointed to by <i>DirID</i> , Merge DICOM Toolkit found an invalid item ID contained in a directory record offset attribute.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Dir_Open_MRDR	MC_Dir_Reference_MRDR

MC_Dir_Root_Count

Returns the total number of items contained within a DICOMDIR.

Synopsis

```
#include "mc3media.h"
MC_STATUS MC_Dir_Root_Count (
    int DirID,
    int *Count
)
DirID The identifier assigned to this DICOMDIR object by the
    MC_Create_File or MC_Create_Empty_File functions.
Count Upon successful completion, the total number of items
    contained with a DICOMDIR is returned here.
```

Remarks

The count returned by **MC_Dir_Root_Count** is maintained internally whenever items and entities are added or deleted from a DICOMDIR. Because this count is maintained internally, there is no performance penalty associated with this function call.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

[©] Copyright Merge Healthcare Solutions Inc. 2023

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirlD</i> value is not a valid DICOMDIR object ID.
MC_UNABLE_TO_GET_ITEM_ID	The <i>DirlD</i> value does not contain a reference to a valid root item.
MC_NULL_POINTER_PARM	The <i>Count</i> parameter was NULL.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file for the directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Dir_Entity_Count	MC_Dir_Item_Count

MC_Dir_Root_Entity

Retrieves a pointer to the root entity of a specified DICOMDIR

Synopsis

```
#include ``mc3media.h"
MC_STATUS MC_Dir_Root_Entity (
       int DirID,
       int *RootEntityID,
       int *ItemID,
       char **ItemType,
       int *IsLast
)
 DirlD
                 The identifier assigned to this DICOMDIR object by the
                 MC_Create_File or MC_Create_Empty_File functions.
                 Upon successful completion, the root entity object identifier
 RootEntityID
                 is returned here.
 ItemID
                 Upon successful completion, the item object identifier of the
                 first directory record in the root entity is returned here.
 ItemName
                 Upon successful completion, a pointer to the string name of
                 the item associated with the item object Item/D is returned
                 here.
 lsLast
                 Upon successful completion, this parameter is set to true
                 (non-zero) if the first record in the directory entity
                  RootEntityID is also the last record (i.e., it is the only element)
```

Remarks

MC_Dir_Root_Entity returns the identifier of the root entity of the DICOMDIR. It also returns the identifier, *Item/D*, and type, *ItemName*, of the first directory record in *RootEntity/D*.

The item type returned in *ItemName* can be used to create another record with **MC_Dir_Add_Record** or a new record in a new entity with **MC_Dir_Add_Entity** If the root entity is empty, *Item/D* is set to zero, and *ItemName* is set to NULL.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirID</i> value is not a valid DICOMDIR object ID.
MC_NULL_POINTER_PARM	The <i>RootEntityID</i> , ItemID, or ItemName parameter was NULL.
MC_UNABLE_TO_GET_ITEM_ID	While trying to traverse the DICOMDIR object pointed to by <i>DirlD</i> , Merge DICOM Toolkit found an invalid item ID contained in a directory record offset attribute.
See Also	
MC_Dir_First_Record	MC_Dir_Next_Entity MC_Dir_Next_Record

MC_Dir_Sort

Sorts a DICOMDIR. See remarks section for specific ordering.

Synopsis

DirlD

The ID number of a DICOMDIR to sort

Remarks

Each DICOMDIR Entity will first sort by record type in the following order:

PATIENT STUDY SERIES IMAGE OVERLAY MODALITY LUT VOI LUT CURVE TOPIC VISIT RESULTS INTERPRETATION STUDY COMPONENT STORED PRINT RT DOSE RT STRUCTURE SET RT PLAN RT TREAT RECORD PRESENTATION WAVEFORM SR DOCUMENT PRIVATE MRDR

Within each record type the record entries shall be sorted in the in the following order:

PATIENT

Sorted first by Patient Name (0010,0010) ascending alphabetically if present

If Patient Name not present these records shall be placed at the end of the list.

Second the records within each unique patient name shall be sorted by the Patient ID.

STUDY

Sorted first by Study Date ascending

Sorted second by Study Time ascending within each Study Date

SERIES

Sorted first by Modality ascending alphabetically.

Sorted within each unique Modality by series number.

IMAGE

Sorted by instance number.

OVERLAY

Sorted by overlay number.

MODALITY LUT

Sorted by Lookup table number.

VOI LUT

Sorted by lookup table number.

CURVE

Sorted by curve number.

TOPIC

© Copyright Merge Healthcare Solutions Inc. 2023

Sorted by Topic Title

VISIT

Sorted by Admitting Date if present. Those records without an admitting date shall be grouped as entered by the application as the end of the list.

RESULTS

Instance Creation Date if present. Those records without an instance creation date will be placed at the end of the list in the order as supplied by the application.

INTERPRETATION

Sorted by the Interpretation transcription date.

STUDY COMPONENT Shall be sorted by Modality

STORED PRINT Sorted by Instance Number if present.

RT DOSE

Sorted by Instance Number

RT STRUCTURE SET Sorted by Instance Number

RT PLAN

Sorted by Instance Number

RT TREAT RECORD Sorted by Instance Number

PRESENTATION

First sorted by Presentation Creation Date ascending

Second sorted by Presentation Creation Time ascending within each unique date

WAVEFORM

Sorted first sorted by Content Date ascending.

Second sorted by Content Time ascending within each unique Content Date.

SR DOCUMENT

Sorted first sorted by Content Date ascending.

Second sorted by Content Time ascending within each unique Content Date.

Sorted third by Instance number.

PRIVATE Not Sorted

MRDR

Not Sorted.

Calls to the library to retrieve records will receive them in sorted order after calling this function. Also, when the DICOMDIR is written to a file, the records will be streamed sorted, depth first. The sort is not maintained if new items are added.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_INVALID_DICOMDIR_ID	DirID value is not a valid DICOMDIR object ID.
MC_INVALID_LOWER_DIR_RECORD	While trying to traverse the DICOMDIR object pointed to by <i>DirID</i> , Merge DICOM Toolkit found an invalid item ID contained in a lower-level directory entity attribute.
MC_UNABLE_TO_GET_ITEM_ID	While trying to traverse the DICOMDIR object pointed to by <i>DirID</i> , Merge DICOM Toolkit found an invalid item ID contained in a directory record offset attribute.
MC_VALUE_OUT_OF_RANGE	While trying to traverse the DICOMDIR object pointed to by <i>DirID</i> , Merge DICOM Toolkit found an item ID that contained an attribute with an invalid value.
MC_SYSTEM_ERROR	An unrecoverable error occurred (i.e. no memory available, no disk space, etc). A log has been written to the log file describing the problem.

MC_Dir_Stream_Directory_Record

Stores a DICOMDIR directory record internally in stream format to reduce memory usage.

Synopsis
<pre>#include ``mc3media.h"</pre>
MC_STATUS MC_Dir_Stream_Directory_Record (
int DirID,
<pre>int ItemID</pre>
)
<i>DirlD</i> The identifier assigned to this DICOMDIR object by the MC_Create_File or MC_Create_Empty_File functions.

ItemID The item object identifier of the directory record whose storage method will be changed.

Remarks

MC_Dir_Stream_Directory_Record changes the internal method for storing a directory record. When called, Merge DICOM Toolkit changes the storage method for the specified directory record to a DICOM byte stream instead of Merge DICOM Toolkit's internal structure. Storing as a DICOM byte stream reduces the memory required by Merge DICOM Toolkit. This can be especially useful when creating and working with large DICOMDIRs.

When a directory record is stored as a DICOM byte stream, Merge DICOM Toolkit will automatically change it back to its internal format if a tag is accessed within the directory record.

When reading a DICOMDIR from disk, the **DICOMDIR_STREAM_STORAGE** configuration option can also be used to change how directory records are stored. When enabled, this option will force all directory records in a DICOMDIR to be stored as a DICOM byte stream.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>DirlD</i> value is not a valid DICOMDIR object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid item or record object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_Duplicate_Message

Creates a message identical to the source message. If the transfer syntax on the destination differs from the source, the new message will be changed to the new transfer syntax. This includes changing from one compression transfer syntax to another. This function will also register the compressor/decompressor with the new message.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Duplicate_Message (
    int SourceMsgID,
    int *DestMsgID,
    TRANSFER_SYNTAX DestMsgTransferSyntax,
    MC_STATUS (*YourDestMsgCompressionCallback),
    MC_STATUS (*YourDestMsgDecompressionCallback)
)
```

The identifier assigned to the source message by the MC_Open_Message or MC_Open_Empty_Message function.
Upon successful completion, the message object identifier is returned here.
The DICOM transfer syntax to be used for the destination message. One of the enumerated TRANSFER_SYNTAX types defined in "mc3msg.h".
Name of a function to be registered as the Compression callback function for the destination message. To use the built in compressor, set this to MC_Standard_Compressor . To use the build in RLE compressor, set this to MC_RLE_Compressor .
Name of a function to be registered as the Decompression callback function for the destination message. To use the built in decompressor, set this to MC_Standard_Decompressor . To use the built in RLE decompressor, set this to MC_RLE_Decompressor .

The functions must be prototyped as follows:

int CBisFirst,
int CBisLast
int CBrelease

```
MC_STATUS YourDestMsg(De)CompressionCallback (
    int CBmsgID,
    void** CBcontext,
    unsigned long CBdataLength,
    void* CBdataValue,
    unsigned long* CBoutdataLength,
    void** CBoutdataValue,
```

)

CBmsglD	The identifier assigned to this message object by the MC_Open_Message , MC_Open_Empty_Message functions.
CBcontext	A data structure that contains user data and has to be preserved between compression calls. The first call to the callback function should initialize this structure.
CBdataLength	The length of the incoming data pointed to by CBdataValue.
CBdataValue	Pointer to incoming data.

CBoutdataLengt h	The length of the outgoing data pointed to by CBoutdataValue.
CBoutdataValue	Pointer to the outgoing data.
CBisFirst	This is TRUE (not zero) the first time <i>YourDestMsg(De)CompressionCallback</i> is being called.
CBisLast	This is TRUE (not zero) the last time <i>YourDestMsg(De)CompressionCallback</i> is being called.
CBrelease	This is TRUE (not zero) if the callback should release all the context memory and return.

Remarks

Registering **MC_Standard_Compressor** and/or **MC_Standard_Decompressor** for the destination message will use the standard compressor/decompressor built into the library. You may also register your own (de)compression callbacks for the destination message. The source message can have a decompressor/compressor registered prior to duplication by calling **MC_Register_Compression_Callbacks** for the source message.

The **MC_Duplicate_Message** function is used to duplicate a message, however the transfer syntax may be different on the destination message. The following syntaxes may be switched when using the built in compressor/decompressor **MC_Standard_Compressor/MC_Standard_Decompressor**:

IMPLICIT_LITTLE_ENDIAN IMPLICIT_BIG_ENDIAN EXPLICIT_LITTLE_ENDIAN EXPLICIT_BIG_ENDIAN DEFLATED_EXPLICIT_LITTLE_ENDIAN ENCAPSULATED_UNCOMPRESSED_ELE JPEG_BASELINE JPEG_EXTENDED_2_4 JPEG_LOSSLESS_HIER_14 JPEG_2000 JPEG_2000_LOSSLESS_ONLY

NOTE:Only the transfer syntax may be different when using the built in
compressor/decompressorMC_Standard_Compressor/MC_Standard_Decompressor.

Registering **MC_RLE_Compressor** and/or **MC_RLE_Decompressor** will use Merge DICOM Toolkit's build in RLE compressor and decompressor.

NOTE: The built in compressor/decompressor MC_Standard_Compressor/MC_Standard_Decompressor is only available on platforms supported by the Pegasus libraries from Accusoft (Windows, Solaris, and Linux). The build in compressor/decompressor MC_RLE_Compressor/MC_RLE_Decompressor is supported on all Merge DICOM Toolkit platforms.

Performance Tuning

The destination message should be freed via **MC_Free_Message** when it is no longer needed.

If an encapsulated transfer syntax is being duplicated into a message with the same encapsulated transfer syntax, it is possible to bypass the compressor/decompressor by ensuring that ALL of the following cases are true:

- 1. The source message does not have a decompressor registered
- 2. A decompressor is not being passed in for the destination message.
- 3. The transfer syntaxes are identical

YourDestinationMsg(De)CompressionCallback

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

CBisFirst is set to TRUE the first time *YourDesMsg(De)CompressionCallback* is called. This provides a clear mechanism for the function to know it is being called the first time.

CBisLast is set to TRUE the last time *YourDestMsg(De)CompressionCallback* is called. This provides a clear mechanism for the function to know it is being called the last time.

YourDestMsg(De)CompressionCallback must return MC_NORMAL_COMPLETION if all went well. If not, it must return MC_CANNOT_COMPLY. Both of these are defined in "mc3msg.h".

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_COMPRESSOR_REQUIRED	A compression function must be passed in for the destination message.
MC_DECOMPRESSOR_REQUIRED	A decompression function must be registered with the source message.
MC_INVALID_MESSAGE_ID	<i>SourceMsgID</i> is not a valid message object ID.
MC_CALLBACK_CANNOT_COMPLY	<i>YourDestMsg(De)Compression</i> <i>Callback</i> returned with MC_CANNOT_COMPLY, or the callback function registered for this attribute, and providing data, returned MC_CANNOT_COMPLY.

See Also

MC_Free_Message MC_Register_Compression_Callbacks

MC_Empty_File

Sets all the attributes in a DICOM file to empty

Synopsis	
<pre>#include ``</pre>	mc3media.h"
—	MC_Empty_File (<i>FileID</i>
FileID	The identifier assigned to this item object by the MC_Create_File function.

Remarks

MC_Empty_File sets the value of each attribute in the file identified by *FileID* to empty" (i.e. value length of zero). This is equivalent to calling **MC_Set_Value_To_Empty** for each attribute in an item. This function is useful for client applications which reuse a file and want to ensure there are no attribute values remaining from the last instance of the file.

NOTE: If the file contains an attribute with a value representation of SQ, setting the file values to empty causes the SQ attribute's sequence item objects to be freed.

The DICOM prefix for the file is set to "DICM". All bytes in the file preamble are set to 00H.

When the file object identified by *FileID* was a DICOMDIR object, after each attribute is set to empty, an empty root directory entity is created in the object and the root directory record offset attributes are initialized. If the file object was not a DICOMDIR object, but it is to be used as a DICOMDIR object, **MC_Set_Service_Command** should be called for the file object before the **MC_Dir_...** functions are called for the object.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
See Also	
MC_Empty_Message	MC_Empty_Item MC_Set_Value_To_Empty MC_Set_pValue_To_Empty

MC_Empty_Item

Sets all the attributes in an item to empty.

Synopsis	
<pre>#include ``mc3msg.h"</pre>	
—	C_Empty_Item (temID
ltemID	The identifier assigned to this item object by the MC_Open_Item function.
Domorko	

Remarks

MC_Empty_Item sets the value of each attribute in the item identified by *ItemID* to "empty" (i.e. value length of zero). This is equivalent to calling **MC_Set_Value_To_Empty** for each attribute in an item. This function is useful for client applications which reuse an item and want to ensure there are no attribute values remaining from the last instance of the item.

NOTE:	If the item contains an attribute with a value representation of SQ,
	setting the item values to empty causes the SQ attribute's sequence
	item objects to be freed (i.e. MC_Free_Item is automatically called for
	any items used by the item being emptied).

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid item object ID.
See Also	
MC_Empty_Message MC_Set_Value_To_Empty MC_Set_pValue_To_Empty	

MC_Empty_Message

Sets all the attributes in a message to empty.

Synopsis

[©] Copyright Merge Healthcare Solutions Inc. 2023

MessageIDThe identifier assigned to this object by theMC_Open_Message function or the MC_Open_Item function.

Remarks

MC_Empty_Message sets the value of each attribute in the message identified by *MessageID* to "empty" (i.e. value length of zero). This is equivalent to calling **MC_Set_Value_To_Empty** for each attribute in a message. This function is useful for client applications which reuse a message and want to ensure there are no attribute values remaining from the last instance of the message.

NOTE: If the message contains an attribute with a value representation of SQ, setting the message values to empty causes the SQ attribute's sequence item objects to be freed (i.e. **MC_Free_Item** is automatically called for any items used by the message being emptied).

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
See Also	
MC_Empty_Item MC_Set_Value_To_Empty MC_Set_pValue_To_Empty	

MC_Enable_Unicode_Conversion

Enable/disable Unicode conversion functions.

Synopsis

```
#include "mc3msg.h"
MC_STATUS MC_Enable_Unicode_Conversion (
    int Enable
)
Enable Set to non-zero to load the Unicode conversion library and
    enable its conversion functions. Set to zero to unload the
    Unicode conversion library and disable its conversion
    functions.
```

Remarks

MC_Enable_Unicode_Conversion set to non-zero initializes the Unicode conversion library by loading the Unicode shared objects and its dependency files as specified in Platform Notes. When success, all Unicode conversion functions are available for use. When sets to zero, the function

unloads the Unicode conversion library shared objects, its dependency files and disable all Unicode conversion functions.

```
NOTE: If enabling Unicode conversion fails, check the location of shared objects or dependency files, or loading path as specified in the Platform Notes.
```

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION MC_CANNOT_COMPLY	The function completed normally. The Unicode library fails to initialize or load.
See Also	
MC_Byte_To_Unicode	MC_Unicode_To_Byte MC_Get_Value_To_UnicodeString

MC_Get_Next_Value_To_UnicodeString MC_Set_Value_From_UnicodeString MC_Set_Next_Value_From_UnicodeString

MC_Error_Message

Returns a string description of a Merge DICOM Toolkit status code.

Synopsis

#include ``mc3msg.h"

)

StatusCode A Merge DICOM Toolkit status code.

Remarks

It is often useful to log a descriptive message when an error status is received. **MC_Error_Message** returns a string message describing a given **MC_STATUS** code.

Return Value

Pointer to the descriptive string.

MC_File_To_Message

Converts a file object into a message object.

Synopsis

Remarks

MC_File_To_Message changes the file object pointed to by *FileID* into a message object. In the process, the DICOM File Meta Information attributes are removed from the object, and the "command type" attributes used by most DICOM services are added to the new message object. While Merge DICOM Toolkit sets the values of many of these "command type" attributes automatically, some services require the application to set them. Refer to the description of **MC_Send_Request_Message** for more information).

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
See Also	

MC_Message_To_File

MC_Free_Item

Releases a Merge DICOM Toolkit item object.

Synopsis

Remarks

The **MC_Free_Item** function frees the system resources used by the item object. The variable pointed to by *Item/D* is set to -1 (an invalid item ID).

[©] Copyright Merge Healthcare Solutions Inc. 2023

NOTE:	If the item object being freed contains an attribute with a value
	representation of SQ, freeing the item causes the SQ attribute's
	sequence item objects to be freed also (i.e. MC_Free_Item is
	automatically called for any items used by the item identified by
	<i>ItemID</i>). If any of the items thus freed also contain sequences of items,
	those item in turn are recursively freed.

NOTE: The **ELIMINATE_ITEM_REFERENCES** configuration value will determine if a freed item object that is in use as a value in another sequence of items should be cleared.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ITEM_ID	The <i>*ItemID</i> value is not a valid item object ID.
MC_NULL_POINTER_PARM	<i>ItemID</i> was NULL.
See Also	
MC_Free_File MC_Free_Message MC_Open_Item	

MC_Free_File

Releases a Merge DICOM Toolkit file object

```
FileID
```

A pointer to the identifier assigned to this object by the **MC_Create_File** or **MC_Create_Empty_File** function.

Remarks

The **MC_Free_File** function frees the system resources used by the file object. The variable pointed to by *FileID* is set to -1 (an invalid file ID).

NOTE:

If the file contains an attribute with a value representation of SQ, freeing the file causes the SQ attribute's sequence item objects to be freed also. If any of the items thus freed also contain sequences of items, those items in turn are recursively freed.

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

NOTE:

The **ELIMINATE_ITEM_REFERENCES** configuration value will determine if a freed item object that is in use as a value in another sequence of items should be cleared.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>*FileID</i> value is not a valid file object ID.
MC_NULL_POINTER_PARM	<i>FileID</i> was NULL.
See Also	

See Also

MC_Free_Item

MC_Free_Message

MC_Free_Message

Releases a Merge DICOM Toolkit message object.

Synopsis

MessageIDA pointer to The identifier assigned to this object by the
MC_Open_Message function or the MC_Open_Item function.

Remarks

The **MC_Free_Message** function frees the system resources used by the message object. The variable pointed to by *MessageID* is set to -1 (an invalid message ID);

NOTE: If the message object being freed contains an attribute with a value representation of SQ, freeing the message causes the SQ attribute's sequence item objects to be freed also (i.e. **MC_Free_Item** is automatically called for any items used by the message identified by *MessageID*).

NOTE: If any of the items thus freed also contain sequences of items, those item in turn are recursively freed.

NOTE:

The **ELIMINATE_ITEM_REFERENCES** configuration value will determine if a freed item object that is in use as a value in another sequence of items should be cleared.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	The <i>*MessageID</i> value is not a valid message object ID.
MC_NULL_POINTER_PARM	MessagelD was NULL.
• · · ·	

See Also

MC_Open_Message MC_Free_Item MC_Free_File

MC_FreeService

Releases a Merge DICOM Toolkit service that was dynamically created using MC_NewService.

Synopsis

)

ServiceName Service name assigned by the application in the MC_NewService call

Remarks

The MC_FreeService function frees the system resources used by the service objects.

NOTE: If the service being freed is already in use in a service list an error is returned. The service list must be freed prior to freeing the service.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_SERVICE_IN_USE	The service is in use in a service list
MC_INVALID_SERVICE_NAME	The service name does not represent a name currently defined.

See Also

MC_NewService MC_NewProposedServiceList MC_FreeServiceList

MC_FreeServiceList

Releases a Merge DICOM Toolkit service list.

Synopsis

#include ``mergecom.h"

```
MC_STATUS MC_FreeServiceList (
char *ServicelistName
```

)

ServiceListName Service list name assigned by the application in the MC_NewProposedServiceList call

Remarks

The **MC_FreeServiceList** function frees the system resources used by service list objects that were dynamically created using the MC_NewProposedServiceList function.

NOTE: Extreme care must be taken to assure that no association is currently using the service list.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SERVICE_LIST_NAME	The service list name does not represent a name currently defined.
Cas Alas	

See Also

MC_NewService

MC_NewProposedServiceList

MC_FreeService

MC_FreeSyntaxList

Releases a Merge DICOM Toolkit message object.

Synopsis

Remarks

The MC_FreeSyntaxList function frees the system resources used by the syntax list objects.

NOTE: If the syntax list is currently in use in a service an error is returned. The service must be freed prior to freeing the syntax list.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SYNTAX_LIST_NAME	The syntax list name does not represent a name currently defined.
MC_SYNTAX_IN_USE	The syntax list is in use by a service definition.
See Also	
MC_NewSyntaxList MC_ServiceServiceList	

MC_FreeService

MC_Get_Association_Info

Retrieves information about a given association, including the number of proposed and accepted services, the remote application title and remote host name.

Synopsis

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
AssociationID
                  The identifier assigned to this object by the
                  MC_Open_Association, MC_Open_Secure_Association,
                  MC_Process_Association_Request,
                  MC_Process_Secure_Association_Request,
                  MC_Wait_For_Association or
                  MC_Wait_For_Secure_Association function.
                  A structure of type AssocInfo you have declared that is filled
 AssociationInf
                 with information about the association. AssocInfo is defined
 0
                  in mergecom.h as:
typedef enum {
   NO USER IDENTITY = 0,
   USERNAME = 1,
   USERNAME AND PASSCODE = 2,
   KERBEROS SERVICE TICKET = 3,
   SAML ASSERTION = 4
} USER IDENTITY TYPE;
typedef struct MC Assoc Info {
   char RemoteApplicationTitle[20]; /* 16-characters max */
   char RemoteHostName[40];
                                    /* Network node name */
                                    /* the TCP socket the
   MC_SOCKET Tcp_socket;
   char RemoteIPAddress[40];
char Local ****
                                        association is using */
   char RemoteIPAddress[40]; /* Network IP address */
char LocalApplicationTitle[20]; /* 16-characters max */
   char RemoteImplementationClassUID[66]; /* 64-characters max */
   char RemoteImplementationVersion[20]; /* 16-characters max */
   unsigned long LocalMaximumPDUSize;
   unsigned long RemoteMaximumPDUSize;
   unsigned short MaxOperationsInvoked; /* Negotiated Max operations
                                           invoked by the assoc
                                           requestor */
   unsigned short MaxOperationsPerformed; /* Negotiated Max operations
                                            erformed by the assoc
                                            requestor */
   USER IDENTITY TYPE UserIdentityType; /* User Identity negotiation
                                            type as defined in DICOM
                                            Supplement 99 */
   unsigned char PositiveResponseRequested; /* Set according to if a
                                               positive response is
                                               requested when User
                                               Identity is specified */
   unsigned char PositiveResponseReceived; /* Set for association
                                              requestors if a positive
                                              response was received from
                                              the association acceptor */
} AssocInfo;
```

Remarks

MC_Get_Association_Info retrieves information about an established association identified by *AssociationID*. This association information is returned in the structure pointed to by *AssociationInfo* detailed above. Specific information regarding the accepted services is not specified in this structure and needs to be obtained through the **MC_Get_First_Acceptable_Service** and **MC_Get_Next_Acceptable_Service** calls.

This function is to be called after an MC_Open_Association, MC_Open_Association_With_Identity, MC_Open_Secure_Association, MC_Process_Association_Request, MC_Process_Secure_Association_Request, MC_Wait_For_Association or MC_Wait_For_Secure_Association call, and before any MC_Read_Message, MC_Send_Request_Message, MC_Send_Request, MC_Send_Response_Message or MC_Send_Response calls.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_NULL_POINTER_PARM	The AssociationInfo pointer had a null value.
MC_STATE_VIOLATION	The call was made after messages have been exchanged over the association.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Get_First_Acceptable_Service MC_Library_Initialization MC_Get_Listen_Socket_For_Port	MC_Get_Next_Acceptable_Service MC_Get_Listen_Socket

MC_Get_Attribute_Info

Retrieves information about a given attribute in a message.

Synopsis

```
#include "mc3msg.h"
MC_STATUS MC_Get_Attribute_Info (
    int MsgFileItemID,
    unsigned long Tag,
    MC_VR *ValueRep,
    int *NumValues
```

)

MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
Tag	A tag identifying a message attribute.
ValueRep	The attribute's value representation code is returned here.

	Possible return codes (defined in "mc3msg.h") are:	
	AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR,	
	UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL,	
	OD, OF, SQ, UNKNOWN_VR	
NumValues	The number of values assigned to the attribute is placed here. (Note: will be one (1) if the attribute's value is NULL.)	

Remarks

MC_Get_Attribute_Info retrieves information about a message attribute identified by *Tag.* The attribute's Value Representation code is returned at *ValueRep*, and the number of values stored for the attribute is returned at *NumValues. ValueRep* and *NumValues* parameters may be NULL, in which case the corresponding information is not returned. If the attribute is "empty", *NumValues* will contain zero (0). If the attribute's value is NULL, *NumValues* will contain one (1).

MC_Get_Attribute_Info sets an internal "attribute pointer" such that **MC_Get_Next_Attribute** will retrieve the attribute following this one.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
See Also	
MC_Get_First_Attribute MC_Get_Next_Attribute	MC_Get_Tags_Dict_Info MC_Get_pAttribute_Info

MC_Get_Bool_Config_Value

Used to get the value of a boolean toolkit configuration parameter at runtime.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Get_Bool_Config_Value (
    BoolParm Parm,
    int *Value
)
Parm An enumerated constant identifying the boolean
    configuration parameter to get. Parm can have any of the
    following values:
    ACCEPT_ANY_APPLICATION_TITLE
    ACCEPT_ANY_CONTEXT_NAME
    ACCEPT_ANY_HOSTNAME
```

ACCEPT_ANY_PRESENTATION_CONTEXT ACCEPT DIFFERENT IC UID ACCEPT DIFFERENT VERSION ACCEPT_MULTIPLE_PRES_CONTEXTS ACCEPT RELATED GENERAL SERVICES ACCEPT_STORAGE_SERVICE_CONTEXTS ALLOW COMMA IN DS FL FD STRINGS ALLOW EMPTY PDV LENGTH ALLOW INVALID PRIVATE ATTRIBUTES ALLOW INVALID PRIVATE CREATOR CODES ALLOW_LIBRARY_EXCEPTION_HANDLER ALLOW_OUT_OF_RANGE_BITS_JPEG_LOSSLESS ATT 00081190 USE UT VR ATT 00287FE0 USE UT VR ATT_0040E010_USE_UT_VR ATT 0074100A USE ST VR AUTO ECHO SUPPORT BLANK FILL LOG FILE CALCULATE DEFINED LENGTH FOR CB COMPRESSION ALLOW FRAGS COMPRESSION J2K LOSSY USE QUALITY COMPRESSION USE HEADER QUERY COMPRESSION_WHEN_J2K_USE_LOSSY CREATE OFFSET TABLE DEFLATE ALLOW FLUSH DICOMDIR STREAM STORAGE DUPLICATE ENCAPSULATED ICON ELIMINATE ITEM REFERENCES EMPTY PRIVATE CREATOR CODES EXPLICIT VR TO UN FOR LENGTH GT 64K EXPORT EMPTY PRIVATE CREATOR CODES EXPORT GROUP LENGTHS TO MEDIA EXPORT GROUP LENGTHS TO NETWORK EXPORT_PRIVATE_ATTRIBUTES_TO_MEDIA EXPORT PRIVATE ATTRIBUTES TO NETWORK EXPORT UN VR TO MEDIA EXPORT UN VR TO NETWORK EXPORT UNDEFINED LENGTH SQ EXPORT_UNDEFINED_LENGTH_SQ_IN_DICOMDIR FORCE OPEN EMPTY ITEM HARD_CLOSE_TCP_IP_CONNECTION INSURE EVEN UID LENGTH LIST UN ATTRIBUTES LOG FILE BACKUP MSG_FILE_ITEM_OBJ_TRACE NETWORK CAPTURE PRIVATE SYNTAX 1 ENCAPSULATED PRIVATE_SYNTAX_1_EXPLICIT_VR PRIVATE SYNTAX 1 LITTLE ENDIAN PRIVATE SYNTAX 2 ENCAPSULATED PRIVATE_SYNTAX_2_EXPLICIT_VR PRIVATE_SYNTAX_2_LITTLE_ENDIAN **REJECT INVALID VR** RELEASE SQ ITEMS REMOVE_PADDING_CHARS

REMOVE_SINGLE_TRAILING_SPACE RETURN COMMA IN DS FL FD STRINGS **REWRITE CAPTURE FILES** SEND_ECHO_PRIORITY SEND_LENGTH_TO_END SEND MSG ID RESPONSE SEND RECOGNITION CODE SEND RESPONSE PRIORITY SEND SOP CLASS UID SEND SOP INSTANCE UID TCPIP DISABLE NAGLE TOLERATE_INVALID_IN_DEFAULT_CHARSET UPDATE GROUP 0028 ON DUPLICATE USE FREE DATA CALLBACK These names are the same as those given to the parameters in the toolkit configuration files. The boolean value to which Parm is set is returned here.

Remarks

Value

The Merge DICOM Toolkit Library accesses several configuration files at startup. This call allows your application to get boolean configurable parameters specified in these files at runtime. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Value</i> was NULL.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Get_Int_Config_Value MC_Get_Log_Destination MC_Get_Long_Config_Value MC_Get_String_Config_Value	MC_Set_Int_Config_Value MC_Set_Log_Destination MC_Set_Long_Config_Value MC_Set_String_Config_Value MC_Set_Bool_Config_Value

MC_Get_Encapsulated_Value_To_Function

Retrieves the first frame of encapsulated pixel data from a message. If a decompressor is registered, then the data will be decompressed.

Synopsis

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
unsigned long Tag,
       void *UserInfo,
       void *YourGetFunction,
)
 MsgFileItemID
                     The identifier assigned to this object by the
                     MC_Open_Message, MC_Open_Empty_Message,
                     MC_Create_File, MC_Create_Empty_File, or
                     MC_Open_Item functions.
 Tag
                     A tag identifying a message attribute containing pixel
                     data.
 UserInfo
                     Address of data which will be passed on to
                      YourGetFunction each time it is called. This may be
                     NULL.
 YourGetFunction
                     Name of a function which will be called repeatedly to
                     provide blocks of data of the attribute's value
The function must be prototyped as follows:
MC STATUS YourGetFunction (
       int CBMsgFileItemID,
       unsigned long CBtag,
       void *CBuserInfo,
       int CBdataSize,
       void *CBdataBuffer,
       int CBisFirst,
       int CBisLast
)
 CBMsgFileItemID
                      The identifier assigned to this object by the
                      MC_Open_Message, MC_Open_Empty_Message,
                      MC_Create_File, MC_Create_Empty_File, or
                      MC_Open_Item functions.
 CBtag
                      DICOM tag which identifies the attribute.
 CBuserInfo
                      Address of user data which is being passed from the
                      MC_Get_Encapsulated_Value_To_Function function.
                      This may be NULL.
 CBdataSize
                      The number of bytes in CBdataBuffer.
 CBdataBuffer
                      Address of the buffer containing a portion of the
                      attribute's value.
 CBisFirst
                      This is TRUE (not zero) the first time YourGetFunction
                      is being called for this attribute.
 CBisLast
                      This is TRUE (not zero) the last time YourGetFunction
                      is being called for this attribute.
```

Remarks

The **MC_Get_Encapsulated_Value_To_Function** function is used to fetch the value of the first frame of an attribute which has a value representation of OB or OW and is encapsulated. If a decompressor is registered, the data will be decompressed. Such attributes tend to have values of great length. To

accommodate this, one passes the **MC_Get_Encapsulated_Value_To_Function** function the name of a function (*YourGetFunction*) which Merge DICOM Toolkit, in turn, calls. This "callback" function is called repeatedly to provide blocks of the attribute's data value. If a decompressor is registered for this message, the data will be passed back decompressed. An optional *UserInfo* parameter may be used to pass information between the **MC_Get_Encapsulalated_Value_To_Function** caller and *YourGetFunction* which receives the data in its *CBuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

MC_Get_Encapsulated_Value_To_Function sets an internal frame pointer such that MC_Get_Next_Encapsulated_Value_To_Function will retrieve the next frame following this one.

YourGetFunction

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

CBisFirst is set to TRUE the first time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the first time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often open a file at this time.

CBisLast is set to TRUE the last time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the last time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often close a file at this time.

YourGetFunction must return MC_NORMAL_COMPLETION if all went well. If not, it must return MC_CANNOT_COMPLY. Both of these are defined in "mc3msg.h".

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally, and there are no more frames in the message.
MC_INVALID_TRANSFER_SYNTAX	The message's transfer syntax is non-encapsulated.
MC_END_OF_FRAME	The function completed normally, and there are more frames in the message.
MC_NULL_POINTER_PARM	YourGetFunction is NULL
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INCOMPATIBLE_VR	The attribute does not contain OB or OW data
MC_EMPTY_VALUE	No value has been assigned to this attribute yet.
MC_NULL_VALUE	The attribute has a value of NULL.
MC_NO_MORE_VALUES	The attribute has no more values.
MC_CALLBACK_CANNOT_COMPLY	YourGetFunction returned MC_CANNOT_COMPLY.
MC_COMPRESSION_FAILURE	There was an error while trying to decompress the data.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_MISSING_DELIMITER

The expected delimiter can not be detected

See Also

MC_Get_Next_Encapsulated_Value_To_Function MC_Get_Frame_To_Function MC_Get_Offset_Table_To_Function MC_Register_Compression_Callbacks

MC_Get_Enum_From_Transfer_Syntax

Converts the DICOM UID for a transfer syntax into an enumerated value.

Synopsis

"mc3msg.h" is returned:

IMPLICIT_LITTLE_ENDIAN IMPLICIT_BIG_ENDIAN EXPLICIT_LITTLE_ENDIAN EXPLICIT_BIG_ENDIAN DEFLATED_EXPLICIT_LITTLE_ENDIAN ENCAPSULATED_UNCOMPRESSED_ELE RLE JPEG_BASELINE JPEG_EXTENDED_2_4 JPEG_EXTENDED_3_5 JPEG_SPEC_NON_HIER_6_8 JPEG_SPEC_NON_HIER_7_9 JPEG_FULL_PROG_NON_HIER_10_12 JPEG_FULL_PROG_NON_HIER_11_13 JPEG_LOSSLESS_NON_HIER_14 JPEG_LOSSLESS_NON_HIER_15 JPEG_EXTENDED_HIER_16_18 JPEG_EXTENDED_HIER_17_19 JPEG_SPEC_HIER_20_22 JPEG_SPEC_HIER_21_23 JPEG_FULL_PROG_HIER_24_26 JPEG_FULL_PROG_HIER_25_27 JPEG_LOSSLESS_HIER_28 JPEG_LOSSLESS_HIER_29 JPEG_LOSSLESS_HIER_14 JPEG_2000_LOSSLESS_ONLY JPEG_2000 JPEG_LS_LOSSLESS JPEG_LS_LOSSY HEVC_H265_M10P_LEVEL_5_1 HEVC_H265_MP_LEVEL_5_1 JPIP_REFERENCED JPIP_REFERENCED_DEFLATE MPEG2_MPML MPEG2_MPHL MPEG4_AVC_H264_HP_LEVEL_4_1 MPEG4_AVC_H264_BDC_HP_LEVEL_4_1 MPEG4_AVC_H264_HP_LEVEL_4_2_2D MPEG4_AVC_H264_HP_LEVEL_4_2_3D MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2 JPEG_2000_MC_LOSSLESS_ONLY JPEG_2000_MC SMPTE_ST_2110_20_UNCOMPRESSED_PROGRESSIVE_ACTI **VE_VIDEO** SMPTE_ST_2110_20_UNCOMPRESSED_INTERLACED_ACTIV E_VIDEO

SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO PRIVATE_SYNTAX_1 PRIVATE_SYNTAX_2

Remarks

MC_Get_Enum_From_Transfer_Syntax converts a DICOM transfer syntax UID into a Merge DICOM Toolkit enumerated value. The UID values used by **MC_Get_Enum_From_Transfer_Syntax** are specified in Merge DICOM Toolkit's configuration files. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TRANSFER_SYNTAX	<i>Uid</i> is not registered in the Merge DICOM Toolkit configuration files.

See Also

MC_Get_Transfer_Syntax_From_Enum MC_Get_MergeCOM_Service MC_Get_UID_From_MergeCOM_Service

MC_Get_File_Length

Gets the length in bytes of a file object.

```
Synopsis
#include ``mc3media.h"
MC_STATUS MC_Get_File_Length (
    int FileID,
    unsigned long *FileLength
)
FileID The identifier assigned to this object by the MC_Create_File
    or MC_Create_Empty_File function.
FileLength The file object size is returned here.
```

Remarks

MC_Get_File_Length returns the number of bytes it would take to write a file object in DICOM Part 10 format. This function assumes the transfer syntax UID is contained in the group 2 elements. Note that a possible value for *FileLength* is undefined length (i.e., 0xffffffff), when the file object is encoded in an encapsulated transfer syntax. Also note that the file length may change. The padding is specified when calling **MC_Write_File**.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
MC_INVALID_TRANSFER_SYNTAX	The transfer syntax UID contained in the file objects group 2 elements was incorrect.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_NULL_POINTER_PARM	FileLength was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment.

MC_Get_File_Preamble

Gets the preamble for a file object.

Synopsis

)

```
#include ``mc3media.h"
MC STATUS MC Get File Preamble (
       int FileID,
       char *Preamble
 FileID
                 The identifier assigned to this object by the MC_Create_File
                 or MC_Create_Empty_File function.
                 A pointer to a user allocated 128 byte buffer in which the
 Preamble
                 preamble associated with the file object FileID is copied.
```

Remarks

MC_Get_File_Preamble retrieves the preamble associated with a file object. The function stores the DICOM file preamble for the file object File/D in the buffer pointed to by Preamble.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
MC_NULL_POINTER_PARM	Preamble was NULL.
See Also	
MC_Set_File_Preamble	

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Get_Filename

Gets the filename associated with a file object.

Synopsis

```
#include ``mc3media.h"
MC_STATUS MC_Get_Filename (
    int FileID,
    char *FileName,
    int FileSize,
```

```
)
```

FileID	The identifier assigned to this object by the	
	MC_Create_Empty_File or MC_Create_File function.	
Filename	A pointer to a string where the filename will be returned.	
FileSize	The length in bytes of <i>Filename</i> .	

Remarks

The MC_Get_Filename function retrieves the filename associated with a file object.

The filename is passed to the user's callback function when the **MC_Write_File** function is called.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
MC_NULL_POINTER_PARM	<i>Filename</i> was NULL.
MC_BUFFER_TOO_SMALL	Filename was too small to contain the complete filename.
See Also	
MC_Write_File	MC_Empty_File

MC_Get_First_Acceptable_Service

Retrieves the first of possibly many services that have been accepted over an association.

Synopsis

```
#include ``mergecom.h"
```

```
MC_STATUS MC_Get_First_Acceptable_Service (
    int AssocID,
        ServiceInfo *AserviceInfo
)
```

AssocID	MC_Open_Associat MC_Process_Assoc	ed to this object by the tion, MC_Open_Secure_Association, tiation_Request, re_Association_Request,
	MC_Wait_For_Asso MC_Wait_For_Secu	ciation or re_Association function.
AserviceInfo	A structure of type S	ServiceInfo you have declared that is filled but the accepted service. ServiceInfo is
typedef struct char Servic	<pre>MC_Service_Info { ceName[50]; </pre>	/* Merge DICOM Toolkit service name */
TRANSFER_SY	YNTAX SyntaxType; ,	/* transfer syntax negotiated for the service */
ROLE_TYPE H	RoleNegotiated;	<pre>/* The role negotiated for the service */</pre>
int Present char SOPCla		/* Presentation Context ID */ /* SOP Class UID associated with ServiceName */
char Servic	ceClassUID[66]; ,	<pre>/* Optional Service Class UID for the service if defined in the association request */</pre>
int NumberH	RelatedSOPClasses;	/* Number of related general SOP Classes contained in the association request */
RelatedSOPC	Class RelatedSOPCla	asses[10]; /* Array containing Related general SOP classes for this negotiated service */
<pre>} ServiceInfo;</pre>		

"TRANSFER_SYNTAX" and "ROLE_TYPE" are defined in mc3msg.h.

Remarks

The Merge DICOM Toolkit service name and transfer syntax negotiated is returned for the first service negotiated for and acceptable to both sides. These values are returned in the structure detailed above.

MC_Get_First_Acceptable_Service sets an internal "service pointer" such that **MC_Get_Next_Acceptable_Service** will retrieve the service following this one.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Meaning
The function completed normally.
AssocID is not a valid association object ID.
One or more of the pointer-type parameters was NULL.
Call was made after reading or sending a message.
The library has not been properly initialized.
No services were accepted.

See Also

MC_Open_Association MC_Wait_For_Association MC_Get_Next_Acceptable_Service MC_Open_Secure_Association MC_Wait_For_Secure_Association

MC_Get_First_Attribute

Retrieves information about the first attribute in a message.

Synopsis

```
#include ``mc3msg.h"
```

```
MC STATUS MC Get First Attribute (
       int MsgFileItemID,
       unsigned long *Tag,
       MC VR *ValueRep,
       int *NumValues
)
 MsgFileItemI
                The identifier assigned to this object by the
                MC_Open_Message, MC_Open_Empty_Message,
 D
                MC_Create_File, MC_Create_Empty_File, or
                MC_Open_Item functions.
 Tag
                The attribute's tag is returned here.
 ValueRep
                The attribute's value representation code is returned here.
                Possible return codes (defined in "mc3msg.h") are:
                AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR,
                UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL,
                OD, OF, SQ, UNKNOWN_VR
 NumValues
                The number of values assigned to the attribute is returned
                here. (Note: Will return one (1) if the attribute's value is NULL.)
```

Remarks

MC_Get_First_Attribute retrieves information about the first attribute in a message. The attribute's tag is returned at *Tag*, the attribute's Value Representation code is returned at *ValueRep* and the number of values stored for the attribute is returned at *NumValues*. Any of the *Tag, ValueRep* or *NumValues* parameters may be NULL, in which case the corresponding information is not returned. If the attribute is "empty" *NumValues* will contain zero (0). If the attribute's value is NULL, *NumValues* will contain 1.

MC_Get_First_Attribute sets an internal "attribute pointer" such that **MC_Get_Next_Attribute** will retrieve the attribute following this one.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value

Meaning

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_MESSAGE_EMPTY	There are no attributes in the message.
See Also	
MC_Get_Next_Attribute	MC_Get_Attribute_Info

MC_Get_Frame_To_Function

Retrieves the requested frame of encapsulated or non-encapsulated pixel data from a message. For encapsulated data, if a decompressor is registered, then the data will be returned decompressed.

Synopsis

```
#include ``mc3msg.h"
```

MC_Get_pAttribute_Info

```
MC_STATUS MC_Get_Frame_To_Function (
    int MsgFileItemID,
    int Frame,
    void *UserInfo,
    void *YourGetFunction(),
```

```
)
```

MsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
Frame	An integer identifying a frame number of encapsulated pixel data (starting from 0).
UserInfo	Address of data which will be passed on to <i>YourGetFunction</i> each time it is called. This may be NULL.
YourGetFunction	Name of a function which will be called repeatedly to provide blocks of data of the attribute's value

The function must be prototyped as follows:

```
MC_STATUS YourGetFunction (
    int CBMsgFileItemID,
    unsigned long CBtag,
    void *CBuserInfo,
    int CBdataSize,
    void *CBdataBuffer,
    int CBisFirst,
    int CBisLast
```

```
)
```

CBMsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
CBtag	DICOM tag which identifies the attribute.
CBuserInfo	Address of user data which is being passed from the MC_Get_Frame_To_Function function. This may be NULL.
CBdataSize	The number of bytes in CBdataBuffer.
CBdataBuffer	Address of the buffer containing a portion of the attribute's value.
CBisFirst	This is TRUE (not zero) the first time <i>YourGetFunction</i> is being called for this attribute.
CBisLast	This is TRUE (not zero) the last time <i>YourGetFunction</i> is being called for this attribute.

Remarks

The **MC_Get_Frame_To_Function** function is used to fetch an encapsulated or non-encapsulated frame of a pixel data attribute. In the encapsulated case, the functionality is similar to **MC_Get_Encapsulated_Value_To_Function** and **MC_Get_Next_Encapsulated_Value_To_Function** except that the user can retrieve the encapsulated frame by its number (starting from 0).

As pixel data attributes usually have values of great length the most effective way to handle it is to use **MC_Open_File_Upto_Tag_Bypass_Value** and the callback mechanism with **MC_Register_Callback_Function** to retrieve the attribute values from media.

The (*YourGetFunction*) function is used to provide repeatedly the blocks of the attribute's data value. If a decompressor is registered for this message, the data will be passed back decompressed. An optional *UserInfo* parameter may be used to pass information between the

MC_Get_Frame_To_Function caller and *YourGetFunction* which receives the data in its *CBuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

YourGetFunction

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

CBisFirst is set to TRUE the first time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the first time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often open a file at this time.

CBisLast is set to TRUE the last time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the last time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often close a file at this time.

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

YourGetFunction must return MC_NORMAL_COMPLETION if all went well. If not, it must return MC_CANNOT_COMPLY. Both of these are defined in "mc3msg.h".

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally, and there are no more frames in the message.
MC_INVALID_TRANSFER_SYNTAX	The message's transfer syntax is non-encapsulated.
MC_END_OF_FRAME	The function completed normally, and there are more frames in the message.
MC_NULL_POINTER_PARM	YourGetFunction is NULL
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INCOMPATIBLE_VR	The attribute does not contain OB or OW data
MC_EMPTY_VALUE	No value has been assigned to this attribute yet.
MC_NULL_VALUE	The attribute has a value of NULL.
MC_NO_MORE_VALUES	The attribute has no more values.
MC_CALLBACK_CANNOT_COMPLY	YourGetFunction returned MC_CANNOT_COMPLY.
MC_COMPRESSION_FAILURE	There was an error while trying to decompress the data.
MC_MISSING_DELIMITER	The expected delimiter can not be detected

See Also

MC_Get_Encapsulated_Value_To_Function MC_Get_Next_Encapsulated_Value_To_Function MC_Open_File_Upto_Tag_Bypass_Value MC_Register_Compression_Callbacks

MC_Get_Int_Config_Value

Used to get the value of an integer toolkit configuration parameter at runtime.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Get_Int_Config_Value (
    IntParm Parm,
    int *Value
)
Parm An enumerated constant identifying the integer
    configuration parameter to get. Parm can have any of the
    following values:
        ARTIM_TIMEOUT
```

ASSOC_REPLY_TIMEOUT COMPRESSION CHROM FACTOR COMPRESSION J2K LOSSY QUALITY COMPRESSION_J2K_LOSSY_RATIO COMPRESSION LUM FACTOR CONNECT_TIMEOUT DEFLATE COMPRESSION LEVEL DESIRED LAST PDU SIZE FLATE GROW OUTPUT BUF SIZE IGNORE JPEG BAD SUFFIX INACTIVITY_TIMEOUT LARGE_DATA_SIZE LIST SQ DEPTH LIMIT LIST VALUE LIMIT LOG_FILE_LINE_LENGTH LOG FILE SIZE LOG_MEMORY_SIZE MAX PENDING CONNECTIONS NUM HISTORICAL LOG FILES NUMBER OF CAP FILES **OBOW BUFFER SIZE** PEGASUS NUMBER OF THREADS RELEASE_TIMEOUT TCPIP_KEEP_ALIVE_TIME TCPIP KEEP ALIVE INTERVAL TCPIP_LISTEN_PORT TCPIP RECEIVE BUFFER SIZE TCPIP_SEND_BUFFER_SIZE WORK BUFFER SIZE WRITE TIMEOUT

These names are the same as those given to the parameters in the toolkit configuration files. A description of the options can be found in Appendix B.

Value Parm's integer value is returned here.

Remarks

The Merge DICOM Toolkit Library accesses several configuration files at startup. This call allows your application to get integer configurable parameters specified in these files at runtime. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Value</i> was NULL.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_MUST_BE_POSITIVE	Parameter value cannot be negative.

MC_LIBRARY_NOT_INITIALIZED

The library has not been properly initialized.

See Also

MC_Set_Int_Config_Value MC_Get_Log_Destination MC_Get_Long_Config_Value MC_Get_String_Config_Value

MC_Set_Log_Destination MC_Set_Long_Config_Value MC_Set_String_Config_Value

MC_Get_Listen_Socket

Returns the id of the socket being listened on

Synopsis

AlistenSocket The socket being listened on.

Remarks

MC_Get_Listen_Socket returns the DICOM listen socket. It will return the current listen port used by calls to MC_Wait_For_Association or MC_Wait_For_Secure_Association. MC_Get_Listen_Socket_For_Port must be used when listening is done with MC_Wait_For_Association_On_Port or MC_Wait_For_Secure_Association_On_Port.

In specialized cases where the server application is waiting on several asynchronous events, not just the association event, the **MC_Get_Listen_Socket** call can be made to request the file descriptor for the DICOM listen socket. In this way the server application can do a select() system call on this and other file descriptors. When the select returns with an event on the DICOM listen socket descriptor, the application can call **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** and get an immediate response.

Similarly, once the association is established, both the client and server applications can use the **MC_Get_Association_Info** call to get the file descriptor for the socket over which message exchange will occur. Again, select() can be used to wait asynchronously for a DICOM request or response message.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>AlistenSocket</i> was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment.

See Also

MC_Get_Association_Info

MC_Get_Listen_Socket_For_Port

Returns the id of the socket being listened on for a specific port

Synopsis

```
#include ``mergecom.h"
```

```
MC_STATUS MC_Get_Listen_Socket_For_Port (
    int Port
    int *ListenSocket
)
Port The port number to check for the listen socket
```

ListenSocket The socket being listened on.

Remarks

MC_Get_Listen_Socket_For_Port returns the DICOM listen socket for a specific port. It can be used to get the default listen port from a call to MC_Wait_For_Association or MC_Wait_For_Secure_Association. It can also be used to get the listen socket for a specific listen port from calls to MC_Wait_For_Association_On_Port and MC_Wait_For_Secure_Association_On_Port.

In specialized cases where the server application is waiting on several asynchronous events, not just the association event, the **MC_Get_Listen_Socket_For_Port** call can be made to request the file descriptor for the DICOM listen socket. In this way the server application can do a select() system call on this and other file descriptors. When the select returns with an event on the DICOM listen socket descriptor, the application can call **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** and get an immediate response.

Similarly, once the association is established, both the client and server applications can use the **MC_Get_Association_Info** call to get the file descriptor for the socket over which message exchange will occur. Again, select() can be used to wait asynchronously for a DICOM request or response message.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>LlistenSocket</i> was NULL.
MC_INVALID_PORT_NUMBER	Port is not currently being listened on.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment.

See Also

MC_Get_Association_Info

MC_Get_Log_Destination

Used to get the value of a integer toolkit configuration parameter at runtime.

Synopsis #include ``mc3msg.h" MC STATUS MC Get Log Destination (LogParm Parm, int *Destination) Parm An enumerated constant identifying the class of logging that is to be retreived. Parm can have any of the following values: ERROR_DESTINATIONS, WARNING_DESTINATIONS, INFO_DESTINATIONS, T1_DESTINATIONS, T2_DESTINATIONS, T3_DESTINATIONS, T4_DESTINATIONS, T5_DESTINATIONS, T6_DESTINATIONS, **T7_DESTINATIONS, T8_DESTINATIONS T9_DESTINATIONS** A defined term specifying where the logging is being Destination directed. Value can have any of the following values: File_Destination, Memory_Destination, Screen_Destination, Bitbucket_Destination These values can also be OR'ed together to indicate multiple destinations.

Remarks

This call allows you to retrieve the logging of error, warning, and info messages at runtime. The [DEFAULT_LIBRARY] section of the Merge DICOM Toolkit initialization file contains the setting used at startup. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value

Meaning

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Value</i> was NULL.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Set_Int_Config_Value	MC_Get_Int_Config_Value MC_Set_Long_Config_Value
	MC_Get_Long_Config_Value
MC_Set_String_Config_Value	MC_Get_String_Config_Value

MC_Get_Long_Config_Value

MC_Set_Log_Destination

Used to get the value of a long integer toolkit configuration parameter at runtime.

Synopsis

#include ``mc3msg.h" MC STATUS MC Get Long Config Value (LongParm Parm, long int *Value) Parm An enumerated constant identifying the long integer configuration parameter to get. Parm can have only the following values: CAPTURE FILE SIZE, PDU_MAXIMUM_LENGTH, CALLBACK_MIN_DATA_SIZE, PIXEL_BUFFER_SIZE These names are the same as those given to the parameters in the toolkit configuration files. Value Parm's long integer value is returned here.

Remarks

The Merge DICOM Toolkit Library accesses several configuration files at startup. This call allows your application to get long integer configurable parameters specified in these files at runtime. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Value</i> was NULL.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.

O Copyright Merge Healthcare Solutions Inc. 2023

See Also

MC_Get_Int_Config_Value MC_Get_Log_Destination MC_Set_Long_Config_Value MC_Set_String_Config_Value MC_Set_Int_Config_Value MC_Set_Log_Destination MC_Get_String_Config_Value

MC_Get_MergeCOM_Service

Returns the Merge DICOM Toolkit service name associated with a DICOM SOP Class UID.

Synopsis

Remarks

MC_Get_MergeCOM_Service returns the service name for *UID*. This function is often useful in conjunction with **MC_Set_Service_Command** to set the message type for a message before validation or sending across the network.

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these SOP Class UIDs.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>UID</i> or <i>ServiceName</i> was NULL.
MC_BUFFER_TOO_SMALL	<i>ServiceName</i> was not large enough to contain the Merge DICOM Toolkit service name.
MC_INVALID_SOP_CLASS_UID	<i>UID</i> does not contain a Merge DICOM Toolkit supported SOP Class UID.

See Also

MC_Set_Service_CommandMC_Get_Tag_InfoMC_Get_pTag_InfoMC_Get_Enum_From_Transfer_SyntaxMC_Get_Transfer_Syntax_From_EnumMC_Get_UID_From_MergeCOM_Service

MC_Get_MergeCOM_Service_From_UID

Converts a DICOM SOP Class UID into a Merge DICOM Toolkit service name.

Synopsis

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_UNKNOWN_SERVICE	<i>ServiceName</i> was not registered in the Merge DICOM Toolkit configuration files.
MC_BUFFER_TOO_SMALL	<i>Uid</i> is not large enough to contain the unique identifier.
See Also	
MC_Get_MergeCOM_Service MC_Get_Enum_From_Transfer_Syntax MC_Get_Transfer_Syntax_From_Enum	

MC_Get_UID_From_MergeCOM_Service

MC_Get_Message_Service

Returns the service name and command ID association with a message.

Synopsis

#include ``mergecom.h"

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
      MC_STATUS MC_Get_Message_Service (
int MessageID,
char **ServiceName,
MC_COMMAND *Command

      MessageID
      The message object's identification number.

      ServiceName
      The name of the service associated with the message is
returned here.

      Command
      The command associated with this message is returned here.

      Command
      The command associated with this message is returned here.

      MessageID
      The command associated with this message is returned here.

      MessageID
      The command associated with this message is returned here.

      MessageID
      The messageID

      MessageID
      The command associated with this message is returned here.

      MessageID
      The messageID

      MessageID
      The command associated with this message is returned here.

      MessageID
      The messageID

      MessageID
      The messageID

      MessageID
      The command associated with this message is returned here.

      MessageID
      The messageID

      MessageID
      The mes
```

MC_Get_Message_Service returns the service name and command which were specified when the message identified by *MessageID* was opened. This function is often useful in callback functions which are supporting more than one service or command. Note well that this memory is de-allocated when the **MC_Free_Message** API call is made for *MessageID*.

C_STORE_RQDICOM Composite Store Service RequestC_STORE_RSPDICOM Composite Store Service Response
C_STORE_RSP DICOM Composite Store Service Response
C_ECHO_RQ DICOM Verification Service Request
C_ECHO_RSP DICOM Verification Service Response
C_FIND_RQ DICOM Composite Find Service Request
C_FIND_RSP DICOM Composite Find Service Response
C_CANCEL_FIND_RQ Cancel DICOM Composite Find Service Request
C_GET_RQ DICOM Composite Get Service Request
C_GET_RSP DICOM Composite Get Service Response
C_CANCEL_GET_RQ Cancel DICOM Composite Get Service Request
C_MOVE_RQ DICOM Composite Move Service Request
C_MOVE_RSP DICOM Composite Move Service Response
C_CANCEL_MOVE_RQ Cancel DICOM Composite Move Service Request
N_EVENT_REPORT_RQ DICOM Normalized Report Service Request
N_EVENT_REPORT_RSP DICOM Normalized Report Service Response
N_GET_RQ DICOM Normalized Get Service Request

The following commands are supported by the Merge DICOM Toolkit:

[©] Copyright Merge Healthcare Solutions Inc. 2023

N_GET_RSP	DICOM Normalized Get Service Request
N_SET_RQ	DICOM Normalized Set Service Request
N_SET_RSP	DICOM Normalized Set Service Response
N_ACTION_RQ	DICOM Normalized Action Service Request
N_ACTION_RSP	DICOM Normalized Action Service Response
N_CREATE_RQ	DICOM Normalized Create Service Request
N_CREATE_RSP	DICOM Normalized Create Service Response
N_DELETE_RQ	DICOM Normalized Delete Service Request
N_DELETE_RSP	DICOM Normalized Delete Service Response

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these commands.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	ServiceName or Command was NULL.
MC_INVALID_MESSAGE_ID	<i>MessageID</i> is not a valid message object ID.
See Also	

MC_Register_Callback_Function

MC_Get_Message_Transfer_Syntax

Returns the transfer syntax over which a message was received over the network.

Synopsis #include "mergecom.h" MC_STATUS MC_Get_Message_Transfer_Syntax (int MessageID, TRANSFER_SYNTAX *TransferSyntax) MessageID The message object's identification number. TransferSyntax The name of the transfer syntax associated with the message is returned here. It will be one of the enumerated TRANSFER_SYNTAX types defined in "mc3msg.h" (see MC_Get_Transfer_Syntax_From_Enum).

MC_Get_Message_Transfer_Syntax returns the DICOM transfer syntax over which a message was received on the network. This function is often useful when encapsulated (compressed) and standard transfer syntaxes are both negotiated for a service. The transfer syntax is needed in this case to determine how the pixel data in a message should be decoded

Reference the *Merge DICOM Toolkit User's Manual* for a discussion on negotiating multiple transfer syntaxes for a service.

MC_Get_Message_Transfer_Syntax can also be used to get the transfer syntax of a DICOM file object. When used with a DICOM file object, the routine will return the transfer syntax set for the tag (0002,0010) Transfer Syntax UID in the file object.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>TransferSyntax</i> was NULL.
MC_INVALID_MESSAGE_ID	MessageID is not a valid message object ID.
MC_INVALID_TRANSFER_SYNTAX	The transfer syntax of the DICOM file or message is not set.
See Also	
MC_Set_Message_Transfer_Syntax	MC_Read_Message MC_Reset_Message_Transfer_Syntax

MC_Get_Meta_ServiceName

Returns the meta service name corresponding to the presentation context in which a message was received over the network.

```
#include ``mergecom.h"
MC STATUS MC Get Meta ServiceName (
       int AssociationID,
       char *Value,
       int BufferSize
)
 AssociationID An association identification number returned by a
                MC_Wait_For_Association,
                MC_Process_Association_Request,
                MC_Process_Secure_Association_Request or
                MC_Wait_For_Secure_Association call.
 Value
                The name of the meta service of the most recent request
                message is returned here.
 BufferSize
                The size of the Value buffer in bytes
```

MC_Get_Meta_ServiceName returns the DICOM meta service name for a request message that was received on the network. This function is only useful on the receiving end of a

MC_Send_Request_Message_For_Service or MC_Send_Request_For_Service. An example is the Filmbox when both BASIC_GRAYSCALE_PRINT_MANAGEMENT and

BASIC_COLOR_PRINT_MANAGEMENT are negotiated. The

MC_Send_Request_Message_For_Service would specify which of the services should be used. On the receiving end, **MC_Read_Message** or **MC_Read_To_Stream** returns BASIC_FILM_BOX as the service name, but does not contain any information on the meta SOP. This function will return the meta SOP based on the most recent received message's presentation context.

This function should only be used immediately after receiving a request message.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	The association ID passed in was not valid.
MC_NULL_POINTER_PARM	<i>Value</i> was NULL.
MC_BUFFER_TOO_SMALL	<i>Value</i> is not large enough.
MC_NOT_META_SOP	The most recent received message was not sent on a presentation context that belongs to a Meta SOP.
MC_NOT_FOUND	The meta SOP for the most recently received message could not be found.

See Also

MC_Read_Message MC_Read_To_Stream MC_Send_Request_Message_For_Service MC_Send_Request_For_Service

MC_Get_Negotiation_Info

Retrieves received extended negotiation information.

Synopsis

```
#include "mergecom.h"
MC STATUS MC Get Negotiation Info (
```

int AssociationID, char *ServiceName, void **ExtInfoBuffer, int *ExtInfoLength

AssociationID	An association identification number returned by a MC_Wait_For_Association,
	MC_Process_Association_Request, MC_Process_Secure_Association_Request or
	MC_Wait_For_Secure_Association call.
ServiceName	The name given to a valid DICOM service.
ExtInfoBuffer	A pointer to the buffer containing extended negotiation information will be returned here.
ExtInfoLength	The number of bytes contained in the <i>ExtInfoBuffer</i> will be returned here.

The DICOM standard allows application entities to exchange "extended negotiation information" when establishing an association. The contents of the negotiation information must be known to both the association requestor application and the association acceptor application. Such extended negotiation is not often used for DICOM services, but some services may require it.

MC_Get_Negotiation_Info allows the caller to retrieve any extended negotiation information which may have been received during a **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** call. If the remote application sent negotiation information for *ServiceName*, the negotiation information is returned at *ExtInfoBuffer* and its length at *ExtInfoLength*. If no negotiation information was received, the function will return a status of **MC_EXT_INFO_UNAVAILABLE**.

MC_Set_Negotiation_Info_For_Association may be used to set the extended negotiation information before calling **MC_Accept_Association** to accept the association. Merge DICOM Toolkit will only return extended negotiation information to the remote application that has been registered with **MC_Set_Negotiation_Info_For_Association**.

Note also that **MC_Set_Negotiation_Info** may be used to set extended negotiation for all associations accepted with **MC_Accept_Association**. Any service with extended negotiation information in the association request will have the extended negotiation information registered in **MC_Set_Negotiation_Info** returned by **MC_Accept_Association**. Use of **MC_Set_Negotiation_Info** is deprecated.

Use **MC_Clear_Negotiation_Info** to remove extended information registered using the **MC_Set_Negotiation_Info** function call.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association identifier.
MC_NULL_POINTER_PARM	Either <i>ServiceName</i> or <i>ExtInfoBuffer</i> or <i>ExtInfoLength</i> was NULL.

MC_UNKNOWN_SERVICE

ServiceName was not registered in the Merge DICOM Toolkit configuration files.

MC_EXT_INFO_UNAVAILABLE

No extended negotiation information was received for *ServiceName*.

See Also

MC_Set_Negotiation_Info MC_Clear_Negotiation_Info

MC_Set_Negotiation_Info_For_Association

MC_Get_Next_Acceptable_Service

Retrieves additional services that have been accepted over an association, after **MC_Get_First_Acceptable_Service** has been called to retrieve the first.

```
#include ``mergecom.h"
MC STATUS MC Get Next Acceptable Service (
       int AssocID,
      ServiceInfo *AserviceInfo
)
 AssocID
               The identifier assigned to this object by the
               MC_Open_Association, MC_Open_Secure_Association,
               MC_Process_Association_Request,
               MC_Process_Secure_Association_Request,
               MC_Wait_For_Association or
               MC_Wait_For_Secure_Association function.
 AserviceInfo
               A structure of type ServiceInfo you have declared that is filled
               with information about the accepted service. ServiceInfo is
               defined in mergecom.h as:
typedef struct MC_Service_Info {
    char ServiceName[50]; /* Merge DICOM Toolkit service name */
    TRANSFER_SYNTAX SyntaxType; /* transfer syntax negotiated
                                      for the service */
   ROLE_TYPE RoleNegotiated; /* The role negotiated for the
                                      service */
           PresentationContextID; /* Presentation Context ID */
    int.
                                   /* SOP Class UID associated
    char SOPClassUID[66];
                                      with ServiceName */
                                  /* Optional Service Class UID
    char ServiceClassUID[66];
                                      for the service if defined
                                      in association request */
    int
           NumberRelatedSOPClasses; /* Number of related general
                                      SOP Classes contained in
                                      the association request */
    RelatedSOPClass RelatedSOPClasses[10]; /* Array containing
                                      Related general SOP classes
                                       for this negotiated
                                       service */
} ServiceInfo;
```

"TRANSFER_SYNTAX" and "ROLE_TYPE" are defined in mc3msg.h.

Remarks

The Merge DICOM Toolkit service name and transfer syntax negotiated is returned for the first service negotiated for and acceptable to both sides. These values are returned in the structure detailed above.

When there are no more services in the list of accepted services MC_END_OF_LIST is returned.

MC_Get_Next_Acceptable_Service sets an internal "service pointer" such that MC_Get_Next_Acceptable_Service can be called again to retrieve the service following the current one.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssocID is not a valid association object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_STATE_VIOLATION	Call was made after reading or sending a message.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_END_OF_LIST	No more services in list of accepted services.
See Also	
MC_Open_Association MC_Wait_For_Association MC_Get_First_Acceptable_Service MC_Process_Secure_Association_Re	MC_Open_Secure_Association MC_Wait_For_Secure_Association MC_Process_Association_Request equest

MC_Get_Next_Attribute

Retrieves information about the next attribute in a message.

```
#include ``mc3msg.h"
MC_STATUS MC_Get_Next_Attribute (
    int MsgFileItemID,
    unsigned long *Tag,
    MC_VR *ValueRep,
    int *NumValues
)
MsgFileItem/ The identifier assigned to this object by the
D MC_Open_Message, MC_Open_Empty_Message,
    MC_Create_File, MC_Create_Empty_File, or
    MC_Open_Item functions.
Tag The attribute's tag is returned here.
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

ValueRep	The attribute's value representation code is returned here.
	Possible return codes (defined in "mc3msg.h") are:
	AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR,
	UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL,
	OD, OF, SQ, UNKNOWN_VR
NumValues	The number of values assigned to the attribute is returned
	here. (Note: Will be one (1) if the attribute's value is NULL.)

MC_Get_Next_Attribute retrieves information about the next attribute in a message. MC_Get_First_Attribute, MC_Get_Attribute_Info or MC_Get_pAttribute_Info must be used first to set a current attribute location. The attribute's tag is returned at *Tag*, the attribute's Value Representation code is returned at *ValueRep*, and the number of values stored for the attribute is returned at *NumValues*. Any of the *Tag, ValueRep* or *NumValues* parameters may be NULL, in which case the corresponding information is not returned. If the attribute is "empty", *NumValues* will contain zero (0). If the attribute's value is NULL, *NumValues* will contain one (1).

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_MESSAGE_EMPTY	The message contains no attributes.
MC_NO_MORE_ATTRIBUTES	There are no more attributes in the message.
See Also	
MC_Get_First_Attribute MC_Get_pAttribute_Info	MC_Get_Attribute_Info

MC_Get_Next_Encapsulated_Value_To_Function

Retrieves the next frame of encapsulated pixel data from a message. If a decompressor is registered, then the data will be decompressed.

MsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
Tag	A tag identifying a message attribute containing pixel data.
UserInfo	Address of data which will be passed on to <i>YourGetFunction</i> each time it is called. This may be NULL.
YourGetFunction	Name of a function which will be called repeatedly to provide blocks of data of the attribute's value

The function must be prototyped as follows:

```
MC STATUS YourGetFunction (
       int CBMsgFileItemID,
       unsigned long CBtag,
       void *CBuserInfo,
       int CBdataSize,
       void *CBdataBuffer,
       int CBisFirst,
       int CBisLast
)
 CBMsgFileItemI
                    The message identifier assigned to this object by the
 D
                    MC_Open_Message, MC_Open_Empty_Message
                   functions.
 CBtag
                    DICOM tag which identifies the attribute.
 CBuserInfo
                    Address of user data which is being passed from the
                    MC_Get_Next_Encapsulated_Value_To_Function
                    function. This may be NULL.
 CBdataSize
                    The number of bytes in CBdataBuffer.
 CBdataBuffer
                    Address of the buffer containing a portion of the
                    attribute's value.
                    This is TRUE (not zero) the first time YourGetFunction is
 CBisFirst
                    being called for this attribute.
 CBisLast
                    This is TRUE (not zero) the last time YourGetFunction is
                    being called for this attribute.
```

Remarks

The **MC_Get_Next_Encapsulated_Value_To_Function** function is used to fetch the value of the next frame of an attribute which has a value representation of OB or OW and is encapsulated. If a decompressor is registered, the data will be decompressed. Such attributes tend to have values of great length. To accommodate this, one passes the

MC_Get_Next_Encapsulated_Value_To_Function function the name of a function (*YourGetFunction*) which Merge DICOM Toolkit, in turn, calls. This "callback" function is called repeatedly to provide blocks of the attribute's data value. If a decompressor is registered for this

message, the data will be passed back decompressed. An optional *UserInfo* parameter may be used to pass information between the **MC_Get_Next_Encapsulalated_Value_To_Function** caller and *YourGetFunction* which receives the data in its *CBuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

MC_Get_Next_Encapsulated_Value_To_Function sets an internal frame pointer such that subsequent calls will retrieve the next frame following this one.

YourGetFunction

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

CBisFirst is set to TRUE the first time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the first time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often open a file at this time.

CBisLast is set to TRUE the last time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the last time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often close a file at this time.

YourGetFunction must return MC_NORMAL_COMPLETION if all went well. If not, it must return MC_CANNOT_COMPLY. Both of these are defined in "mc3msg.h".

Return Value

One of the enumerated **MC_STATUS** codes defined in "**mcstatus.h**":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally, and this was the last frame in the message.
MC_INVALID_TRANSFER_SYNTAX	The message's transfer syntax is non-encapsulated.
MC_END_OF_FRAME	The function completed normally, and there are more frames in the message.
MC_NULL_POINTER_PARM	YourGetFunction is NULL
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INCOMPATIBLE_VR	The attribute does not contain OB or OW data
MC_EMPTY_VALUE	No value has been assigned to this attribute yet.
MC_NULL_VALUE	The attribute has a value of NULL.
MC_NO_MORE_VALUES	The attribute has no more values.
MC_CALLBACK_CANNOT_COMPLY	YourGetFunction returned MC_CANNOT_COMPLY.
MC_COMPRESSION_FAILURE	There was an error while trying to decompress the data.

See Also

MC_Get_Encapsulated_Value_To_Function MC_Get_Frame_To_Function MC_Get_Offset_Table_To_Function MC_Register_Compression_Callbacks

MC_Get_Next_pValue... Functions

Retrieves the next value of a private attribute in a message object

```
#include ``mc3msg.h"
MC STATUS MC Get Next pValue (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC DT DataType,
      MC size t BufferSize,
      void *Value,
      int *ValueSize
)
MC STATUS MC Get Next pValue To Float (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      float *Value
MC_STATUS MC_Get_Next_pValue_To_Double (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      double *Value
MC STATUS MC Get Next pValue To ShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      short int *Value
MC_STATUS MC_Get_Next_pValue_To_UShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned short *Value
MC_STATUS MC_Get_Next_pValue_To_Int (
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      int *Value
)
MC STATUS MC Get Next pValue To UInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned int *Value
MC STATUS MC Get Next pValue To LongInt (
      int MsqFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      long int *Value
MC_STATUS MC_Get_Next_pValue_To_ULongInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long *Value
MC STATUS MC Get Next pValue To LongLong (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      long long *Value
MC STATUS MC Get Next pValue To ULongLong (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long long *Value
MC_STATUS MC_Get_Next_pValue_To_String (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC size t BufferSize,
      char *Value
MC STATUS MC_Get_Next_pValue_To_UnicodeString (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC size t BufferSize,
```

int *Va	lueSize,		
MC_Uhar *Value)			
MsgFileItemI D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.		
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.		
Group	The number identifying the private group containing the private attribute. It must be an odd number.		
ElementByte	The number identifying <i>Group</i> for this <i>Private</i>	g the private attribute within the private <i>Code</i> .	
DataType	One of the enumerated codes identifying the data type of the value in <i>Value</i> . The MC_DT enumerated type is defined in "mc3msg.h". They are:		
	String_Type	Null-terminated character string	
	Int_Type	Binary integer number	
	UInt_Type	Binary unsigned integer number	
	ShortInt_Type	Binary short integer number	
	UShortInt_Type	Binary unsigned short integer number	
	LongInt_Type	Binary long integer number	
	ULongInt_Type	Binary unsigned long integer number	
	LongLong_Type	Binary 64-bit integer number	
	UlongLong_Type	Binary 64-bit unsigned integer	
	number Float_Type	Binary Floating point number	
	Double_Type	Binary Floating point number	
	Buffer_Type	Binary byte value	
Value	The attribute's value will be returned here.		
ValueSize	The size of the returne	d value is returned here.	
BufferSize	The size of the <i>Value</i> buffer in bytes. This parameter is type size_t on 64-bit Windows and int on all other platforms.		

These functions fetch the next value of a multi-valued private attribute. If no value has yet been fetched from this attribute, these functions behave identically to the functions of similar names, with "_Next" removed from the name. If one or more values have already been fetched from this attribute, the next value with be returned.

If MC_Get_Next_pValue is used, the data type of the fetched value is specified by the DataType parameter. The other function names imply the data type requested. For example: MC_Get_Next_pValue_To_Int is the same as calling MC_Get_Next_pValue with DataType specified as Int_Type. Each function will return the value requested in Value, which must be prototyped as the appropriate type. In the case of MC_Get_Next_pValue_To_String and

MC_Get_Next_pValue_To_UnicodeString, an additional parameter (BufferSize) is required to specify

the size of the buffer to receive the string value. For **MC_Get_Next_pValue**, *BufferSize* must be large enough to contain the data type requested.

For MC_Get_Next_pValue_To_UnicodeString, it is required to call MC_Enable_Unicode_Conversion.

Any reasonable conversion will be made from the attribute's value representation to the data type requested. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

Function	May be used to retrieve values from attributes with these Value Representations
MC_Get_Next_pValue_To_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_Int	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_UInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_LongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_UlongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_UlongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_pValue_To_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ
MC_Get_Next_pValue_To_UnicodeString	LO, LT, PN, SH, ST, UC, UT

NOTE: Attributes with a value representation of **SQ** (sequence of items) have values stored internally as integers - each value is the *ItemID* of an item object opened previously using **MC_Open_Item**.

The same rules apply to the **MC_Get_Next_pValue** function, based on the value used in the *DataType* parameter.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value

Meaning

The function completed normally.

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

MC_Get_pValue Functions MC_Get_Next_Value Functions	MC_Get_Value Functions
See Also	
MC_INVALID_DATA_TYPE	The <i>DataType</i> parameter is invalid.
MC_VALUE_OUT_OF_RANGE	The attribute numeric value is too large to be accommodated by the receiving data type. (E.g. Using MC_Get_pValue_To_Int when the value is 123.45; or instances where the value's sign (+ -) would be affected)
MC_INCOMPATIBLE_VALUE	The attribute's value is not consistent with its VR. (This should never happen.)
MC_INCOMPATIBLE_VR	The function was called to retrieve an attribute value whose value representation (VR) is not one of those listed for the function in the table above.
MC_BUFFER_TOO_SMALL	Buffer is not large enough to contain the value. (MC_Get_Next_pValue and MC_Get_Next_pValue_To_String only)
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_NO_MORE_VALUES	There are no more values for this attribute.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_EMPTY_VALUE	The attribute has no value assigned to it.

MC_Get_Next_Validate_Error

Used to retrieve additional validation error blocks created by the **MC_Validate_Message** function.

Synopsis

```
#include "mc3msg.h"
MC_STATUS MC_Get_Next_Validate_Error (
    int MessageID,
    VAL_ERR **ErrorInfo
)
```

MessageID The message object's identification number

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

Errorli	nfo	violation enc MC_NORM	oun	a validation error block describing the Itered is returned here if a status of COMPLETION is returned. The block is the Iefined in "mc3msg.h".
typede	f stru	ct ValErr_stru	ct	
{				
	unsig	ned long Tag;	/*	Tag of attribute with validation violation */
	int Ma	sgItemID;	/*	ID of message or item object containing the attribute */
	int Va	alueNumber;	/*	Value number involved - zero if no value involved */
	MC ST	ATUS Status;	/*	Validation violation status code */
	int Pa	arentMsgID;	/*	ID of parent of message or item object containing the attribute */
	int Ma	sgLevel;	/*	Nesting level in data set hierarchy of the message or item object containing the attribute; */

} VAL_ERR;

MC_Get_Next_Validate_Error returns additional validation error blocks created by a previous call to **MC_Validate_Message**. The next validation violation error block is returned at **ErrorInfo* if another violation exists. If not, this function returns **MC_END_OF_LIST**.

The VAL_ERR data structure is de-allocated when MC_Free_File, MC_Empty_File or MC_Empty_Message are called.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	Another validation violation exists and its information is returned at <i>*ErrorInfo</i> .
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_NULL_POINTER_PARM	<i>ErrorInfo</i> was NULL.
MC_END_OF_LIST	The error list has been exhausted, or no MC_Validate_Message call has been made for this message.

Validation Violations

If a status of **MC_NORMAL_COMPLETION** is returned, the status of the next validation violation is returned at *ErrorInfo->Status*. The violation status codes are defined in **"mcstatus.h"**. They are arranged by violation type below:

INFO MESSAGES:

MC_NO_CONDITION_FUNCTION	The attribute's DICOM type is "1C" or "2C" and no function
	was available to check the condition. A user-defined condition
	function was not correctly linked with the program.

MC_UNABLE_TO_CHECK_CONDIT	ION
	Not enough information is available to check whether or not a
	DICOM type "1C" or "2C" attribute is required.
WARNINGS:	
MC_NOT_ONE_OF_DEFINED_TERM	1S A value set for this attribute is not one of the valid defined
	terms assigned to this attribute.
MC_NON_SERVICE_ATTRIBUTE	The attribute is not one of those defined for the message's service. Note that private attributes will not cause this violation
ERRORS:	
MC_INVALID_VR_CODE	The attribute's value representation is unknown.
MC_REQUIRED_ATTRIBUTE_MISS	
	An attribute which is required for the message service has been deleted from the message object.
MC_REQUIRED_VALUE_MISSING	been deleted nom the message object.
	The attribute is required to have a value and does not.
MC_VALUE_MAY_NOT_BE_NULL	The attribute is DICOM type "1" or type "1C" and it has been encoded with a NULL value.
MC_VALUE_NOT_ALLOWED	This DICOM type "1C" or type "2C" attribute may not have a value under current conditions.
MC_TOO_FEW_VALUES	The attribute is required to have more values set.
MC_TOO_MANY_VALUES	The attribute has more values set than are allowed.
MC_INVALID_ITEM_ID	This sequence of items (SQ) attribute has an invalid value assigned to it.
MC_NOT_ONE_OF_ENUMERATED_	VALUES
	A value set for this attribute is not one of the valid enumerated values assigned to this attribute.
MC_INVALID_VALUE_FOR_VR	A value for this attribute does not conform to the requirements of its value representation.
MC_INVALID_CHARS_IN_VALUE	A value for this attribute does not contain valid characters for its value representation.
See Also	

MC_Validate_Message

MC_Get_Next_Value... Functions

Retrieves the next value of an attribute in a message object

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Get_Next_Value (
    int MsgFileItemID,
    unsigned long Tag,
```

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

```
MC DT DataType,
      MC_size_t BufferSize,
      void *Value,
      int *ValueSize
MC_STATUS MC_Get_Next_Value_To_Float (
      int MsqFileItemID,
      unsigned long Tag,
      float *Value
MC STATUS MC Get Next Value To Double (
      int MsgFileItemID,
      unsigned long Tag,
      double *Value
MC STATUS MC Get Next Value To ShortInt (
      int MsgFileItemID,
      unsigned long Tag,
      short int *Value
MC_STATUS MC_Get_Next_Value_To_UShortInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned short *Value
)
MC STATUS MC Get Next Value To Int (
      int MsgFileItemID,
      unsigned long Taq,
      int *Value
MC STATUS MC Get Next Value To UInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned int *Value
MC STATUS MC Get Next Value To LongInt (
      int MsgFileItemID,
      unsigned long Tag,
      long int *Value
MC_STATUS MC_Get_Next_Value_To_ULongInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long *Value)
MC STATUS MC Get Next Value To LongLong (
      int MsgFileItemID,
      unsigned long Tag,
      long long *Value
)
MC STATUS MC Get Next Value To ULongLong (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long long *Value)
MC_STATUS MC_Get_Next_Value_To_String (
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
int MsgFileItemID,
       unsigned long Tag,
       MC size t BufferSize,
       char *Value
MC STATUS MC Get Next Value To UnicodeString (
       int MsqFileItemID,
       unsigned long Tag,
       MC size t BufferSize,
       int *ValueSize,
       MC Uhar *Value
)
 MsgFileIteml
                The identifier assigned to this object by the
 D
                MC_Open_Message, MC_Open_Empty_Message,
                MC_Create_File, MC_Create_Empty_File, or
                MC_Open_Item functions.
 Tag
                DICOM tag which identifies the attribute.
 DataType
                One of the enumerated codes identifying the data type of the
                value in Value. The MC_DT enumerated type is defined in
                "mc3msg.h". They are:
                String_Type
                                      Null-terminated character string
                Int_Type
                                      Binary integer number
                UInt_Type
                                      Binary unsigned integer number
                                      Binary short integer number
                ShortInt_Type
                UShortInt_Type
                                      Binary unsigned short integer number
                LongInt_Type
                                      Binary long integer number
                ULongInt_Type
                                      Binary unsigned long integer number
                LongLong_Type
                                      Binary 64-bit integer number
                UlongLong_Type
                                      Binary 64-bit unsigned integer
                number Float_Type
                                              Binary Floating point number
                Double_Type
                                      Binary Floating point number
                Buffer_Type
                                      Binary byte value
 Value
                The attribute's value will be returned here.
 ValueSize
                The size of the returned value is returned here.
 BufferSize
                The size of Value buffer in bytes. This parameter is type size_t
                on 64-bit Windows and int on all other platforms.
```

These functions fetch the next value of a multi-valued attribute. If no value has yet been fetched from this attribute, these functions behave identically to the functions of similar names, with "_Next" removed from the name. If one or more values have already been fetched from this attribute, the next value with be returned.

If **MC_Get_Next_Value** is used, the data type of the fetched value is specified by the *DataType* parameter. The other function names imply the data type requested. For example: **MC_Get_Next_Value_To_Int** is the same as calling **MC_Get_Next_Value** with *DataType* specified as **Int_Type**. Each function will return the value requested in *Value*, which must be prototyped as the

appropriate type. In the case of MC_Get_Next_Value_To_String and

MC_Get_Next_Value_To_UnicodeString, an additional parameter (*BufferSize*) is required to specify the size of the buffer to receive the string value. For **MC_Get_Next_Value**, *BufferSize* must be large enough to contain the data type requested.

For MC_Get_Next_Value_To_UnicodeString, it requires to call MC_Enable_Unicode_Conversion first. See documentation for MC_Enable_Unicode_Conversion.

Any reasonable conversion will be made from the attribute's value representation to the data type requested. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

NOTE: Attributes with a value representation of **SQ** (sequence of items) have values stored internally as integers - each value is the *Item/D* of an item object opened previously using **MC_Open_Item**.

Function	May be used to retrieve values from attributes with these Value Representations
MC_Get_Next_Value_To_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_Int	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_UInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_LongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_ULongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_UlongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Next_Value_To_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ
MC_Get_Next_Value_To_UnicodeString	LO, LT, PN, SH, ST, UC, UT

The same rules apply to the **MC_Get_Next_Value** function, based on the value used in the *DataType* parameter.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value

Meaning

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_NORMAL_COMPLETION	The function completed normally.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_NO_MORE_VALUES	There are no more values for this attribute.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_BUFFER_TOO_SMALL	Buffer is not large enough to contain the value. (MC_Get_Next_Value and MC_Get_Next_Value_To_String only)
MC_INCOMPATIBLE_VR	The function was called to retrieve an attribute value whose value representation (VR) is not one of those listed for the function in the table above.
MC_INCOMPATIBLE_VALUE	The attribute's value is not consistent with its VR. (This should never happen.)
MC_VALUE_OUT_OF_RANGE	The attribute numeric value is too large to be accommodated by the receiving data type. (E.g., using MC_Get_Value_To_Int when the value is 123.45; or instances where the value's sign (+ -) would be affected)
MC_INVALID_DATA_TYPE	The <i>DataType</i> parameter is invalid.
MC_CANNOT_COMPLY	The function fails to process the request. Consult Merge DICOM Toolkit log file for detail.
See Also	
MC_Get_pValue Functions MC_Get_Next_Value Functions MC_Enable_Unicode_Conversion	MC_Get_Next_pValue Functions MC_Get_Value_To_Function

MC_Get_Offset_Table_To_Function

Retrieves the basic offset table of encapsulated pixel data from a message.

```
#include "mc3msg.h"
MC_STATUS MC_Get_Offset_Table_To_Function (
    int MsgFileItemID,
    void *UserInfo,
    void *YourGetFunction,
)
```

MsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
UserInfo	Address of data which will be passed on to <i>YourGetFunction</i> each time it is called. This may be NULL.
YourGetFunction	Name of a function which will be called repeatedly to provide blocks of data of the offset table
The function must be p	rototyped as follows:
MC_STATUS YourGet int CBMsgFi unsigned lo void* CBuse int CBdataS void* CBdat int CBisFir int CBisLas) CBMsgEileltem/D	<pre>leItemID, ng CBtag, rInfo, 'ize, aBuffer, st,</pre>
CBMsgFileItemID	MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
CBtag	DICOM tag which identifies the attribute.
CBuserInfo	Address of user data which is being passed from the MC_Get_Offset_Table_To_Function function. This may be NULL.
CBdataSize	The number of bytes in <i>CBdataBuffer</i> .
CBdataBuffer	Address of the buffer containing a portion of the attribute's value.
CBisFirst	This is TRUE (not zero) the first time <i>YourGetFunction</i> is being called for this attribute.
CBisLast	This is TRUE (not zero) the last time <i>YourGetFunction</i> is being called for this attribute.

The **MC_Get_Offset_Table_To_Function** function is used to fetch a basic offset table of an encapsulated pixel data attribute.

The (*YourGetFunction*) function is used to provide a byte array which contains an offset table data. This byte array needs to be converted to array of integers, where every integer consists of four bytes. The first integer is the offset table length and all subsequential integer values are the corresponding frame offsets.

YourGetFunction

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

CBisFirst is set to TRUE the first time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the first time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often open a file at this time.

CBisLast is set to TRUE the last time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the last time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often close a file at this time.

YourGetFunction must return **MC_NORMAL_COMPLETION** if all went well. If not, it must return **MC_CANNOT_COMPLY**. Both of these are defined in "mc3msg.h".

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally, and there are no more frames in the message.
MC_NULL_POINTER_PARM	YourGetFunction is NULL
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_EMPTY_VALUE	No value has been assigned to this attribute yet.
MC_CALLBACK_CANNOT_COMPLY	YourGetFunction returned MC_CANNOT_COMPLY.

See Also

MC_Get_Encapsulated_Value_To_Function MC_Get_Next_Encapsulated_Value_To_Function MC_Get_Frame_To_Function MC_Register_Compression_Callbacks

MC_Get_pAttribute_Info

Retrieves information about a given private attribute in a message.

Synopsis

#include "mc3msg.h"		
<pre>MC_STATUS MC_Get_pAttribute_Info (int MsgFileItemID, char *PrivateCode, unsigned short Group, unsigned char ElementByte, MC_VR *ValueRep, int *NumValues)</pre>		
MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.	
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.	
Group	The number identifying the private group containing the private attribute. It must be an odd number.	
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .	
ValueRep	The attribute's value representation code is returned here. Possible return codes (defined in "mc3msg.h") are: AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR, UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL, OD, OF, SQ, UNKNOWN_VR	
NumValues	The number of values assigned to the attribute is placed here. (Note: will be one (1) if the attribute's value is NULL.)	

Remarks

MC_Get_pAttribute_Info retrieves information about a message attribute identified by *ElementByte* for the *PrivateCode* within the given *Group*. The attribute's Value Representation code is returned at *ValueRep*, and the number of values stored for the attribute is returned at *NumValues*. If the attribute is "empty", *NumValues* will contain zero (0). If the attribute's value is NULL, *NumValues* will contain one (1).

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value

Meaning

MC_NORMAL_COMPLETION

The function completed normally.

MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_PRIVATE_CODE	The message does not contain attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
See Also	
MC_Get_First_Attribute MC_Get_Attribute_Info	MC_Get_Next_Attribute

MC_Get_pTag_Info

Retrieves a descriptive string for a given private tag.

Synopsis

PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group containing the private attribute. It must be an odd number.
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .
Name	The tag's name is returned here.
NameLength	The length in bytes of <i>Name</i> .

Remarks

MC_Get_pTag_Info retrieves information about an attribute identified by *ElementByte* for the *PrivateCode* within the given *Group*. A descriptive test string for the attribute is returned at *Name*.

This function will only work with private attributes added to the Merge DICOM Toolkit's data dictionary through use of an extended toolkit.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TAG	The Merge DICOM Toolkit data dictionary does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_DDFILE_ERROR	An error occurred while trying to access the Merge DICOM Toolkit data dictionary. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_MISSING_CONFIG_PARM	This error occurs when the toolkit can not obtain the name of the data dictionary.
MC_BUFFER_TOO_SMALL	<i>Name</i> is not large enough to contain the descriptive string.
See Also	
MC_Get_Tag_Info	MC_Get_Attribute_Info MC_Get_pAttribute_Info

MC_Get_pTags_Dict_Info

Retrieves all dictionary information for a given private tag.

Synopsis

```
#include ``mc3msg.h"
```

```
MC_STATUS MC_Get_pTags_Dict_Info (
    char *PrivateCode,
    unsigned short Group,
    unsigned char ElementByte,
    char *Name,
    int NameSize,
    MC_VR* ValueRep,
    unsigned short *ValueMult_low,
    unsigned short *ValueMult_high
```

)

PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group containing the private attribute. It must be an odd number.
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .
Tag	A tag identifying a message attribute.
Name	A buffer to contain the tag's name.
NameSize	The size, in bytes, of <i>Name</i>

ValueRep	The attribute's value representation code is returned here.
	Possible return codes (defined in "mc3msg.h") are: AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR, UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL, OD, OF, SQ, UNKNOWN_VR
ValueMult_low	The minimum multiplicity will be returned here
ValueMult_high	The maximum multiplicity will be returned here

MC_Get_pTags_Dict_Info retrieves information about a private attribute identified by *ElementByte* for the *PrivateCode* within the given *Group*. The attribute's name is returned at *Name*, the alue representation code is returned at *ValueRep*, and the value multiplicities are returned in *ValueMult_low* and *ValueMult_high*.

This function will only work with private attributes added to the Merge DICOM Toolkit's data dictionary through use of an extended toolkit.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Name, ValueRep, ValueMult_low, ValueMult_High, and/</i> or <i>PrivateCode</i> are invalid.
MC_BUFFER_TOO_SMALL	NameSize is not large enough to contain Name.
MC_MISSING_CONFIG_PARM	This error occurs when the toolkit can not obtain the name of the data dictionary.
MC_DDFILE_ERROR	An error occurred while trying to access the Merge DICOM Toolkit data dictionary. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TAG	The Merge DICOM Toolkit data dictionary does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
See Also	
MC_Get_pTag_Info	MC_Get_Tags_Dict_Info

MC_Get_pValue... Functions

Retrieves the first or only value of a private attribute in a message object.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Get_pValue (
    int MsgFileItemID,
    char *PrivateCode,
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
unsigned short Group,
      unsigned char ElementByte,
      MC DT DataType,
      MC_size_t BufferSize,
      void *Value,
      int *ValueSize
MC_STATUS MC_Get_pValue_To_Float (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      float* Value
MC STATUS MC Get pValue To Double (
(
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      double *Value
MC STATUS MC_Get_pValue_To_ShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      short int *Value
MC STATUS MC Get pValue To UShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned short *Value
MC STATUS MC Get pValue To Int (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      int *Value
MC STATUS MC Get pValue To UInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned int *Value
)
MC_STATUS MC_Get_pValue_To_LongInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
```

```
long int *Value
)
MC STATUS MC Get pValue To ULongInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long *Value
)
MC STATUS MC_Get_pValue_To_LongLong (
(
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      long long *Value
)
MC STATUS MC Get pValue To ULongLong (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long long *Value
)
MC STATUS MC Get pValue To String (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC size t BufferSize,
      char *Value
MC STATUS MC Get pValue To UnicodeString (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC size t BufferSize,
      int *ValueSize,
      MC_Uhar *Value
MC_STATUS MC_Get_pValue_To_Buffer (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC size t BufferSize,
      void *Value,
      int *ValueSize
)
              The identifier assigned to this object by the
 MsgFileIteml
 D
              MC_Open_Message, MC_Open_Empty_Message,
              MC_Create_File, MC_Create_Empty_File, or
              MC_Open_Item functions.
```

PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.	
Group	The number identifying the private group containing the private attribute. It must be an odd number.	
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .	
DataType	One of the enumerated codes identifying the data type which should be used for the returned value in <i>Value</i> . The MC_DT enumerated type is defined in "mc3msg.h". They are:	
	String_Type	Null-terminated character string
	Int_Type	Binary integer number
	UInt_Type	Binary unsigned integer number
	ShortInt_Type	Binary short integer number
	UShortInt_Type	Binary unsigned short integer number
	LongInt_Type	Binary long integer number
	ULongInt_Type	Binary unsigned long integer number
	LongLong_Type	Binary 64-bit integer number
	UlongLong_Type	Binary 64-bit unsigned integer
	number Float_Type	Binary Floating point number
	Double_Type	Binary Floating point number
	Buffer_Type	Binary byte value
Value	The attribute's value will be returned here.	
BufferSize	The size of the <i>Value</i> buffer in bytes. This parameter is type size_t on 64-bit Windows and int on all other platforms.	
ValueSize	The size (in bytes) of the value returned in the <i>Value</i> buffer.	

These functions fetch the first (or only) value of a private attribute identified by *ElementByte* for the *PrivateCode* within the given private *Group*. If more than one value exists for this attribute, the first one will be returned. If **MC_Get_pValue** is used, the data type of the fetched value is specified by the *DataType* parameter. The other function names imply the data type requested. For example, **MC_Get_pValue_To_Int** is the same as calling **MC_Get_pValue** with *DataType* specified as **Int_Type**. Each function will return the value requested in *Value*, which must be prototyped as the appropriate type. For **MC_Get_pValue_To_String**, **MC_Get_pValue_To_UnicodeString** and **MC_Get_pValue_To_Buffer**, an additional parameter (*BufferSize*) is required to specify the size of the buffer to receive the string value. For **MC_Get_pValue**, *BufferSize* must be large enough to contain the data type requested.

For **MC_Get_pValue_To_UnicodeString**, it requires to call **MC_Enable_Unicode_Conversion** first. See documentation for **MC_Enable_Unicode_Conversion**.

MC_Get_pValue_To_Buffer is used to retrieve the value of a private attribute of a binary value representation of UN, OB, OW, OV, OL, OD or OF, or a text value representation of UR or UT, or an attribute whose value representation is unknown. This may occur if **MC_Stream_To_Message** is used and one or more of the stream attributes is unknown. Using **MC_Get_pValue_To_Buffer** for attributes

whose Value Representation is known will result in an error. The attribute's value is simply copied (memcpy) to the *Value* buffer and its length is placed at *ValueSize*.

Any reasonable conversion will be made from the attribute's value representation to the data type requested. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

NOTE: Attributes with a value representation of **SQ** (sequence of items) have values stored internally as integers - each value is the *Item/D* of an item object opened previously using **MC_Open_Item**.

Function	May be used to retrieve values from attributes with these Value Representations
MC_Get_pValue_To_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_Int	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_UInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_LongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_ULongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_UlongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_pValue_To_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ
MC_Get_pValue_To_UnicodeString	LO, LT, PN, SH, ST, UC, UT
MC_Get_pValue_To_Function	UNKNOWN_VR, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR, UT
MC_Get_pValue_To_Buffer	UNKNOWN_VR, OB, OW , OV, OL, OD, OF, UR, UT

The same rules apply to the **MC_Get_pValue** function, based on the value used in the *DataType* parameter. The **MC_Get_pValue_To_Function** call is described in its own section.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_BUFFER_TOO_SMALL	Buffer is not large enough to contain the value. (MC_Get_pValue, MC_Get_pValue_To_Buffer and MC_Get_pValue_To_String only)
MC_INCOMPATIBLE_VR	The function was called to retrieve an attribute value whose value representation (VR) is not one of those listed for the function in the table above.
MC_INCOMPATIBLE_VALUE	The attribute's value is not consistent with its VR. (This should never happen.)
MC_VALUE_OUT_OF_RANGE	The attribute numeric value is too large to be accommodated by the receiving data type. (E.g., using MC_Get_pValue_To_Int when the value is 123.45; or instances where the value's sign(+ -) would be affected)
MC_INVALID_DATA_TYPE	The <i>DataType</i> parameter is invalid.
See Also	
MC_Get_pValue_To_Function MC_Get_Next_pValue Functions MC_Get_Next_Value Functions	MC_Get_Value_To_Function MC_Get_Value Functions

MC_Get_pValue_Count

Returns the number of values assigned to a private attribute in a message object.

 $[\]ensuremath{\textcircled{O}}$ Copyright Merge Healthcare Solutions Inc. 2023

Merge DICOM Toolkit™ 5.16.0 C/C++ REFERENCE MANUAL

```
unsigned short Group,
unsigned char ElementByte,
int *CountPtr
```

)

MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group containing the private attribute. It must be an odd number.
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .
CountPtr	The number of values the attribute contains is returned here.

Remarks

MC_Get_pValue_Count returns the number of values assigned to a private attribute identified by *ElementByte* for the *PrivateCode* within the given *Group*. The number of returned at **CountPtr.*

Only certain attributes are allowed to contain multiple values and be conformant to the DICOM standard. The developer usually knows ahead of time which attributes will contain multiple values and which will not. Some attributes, such as those with value representations of OB or OW, may never contain more than one value.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>CountPtr</i> was NULL.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_NULL_VALUE	The attribute contains a NULL value (i.e. its value is of length zero).
MC_EMPTY_VALUE	No value has been assigned to the attribute.
See Also	
MC_Get_Value_Count	

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Get_pValue_Length

Returns the length of a private attribute value.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Get_pValue_Length (
       int MsgFileItemID,
       char *PrivateCode,
       unsigned short Group,
       unsigned char ElementByte,
       int ValueNumber,
       unsigned long *Length
)
 MsgFileItemI
                The identifier assigned to this object by the
                MC_Open_Message, MC_Open_Empty_Message,
 D
                MC_Create_File, MC_Create_Empty_File, or
                MC_Open_Item functions.
                The code string which identifies which block in the private
 PrivateCode
                Group "owns" the attribute.
 Group
                The number identifying the private group containing the
                private attribute. It must be an odd number.
 ElementByte
                The number identifying the private attribute within the private
                Group for this PrivateCode.
                Retrieve the length of this attribute value.
 ValueNumber
 Length
                The value's length is returned here.
```

Remarks

MC_Get_pValue_Length returns the length of a given value of an attribute in a message object. The size, in bytes, is returned at **Length*.

The function returns the length of value *ValueNumber*. The first value is number 1. Only certain attributes are allowed to contain multiple values and be conformant to the DICOM standard. The developer usually knows ahead of time which attributes will contain multiple values and which will not. Some attributes, such as those with value representations of OB or OW, may never contain more than one value.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_NULL_POINTER_PARM	<i>Length</i> or <i>PrivateCode</i> was NULL.

MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_NULL_VALUE	The attribute contains a NULL value (i.e. its value is of length zero).
MC_EMPTY_VALUE	No value has been assigned to the attribute.
MC_INVALID_VALUE_NUMBER	<i>ValueNumber</i> was negative or the attribute does not have that many values.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_TEMP_FILE_ERROR	File I/O error accessing value information in temporary files.
See Also	
MC_Get_Value_Length	

MC_Get_pValue_To_Function

Retrieves the value of a private attribute which has a binary value representation of OB, OW, OV, OL, OD or OF, or a numeric value representation of AT, SS, US, SL, UL, SV, UV, FL or FD, or a text value representation of UR or UT, from a message object.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Get_pValue_To_Function (
    int MsgFileItemID,
    char *PrivateCode,
    unsigned short Group,
    unsigned char ElementByte,
    void *UserInfo,
    MC_STATUS (*YourGetFunction)()
)
```

MsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group containing the private attribute. It must be an odd number.
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .
UserInfo	Address of data which will be passed on to <i>YourGetFunction</i> each time it is called. This may be NULL.
YourGetFunction	Name of a function which will be called repeatedly to provide the function with blocks of the attribute's value.

The function must be prototyped as follows:

```
MC_STATUS YourGetFunction (
```

```
int CBMsgFileItemID,
unsigned long CBtag,
void* CBuserInfo,
int CBdataSize,
void* CBdataBuffer,
int CBisFirst,
int CBisLast
```

```
)
```

CBMsgFileItemI D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
CBtag	DICOM tag which identifies the attribute.
CBuserInfo	Address of user data which is being passed from the MC_Get_pValue_To_Function function. This may be NULL.
CBdataSize	The number of bytes in <i>CBdataBuffer</i> .
CBdataBuffer	Address of the buffer containing a portion of the attribute's value.
CBisFirst	This is TRUE (not zero) the first time <i>YourGetFunction</i> is being called for this attribute.
CBisLast	This is TRUE (not zero) the last time <i>YourGetFunction</i> is being called for this attribute.

The **MC_Get_pValue_To_Function** function is used to fetch the value of a private attribute which has a value representation of UN, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR or UT. Such attributes tend to have values of great length. To accommodate this, one uses the **MC_Get_pValue_To_Function** function to specify the name of a function (*YourGetFunction*) which Merge DICOM Toolkit, in turn, calls. This "callback" function is called repeatedly to provide blocks of the attribute's data value.

An optional *UserInfo* parameter may be used to pass information between the **MC_Get_pValue_To_Function** caller and *YourGetFunction* which receives the data in its *CBuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

YourGetFunction

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

CBisFirst is set to TRUE the first time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the first time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often open a file at this time.

CBisLast is set to TRUE the last time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the last time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often close a file at this time.

YourGetFunction must return MC_NORMAL_COMPLETION if all went well. If not, it must return MC_CANNOT_COMPLY. Both of these are defined in "mc3msg.h".

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INCOMPATIBLE_VR	The attribute's value representation was not OB, OW, or OF.
MC_EMPTY_VALUE	No value has been assigned to this attribute yet.
MC_CALLBACK_CANNOT_COMPLY	YourGetFunction returned with MC_CANNOT_COMPLY,

See Also

MC_Get_Value	MC_Get_Value_To_Function
MC_Get_Value_To_Float	MC_Get_Value_To_Double
MC_Get_Value_To_ShortInt	MC_Get_Value_To_UShortInt
MC_Get_Value_To_LongInt	MC_Get_Value_To_Int
MC_Get_Value_To_ULongInt	MC_Get_Value_To_Uint
MC_Get_Value_To_NULL	MC_Get_Value_To_Empty
MC_Get_pValue	
MC_Get_pValue_To_Float	MC_Get_pValue_To_Double
MC_Get_pValue_To_ShortInt	MC_Get_pValue_To_UShortInt
MC_Get_pValue_To_LongInt	MC_Get_pValue_To_Int
MC_Get_pValue_To_ULongInt	MC_Get_pValue_To_Uint
MC_Get_pValue_To_NULL	MC_Get_pValue_To_Empty
MC_Get_Next_Value	
MC_Get_Next_Value_To_Float	MC_Get_Next_Value_To_Double
MC_Get_Next_Value_To_ShortInt	MC_Get_Next_Value_To_UShortInt
MC_Get_Next_Value_To_LongInt	MC_Get_Next_Value_To_Int
MC_Get_Next_Value_To_ULongInt	MC_Get_Next_Value_To_UInt
MC_Get_Next_pValue	
MC_Get_Next_pValue_To_Float	MC_Get_Next_pValue_To_Double
MC_Get_Next_pValue_To_ShortInt	MC_Get_Next_pValue_To_UShortInt
MC_Get_Next_pValue_To_LongInt	MC_Get_Next_pValue_To_Int
MC_Get_Next_pValue_To_ULongInt	MC_Get_Next_pValue_To_UInt

MC_Get_Stream_Length

Returns the size of a DICOM stream which would result from using the **MC_Message_To_Stream** function.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Get_Stream_Length (
    int MessageID,
    unsigned long StartTag,
    unsigned long StopTag,
    unsigned long *LengthPtr,
    TRANSFER_SYNTAX SyntaxType
)
MessageID The identifier assigned to this object by the
    MC_Open_Message function or the MC_Open_Item function.
StartTag The DICOM tag identifying the first attribute which would be
    streamed.
```

-	
StopTag	The DICOM tag identifying the last attribute which would be streamed
LengthPtr	Location where Merge DICOM Toolkit will return the stream length.
	streamed Location where Merge DICOM Toolkit will return the stream length. Specify which DICOM transfer syntax is to be assumed for the stream data. Use one of the enumerated TRANSFER_SYNTAX types defined in "mc3msg.h": IMPLICIT_LITTLE_ENDIAN IMPLICIT_BIG_ENDIAN EXPLICIT_BIG_ENDIAN EXPLICIT_BIG_ENDIAN DEFLATED_EXPLICIT_LITTLE_ENDIAN ENCAPSULATED_UNCOMPRESSED_ELE RLE JPEG_BASELINE JPEG_EXTENDED_2_4 JPEG_SPEC_NON_HIER_6_8 JPEG_SPEC_NON_HIER_7_9 JPEG_SPEC_NON_HIER_10_12 JPEG_FULL_PROG_NON_HIER_113 JPEG_LOSSLESS_NON_HIER_15 JPEG_LOSSLESS_NON_HIER_16_18 JPEG_SPEC_HIER_20_22 JPEG_SPEC_HIER_21_23 JPEG_FULL_PROG_HIER_24_26 JPEG_SPEC_HIER_24_26 JPEG_LOSSLESS_HIER_28 JPEG_LOSSLESS_HIER_24 JPEG_LOSSLESS_HIER_14 JPEG_LOSSLESS_HIER_14 JPEG_2000 JPEG_LS_LOSSLESS
	JPEG_LS_LOSSY HEVC_H265_M10P_LEVEL_5_1 HEVC_H265_MP_LEVEL_5_1
	JPIP_REFERENCED JPIP_REFERENCED_DEFLATE MPEG2_MPML
	MPEG2_MPHL MPEG4_AVC_H264_HP_LEVEL_4_1 MPEG4_AVC_H264_BDC_HP_LEVEL_4_1 MPEG4_AVC_H264_HP_LEVEL_4_2_2D MPEG4_AVC_H264_HP_LEVEL_4_2_3D

MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2 JPEG_2000_MC_LOSSLESS_ONLY JPEG_2000_MC SMPTE_ST_2110_20_UNCOMPRESSED_PROGRESSIVE_A CTIVE VIDEO SMPTE_ST_2110_20_UNCOMPRESSED_INTERLACED_AC TIVE VIDEO SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO PRIVATE_SYNTAX_1 PRIVATE_SYNTAX_2

Remarks

MC_Get_Stream_Length returns the size of a DICOM stream which would result from using the MC_Message_To_Stream function for the message identified by MessageID. The calculation is based on the assumption that the stream would be encoded using SyntaxType DICOM transfer syntax. It is also assumed that the stream would contain attributes with tags greater than or equal to StartTag and less than or equal to StopTag. The length is returned to the unsigned long at *LengthPtr.

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_INVALID_TRANSFER_SYNTAX	<i>SyntaxType</i> is not one of the TRANSFER_SYNTAX codes defined in "mc3msg.h".
MC_NULL_POINTER_PARM	<i>LengthPtr</i> was NULL.
See Also	

MC_Message_To_Stream

MC_Get_String_Config_Value

Used to get the value of a character string toolkit configuration parameter at runtime.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Get_String_Config_Value (
      StringParm Parm,
      int ValueLength,
      char *Value
)
 Parm
```

An enumerated constant identifying the character string configuration parameter to get. Parm can have any of the following values:

MERGECOM 3 PROFILE MERGECOM 3 SERVICES MERGECOM_3_APPLICATIONS CAPTURE FILE COMPRESSION RGB TRANSFORM FORMAT DECODER_TAG_FILTER DEFLATED EXPLICIT LITTLE ENDIAN SYNTAX DICTIONARY_ACCESS DICTIONARY_FILE ENCAPSULATED UNCOMPRESSED ELE SYNTAX EXPLICIT BIG ENDIAN SYNTAX EXPLICIT LITTLE ENDIAN SYNTAX HEVC H265 M10P LEVEL 5 1 SYNTAX HEVC_H265_MP_LEVEL_5_1_SYNTAX IMPLEMENTATION CLASS UID IMPLEMENTATION VERSION IMPLICIT BIG ENDIAN SYNTAX IMPLICIT_LITTLE_ENDIAN_SYNTAX INITIATOR NAME **IP_TYPE** JPEG_2000_LOSSLESS_ONLY_SYNTAX JPEG 2000 MC LOSSLESS ONLY SYNTAX JPEG 2000 MC SYNTAX JPEG 2000_SYNTAX JPEG BASELINE_SYNTAX JPEG EXTENDED 2 4 SYNTAX JPEG EXTENDED 3 5 SYNTAX JPEG EXTENDED HIER 16 18 SYNTAX JPEG EXTENDED HIER 17 19 SYNTAX JPEG FULL PROG HIER 24 26 SYNTAX JPEG_FULL_PROG_HIER_25_27_SYNTAX JPEG FULL PROG NON HIER 10 12 SYNTAX JPEG FULL PROG NON HIER 11 13 SYNTAX JPEG LOSSLESS HIER 14 SYNTAX JPEG LOSSLESS HIER 28 SYNTAX JPEG_LOSSLESS_HIER_29_SYNTAX JPEG LOSSLESS NON HIER 14 SYNTAX JPEG LOSSLESS NON HIER 15 SYNTAX JPEG LS LOSSLESS SYNTAX JPEG LS LOSSY SYNTAX JPEG SPEC HIER 20 22 SYNTAX JPEG_SPEC_HIER_21_23_SYNTAX JPEG SPEC NON HIER 6 8 SYNTAX JPEG SPEC NON HIER 7 9 SYNTAX JPIP REFERENCED DEFLATE SYNTAX JPIP REFERENCED SYNTAX LARGE DATA_STORE LICENSE LOCAL APPL CONTEXT NAME LOG FILE MPEG2_MPHL_SYNTAX MPEG2 MPML SYNTAX MPEG4_AVC_H264_BDC_HP_LEVEL_4_1_SYNTAX MPEG4_AVC_H264_HP_LEVEL_4_1_SYNTAX MPEG4_AVC_H264_HP_LEVEL_4_2_2D_SYNTAX

	MPEG4_AVC_H264_HP_LEVEL_4_2_3D_SYNTAX MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2_SYNTAX MSG_INFO_FILE NULL_TYPE3_VALIDATION PEGASUS_DISP_REG_NAME PEGASUS_OP_D2SEPLUS_NAME PEGASUS_OP_D2SEPLUS_REGISTRATION PEGASUS_OP_J2KE_NAME PEGASUS_OP_J2KE_REGISTRATION PEGASUS_OP_J2KE_REGISTRATION PEGASUS_OP_J2KP_REGISTRATION PEGASUS_OP_JLSE_REGISTRATION PEGASUS_OP_JLSE_REGISTRATION PEGASUS_OP_JLSP_REGISTRATION PEGASUS_OP_JLSP_REGISTRATION PEGASUS_OP_JLSP_REGISTRATION PEGASUS_OP_LIE3PLUS_REGISTRATION PEGASUS_OP_LIE3PLUS_REGISTRATION PEGASUS_OP_LIE3PLUS_REGISTRATION PEGASUS_OP_LIP3PLUS_REGISTRATION PEGASUS_OP_LIP3PLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OP_SE2DPLUS_REGISTRATION PEGASUS_OPCODE_PATH PRIVATE_SYNTAX_1_SYNTAX PRIVATE_SYNTAX_2_SYNTAX RECEIVER_NAME RLE_SYNTAX SMPTE_ST_2110_20_UNCOMPRESSED_PROGRESSIVE_ ACTIVE_VIDEO_SYNTAX SMPTE_ST_2110_20_UNCOMPRESSED_INTERLACED_A CTIVE_VIDEO_SYNTAX SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO_SYNTAX TEMP_FILE_DIRECTORY UNKNOWN_VR_CODE These names are the same as those given to the
	parameters in the toolkit configuration files.
ValueLength	The length of <i>Value</i> .
	-

Value

The Merge DICOM Toolkit Library accesses several configuration files at startup. This call allows your application to get character string configurable parameters specified in these files at runtime. Please see the detailed description of toolkit configuration elsewhere in this manual.

A character string to copy the value of Parm into.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.	

MC_NULL_POINTER_PARM	<i>Value</i> was a null pointer.
MC_BUFFER_TOO_SMALL	<i>Value</i> is not large enough to contain the configuration string.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Alee	

See Also

MC_Get_Int_Config_Value MC_Get_Log_Destination MC_Get_Long_Config_Value MC_Set_String_Config_Value MC_Set_Int_Config_Value MC_Set_Log_Destination MC_Set_Long_Config_Value

MC_Get_Tag_Info

Retrieves a descriptive string for a given tag.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Get_Tag_Info (
    unsigned long Tag,
    char *Name,
    int NameLength
```

```
)
```

Tag	A tag identifying a message attribute.
Name	The tag's name is returned here.

NameLength The length in bytes of *Name*.

Remarks

MC_Get_Tag_Info retrieves information about an attribute identified by *Tag*. A descriptive text string for the attribute is returned at *Name*.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_INVALID_TAG	The attribute identified by <i>Tag</i> is not a valid DICOM attribute.	
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.	
MC_DDFILE_ERROR	An error occurred while trying to access the Merge DICOM Toolkit data dictionary. A message describing the error has been written to the Merge DICOM Toolkit log file.	
MC_MISSING_CONFIG_PARM	This error occurs when the toolkit can not obtain the name of the data dictionary.	
MC_BUFFER_TOO_SMALL	<i>Name</i> is not large enough to contain the descriptive string.	

See Also

```
MC_Get_pTag_Info
```

MC_Get_Attribute_Info MC_Get_pAttribute_Info

MC_Get_Tags_Dict_Info

Retrieves all dictionary information for a given tag.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Get_Tags_Dict_Info (
    unsigned long Tag,
    char *Name,
    int NameSize,
    MC_VR *ValueRep,
    unsigned short *ValueMult_low,
    unsigned short *ValueMult_high
)
```

)

Tag	A tag identifying a message attribute.		
Name	A buffer to contain the tag's name.		
NameSize	The size, in bytes, of <i>Name</i>		
ValueRep	The attribute's value representation code is returned here.		
	Possible return codes (defined in "mc3msg.h") are: AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR, UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL, OD, OF, SQ, UNKNOWN_VR		
ValueMult_low	The minimum multiplicity will be returned here		
ValueMult_high	The maximum multiplicity will be returned here		

Remarks

MC_Get_Tags_Dict_Info retrieves information about an attribute identified by *Tag*. The attribute's Name is returned at *Name, v*alue representation code is returned at *ValueRep*, and the value multiplicities are returned in *ValueMult_low* and *ValueMult_high*. The values returned are from the dictionary file.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Name, ValueRep, ValueMult_low, ValueMult_High, and/</i> or <i>Tag</i> are invalid.
MC_BUFFER_TOO_SMALL	NameSize is not large enough to contain Name.
MC_MISSING_CONFIG_PARM DICTIONARY_FILE is invalid	
MC_DDFILE_ERROR	Data Dictionary problem

© Copyright Merge Healthcare Solutions Inc. 2023

MC_INVALID_TAG

Tag is not found

See Also

MC_Get_Attribute_Info

MC_Get_Tag_Keyword

Generates attribute's keyword for a given tag.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Get_Tag_Keyword (
    unsigned long Tag,
    char *Keyword,
    int Length
)
```

Tag	A tag identifying a message attribute.
Keyword	A buffer to contain the tag's keyword.
Length	The size, in bytes, of <i>Keyword</i>

Remarks

MC_Get_Tag_Keyword generates a keyword for standard attributes based on the name of the data element as described by section 6 part 6 in the DICOM standard. The attribute's Keyword is returned at *Keyword* and the length of the keyword buffer is returned at *Length*.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_INVALID_TAG	Tag is not found or private tag	
MC_NULL_POINTER_PARM	<i>Keyword, Length</i> and/or <i>Tag</i> are invalid.	
MC_BUFFER_TOO_SMALL	Keyword Size is not large enough to contain Keyword.	
MC_MISSING_CONFIG_PARM	DICTIONARY_FILE is invalid	
MC_DDFILE_ERROR	Data Dictionary problem	
MC_INVALID_TAG	Tag is not found	

MC_Get_Transfer_Syntax_From_Enum

Gets the DICOM UID for a transfer syntax.

Synopsis

#include ``mc3msg.h"

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_STATUS MC_Get_Transfer_Syntax_From_Enum (
TRANSFER_SYNTAX SyntaxType,					
char *Uid, int UidLength					
)	allengen				
SyntaxType	Specify the DICOM transfer syntax for which to find a UID.				
	Use one of the enumerated TRANSFER_SYNTAX types				
	defined in mc3msg.h:				
	IMPLICIT_LITTLE_ENDIAN				
	IMPLICIT_BIG_ENDIAN				
	EXPLICIT_LITTLE_ENDIAN				
	EXPLICIT_BIG_ENDIAN				
	DEFLATED_EXPLICIT_LITTLE_ENDIAN				
	ENCAPSULATED_UNCOMPRESSED_ELE				
	RLE				
	JPEG_BASELINE				
	JPEG_EXTENDED_2_4				
	JPEG_EXTENDED_3_5				
	JPEG_SPEC_NON_HIER_6_8				
	JPEG_SPEC_NON_HIER_7_9				
	JPEG_FULL_PROG_NON_HIER_10_12				
	JPEG_FULL_PROG_NON_HIER_11_13				
	JPEG_LOSSLESS_NON_HIER_14				
	JPEG_LOSSLESS_NON_HIER_15				
	JPEG_EXTENDED_HIER_16_18				
	JPEG_EXTENDED_HIER_17_19				
	JPEG_SPEC_HIER_20_22				
	JPEG_SPEC_HIER_21_23				
	JPEG_FULL_PROG_HIER_24_26				
	JPEG_FULL_PROG_HIER_25_27				
	JPEG_LOSSLESS_HIER_28				
	JPEG_LOSSLESS_HIER_29				
	JPEG_LOSSLESS_HIER_14				
	JPEG_2000_LOSSLESS_ONLY				
	JPEG_2000				
	JPEG_LS_LOSSLESS				
	JPEG_LS_LOSSY				
	HEVC_H265_M10P_LEVEL_5_1				
	HEVC_H265_MP_LEVEL_5_1				
	JPIP_REFERENCED				
JPIP_REFERENCED_DEFLATE					
	MPEG2_MPML				
	MPEG2_MPHL				
	MPEG4_AVC_H264_HP_LEVEL_4_1				
	MPEG4_AVC_H264_BDC_HP_LEVEL_4_1				
	MPEG4_AVC_H264_HP_LEVEL_4_2_2D				

	MPEG4_AVC_H264_HP_LEVEL_4_2_3D
	MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2,
	JPEG_2000_MC_LOSSLESS_ONLY
	JPEG_2000_MC
	SMPTE_ST_2110_20_UNCOMPRESSED_PROGRESSIVE_ACTI
	VE_VIDEO
	SMPTE_ST_2110_20_UNCOMPRESSED_INTERLACED_ACTIV
	E_VIDEO
	SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO
	PRIVATE_SYNTAX_1
	PRIVATE_SYNTAX_2
Uid	The transfer syntax's UID is returned here.
UidLength	The length in bytes of <i>Uid</i> .

MC_Get_Transfer_Syntax_From_Enum converts a Merge DICOM Toolkit enumerated value that represents a DICOM transfer syntax into the UID for that transfer syntax. The UID values used by **MC_Get_Transfer_Syntax_From_Enum** are specified in Merge DICOM Toolkit's configuration files. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL. <i>SyntaxName</i> is not registered in the Merge DICOM Toolkit configuration files.	
MC_INVALID_TRANSFER_SYNTAX		
MC_BUFFER_TOO_SMALL	<i>Uid</i> is not large enough to contain the unique identifier.	
See Also		
MC_Get_Enum_From_Transfer_Syntax MC_Get_MergeCOM_Service MC_Get_UID_From_MergeCOM_Service		

MC_Get_UID_From_MergeCOM_Service

Converts a Merge DICOM Toolkit service name into a DICOM SOP Class UID.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Get_UID_From_MergeCOM_Service (
    char *ServiceName,
    char *Uid,
```

)

int UidLength

ServiceName	String name of a DICOM service
Uid	The service's SOP Class UID is returned here.
UidLength	The length in bytes of <i>Uid</i> buffer.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_UNKNOWN_SERVICE	<i>ServiceName</i> was not registered in the Merge DICOM Toolkit configuration files.
MC_BUFFER_TOO_SMALL	<i>Uid</i> is not large enough to contain the unique identifier.
See Also	

MC_Get_MergeCOM_Service MC_Get_MergeCOM_Service_From_UID MC_Get_Enum_From_Transfer_Syntax MC_Get_Transfer_Syntax_From_Enum

MC_Get_User_Identity_Info

Returns the contents of a user identity association negotiation related field.

```
Synopsis
#include ``mergecom.h"
typedef enum {
      PRIMARY FIELD,
      SECONDARY FIELD,
      SERVER RESPONSE FIELD
} USER_IDENTITY_FIELD_TYPE;
MC STATUS MC_Get_User_Identity_Info (
      int AssocID,
      USER_IDENTITY_FIELD_TYPE FieldType,
      void *Field,
      unsigned short FieldLength
)
 AssocID
               The identifier assigned to this object by the
               MC_Open_Association, MC_Open_Secure_Association,
               MC_Process_Association_Request,
               MC_Process_Secure_Association_Request,
               MC_Wait_For_Association or
               MC_Wait_For_Secure_Association function.
```

FieldType	An enumerated value specifying the User Identity related field to retrieve.
Field	A buffer into which the user identity field specified by <i>FieldType</i> is copied into. Note that if <i>Field</i> is a text field, a NULL termination character will not be included in the value returned in this parameter.
FieldLength	The length of <i>Field</i> in bytes.

MC_Get_User_Identity_Info returns the value of a field used in user identity association negotiation as defined in DICOM Supplement 99. It is used by an application accepting associations to determine the user identity related fields in an association request (the **PRIMARY_FIELD** and **SECONDARY_FIELD**), and it is used by applications opening associations to retrieve the user identity server response, if applicable (the **SERVER_RESPONSE_FIELD**).

The structure returned by the **MC_Get_Association_Info** routine contains general information about the user identity negotiation such as the type of user identity information contained in the association negotiation, and if a server response has been requested in the association negotiation. For user identity types that contain additional fields, it is recommended that the

MC_Get_User_Identity_Length routine be first used to determine the length of a specific user identity field. The **MC_Get_User_Identity_Info** routine can then be used to retrieve the actual field's value.

This function is to be called after an MC_Open_Association, MC_Open_Secure_Association, MC_Wait_For_Association or MC_Wait_For_Secure_Association call, and before any MC_Read_Message, MC_Send_Request_Message, MC_Send_Request, MC_Send_Response_Message or MC_Send_Response calls.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Field</i> was NULL.
MC_BUFFER_TOO_SMALL	<i>Field</i> was not large enough to contain the user identity field requested.
MC_INVALID_ASSOC_ID	AssocID is not a valid association object ID.
MC_STATE_VIOLATION	The call was made after messages have been exchanged over the association.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Open_Association_With_Identity MC_Accept_Association_With_Identity MC_Get_User_Identity_Length MC_Get_Association_Info

MC_Get_User_Identity_Length

Returns the length of a user identity association negotiation related field.

Synopsis

```
#include ``mergecom.h"
typedef enum {
    PRIMARY FIELD,
    SECONDARY FIELD,
    SERVER RESPONSE FIELD
} USER IDENTITY FIELD TYPE;
MC STATUS MC Get User Identity Length (
       int AssocID,
      USER IDENTITY FIELD TYPE FieldType,
      unsigned short *FieldLength
)
 AssocID
                The identifier assigned to this object by the
                MC_Open_Association, MC_Open_Secure_Association,
                MC_Process_Association_Request,
                MC_Process_Secure_Association_Request,
                MC_Wait_For_Association or
                MC_Wait_For_Secure_Association function.
 FieldType
                An enumerated value specifying the User Identity related field
                to retrieve.
 FieldLength
                The length of FieldType in bytes is returned here.
```

Remarks

MC_Get_User_Identity_Length returns the length of a field used in user identity association negotiation as defined in DICOM Supplement 99. It is used by an application accepting associations to determine the length of the user identity related fields in an association request (the **PRIMARY_FIELD** and **SECONDARY_FIELD**), and it is used by applications opening associations to determine the length of the user response, if applicable (the **SERVER_RESPONSE_FIELD**).

The structure returned by the **MC_Get_Association_Info** routine contains general information about the user identity negotiation such as the type of user identity information contained in the association negotiation, and if a server response has been requested in the association negotiation. For user identity types that contain additional fields, it is recommended that the

MC_Get_User_Identity_Length routine be first used to determine the length of a specific user identity field. The **MC_Get_User_Identity_Info** routine can then be used to retrieve the actual field's value.

This function is to be called after an MC_Open_Association, MC_Open_Secure_Association, MC_Wait_For_Association or MC_Wait_For_Secure_Association call, and before any MC_Read_Message, MC_Send_Request_Message, MC_Send_Request, MC_Send_Response_Message or MC_Send_Response calls.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	FieldLength was NULL.
MC_INVALID_ASSOC_ID	AssocID is not a valid association object ID.
MC_STATE_VIOLATION	The call was made after messages have been exchanged over the association.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Open_Association_With_Identity MC_Accept_Association_With_Identity MC_Get_User_Identity_Info MC_Get_Association_Info

MC_Get_Value... Functions

Retrieves the first or only value of an attribute in a message object.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Get Value (
      int MsgFileItemID,
      unsigned long Tag,
      MC_DT DataType,
      MC size t BufferSize,
      void *Value,
      int *ValueSize
MC STATUS MC Get Value To Float (
      int MsgFileItemID,
      unsigned long Tag,
      float *Value
)
MC STATUS MC Get Value To Double (
      int MsgFileItemID,
      unsigned long Tag,
      double *Value
)
```

```
MC_STATUS MC_Get_Value_To_ShortInt (
      int MsgFileItemID,
      unsigned long Tag,
      short int *Value
MC STATUS MC Get Value To UShortInt (
      int MsqFileItemID,
      unsigned long Tag,
      unsigned short *Value
MC STATUS MC Get Value To Int (
      int MsgFileItemID,
      unsigned long Tag,
      int *Value
MC STATUS MC Get Value To UInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned int *Value
MC_STATUS MC_Get_Value_To_LongInt (
      int MsgFileItemID,
      unsigned long Tag,
      long int *Value
)
MC STATUS MC Get Value To ULongInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long *Value
MC STATUS MC Get Value To LongLong (
      int MsgFileItemID,
      unsigned long Tag,
      long long *Value
MC STATUS MC Get Value To ULongLong (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long long *Value
MC_STATUS MC_Get_Value_To_String (
      int MsgFileItemID,
      unsigned long Tag,
      MC size t BufferSize,
      char *Value
MC STATUS MC Get Value To UnicodeString (
      int MsgFileItemID,
      unsigned long Tag,
      MC_size_t BufferSize,
      int *ValueSize,
      MC Uhar *Value
MC STATUS MC Get Value To Buffer (
      int MsgFileItemID,
      unsigned long Tag,
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

<pre>MC_size_t BufferSize, char *Value, int *ValueSize</pre>		
)		
MsgFileItemID	• •	MC_Open_Empty_Message, Create_Empty_File, or
Tag	DICOM tag which iden	tifies the attribute.
DataType	One of the enumerated codes identifying the data type which should be used for the returned value in <i>Value</i> . The MC_DT enumerated type is defined in "mc3msg.h". They are:	
	String_Type Int_Type UInt_Type ShortInt_Type UShortInt_Type number LongInt_Type ULongInt_Type LongLong_Type UlongLong_Type number Float_Type Double_Type Buffer_Type	Null-terminated character string Binary integer number Binary unsigned integer number Binary short integer number Binary unsigned short integer Binary long integer number Binary unsigned long integer number Binary 64-bit integer number Binary 64-bit unsigned integer Binary Floating point number Binary Floating point number Binary byte value
Value	The attribute's value w	
BufferSize	The size of the <i>Value</i> buffer in bytes. This parameter is type size_t on 64-bit Windows and int on all other platforms.	
ValueSize		ne value returned in the <i>Value</i> buffer. If 9e , the length of the string (not ull) is returned.

These functions fetch the first (or only) value of an attribute with the given *Tag.* If more than one value exists for this attribute, the first one will be returned. If **MC_Get_Value** is used, the data type of the fetched value is specified by the *DataType* parameter. The other function names imply the data type requested. For example: **MC_Get_Value_To_Int** is the same as calling **MC_Get_Value** with *DataType* specified as **Int_Type**. Each function will return the value requested in *Value*, which must be prototyped as the appropriate type. For **MC_Get_Value_To_String**,

MC_Get_Value_To_UnicodeString and MC_Get_Value_To_Buffer, an additional parameter (*BufferSize*) is required to specify the size of the buffer to receive the string value. For MC_Get_Value, *BufferSize* must be large enough to contain the data type requested.

For **MC_Get_Value_To_UnicodeString**, it requires to call **MC_Enable_Unicode_Conversion** first. See documentation for **MC_Enable_Unicode_Conversion**.

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Get_Value_To_Buffer is used to retrieve the value of a standard attribute of a binary value representation of UN, OB, OW, OV, OL, OD or OF, or a text value representation of UR or UT, or an attribute whose value vepresentation is unknown. Using **MC_Get_Value_To_Buffer** for attributes with other Value Representations will result in an error. The attribute's value is simply copied (memcpy) to the *Value* buffer and its length is placed at *ValueSize*.

Any reasonable conversion will be made from the attribute's value representation to the data type requested. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

NOTE: Attributes with a value representation of **SQ** (sequence of items) have values stored internally as integers - each value is the *Item/D* of an item object opened previously using **MC_Open_Item**.

Function	May be used to retrieve values from attributes with these Value Representations
MC_Get_Value_To_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_Int	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_UInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_LongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_UlongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_UlongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Get_Value_To_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ
MC_Get_Value_To_UnicodeString	LO, LT, PN, SH, ST, UC, UT
MC_Get_Value_To_Function	UNKNOWN_VR, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR, UT
MC_Get_Value_To_Buffer	UNKNOWN_VR, OB, OW, OV, OL, OD, OF, UR, UT

The same rules apply to the **MC_Get_Value** function, based on the value used in the *DataType* parameter. The **MC_Get_Value_To_Function** call is described in its own section.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. it's length is zero).
MC_BUFFER_TOO_SMALL	Buffer is not large enough to contain the value. (MC_Get_Value, MC_Get_Value_To_Buffer and MC_Get_Value_To_String only)
MC_INCOMPATIBLE_VR	The function was called to retrieve an attribute value whose value representation (VR) is not one of those listed for the function in the table above.
MC_INCOMPATIBLE_VALUE	The attribute's value is not consistent with its VR. (This should never happen.)
MC_VALUE_OUT_OF_RANGE	The attribute numeric value is too large to be accommodated by the receiving data type. (E.g. Using MC_Get_Value_To_Int when the value is 123.45; or instances where the value's sign(+ -) would be affected)
MC_INVALID_DATA_TYPE	The <i>DataType</i> parameter is invalid.
MC_CANNOT_COMPLY	The function fails to process the request. Consult Merge DICOM Toolkit log file for detail.
See Also	
MC_Get_pValue Functions MC_Get_Next_pValue Functions MC_Get_Next_Value Functions	MC_Get_pValue_To_Function MC_Get_Value_To_Function MC_Enable_Unicode_Conversion

MC_Get_Value_Count

Returns the number of values assigned to an attribute in a message object.

Synopsis

```
#include "mc3msg.h"
MC_STATUS MC_Get_Value_Count (
    int MsgFileItemID,
    unsigned long Tag,
    int *CountPtr
)
```

MsgFileItemID	The identifier assigned to this object by the
	MC_Open_Message, MC_Open_Empty_Message,
	MC_Create_File, MC_Create_Empty_File, or
	MC_Open_Item functions.
Tag	DICOM tag which identifies the attribute.
CountPtr	The number of values the attribute contains is returned here.

MC_Get_Value_Count returns the number of values assigned to an attribute in a message object. The number is returned at **CountPtr.*

Only certain attributes are allowed to contain multiple values and be conformant to the DICOM standard. The developer usually knows ahead of time which attributes will contain multiple values and which will not. Some attributes, such as those with value representations of OB or OW, may never contain more than one value.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>CountPtr</i> was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_NULL_VALUE	The attribute contains a NULL value (i.e. its value is of length zero).
MC_EMPTY_VALUE	No value has been assigned to the attribute.
See Also	
MC_Get_pValue_Count	

MC_Get_Value_Length

Returns the length of an attribute value.

Synopsis

MsgFileItemID	The identifier assigned to this object by the
	MC_Open_Message, MC_Open_Empty_Message,
	MC_Create_File, MC_Create_Empty_File, or
	MC_Open_Item functions.
Tag	DICOM tag which identifies the attribute.
ValueNumber	Retrieve the length of this attribute value.
Length	The value's length is returned here.

MC_Get_Value_Length returns the length of a given value of an attribute in a message object. The size, in bytes, is returned at **Length*.

The function returns the length of value *ValueNumber*. The first value is number 1. Only certain attributes are allowed to contain multiple values and be conformant to the DICOM standard. The developer usually knows ahead of time which attributes will contain multiple values and which will not. Some attributes, such as those with value representations of OB or OW, may never contain more than one value.

For pixel data tag which has a default value representation of OW, the returned length can be odd or even depending on the actual size of pixel data and the effective VR. For example, if pixel data is generated by uncompressing encapsulated data with 8 bits allocation per pixel, the effective VR is set to OB. If actual uncompressed data size is odd, the function returns an odd size length (No padding has been applied). See **MC_Get_Attribute_Info** for retrieving the effective VR of an attribute. However, if pixel data is retrieved from a file or stream, the length is always even size due to padding.

If the attribute has a value representation of SQ, the number of items in the sequence is returned in **Length*.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	Length was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_NULL_VALUE	The attribute contains a NULL value (i.e. its value is of length zero).
MC_EMPTY_VALUE	No value has been assigned to the attribute.
MC_INVALID_VALUE_NUMBER	<i>ValueNumber</i> was negative, or the attribute does not have that many values.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_TEMP_FILE_ERROR

File I/O error accessing value information in temporary files.

See Also

MC_Get_pValue_Length

MC_Get_Value_To_Function

Retrieves the value of an attribute which has a binary value representation of OB, OW, OV, OL, OD or OF, or a numeric value representation of AT, SS, US, SL, UL, SV, UV, FL or FD, or a text value representation of UR or UT, from a message object.

Synopsis

```
MsgFileItemIDThe identifier assigned to this object by the<br/>MC_Open_Message, MC_Open_Empty_Message,<br/>MC_Create_File, MC_Create_Empty_File, or<br/>MC_Open_Item functions.TagDICOM tag which identifies the attribute.UserInfoAddress of data which will be passed on to<br/>YourGetFunction each time it is called. This may be<br/>NULL.YourGetFunctionName of a function which will be called repeatedly to<br/>provide the function with blocks of the attribute's value.
```

The function must be prototyped as follows:

```
MC STATUS YourGetFunction (
       int CBMsqFileItemID,
      unsigned long CBtag,
      void* CBuserInfo,
      int CBdataSize,
      void* CBdataBuffer,
      int CBisFirst,
      int CBisLast
)
 CBMsgFileItemID
                    The identifier assigned to this object by the
                    MC_Open_Message, MC_Open_Empty_Message,
                     MC_Create_File, MC_Create_Empty_File, or
                     MC_Open_Item functions.
 CBtag
                     DICOM tag which identifies the attribute.
```

CBuserInfo	Address of user data which is being passed from the MC_Get_Value_To_Function function. This may be NULL.
CBdataSize	The number of bytes in <i>CBdataBuffer</i> .
CBdataBuffer	Address of the buffer containing a portion of the attribute's value.
CBisFirst	This is TRUE (not zero) the first time <i>YourGetFunction</i> is being called for this attribute.
CBisLast	This is TRUE (not zero) the last time <i>YourGetFunction</i> is being called for this attribute.

The **MC_Get_Value_To_Function** function is used to fetch the value of an attribute which has a value representation of UN, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR or UT. Such attributes tend to have values of great length. This function can also be used to fetch the value of an attribute which has a numeric value representation (SL, SS, SV, UL, US, UV, AT, FL, FD).

To accommodate this, one uses the **MC_Get_Value_To_Function** function to specify the name of a function (*YourGetFunction*) which Merge DICOM Toolkit, in turn, will call. This "callback" function is called repeatedly to provide blocks of the attribute's data value. An optional *UserInfo* parameter may be used to pass information between the **MC_Get_Value_To_Function** caller and *YourGetFunction* which receives the data in its *CBuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

YourGetFunction

Each block of data is returned in *CBdataBuffer* and the number of bytes in the block is specified by *CBdataSize*.

CBisFirst is set to TRUE the first time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the first time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often open a file at this time.

CBisLast is set to TRUE the last time *YourGetFunction* is called. This provides a clear mechanism for the function to know it is being called the last time for the attribute identified by *CBtag* in the message identified by *CBMsgFileItemID*. Users often close a file at this time.

YourGetFunction must return **MC_NORMAL_COMPLETION** if all went well. If not, it must return **MC_CANNOT_COMPLY**. Both of these are defined in "mc3msg.h".

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .

MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INCOMPATIBLE_VR	The attribute's value representation was not OB, OW, OF, SL, SS, UL, US, AT, FL or FD.
MC_EMPTY_VALUE	No value has been assigned to this attribute yet.
MC_CALLBACK_CANNOT_COMPLY	
	YourGetFunction returned with MC_CANNOT_COMPLY.
See Also	
MC_Get_Value Functions	MC_Get_pValue Functions
MC_Get_pValue_To_Function	MC_Get_Next_Value Functions
MC_Get_Next_pValue Functions	

MC_Get_Version_String

Retrieves a descriptive string containing the toolkit version.

Synopsis

```
char *Version
```

)

VersionLengt The length in bytes of Version.

h

Version The Merge DICOM Toolkit version is returned here.

Remarks

MC_Get_Version_String retrieves a text string containing Merge DICOM Toolkit's version.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>Version</i> was NULL.
MC_BUFFER_TOO_SMALL	Version is not large enough to contain the descriptive string.
See Also	

MC_Library_Initialization

MC_Json_To_Message

Reads attribute values from a DICOM JSON Model string into a message, file or item object.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Synopsis

MessageID

```
#include "mc3msg.h"
MC_STATUS MC_Json_To_Message (
    int MessageID,
    void *UserInfo,
    MC STATUS (*YourGetJsonFunction)()
```

)

The identifier of a message, file or item object.

UserInfo	Address of data which will be passed on to
	YourGetJsonFunction each time it is called.
	This may be NULL.
YourGetJsonFunction	Name of a function which will be called

repeatedly to get the JSON string.

The function must be prototyped as follows:

```
MC_STATUS YourGetJsonFunction (
    int CBmessageID,
    void* CBuserInfo,
    int* CBdataSize,
    void** CBdataBuffer,
    int CBisFirst,
    int* CBisLast
```

```
)
```

```
CBmessageID
                 The identifier assigned to the message object by the
                 MC_Open_Message function.
                 Address of data which is being passed from the
CBuserInfo
                 MC_Json_To_Message function. This may be NULL.
CBdataSize
                 Set *CBdataSize to the number of bytes you are providing.
CBdataBuffer
                 Set * CBdataBuffer to the address of the data you are
                 providing.
CBisFirst
                 Set to TRUE (not zero) by the toolkit on the first call.
CBisLast
                 Set *CBisLast to TRUE (not zero) when you are returning
                 with the last block of JSON data.
```

Remarks

MC_Json_To_Message requests that the DICOM JSON Model string buffer be converted into a DICOM message. The JSON data is requested from the user in blocks by calling *YourGetJsonFunction* repeatedly until the entire JSON content has been received by the Merge DICOM Toolkit.

YourGetJsonFunction

YourGetJsonFunction will be called repeatedly to get blocks of data from it. Merge DICOM Toolkit sets *CBisFirst* to TRUE(non-zero) the first time it calls **YourGetJsonFunction** for this message and **YourGetJsonFunction** should set **CBisLast* to TRUE(non-zero) when it gives Merge DICOM Toolkit the final block of data to be converted.

***CBdataBuffer* is set to the address of a buffer containing the data block to be converted, and *CBdataSize* is set to the number of bytes to be placed at ***CBdataBuffer.*

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	YourGetJsonFunction parameter was NULL.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message, file or item object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CALLBACK_CANNOT_COMPLY	YourGetJsonFunction returned a value other than MC_NORMAL_COMPLETION.
MC_LIBRARY_NOT_INITIALIZED	This call was made prior to the initialization of the library.
See Also	
MC_Message_To_Json	

MC_Library_Initialization

Prepares the Merge DICOM Toolkit library and provides optional information to the library.

Synopsis

```
#include "mc3msg.h"
MC_STATUS MC_Library_Initialization (
    void *(*CfgFunction) (void),
    void *(DictionaryFunction) (void),
    void *(*FutureFunction) (void)
)

CfgFunction The name of the function generated by the genconf
    utility, or NULL. (see below)
DictionaryFunction The name of the function generated by the gendict
    utility, or NULL. (see below)
FutureFunction NULL. This argument is reserved for future use.
```

Remarks

MC_Library_Initialization prepares the library to accept other API function calls. If any of the three parameters are set to NULL, no special initialization options have been selected. If a parameter is not NULL, **MC_Library_Initialization** calls the functions specified to initialize configuration or dictionary data structures.

MC_Library_Initialization must be the first library function call made!

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_CONFIG_INFO_ERROR	An error has occurred while trying to initialize configuration information. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_DDFILE_ERROR	An error has occurred while trying to access the Merge DICOM Toolkit data dictionary. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_ALREADY_INITIALIZE	D
	The MC_Library_Initialization has already been called and the library is already initialized.
MC_MISSING_CONFIG_PARM	This error occurs when the toolkit can not obtain the name of the data dictionary.
MC_NO_FILE_SYSTEM	This error occurs only in toolkits running on platforms that do not have a file system. When this error occurs, no configuration function has been specified for the initialization of the configuration information.
MC_NO_MERGE_INI	This error occurs when the toolkit can not open the toolkit initialization file. See the section titled "Configuration" in this manual and the <i>Merge DICOM Toolkit Users Manual</i> for a detailed discussion of the initialization file(s).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Library_Reset MC_Set_MergeINI	MC_Library_Release

MC_Library_Release

Releases all resources used by the Merge DICOM Toolkit library.

Synopsis

#include ``mc3msg.h"

MC_STATUS MC_Library_Release (void)

Remarks

MC_Library_Release releases all resources used by the library. This option is normally used before exiting an application. **MC_Library_Initialization** must be called before the library can be used again.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_LIBRARY_NOT_INITIALIZED	The library has not been initialized by a call to MC_Library_Initialization .
See Also	
MC_Library_Initialization MC_Set_MergeINI	MC_Library_Reset

MC_Library_Reset

Sets the Merge DICOM Toolkit library back to its initial state.

Synopsis

#include ``mc3msg.h"

MC_STATUS MC_Library_Reset (void)

Remarks

MC_Library_Reset resets the library to its initial state. This call is normally used when the library is used for an embedded application. The same options specified by the **MC_Library_Initialization** function call will be in effect. This function is <u>not</u> normally called.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_LIBRARY_NOT_INITIALIZED	The library has not been initialized by a call to MC_Library_Initialization .
MC_CONFIG_INFO_ERROR	An error occurred while trying to initialize configuration information. A message describing the error is written to the Merge DICOM Toolkit log file.
MC_DDFILE_ERROR	An error occurred while trying to access the Merge DICOM Toolkit data dictionary. A message describing the error is written to the Merge DICOM Toolkit log file.
MC_MISSING_CONFIG_PARM	This error occurs when the toolkit cannot obtain the name of the data dictionary.
MC_NO_FILE_SYSTEM	This error occurs only in toolkits running on platforms that do not have a file system. When this error occurs, no configuration function has been specified for the initialization of the configuration information.

MC_NO_MERGE_INI	This error occurs when the toolkit cannot open the toolkit initialization file. Refer to <i>Configuration</i> in this manual and the <i>Merge DICOM Toolkit Users Manual</i> for more initialization-file details.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error is written to the Merge DICOM Toolkit log file.
See Also	
MC_Library_Initialization MC_Set_MergeINI	MC_Library_Release

MC_List_File (All toolkits except Windows versions)

Prepares a listing of the current contents of a file object.

Synopsis

```
#include ``mc3media.h"
void MC_List_File (
    int FileID,
    FILE *StreamHandle
)

FileID The identifier assigned to this object by the MC_Create_File
    or MC_Create_Empty_File function.
StreamHandle The handle of an open stream. If NULL, the listing will be
    directed to STDOUT.
```

Remarks

MC_List_File prepares a report listing the current contents of the file object identified by *FileID*. The filename and preamble associated with the file object will also be listed. The report will be written to the stream identified by *StreamHandle*. Use NULL to direct the report to STDOUT.

<u>NOTE:</u>	If the file object contains an attribute with a value representation of SQ (sequence of items), each item in the sequence will be listed. Each sequence of items is indented in the listing four spaces to the right of its owner DICOMDIR, file, message, or item.
Return Value	
There is no return value.	
See Also	
MC_List_Message	MC_List_Item

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_List_File (Windows toolkit versions)

Prepares a listing of the current contents of a file object.

Synopsis

<pre>#include ``mc3media.h"</pre>		
<pre>void MC_List_File (int FileID, char *Afilename</pre>		
)		
FileID	The identifier assigned to this object by the MC_Create_File or MC_Create_Empty_File function.	
Afilename	The name of file to open. If NULL and using the static library, the listing will be directed to STDOUT. If NULL and using the DLL library, an exception will occur.	

Remarks

Because FILE* variables cannot be passed to a DLL, we have changed this function on Windows platforms to pass a filename instead. With the static library, the output will be sent to stdout when the Afilename parameter is set to NULL. If it is set to NULL when using the DLL, an exception error will occur. When Afilename is set to a filename, the message, item, or file object is listed to this text file.

MC_List_File prepares a report listing the current contents of the file object identified by *FileID*. The filename and preamble associated with the file object will also be listed. The report will be written to the file identified by *Afilename*. Use NULL to direct the report to STDOUT **only** when linked with the static toolkit library.

NOTE:

If the file object contains an attribute with a value representation of SQ (sequence of items), each item in the sequence will be listed. Each sequence of items is indented in the listing four spaces to the right of its owner DICOMDIR, file, message, or item.

Return Value

There is no return value.

See Also

MC_List_Message

MC_List_Item

MC_List_Item (All toolkits except Windows versions)

Creates a list of the current contents of an item object.

Synopsis

[©] Copyright Merge Healthcare Solutions Inc. 2023

FILE	*StreamHandle
	DELCAIMIANALC

ItemID	The identifier assigned to this message object by the MC_Open_Item function.
StreamHandle	The handle of an open stream. If NULL, the listing will be directed to STDOUT.

)

MC_List_Item prepares a report listing the contents of the current contents of the item object identified by *ItemID*. The report will be written to the stream identified by *StreamHandle*. Use NULL to direct the report to STDOUT.

NOTE: An item is listed automatically if the owning message object is listed by **MC_List_Message**, or if the owning item object is listed by **MC_List_Item**.

Return Value

There is no return value.

See Also

MC_List_Message

MC_List_Item (Windows toolkit versions)

Prepares a listing of the current contents of an item object.

Synopsis

Remarks

Because FILE* variables cannot be passed to a DLL, we have changed this function on Windows platforms to pass a filename instead. With the static library, the output will be sent to stdout when the Afilename parameter is set to NULL. If it is set to NULL when using the DLL, an exception error will occur. When Afilename is set to a filename, the message, item, or file object is listed to this text file.

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

MC_List_Item prepares a report listing the contents of the current contents of the item object identified by *ItemID*. The report will be written to the stream identified by *Afilename*. Use NULL to direct the report to STDOUT **only** when linked with the static toolkit library.

NOTE: An item is listed automatically if the owning message object is listed by **MC_List_Message**, or if the owning item object is listed by **MC_List_Item**.

Return Value

There is no return value.

See Also

MC_List_Message

MC_List_Message (All toolkits except Windows versions)

Prepares a listing of the current contents of a message object.

Synopsis

#include ``mc3msg.h"

```
void MC_List_Message (
    int MessageID,
    FILE *StreamHandle
```

)

MessageID	The identifier assigned to this object by the MC_Open_Message function or the MC_Open_Item function.
StreamHandle	The handle of an open stream. If NULL, the listing will be directed to STDOUT.

Remarks

MC_List_Message prepares a report listing the contents of the current contents of the message object identified by *MessageID*. The report will be written to the stream identified by *StreamHandle*. Use NULL to direct the report to STDOUT.

NOTE: If the message object contains an attribute with a value representation of SQ (sequence of items), each item in the sequence will be listed. Each sequence of items is indented in the listing four spaces to the right of its owning message or item.

Return Value

There is no return value.

See Also

MC_List_Item

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_List_Message (Windows toolkit versions)

Prepares a listing of the current contents of a message object.

Synopsis

<pre>#include ``mc3msg.h"</pre>		
<pre>void MC_List_Message (int MessageID, char *Afilename</pre>		
)		
MessagelD	The identifier assigned to this object by the MC_Open_Message function or the MC_Open_Item function.	
Afilename	The name of file to open. If NULL and using the static library, the listing will be directed to STDOUT. If NULL and using the DLL library, an exception will occur.	

Remarks

Because FILE* variables cannot be passed to a DLL, we have changed this function on Windows platforms to pass a filename instead. With the static library, the output will be sent to stdout when the Afilename parameter is set to NULL. If it is set to NULL when using the DLL, an exception error will occur. When Afilename is set to a filename, the message, item, or file object is listed to this text file.

MC_List_Message prepares a report listing the contents of the current contents of the message object identified by *MessageID*. The report will be written to the stream identified by *Afilename*. Use NULL to direct the report to STDOUT **only** when linked with the static toolkit library.

NOTE: If the message object contains an attribute with a value representation of SQ (sequence of items), each item in the sequence will be listed. Each sequence of items is indented in the listing four spaces to the right of its owning message or item.

Return Value

There is no return value.

See Also

MC_List_Item

MC_MemoryLog_To_Function

Retrieves the messages that have been placed in the memory log buffer.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_MemoryLog_To_Function (
    int AclearLog,
    MC_STATUS (*UserFunction)()
)
```

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

AclearLog	Memory log will be cleared if this is true.
UserFunction	The name of a function that will be called once for each
	message (line) in the memory buffer.

UserFunction must be prototyped as follows:

<pre>MC_STATUS UserFunction (</pre>		
LogLine	Address of the log message passed to the user function.	
CBisFirst	This is TRUE (not zero) the first time <i>UserFunction</i> is being called.	
CBisLast	This is TRUE (not zero) the last time <i>UserFunction</i> is being called.	

Remarks

The **MC_MemoryLog_To_Function** function is used to fetch the contents of the memory log. The user specifies the name of a function (*UserFunction*) which Merge DICOM Toolkit, in turn, calls. This "callback" function is called repeatedly to provide lines from the memory log.

UserFunction

Each message (or line) in the memory log is returned in *LogLine*.

CBisFirst is set to TRUE the first time *UserFunction* is called. This provides a clear mechanism for the function to know it is being called the first time. Users often open a file at this time.

CBisLast is set to TRUE the last time *UserFunction* is called. This provides a clear mechanism for the function to know it is being called the last time. Users often close a file at this time.

UserFunction must return MC_NORMAL_COMPLETION if all went well. If not, it must return MC_CANNOT_COMPLY. Both of these are defined in "mc3msg.h".

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	UserFunction parameter was NULL.
MC_LOG_EMPTY	The memory log buffer has not been created yet or is empty.
MC_CALLBACK_CANNOT_COMPLY	UserFunction returned with MC_CANNOT_COMPLY.

See Also

MC_Set_Log_Destination

MC_Register_MemoryLog_Function MC_Register_Enhanced_MemoryLog_Function

MC_Message_To_File

Changes a message object into a file object

Synopsis

<pre>#include ``mc3media.h"</pre>		
MC_STATUS MC_M int Mess char *Fi		
MessagelD	The identifier assigned to this object by the MC_Open_Message function.	
Filename	A pointer to a string containing the filename to be associated with the new file object.	

Remarks

MC_Message_To_File changes the message object pointed to by *MessageID* into a file object. In the process, the "command type" attributes are removed from the message object, and the DICOM File Meta Information attributes are added to the new file object. These new attributes must be given values by the user. If the message was opened as an empty message and the command and service were not set, the command and service must be set for the converted file object before validating.

The DICOM prefix for the file is set to "DICM". All bytes in the file preamble are set to 00H.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_NULL_POINTER_PARM	<i>Filename</i> has a value of NULL.
See Also	

MC_File_To_Message

MC_Message_To_Json

Converts a DICOM message, file or item into a JSON string based on DICOM JSON Model and passes the converted data to the user.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Message_To_Json (
    int MessageID,
    void *UserInfo,
    JSON OPTIONS JsonOptions,
```

MC_STATUS (*Yourk	ReceiveJsonFunction)()
)	
MessageID	The identifier of a message, file or item object.
UserInfo	Address of data which will be passed on to <i>YourReceiveJsonFunction</i> each time it is called. This may be NULL.
JsonOptions	Combination of the JSON configuration flags that are defined in the JSON_OPTIONS enumerator list:
	JsonOptDefault
	JsonOptIncludeBulks
	JsonOptIncludePixelData
	JsonOptExcludeSequences
	JsonOptExcludePrivateAttributes
	JsonOptIndentation
YourReceiveJsonFunction	Pointer to a function which will be called repeatedly by the toolkit to receive the JSON string.

The function must be prototyped as follows:

```
MC_STATUS YourReceiveJsonFunction (
    int CBmessageID,
    void *CBuserInfo,
    int CBdataSize,
    void *CBdataBuffer,
    int CBisFirst,
    int CBisLast
)
```

CBmessagelD	The identifier assigned to the message object by the MC_Open_Message function.
CBuserInfo	Address of data which is being passed from the MC_Message_To_Json function. This may be NULL.
CBdataSize	The number of bytes of JSON data being provided to you in <i>CBdataBuffer</i> .
CBdataBuffer	The buffer containing JSON data from the message object.
CBisFirst	Is TRUE (not zero) when Merge DICOM Toolkit is providing the first block of JSON data.
CBisLast	Is TRUE (not zero) when Merge DICOM Toolkit is providing the last block of JSON data.

Remarks

MC_Message_To_Json requests that the contents of the message identified by *MessageID* be converted into a JSON string. The converted message is passed to the user in blocks by calling *YourReceiveJsonFunction* repeatedly until the entire converted message has been transferred.

MC_Message_To_Json can pass data to *YourReceiveJsonFunction* by specifying the data's address in *UserInfo*. Merge DICOM Toolkit passes the address to *YourReceiveJsonFunction* in *CBuserInfo* each time it is called. *UserInfo* may be NULL.

JsonOptions specify the configuration flags for the DICOM to JSON conversion. The enumerated *JSON_OPTIONS* flags are defined in **"mc3msg.h"**:

Value	Meaning
JsonOptDefault	Use the default JSON conversion settings.
JsonOptIncludeBulks	Store bulk attributes (VR is OB or OW) in the JSON.
JsonOptIncludePixelData	Store Pixel Data buffer in the JSON.
JsonOptExcludeSequences	Do not store Sequence attributes in the JSON.
JsonOptExcludePrivateAttributes	Do not store Private attributes in the JSON.
JsonOptIndentation	Use default indentation in the generated JSON.

YourReceiveJsonFunction

YourReceiveJsonFunction will be called repeatedly to pass blocks of data to it. Merge DICOM Toolkit sets *CBisFirst* to TRUE(non-zero) the first time it calls *YourReceiveJsonFunction* for this message and it sets *CBisLast* to TRUE(non-zero) when it calls *YourReceiveJsonFunction* with the final block of converted data.

CBdataBuffer is set to the address of a buffer containing the converted data block, and *CBdataSize* is set to the number of bytes placed at *CBdataBuffer*. *The data buffer is valid only for the duration of the call to YourReceiveJsonFunction*.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The YourReceiveJsonFunction parameter was NULL.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message, file or item object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_MESSAGE_EMPTY	The message has no attributes in it.
MC_CALLBACK_CANNOT_COMPLY	YourReceiveJsonFunction returned a value other than MC_NORMAL_COMPLETION.
MC_LIBRARY_NOT_INITIALIZED	This call was made prior to the initialization of the library.
See Also	
MC_Json_To_Message	

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Message_To_SR

Converts a message created with the toolkit's message API calls (MC_Open_Message, etc) into an SR tree management object.

Synopsis

#include ``mc3sra.h"
MC_STATUS MC_Message_To_SR (
 int MsgID
)
Msg/D The identifier assigned to a SR message object by the
 MC_Open_Message functions.

Remarks

MC_Message_To_SR converts a message that was created with the toolkit's message API calls into an SR tree management object.

For example, a STANDARD_BASIC_TEXT_SR, C_STORE_RQ message has been created and manipulated by the toolkit messaging API calls, **MC_Message_To_SR** can be used to convert this message object into an SR tree management object. Once this function call has been issued, further changes to the SR toolkit message can be accomplished through the use of the MC_SR... API calls.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	The <i>MsgID</i> value is not a valid object ID.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file for the directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child	MC_SR_Get_First_Child
MC_SR_Get_Next_Child	MC_SR_Get_Root
MC_SR_Delete_Child	

MC_Message_To_Stream

Request that the values of a message object be returned as a DICOM stream.

Synopsis

#include ``mc3msg.h"

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
MC_STATUS MC_Message_To_Stream (
       int MessageID,
       unsigned long StartTag,
       unsigned long StopTag,
       TRANSFER_SYNTAX SyntaxType,
       void *UserInfo,
       MC STATUS (*YourReceiveStreamFunction) ()
)
 MessageID
                            The identifier assigned to this object by the
                            MC_Open_Message function or the
                            MC_Open_Item function.
                            The DICOM tag identifying the first attribute
 StartTag
                            which should be streamed.
                            The DICOM tag identifying the last attribute
 StopTag
                            which should be streamed
 SyntaxType
                            Specify which DICOM transfer syntax is to be
                            used to encode the stream data. Use one of the
                            enumerated TRANSFER_SYNTAX types
                            defined in "mc3msg.h":
                            IMPLICIT_LITTLE_ENDIAN
                            IMPLICIT_BIG_ENDIAN
                            EXPLICIT_LITTLE_ENDIAN
                            EXPLICIT_BIG_ENDIAN
                            DEFLATED_EXPLICIT_LITTLE_ENDIAN
                            ENCAPSULATED_UNCOMPRESSED_ELE
```

RLE JPEG_BASELINE JPEG_EXTENDED_2_4 JPEG_EXTENDED_3_5 JPEG_SPEC_NON_HIER_6_8 JPEG_SPEC_NON_HIER_7_9 JPEG_FULL_PROG_NON_HIER_10_12 JPEG_FULL_PROG_NON_HIER_11_13 JPEG_LOSSLESS_NON_HIER_14 JPEG_LOSSLESS_NON_HIER_15 JPEG_EXTENDED_HIER_16_18 JPEG_EXTENDED_HIER_17_19 JPEG_SPEC_HIER_20_22 JPEG_SPEC_HIER_21_23 JPEG_FULL_PROG_HIER_24_26 JPEG_FULL_PROG_HIER_25_27 JPEG_LOSSLESS_HIER_28 JPEG_LOSSLESS_HIER_29 JPEG_LOSSLESS_HIER_14 JPEG_2000_LOSSLESS_ONLY JPEG_2000 JPEG_LS_LOSSLESS JPEG_LS_LOSSY HEVC_H265_M10P_LEVEL_5_1 HEVC_H265_MP_LEVEL_5_1 JPIP_REFERENCED JPIP_REFERENCED_DEFLATE MPEG2_MPML MPEG2_MPHL MPEG4_AVC_H264_HP_LEVEL_4_1 MPEG4_AVC_H264_BDC_HP_LEVEL_4_1 MPEG4_AVC_H264_HP_LEVEL_4_2_2D MPEG4_AVC_H264_HP_LEVEL_4_2_3D MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2 JPEG_2000_MC_LOSSLESS_ONLY JPEG_2000_MC SMPTE_ST_2110_20_UNCOMPRESSED_PROG RESSIVE_ACTIVE_VIDEO SMPTE_ST_2110_20_UNCOMPRESSED_INTER LACED_ACTIVE_VIDEO SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO PRIVATE_SYNTAX_1 PRIVATE_SYNTAX_2

UserInfo		Address of data which will be passed on to <i>YourReceiveStreamFunction</i> each time it is called. This may be NULL.
YourReceiveStro n	eamFunctio	Name of a function which will be called repeatedly to provide blocks of streamed DICOM message data.
The function must	be prototyped	d as follows:
- int CBme void *Cl int CBda	essageID, BuserInfo, ataSize, BdataBuffe sFirst,	reamFunction (r,
)		
CBmessagelD		er assigned to the message object by the Message function.
CBuserInfo		data which is being passed from the ge_To_Stream function. This may be NULL.
CBdataSize	The numbe <i>CBdataBuf</i>	r of bytes of stream data being provided to you in <i>fer</i> .
CBdataBuffer	The buffer o	containing stream data from the message object.
CBisFirst	•	ot zero) when Merge DICOM Toolkit is providing ck of stream data.
CBisLast		ot zero) when Merge DICOM Toolkit is providing ck of stream data.

Remarks

MC_Message_To_Stream requests that the contents of the message identified by *MessageID* be "streamed" (i.e. put in the form defined by the DICOM standard). The streamed message is passed to the user in blocks by calling *YourReceiveStreamFunction* repeatedly until the entire streamed message has been transferred.

MC_Message_To_Stream can pass data to *YourReceiveStreamFunction* by specifying the data's address in *UserInfo*. Merge DICOM Toolkit passes the address to *YourReceiveStreamFunction* in *CBuserInfo* each time it is called. *UserInfo* may be NULL.

StartTag and *StopTag* specify which attributes in the message are to be placed in the stream. Any attributes in the message with tags less than *StartTag* or greater than *StopTag* will be ignored. Neither *StartTag* nor *StopTag* need be in the message.

SyntaxType must be set to **one of the values listed above.** The transfer syntax specifies the byte order used in the streamed message, whether or not each attribute's value representation is explicitly encoded in the stream, and how the pixel data is encoded in the message.

[©] Copyright Merge Healthcare Solutions Inc. 2023

NOTE: If the message contains "group length" attributes (i.e. attributes with tags of the form gggg0000: any group, element zero),
 MC_Message_To_Stream will automatically calculate the group length value when the message is streamed.

YourReceiveStreamFunction

Return Value

YourReceiveStreamFunction will be called repeatedly to pass blocks of data to it. Merge DICOM Toolkit sets **CBisFirst* to TRUE (non-zero) the first time it calls *YourReceiveStreamFunction* for this attribute and it sets **CBisLast* to TRUE(non-zero) when it calls *YourReceiveStreamFunction* with the final block of streamed data.

**CBdataBuffer* is set to the address of a buffer containing the stream data block, and **CBdataSize* is be set to the number of bytes placed at **CBdataBuffer*.

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_MESSAGE_EMPTY	The message has no attributes in it.
MC_CALLBACK_CANNOT_COMPLY	,
	<i>YourReceiveStreamFunction</i> returned a value other than MC_NORMAL_COMPLETION.
MC_INVALID_TRANSFER_SYNTAX	An invalid code was used for the <i>SyntaxType</i> parameter.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
See Also	
MC_Stream_To_Message	MC_Get_Stream_Length

MC_Message_To_XML

Converts a DICOM message, file or item into an XML string based on Merge DICOM Model and passes the converted data to the user.

Synopsis

#include ``mc3msg.h"

MC_STATUS MC_Message_To_XML (
 int MessageID,

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
void *UserInfo,
       XML OPTIONS XmlOptions,
       MC STATUS (*YourReceiveXMLFunction)()
)
 MessageID
                            The identifier of a message, file or item object.
 UserInfo
                            Address of data which will be passed on to
                             YourReceiveXMLFunction each time it is called.
                            This may be NULL.
XmlOptions
                            Combination of the XML configuration flags that
                            are defined in the XML_OPTIONS enumerator
                            list:
                              XmlOptDefault
                              XmlOptIncludeBulks
                              XmlOptIncludePixelData
                              XmlOptExcludeSequences
                              XmIOptExcludePrivateAttributes
                              XmlOptBase64Binary
 YourReceiveXMLFunction
                            Pointer to a function which will be called
                            repeatedly by the toolkit to receive the XML
                            string.
```

The function must be prototyped as follows:

```
MC_STATUS YourReceiveXMLFunction (
    int CBmessageID,
    void *CBuserInfo,
    int CBdataSize,
    void *CBdataBuffer,
    int CBisFirst,
    int CBisLast
)
CBmessageID The identifier assigned to the message object by the
    MC_Open_Message function.
CBuserInfo Address of data which is being passed from the
```

```
MC_Message_To_XML function. This may be NULL.CBdataSizeThe number of bytes of XML data being provided to you in<br/>CBdataBuffer.CBdataBufferThe buffer containing XML data from the message object.
```

```
CBisFirstIs TRUE (not zero) when Merge DICOM Toolkit is providing<br/>the first block of XML data.
```

```
CBisLast Is TRUE (not zero) when Merge DICOM Toolkit is providing the last block of XML data.
```

Remarks

MC_Message_To_XML requests that the contents of the message identified by *MessageID* be converted into an XML string. The converted message is passed to the user in blocks by calling *YourReceiveXMLFunction* repeatedly until the entire converted message has been transferred.

MC_Message_To_XML can pass data to *YourReceiveXMLFunction* by specifying the data's address in *UserInfo*. Merge DICOM Toolkit passes the address to *YourReceiveXMLFunction* in *CBuserInfo* each time it is called. *UserInfo* may be NULL.

XmlOptions specify the configuration flags for the DICOM to XML conversion. The enumerated *XML_OPTIONS* flags are defined in **"mc3msg.h"**:

Value	Meaning
XmlOptDefault	Use the default XML conversion settings.
XmlOptIncludeBulks	Store bulk attributes (VR is OB or OW) in the XML.
XmlOptIncludePixelData	Store Pixel Data buffer in the XML.
XmlOptExcludeSequences	Do not store Sequence attributes in the XML.
XmlOptExcludePrivateAttributes	Do not store Private attributes in the XML.
XmlOptBase64Binary	Use Base64 encoding for bulks and UN VR attributes.

YourReceiveXMLFunction

YourReceiveXMLFunction will be called repeatedly to pass blocks of data to it. Merge DICOM Toolkit sets *CBisFirst* to TRUE (non-zero) the first time it calls *YourReceiveXMLFunction* for this message and it sets *CBisLast* to TRUE (non-zero) when it calls *YourReceiveXMLFunction* with the final block of converted data.

CBdataBuffer is set to the address of a buffer containing the converted data block, and *CBdataSize* is set to the number of bytes placed at *CBdataBuffer*. *The data buffer is valid only for the duration of the call to YourReceiveXMLFunction*.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The YourReceiveXMLFunction parameter was NULL.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message, file or item object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_MESSAGE_EMPTY	The message has no attributes in it.
MC_CALLBACK_CANNOT_COMPLY	YourReceiveXMLFunction returned a value other than MC_NORMAL_COMPLETION.

MC_LIBRARY_NOT_INITIALIZED This call was made prior to the initialization of the library.

See Also

MC_XML_To_Message

MC_Message_To_XML_Native

Converts a DICOM message, file or item into an XML string based on Native DICOM Model and passes the converted data to the user.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Message To XML Native (
       int MessageID,
       void *UserInfo,
       XML OPTIONS XmlOptions,
       MC_STATUS (*YourReceiveXMLFunction) ()
)
 MessageID
                            The identifier of a message, file or item object.
 UserInfo
                            Address of data which will be passed on to
                            YourReceiveXMLFunction each time it is called.
                            This may be NULL.
 XmlOptions
                            Combination of the XML configuration flags that
                            are defined in the XML_OPTIONS enumerator
                            list:
                              XmlOptDefault
                              XmlOptIncludeBulks
                              XmlOptIncludePixelData
                              XmlOptExcludeSequences
                              XmlOptExcludePrivateAttributes
                              XmlOptBase64Binary
                              XmlOptIndentation
                            Pointer to a function which will be called
 YourReceiveXMLFunction
                            repeatedly by the toolkit to receive the XML
                            string.
The function must be prototyped as follows:
MC STATUS YourReceiveXMLFunction (
       int CBmessageID,
```

```
void *CBuserInfo,
int CBdataSize,
void *CBdataBuffer,
int CBisFirst,
int CBisLast
)
```

CBmessagelD	The identifier assigned to the message object by the
	MC_Open_Message function.
CBuserInfo	Address of data which is being passed from the
	MC_Message_To_XML_Native function. This may be NULL.
CBdataSize	The number of bytes of XML data being provided to you in
	CBdataBuffer.
CBdataBuffer	The buffer containing XML data from the message object.
CBisFirst	Is TRUE (not zero) when Merge DICOM Toolkit is providing
	the first block of XML data.
CBisLast	Is TRUE (not zero) when Merge DICOM Toolkit is providing
	the last block of XML data.

Remarks

MC_Message_To_XML_Native requests that the contents of the message identified by *MessageID* be converted into an XML string based on Native DICOM Model. The converted message is passed to the user in blocks by calling *YourReceiveXMLFunction* repeatedly until the entire converted message has been transferred.

MC_Message_To_XML_Native can pass data to *YourReceiveXMLFunction* by specifying the data's address in *UserInfo*. Merge DICOM Toolkit passes the address to *YourReceiveXMLFunction* in *CBuserInfo* each time it is called. *UserInfo* may be NULL.

XmlOptions specify the configuration flags for the DICOM to Native DICOM XML conversion. The enumerated *XML_OPTIONS* flags are defined in **"mc3msg.h"**:

Value	Meaning
XmlOptDefault	Use the default XML conversion settings.
XmlOptIncludeBulks	Store bulk attributes (VR is OB or OW) in the XML.
XmlOptIncludePixelData	Store Pixel Data buffer in the XML.
XmIOptExcludeSequences	Do not store Sequence attributes in the XML.
XmIOptExcludePrivateAttributes	Do not store Private attributes in the XML.
XmlOptBase64Binary	Use Base64 encoding for bulks and UN VR attributes.
XmlOptIndentation	Use default indentation in the generated XML.

YourReceiveXMLFunction

YourReceiveXMLFunction will be called repeatedly to pass blocks of data to it. Merge DICOM Toolkit sets *CBisFirst* to TRUE (non-zero) the first time it calls **YourReceiveXMLFunction** for this message and it sets *CBisLast* to TRUE(non-zero) when it calls **YourReceiveXMLFunction** with the final block of converted data.

CBdataBuffer is set to the address of a buffer containing the converted data block, and *CBdataSize* is set to the number of bytes placed at *CBdataBuffer*. *The data buffer is valid only for the duration of the call to YourReceiveXMLFunction*.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The YourReceiveXMLFunction parameter was NULL.
MC_INVALID_MESSAGE_ID	The <i>MessagelD</i> value is not a valid message, file or item object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_MESSAGE_EMPTY	The message has no attributes in it.
MC_CALLBACK_CANNOT_COMPLY	YourReceiveXMLFunction returned a value other than MC_NORMAL_COMPLETION.
MC_LIBRARY_NOT_INITIALIZED	This call was made prior to the initialization of the library.
See Also	

MC_XML_Native_To_Message

MC_NewProposedServiceList MC_NewProposedServiceListAsync

Creates a new service list for use in association negotiation.

Synopsis

```
#include ``mergecom.h"
MC_STATUS MC_NewProposedServiceList (
      char *ServiceListName,
       char *ServiceNameArray[]
)
MC STATUS MC NewProposedServiceListAsync (
      char *ServiceListName,
      char *ServiceNameArray[]
      unsigned short MaxOperationsInvoked,
      unsigned short MaxOperationsPerformed)
 ServiceListName
                         Application supplied name to associate with this list
 ServiceNameArray
                         NULL terminated array of Service Names to be
                         used in this list
 MaxOperationsInvoked
                         The maximum operations invoked by the
                         association requestor. Note that a setting of 0
                         means an unlimited number of operations can be
                         invoked
```

MaxOperationsPerformeThe maximum operations performed by thedassociation regestor. Note that a setting of 0 means
an unlimited number of operations can be
performed

Remarks

MC_NewProposedServiceList and MC_NewProposedServiceListAsync establish a list of services to be used during the negotiation of an association. The ServiceListName can be used in MC_Open_Association, MC_Open_Secure_Association and MC_Wait_For_Association.

This functionality is used to dynamically create list normally found in the **mergecom.app** configuration file. This method of service list generation augments the existing configuration file.

Service names are generated using MC_NewServiceFromName or MC_NewServiceFromUID.

MC_NewProposedServiceListAsync allows setting of a value for Max Operations Performed and Max Operations Invoked during association negotation. These settings allow negotiation of the DICOM Asynchronous Operations Window. The negotiated results for the association requestor can be examined by calling **MC_Get_Association_Info**.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_DUPLICATE_NAME	A Duplicate Name is present in the SeviceNameArray or the request service list name is already in use.
See Also	
MC_Open_Association MC_Wait_For_Association MC_NewServiceFromName MC_FreeServiceList	MC_Open_Secure_Association MC_Wait_For_Secure_Association MC_NewServiceFromUID MC_Get_Association_Info

MC_NewSyntaxList

Creates a new syntax list for use in the creation of services for association negotiation.

Synopsis

```
#include ``mergecom.h"
#include ``mc3msg.h"
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

TRANSFER_SYNTAX SyntaxArray[]

SyntaxListName	Application supplied name to associate with this list
SyntaxArray	NULL terminated array of TRANSFER_SYNTAX-es
	to be used in this list

Remarks

)

MC_NewSyntaxList establishes a list of syntaxes to be used during the negotiation of an association. The SyntaxList name is supplied by the application and is used as a reference when creating service references

This functionality is used to dynamically create list normally found in the **mergecom.app** configuration file. This method of syntax list generation augments the existing configuration file. TRANSFER_SYNTAX-es are an enumerated type found in **mc3msg.h**

Service names are generated using MC_NewServiceFromName or MC_NewServiceFromUID.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_DUPLICATE_NAME	The requested syntax list name is already in use.
MC_DUPLICATE_SYNTAX	A syntax id is duplicated in the array supplied.
See Also	
MC_Open_Association MC_Wait_For_Association MC_NewServiceFromName MC_FreeSyntaxList	MC_Open_Secure_Association MC_Wait_For_Secure_Association MC_NewServiceFromUID

MC_NewService... Functions

Creates a new service description for use in the creation of service lists.

Synopsis

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
MC STATUS MC NewServiceFromUID (
       char *ServiceName,
       char *SOPClassUID,
       char *SyntaxListName,
       int SCURole,
       int SCPRole
MC STATUS MC NewServiceWithExtInfoFromName (
       char *ServiceName,
       char *SOPClassName,
       char *SyntaxListName,
       int SCURole,
       int SCPRole,
       void *ExtInfoBuffer,
       int ExtInfoLength
MC STATUS MC NewServiceWithExtInfoFromUID (
       char *ServiceName,
       char *SOPClassUID,
       char *SyntaxListName,
       int SCURole,
       int SCPRole,
       void *ExtInfoBuffer,
       int ExtInfoLength
)
 ServiceName
                        Application supplied name to associate with this
                        service
 SOPClassName
                        Name as defined in mergecom.srv
 SOPClassUID
                        UID as defined by DICOM and found in
                        mergecom.srv
 SyntaxListName
                        Name of a syntax list as defined by
                        MC_NewSyntaxList or in the mergecom.app. Can
                        be NULL, if so the default syntax lists will be utilized.
 SCURole
                        Sets the scu role negotiation parameters for this
                        service, 0 or 1
 SCPRole
                        Sets the scp role negotiation parameters for this
                        service, 0 or 1
 ExtInfoBuffer
                        A pointer to the buffer containing extended
                        negotiation information.
                        The number of bytes contained in the ExtInfoBuffer.
 ExtInfoLength
```

Remarks

These routines are utilized to create a service with a specific name to be used in a service list. Each service utilitized in a dynamic service list must be created with a call to one of these functions. The service lists created can then be used in subsequent calls to **MC_Open_Association**, **MC_Open_Secure_Association**, **MC_Wait_For_Association**, or **MC_Wait_For_Secure_Association** to specify the services negotiated for an association.

When creating a new service, it can be identified by its Merge DICOM Toolkit service name specified in the mergecom.srv file or by its SOP Class UID depending on which call is utilized. The SCU and SCP roles are specified for the service along with the name of a syntax list to be used. The MC_NewServiceWithExtInfoFromUID and MC_NewServiceWithExtInfoFromName routines can be used to optionally specify extended negotiatoin information for the service.

The application assigned name is utilized by MC_NewProposedServiceList.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_SOP_CLASS_UID	The SOPClassName or SOPClassUID do not reference a defined SOP in mergecom.srv
See Also	
MC_NewSyntaxList	MC_NewProposedServiceList

MC FreeService

MC_Set_Negotiation_Information_For_Association

MC_Open_Association

MC_Open_Association_With_Connect_Timeout MC_Open_Association_With_Identity MC_Open_Association_With_Identity_With_Reject_Info MC_Open_Association_With_Identity_With_Reject_Info_With_Connect_Timeout MC_Open_Association_With_Reject_Info MC_Open_Association_With_Callback MC_Open_Association_With_All_Optional_Parameters Establishes a connection with a remote DICOM application. Synopsis

#include ``mergecom.h"

```
MC_STATUS MC_Open_Association (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char *RemoteHostTCPIPName,
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
char *ServiceList
)
MC STATUS MC Open Association With Connect Timeout (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char *RemoteHostTCPIPName,
      char *ServiceList,
      SecureSocketFunctions *SecureFunctions,
      void *SecureContext,
      int Timeout
)
typedef enum {
    NO USER IDENTITY = 0,
    USERNAME = 1,
    USERNAME AND PASSCODE = 2,
    KERBEROS SERVICE TICKET = 3,
    SAML ASSERTION = 4
} USER_IDENTITY_TYPE;
MC_STATUS MC_Open_Association_With_Identity (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char * RemoteHostTCPIPName,
      char *ServiceList,
      SecureSocketFunctions *SecureFunctions,
      void *SecureContext,
      USER_IDENTITY_TYPE IdentityType,
      unsigned char ResponseRequested,
      void *PrimaryField,
      unsigned short PrimaryFieldLength,
      void *SecondaryField,
      unsigned short SecondaryFieldLength
)
typedef struct {
      unsigned short Result;
      unsigned short Source;
      unsigned short Reason;
} AssocRejectInfo;
MC_STATUS MC_Open_Association_With_Identity_With_Reject_Info (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char *RemoteHostTCPIPName,
      char *ServiceList,
      SecureSocketFunctions *SecureFunctions,
      void *SecureContext,
      USER IDENTITY TYPE IdentityType,
```

```
unsigned char ResponseRequested,
      void *PrimaryField,
      unsigned short PrimaryFieldLength,
      void *SecondaryField,
      unsigned short SecondaryFieldLength,
      AssocRejectInfo RejectInfo
MC STATUS
MC Open Association With Identity With Reject Info With Connect Timeout (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char * RemoteHostTCPIPName,
      char *ServiceList,
      SecureSocketFunctions *SecureFunctions,
      void *SecureContext,
      USER IDENTITY TYPE IdentityType,
      unsigned char ResponseRequested,
      void *PrimaryField,
      unsigned short PrimaryFieldLength,
      void *SecondaryField,
      unsigned short SecondaryFieldLength,
      AssocRejectInfo RejectInfo,
      int Timeout
)
MC STATUS MC Open Association With Reject Info (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char * RemoteHostTCPIPName,
      char *ServiceList
      AssocRejectInfo *RejectInfo
)
MC STATUS MC Open Association With Callback (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char * RemoteHostTCPIPName,
      char *ServiceList,
      MC STATUS (*YourCallback) ()
)
MC STATUS MC Open Association With All Optional Parameters (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char * RemoteHostTCPIPName,
      char *ServiceList,
      SecureSocketFunctions *SecureFunctions,
      void *SecureContext,
      USER_IDENTITY_TYPE IdentityType,
```

```
unsigned char ResponseRequested,
void *PrimaryField,
unsigned short PrimaryFieldLength,
void *SecondaryField,
unsigned short SecondaryFieldLength,
AssocRejectInfo RejectInfo,
int Timeout,
MC_STATUS (*YourCallback)()
```

)

ApplicationID	The application ID returned from the
	MC_Register_Application function.
AssociationID	The identification number of an association object is returned here.
RemoteApplicationTitle	The DICOM Application Title of the remote application.

Each of the following parameters is optional. Use NULL if not used.

RemoteHostPortNumber	The TCP/IP port used by the remote application to "listen" for DICOM associations
RemoteHostTCPIPName	The remote host's TCP/IP Name. This parameter can be a hostname, an IPv4 address, or an IPv6 address.
ServiceList	Name of a service list in the Merge DICOM Toolkit configuration file.
SecureFunctions	An optional pointer to a structure containing functions that will be called by Merge DICOM Toolkit while processing network I/O over the secure connection. See MC_Open_Secure_Association for details.
	•
SecureContext	An optional pointer to application-specific data that Merge DICOM Toolkit passes to the functions declared in <i>SecureFunctions</i> . See MC_Open_Secure_Association for details.
IdentityType	The type of User Identity negotiation to perform as specified in DICOM PS3.7, Section D.3.3.7.
ResponseRequested	If response from the server is required, this field should be set to
	POSITIVE_RESPONSE_REQUESTED, if a response is not required, this field should be set to NO_RESPONSE_REQUESTED.
PrimaryField	A buffer pointing to the primary field used in User Identity Negotiation.

PrimaryFieldLength	The length of <i>PrimaryField</i> . Note that when passing a text string in <i>PrimaryField</i> , the NULL terminator should not be included in the length.	
SecondaryField	A buffer pointing to the secondary field used in User Identity Negotiation. If the secondary field is not used for the type of User Identity specified by <i>IdentityType</i> , this field can be set to NULL.	
SecondaryFieldLength	The length of <i>SecondaryField</i> . Note that when passing a text string in <i>SecondaryField</i> , the NULL terminator should not be included in the length.	
RejectInfo	If the association is rejected, this return argument will contain the result/source/reason codes explaining the reason for rejection (See DICOM PS3.8, Section 9.3.4).	
Timeout	The number of seconds to wait for the network connect to be accepted.	
	If greater than zero, the parameter supersedes the application-wide CONNECT_TIMEOUT configuration setting.	
	If negative or zero, the CONNECT_TIMEOUT configuration setting takes effect.	
YourCallback	Name of a user function which will be called with the association ID as a parameter	
The callback function must be prototyped as follows:		

```
MC_STATUS YourCallback (
    int AssociationID
)
```

```
AssociationID
```

```
The ID assigned to the association object as soon as it is created. This ID can be used to cancel the association request before the connect timeout expires.
```

Remarks

MC_Open_Association establishes a DICOM association connection with a remote DICOM application.

Each application in a DICOM association has a publicly known name or "application title". This application's title was declared in the **MC_Register_Application** call. The application to which we are intending to connect is specified in *RemoteApplicationTitle*. The remote DICOM system waits for association requests on a given TCP/IP port. That port number is specified by *RemoteHostPortNumber*. The TCP/IP name of the remote host is specified by *RemoteHostTCPIPName*.

Starting a DICOM association is a negotiated process. One application provides one or more services, and the other uses one or more of the services provided. The **MC_Open_Association** function lets the remote DICOM process know which services this application wishes to use. This "service list" is

defined in the Merge DICOM Toolkit Application Configuration file, and the name of the appropriate service list is specified by the *ServiceList* parameter.

A few DICOM services require the applications to negotiate application-specific information. This is performed by sharing "extended negotiation information" when the association is negotiated. If one or more of the services in *ServiceList* requires such negotiation, **MC_Set_Negotiation_Info** should be called for each such service to supply the negotiation information.

If the remote system accepts the association request (i.e. it can support one or more of the services in the *ServiceList*), **MC_Open_Association** creates an association object, puts its ID in **AssociationID*, and returns MC_NORMAL_COMPLETION.

If the service requires the remote application to return extended negotiation information, **MC_Get_Negotiation_Info** should be called to retrieve such information. If the information is unacceptable, **MC_Close_Association** should be called.

MC_Open_Association_With_Identity is used when User Identity information is to be negotiated for the association as defined in DICOM Supplement 99. **MC_Open_Association_With_Identity** can be used to open both secure and unsecure associations. See **MC_Open_Secure_Association** for details on opening secure associations.

When negotiating user identity information, the *IdentityType* paramater is set to a specific kind of user identity. The *PrimaryField* and *SecondaryField* parameters are set as defined in DICOM Part 8. For instance, when USERNAME_AND_PASSCODE is specified, the *PrimaryField* has the username and *SecondaryField* has the passcode. Due to the fact that older DICOM applications may not be looking for the user identity information, it is possible to ask for a response from the server so the application can determine if the user identity information was accepted or ignored. The *ResponseRequested* field is used to request this response. After the association has been negotiated, an application can use the **MC_Get_Association_Information** to retrieve the information on if the server has responded to the user identity information in the request. The **MC_Get_User_Identity_Length** and **MC_Get_User_Identity_Info** routines can also be used if information was returned from the remote system.

NOTE:

MC_Open_Association_With_Identity is identical to MC_Open_Association when *IdentityType* is set to NO_USER_IDENTITY, and the *SecureFunctions*, *PrimaryField*, and *SecondaryField* parameters are set to NULL.

MC_ASSOCIATION_REJECTED is returned if the remote system rejects the association request. **AssociationID* is NOT valid if the association is rejected or if an error occurs.

NOTE: That the information registered with the MC_Set_Negotiation_Info call is not changed during the association process. Use
 MC_Clear_Negotiation_Info to prevent the information from being used in subsequent association negotiations for a given service.

[©] Copyright Merge Healthcare Solutions Inc. 2023

NOTE:

Only the first three parameters (*ApplicationID, AssociationID and RemoteApplicationTitle*) for **MC_Open_Association** are required. Each of the others will be defaulted if they are specified as NULL. The default values are obtained from the *RemoteApplicationTitle*'s section in the Merge DICOM Toolkit Application Configuration file.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally. <i>AssociationID</i> has been set to the association's identification number.
MC_ASSOCIATION_REJECTED	The remote system rejected the open association request.
MC_NEGOTIATION_ABORTED	The association was aborted during negotiation
MC_CONFIG_INFO_MISSING	<i>RemoteApplicationTitle</i> or <i>ServiceList</i> could not be found in the mergecom.app configuration file.
MC_CONFIG_INFO_ERROR	The service list contained too many services. DICOM has a limit of 128 presentation contexts/services when negotiating an association.
MC_CONNECTION_FAILED	The remote system could not be connected to at the TCP/IP level. Check that the remote host name and port number have been configured properly.
MC_NULL_POINTER_PARM	AssociationID and/or RemoteApplicationTitle was NULL.
MC_INVALID_APPLICATION_TITLE	RemoteApplicationTitle was not 1-16 bytes long.
MC_INVALID_APPLICATION_ID	<i>ApplicationID</i> does not identify a valid Merge DICOM Toolkit application.
MC_INVALID_PORT_NUMBER	<i>RemoteHostPortNumber</i> was not NULL and pointed at a negative number.
MC_INVALID_SERVICE_LIST_NAME	ServiceList was not NULL and pointed at an empty string.
MC_INVALID_HOST_NAME	<i>RemoteHostTCPIPName</i> was not NULL and it was not 1- 39 bytes long.
MC_UNKNOWN_HOST_NAME	The <i>RemoteHostTCPIPName</i> is unknown to the system.
MC_TIMEOUT	Timeout attempting the association.
MC_SYSTEM_CALL_INTERRUPTED	The operating system interrupted the network call. Retry the connection.
MC_NO_APPLICATIONS_REGISTERED	An application title for this application has not yet be registered.
MC_LIBRARY_NOT_INITIALIZED	This call was made prior to the initialization of the library
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message

describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Register_Application MC_Wait_For_Association MC_Get_Association_Info MC_Get_Negotiation_Info MC_Get_User_Identity_Info MC_Close_Association MC_Open_Secure_Association MC_Wait_For_Secure_Association MC_Set_Negotiation_Info MC_Get_User_Identity_Length MC_Clear_Negotiation_Info MC_Abort_Association

MC_Open_Empty_Item

Creates a new empty Merge DICOM Toolkit item object.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Open_Empty_Item (
```

```
int *ItemIDPtr
```

```
)
```

ItemIDPtr Upon successful completion, the item object identifier is returned here.

Remarks

The **MC_Open_Empty_Item** function creates a empty "item object" which contains no attributes. The resulting item object is given an identification number which is returned in **ItemIDPtr*. All functions dealing with this item must provide this item ID number.

If an item is opened using **MC_Open_Empty_Item**, it is not necessary to add attributes to the item object before setting attribute values. If one of the set value functions (e.g.

MC_Set_Value_From_String) is used for an attribute, the attribute will automatically be added to the item object before the value is set. (Note that this is NOT THE CASE if an item object is opened using **MC_Open_Item**. In that case the item IS associated with a given item name and attributes other than those associated with that item name must be explicitly added to the item before setting values for the added attributes.)

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>ItemIDPtr</i> was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Open_Item

MC_Open_Empty_Message

Creates a new "empty" Merge DICOM Toolkit message object.

Remarks

The **MC_Open_Empty_Message** function creates a "message object" which contains no attributes. The resulting message object is given an identification number which is returned in **MessageIDPtr*. All functions dealing with this message must provide this message ID number.

The opened message object is not associated with any particular DICOM service or command. If the message object is to be used to send a message to a network partner, or if **MC_Validate_Message** is to be called for the message object, **MC_Set_Service_Command** must be called first to associate the message object with a given DICOM service and command.

If a message is opened using **MC_Open_Empty_Message**, it is not necessary to add attributes to the message object before setting attribute values. If one of the set value functions (e.g. **MC_Set_Value_From_String**) is used for an attribute, the attribute will automatically be added to the message object before the value is set. (Note that this is NOT THE CASE if a message object is opened using **MC_Open_Message**. In that case the message IS associated with a given service/command pair and attributes other than those associated with that service and command must be explicitly added to the message before setting values for the added attributes.)

Reference the "DICOM V3.0 Standard" for more information about these commands.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>MessagelDPtr</i> was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Open_Message MC_Free_Message MC_Create_File MC_Free_Item

MC_Open_File MC_Open_File_Bypass_OBOW MC_Open_File_Upto_Tag MC_Open_File_Upto_Tag_Bypass_Value

Requests that the values of a file object be retrieved from a DICOM file

Synopsis

```
#include ``mc3media.h"
MC_STATUS MC_Open_File (
      int ApplicationID,
      int FileID,
      void *UserInfo,
      MC_STATUS (*YourFromMediaFunction())
)
MC STATUS MC Open File Bypass OBOW (
      int ApplicationID,
      int FileID,
      void *UserInfo,
      MC_STATUS (*YourFromMediaFunction())
)
MC STATUS MC Open File Upto Tag (
      int ApplicationID,
      int FileID,
      void *UserInfo,
      unsigned long Tag,
      long *Offset,
      MC STATUS (*YourFromMediaFunction())
)
MC STATUS MC Open File Upto Tag Bypass Value (
      int ApplicationID,
      int FileID,
      void *UserInfo,
      unsigned long Tag,
      MTI BOOLEAN AbypassTagValue,
      long *Offset,
      MC_STATUS (*YourFromMediaFunction())
)
```

ApplicationIDThe identifier assigned to this object by theMC_Register_Application function.

FileID	The identifier assigned to this object by the MC_Create_File or MC_Create_Empty_File functions.
UserInfo	Address of data which will be passed on to <i>YourFromMediaFunction</i> each time it is called. This may be NULL.
AbypassTagValue	Bypass an attribute value if TRUE.
	Note:
	For MC_Open_File_Upto_Tag_Bypass_Value this argument is unused, as the function will <i>by definition</i> bypass the attribute value. Retained for backward compatibility only.
Tag	The attribute after which to stop reading the file from media.
Offset	Upon successful completion, the offset in bytes from the beginning of the file of:
	 the first attribute greater than Tag(for MC_Open_File_Upto_Tag)
	 the data of the attribute at <i>Tag</i> (for MC_Open_File_Upto_Tag_Bypass_Value)
	is returned here.
YourFromMediaF n	<i>Functio</i> Name of a function which will be called repeatedly to request blocks of DICOM file data from media.
The function must	be prototyped as follows:
- char *CB void *CB int *CBd	
CBfilename	String filename associated with the file object
CBuserInfo	Address of data which is being passed from the MC_Open_File function. This may be NULL.
CBdataSizePtr	Set <i>*CBdataSizePtr</i> to the number of bytes you are providing. This must be an even number.
CBdataBufferPt r	Set <i>*CBdataBufferPtr</i> to the address of the data you are providing.
CBisFirst	This is TRUE (non-zero) the first time Merge DICOM Toolkit calls <i>YourFromMediaFunction</i> to request data blocks.

CBisLastPtr Set **CBisLastPtr* to TRUE (non-zero) when you are returning with the last block of file data.

Remarks

MC_Open_File, MC_Open_File_Bypass_OBOW, and MC_Open_File_Upto_Tag, MC_Open_File_Upto_Tag_Bypass_Value request that the contents of a DICOM file be converted and placed into the file object *FileID*. The file is passed to Merge DICOM Toolkit by *YourFromMediaFunction*. Merge DICOM Toolkit repeatedly calls *YourFromMediaFunction* until all of the file has been processed.

The MC_Open_File... functions can pass data to *YourFromMediaFunction* by specifying the data's address in *UserInfo*. Merge DICOM Toolkit passes the address to *YourFromMediaFunction* in *CBuserInfo* each time it is called. *UserInfo* may be NULL.

When **MC_Open_File** retrieves attributes of type OB, OW, OF or OD and a callback is not registered for the attribute, the library stores the values in a configurable location (normally in temporary files). Then the user can use **MC_Get_Value_To_Function** to retrieve the attribute values. If a callback is registered for the attribute, the callback is supplied the data as it is read. See the description of **MC_Register_Callback_Function** for more details.

When **MC_Open_File_Bypass_OBOW** retrieves attributes of type OB, OW, OF or OD and a callback is not registered for the attribute, the library stores the values in a configurable location (normally in temporary files). Then the user can use **MC_Get_Value_To_Function** to retrieve the attribute values.

Performance Tuning MC_Open_File_Bypass_OBOW can be used to increase performance for handling attributes of type OB, OW, OF or OD. When a callback function is registered with MC_Register_Callback_Function for an attribute of type OB, OW, OF or OD, MC_Open_File_Bypass_OBOW will not read in the attribute's value. Instead, the data will be left on media and the offset of the attribute's value from the beginning of the file along with length of the value will be passed to the user's callback function. When the data is needed by the user or Merge DICOM Toolkit, the callback can retrieve it from media. See the description of MC_Register_Callback_Function for more details.

 Performance Tuning
 When MC_Open_File_Upto_Tag retrieves attributes of type OB, OW, OF or OD and a callback is not registered for the attribute, the library stores the values in a configurable location (normally in

temporary files). Then the user can use **MC_Get_Value_To_Function** to retrieve the attribute values. **MC_Open_File_Upto_Tag** can be used to increase performance for handling attributes of type OB,

OW, OF or OD. **MC_Open_File_Upto_Tag** will stop reading the file from media when it reaches the first attribute greater than *Tag*. As a convenience, the offset in bytes from the beginning of the file of the first attribute greater than *Tag* is returned in *Offset*. If the user wants to use *Offset* to access attributes greater then *Tag* directly from media, the user will have to parse the file directly.

MC_Open_File_Upto_Tag_Bypass_Value has the same functionality as MC_Open_File_Upto_Tag except that the attribute values with given *Tag* will not be read, but left on media. It increases performance for handling large size attributes, especially pixel data (7fe0,0010). Instead of reading the attribute values the above described callback mechanism should be used with MC_Register_Callback_Function for that attribute to be retrieved from media when needed. If the file contains any attributes with a value representation of SQ (i.e. the file contains one or more sequences of items), an item object is automatically opened for each item in the file. The ItemID associated with each opened item object is used as the value for the item in the sequence attribute. Later, the **MC_Get_Value...** functions may be used to retrieve the ItemIDs from the SQ attribute. Then, again using the **MC_Get_Value...** functions, the attributes of the ItemID object may be retrieved.

The MC_Open_File... functions will also read in DICOMDIRs. In order for a DICOMDIR to be properly read in, the file object must be associated with the service for a DICOMDIR. This can be accomplished by specifying the service when the object is created, or calling MC_Set_Service_Command on the object before one of the MC_Open_File... functions is called. When one of these conditions is met, the function will resolve the directory record file offset pointers within the DICOMDIR by assigning them the ItemID associated with each opened item object.

The **MC_Open_File...** functions read in the DICOM File Meta Information attributes in explicit VR little endian transfer syntax. The remainder of the file is read in with the transfer syntax specified in the attribute (0002,0010). If this attribute is not found, the entire file is read in using the Explicit VR Little Endian transfer syntax.

- **NOTE:** A runtime configuration parameter determines what will happen if the input file contains an invalid DICOM file prefix. The default is to change any invalid prefix to "DICM". If requested, however, an invalid prefix can be left in the file. If an invalid prefix is read in or wrote out a warning message will be logged.
- NOTE: A runtime configuration parameter determines what will happen if the input file contains an attribute which is not in the service associated with the file. The default is to ignore such attributes (with a warning message logged). If requested, however, such attributes will be added to the file, along with their values. If the Value Representation of the attribute being added cannot be determined, the attribute is given a pseudo Value Representation of "Unknown_VR". The only way to retrieve the value of such attributes is to use the MC_Get_Value_To_Buffer function. To change the value of such attributes, MC_Set_Value_Representation (or MC_Set_PValue_Representation) must first be called to assign a valid Value Representation to the attribute. If MC_Write_File is used, attributes with unknown VRs are simply copied (memcpy) to the stream with no consideration given to byte ordering.

YourFileToFunction

It is the responsibility of *YourFromMediaFunction* to pass blocks of data back to Merge DICOM Toolkit each time it is called. Merge DICOM Toolkit sets *CBFirstCall* to TRUE (non-zero) the first time it calls *YourFromMediaFunction* for this file.

**CBdataBufferPtr* must be set to the address of the buffer containing the stream data block, and **CBdataSizePtr* must be set to the number of bytes placed at **CBdataBufferPtr*.

YourFromMediaFunction must set **CBisLast* to TRUE (non-zero) when it is providing the last block of the streamed message.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Do not reuse data in your buffer Data in the buffer returned by **CBdataBufferPtr* shall not be reused, because with some configurations Merge DICOM Toolkit may do in-place byte swapping to convert data to a native endian format. If your application needs to keep source data unchanged, you must create a copy of your source data chunk and return its address to the caller.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>FileIDPtr, YourFromMediaFunction</i> , or <i>Filename</i> was NULL.
MC_INVALID_APPLICATION_ID	The application ID does not identify a valid Merge DICOM Toolkit application.
MC_INVALID_FILE	An invalid DICOM Prefix was found within the file. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TRANSFER_SYNTAX	An invalid transfer syntax code was found within the file's DICOM File Meta Information.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CALLBACK_CANNOT_COMPLY	
	<i>YourFromMediaFunction</i> returned with a status other than MC_NORMAL_COMPLETION.
MC_CALLBACK_DATA_SIZE_UNEVE	N
	The <i>CBdataSizePtr</i> parameter returned by <i>YourFromMediaFunction</i> was an uneven number.
MC_CALLBACK_PARM_ERROR	A callback function registered by your application returned an empty (NULL) data buffer when the buffers length was specified as being non-zero. See
	MC_Register_Callback_Function for details.
MC_OUT_OF_ORDER_TAG	A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TAG	The file contains an invalid tag.
MC_VALUE_TOO_LARGE	An attribute in the file (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation.
MC_UNEXPECTED_EOD	<i>YourFromMediaFunction</i> stopped without passing the entire value for an attribute.
MC_INVALID_LENGTH_FOR_VR	The value(s) for one of the file attributes was not legal for its value representation.
Any of the status and a raturned by M	C Sat Value may also be returned

Any of the status codes returned by **MC_Set_Value** may also be returned.

[©] Copyright Merge Healthcare Solutions Inc. 2023

See Also

MC_Register_Callback_Function

MC_Write_File MC_Write_File_By_Callback

MC_Open_Item

Creates a new Merge DICOM Toolkit item object.

Synopsis		
<pre>#include ``mc3msg.h"</pre>		
MC_STATUS MC_O int *Ite char *It)	mIDPtr,	
ltemIDPtr	Upon successful completion, the item object identifier is returned here.	
ItemName	String name of the item to be associated with this item object.	

Remarks

The **MC_Open_Item** function creates a "item object" which contains (or will contain) all of the attributes of a named item which will be used in a sequence of items in a message object or in another item object. The resulting item object is given an identification number which is returned in **ItemIDPtr.* All functions dealing with this item must provide this item ID number.

The *ItemName* is used to access configuration information which describes the attributes of the message. If such configuration information is not available, an empty item object is created, and a warning message is logged.

Return Value

One of these enumerated **MC_**STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Open_Empty_Item	MC_Open_Message

MC_Open_Message

Creates a new Merge DICOM Toolkit message object.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Open Message (
       int *MessageIDPtr,
       char *ServiceName,
      MC COMMAND Command
)
 MessageIDPtr
                Upon successful completion, the message object identifier
                is returned here.
 ServiceName
                String name of a DICOM service to be associated with this
                 message object.
 Command
                The command which is to be associated with this message.
                The MC_COMMAND enumerated values are defined in
                 "mc3msg.h".
```

Remarks

The **MC_Open_Message** function creates a "message object" which contains (or will contain) all of the attributes of a DICOM message which will be used for the given *ServiceName* and *Command*. The resulting message object is given an identification number which is returned in **MessageIDPtr*. All functions dealing with this message must provide this message ID number.

The *ServiceName* and *Command* are used to access configuration information which describes the attributes of the message. If such configuration information is not available, an empty message object is created, and a warning message is logged.

MC_Open_Message generates in each message object created the set of "command type" attributes used by most DICOM services. While Merge DICOM Toolkit sets the values of many of these "command type" attributes automatically, some services require the application to set them. (Refer to the description of **MC_Send_Request_Message** for more information.)

MC_Open_Empty_Message should be used if the service and command are not yet known, or if there is no need to validate that values will be set only for attributes assigned to a given service/command pair.

Command	Description
C_STORE_RQ	DICOM Composite Store Service Request
C_STORE_RSP	DICOM Composite Store Service Response
C_ECHO_RQ	DICOM Verification Service Request
C_ECHO_RSP	DICOM Verification Service Response

The following commands are supported by the Merge DICOM Toolkit:

[©] Copyright Merge Healthcare Solutions Inc. 2023

C_FIND_RQ	DICOM Composite Find Service Request
C_FIND_RSP	DICOM Composite Find Service Response
C_CANCEL_FIND_RQ	Cancel DICOM Composite Find Service Request
C_GET_RQ	DICOM Composite Get Service Request
C_GET_RSP	DICOM Composite Get Service Response
C_CANCEL_GET_RQ	Cancel DICOM Composite Get Service Request
C_MOVE_RQ	DICOM Composite Move Service Request
C_MOVE_RSP	DICOM Composite Move Service Response
C_CANCEL_MOVE_RQ	Cancel DICOM Composite Move Service Request
N_EVENT_REPORT_RQ	DICOM Normalized Report Service Request
N_EVENT_REPORT_RQ N_EVENT_REPORT_RSP	DICOM Normalized Report Service Request DICOM Normalized Report Service Response
N_EVENT_REPORT_RSP	DICOM Normalized Report Service Response
N_EVENT_REPORT_RSP N_GET_RQ	DICOM Normalized Report Service Response DICOM Normalized Get Service Request
N_EVENT_REPORT_RSP N_GET_RQ N_GET_RSP	DICOM Normalized Report Service Response DICOM Normalized Get Service Request DICOM Normalized Get Service Request
N_EVENT_REPORT_RSP N_GET_RQ N_GET_RSP N_SET_RQ	DICOM Normalized Report Service Response DICOM Normalized Get Service Request DICOM Normalized Get Service Request DICOM Normalized Set Service Request
N_EVENT_REPORT_RSP N_GET_RQ N_GET_RSP N_SET_RQ N_SET_RSP	DICOM Normalized Report Service Response DICOM Normalized Get Service Request DICOM Normalized Get Service Request DICOM Normalized Set Service Request DICOM Normalized Set Service Response
N_EVENT_REPORT_RSP N_GET_RQ N_GET_RSP N_SET_RQ N_SET_RSP N_ACTION_RQ	DICOM Normalized Report Service Response DICOM Normalized Get Service Request DICOM Normalized Get Service Request DICOM Normalized Set Service Request DICOM Normalized Set Service Response DICOM Normalized Action Service Request
N_EVENT_REPORT_RSP N_GET_RQ N_GET_RSP N_SET_RQ N_SET_RSP N_ACTION_RQ N_ACTION_RSP	DICOM Normalized Report Service Response DICOM Normalized Get Service Request DICOM Normalized Get Service Request DICOM Normalized Set Service Request DICOM Normalized Set Service Response DICOM Normalized Action Service Request DICOM Normalized Action Service Response
N_EVENT_REPORT_RSP N_GET_RQ N_GET_RSP N_SET_RQ N_SET_RSP N_ACTION_RQ N_ACTION_RSP N_CREATE_RQ	DICOM Normalized Report Service Response DICOM Normalized Get Service Request DICOM Normalized Get Service Request DICOM Normalized Set Service Request DICOM Normalized Set Service Response DICOM Normalized Action Service Request DICOM Normalized Action Service Response DICOM Normalized Create Service Request
N_EVENT_REPORT_RSP N_GET_RQ N_GET_RSP N_SET_RQ N_SET_RSP N_ACTION_RQ N_ACTION_RSP N_CREATE_RQ N_CREATE_RSP	DICOM Normalized Report Service Response DICOM Normalized Get Service Request DICOM Normalized Get Service Request DICOM Normalized Set Service Request DICOM Normalized Set Service Response DICOM Normalized Action Service Request DICOM Normalized Action Service Response DICOM Normalized Create Service Response

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these commands.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_COMMAND	Command is not a supported command.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file.

MC_SYSTEM_ERROR

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Open_Empty_Message MC_Free_Item MC_Free_Message

MC_Open_Secure_Association MC_Open_Secure_Association_With_Reject_Info

Establishes a connection with a remote DICOM application over a secure socket connection.

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Open Secure Association (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char * RemoteHostTCPIPName,
      char *ServiceList,
      SecureSocketFunctions *SS functions,
      void *SS application context
)
typedef struct {
      unsigned short Result;
      unsigned short Source;
      unsigned short Reason;
} AssocRejectInfo;
MC STATUS MC Open Secure Association With Reject Info (
      int ApplicationID,
      int *AssociationID,
      const char *RemoteApplicationTitle,
      int *RemoteHostPortNumber,
      char *RemoteHostTCPIPName,
      char *ServiceList,
      SecureSocketFunctions *SS functions,
      void *SS application context,
      AssocRejectInfo RejectInfo
)
 ApplicationID
                       The application ID returned from the
                       MC_Register_Application function.
                       The identification number of an association object
 AssociationID
                       is returned here.
 RemoteApplicationTitle
                       The DICOM Application Title of the remote
                       application.
```

	Pointer to a structure containing functions that will be called by Merge DICOM Toolkit while processing network I/O over the secure connection. (see below)
Each of the following is optio	nal. Use NULL if not used
RejectInfo	If the association is rejected, this return argument will contain the result/source/reason codes explaining the reason for rejection (See DICOM PS3.8, Section 9.3.4).
RemoteHostPortNumbe r	The TCP/IP port used by the remote application to "listen" for DICOM associations
RemoteHostTCPIPName	The remote host's TCP/IP Name. This parameter can be a hostname, an IPv4 address, or an IPv6 address.
ServiceList	Name of a service list in the Merge DICOM Toolkit configuration file.
SS_context	An optional pointer to application-specific data that Merge DICOM Toolkit passes to the functions declared in <i>SS_functions</i> .

Remarks

MC_Open_Secure_Association establishes a secure DICOM association connection with a remote DICOM application. Use **MC_Open_Association** if a <u>secure</u> network connection is not required.

Each application in a DICOM association has a publicly known name or "application title". This application's title was declared in the **MC_Register_Application** call. The application to which we are intending to connect is specified in *RemoteApplicationTitle*. The remote DICOM system waits for association requests on a given TCP/IP port. That port number is specified by *RemoteHostPortNumber*. The TCP/IP name of the remote host is specified by *RemoteHostTCPIPName*.

Starting a DICOM association is a negotiated process. One application provides one or more services, and the other uses one or more of the services provided. The **MC_Open_Secure_Association** function lets the remote DICOM process know which services this application wishes to use. This "service list" is defined in the Merge DICOM Toolkit Application Configuration file, and the name of the appropriate service list is specified by the *ServiceList* parameter.

A few DICOM services require the applications to negotiate application-specific information. This is performed by sharing "extended negotiation information" when the association is negotiated. If one or more of the services in *ServiceList* requires such negotiation, **MC_Set_Negotiation_Info** should be called for each such service to supply the negotiation information.

If the remote system accepts the association request (i.e. it can support one or more of the services in the *ServiceList*), **MC_Open_Secure_Association** creates an association object, puts its ID in **AssociationID*, and returns MC_NORMAL_COMPLETION.

[©] Copyright Merge Healthcare Solutions Inc. 2023

If the service requires the remote application to return extended negotiation information, **MC_Get_Negotiation_Info** should be called to retrieve such information. If the information is unacceptable, **MC_Close_Association** should be called.

NOTE:That the information registered with the MC_Set_Negotiation_Info call
is not changed during the association process. UseMC_Clear_Negotiation_Info to prevent the information from being
used in subsequent association negotiations for a given service.

MC_ASSOCIATION_REJECTED is returned if the remote system rejects the association request. **AssociationID* is NOT valid if the association is rejected or if an error occurs.

NOTE:Only four parameters (ApplicationID, AssociationID,
RemoteApplicationTitle and SS_functions) are required.
RemoteHostPortNumber, RemoteHostTCPIPName and ServiceList
will be defaulted if they are specified as NULL. The default values are
obtained from the RemoteApplicationTitle's section in the Merge
DICOM Toolkit Application Configuration file.

SecureSocketFunctions

When using the **MC_Open_Secure_Association** call, Merge DICOM Toolkit establishes a TCP/IP connection and then calls the functions provided by the *SS_functions* parameter to establish the secure connection and to pass data through the secure connection. The *SS_functions* are responsible for sending and receiving all data through the secure connection, thus allowing them to do so using a secure protocol such as Secure Socket Layer(SSL). Merge DICOM Toolkit closes the underlying TCP/IP connection when all association processing has completed and after it calls the **SS_Session_Shutdown** callback.

The **SecureSocketFunctions** structure is declared in mergecom.h as follows:

```
typedef struct MC Secure Socket Functions Struct
{
   SS STATUS (NOEXP FUNC *SS Session Start) (
       MC SOCKET SocketToUse,
       CONN TYPE
                    ConnectionType,
       void
                    *ApplicationContext,
       void
                    **SecurityContext
   );
   SS STATUS (NOEXP FUNC *SS Read) (
                    *SScontext,
       void
                    *ApplicationContext,
       void
                    *Buffer,
       char
       unsigned int BytesToRead,
       unsigned int *BytesRead,
       int
                    Timeout
   ):
   SS STATUS (NOEXP FUNC *SS Write) (
       void
                    *SScontext,
       void
                    *ApplicationContext,
       char
                    *Buffer,
```

Merge DICOM Toolkit[™] 5.16.0 C/C++ REFERENCE MANUAL

```
unsigned int BytesToWrite,
unsigned int *BytesWritten,
int Timeout
);
void (NOEXP_FUNC *SS_Session_Shutdown) (
void *SScontext,
void *ApplicationContext
);
} SecureSocketFunctions;
```

You must provide valid function pointers for each of the four fields in the **SecureSocketFunctions** structure.

SS_Session_Start

Merge DICOM Toolkit calls the **SS_Session_Start** function just after it has established a TCP/IP connection with the remote host. The *SocketToUse* parameter contains the socket assigned to the connection. <u>Please note that the connection is non-blocking</u>. The *ApplicationContext* parameter is the presented by the *SS_context* parameter of the **MC_Open_Secure_Association** call. *ConnectionType* will be the manifest constant REQUESTER_CONNECTION if the **SS_Session_Start** function is called as a result of a **MC_Open_Secure_Association** call. (It will be ACCEPTOR_CONNECTION if it is called as a result of a **MC_Wait_For_Secure_Association** call.)

The **SS_Session_Start** function is responsible for establishing a secure connection using the socket provided. It is assumed, but not required, that the connection will be a Secure Socket Layer (SSL) or Transport Layer Socket (TLS) connection. A pointer to a context block should be returned at **SecurityContext* if the secure connection is established. Merge DICOM Toolkit will provide this pointer when it calls the other callback functions.

SS_Session_Start must return SS_NORMAL_COMPLETION if the secure connection was successfully established, otherwise SS_ERROR must be returned. If it returns SS_ERROR, the TCP/IP connection will be closed and the **MC_Open_Secure_Association** call will return a status of **MC_NEGOTIATION_ABORTED**.

SS_Session_Shutdown

Merge DICOM Toolkit calls the **SS_Session_Shutdown** function when the association is aborted or closed. It is the responsibility of the **SS_Session_Shutdown** function to gracefully close the secure network connection. Merge DICOM Toolkit closes the TCP/IP socket connection after calling **SS_Session_Shutdown**. The *ApplicationContext* parameter is that presented by the *SS_context* parameter of the **MC_Open_Secure_Association** call, and the *SScontext* parameter is that previously returned by the **SS_Session_Start** function.

SS_Read

Merge DICOM Toolkit calls the **SS_Read** function whenever it needs association data from the secure connection. The *ApplicationContext* parameter is that presented by the *SS_context* parameter of the **MC_Open_Secure_Association** call, and the *SScontext* parameter is that previously returned by the **SS_Session_Start** function. It is the responsibility of the **SS_Read** function to retrieve into *Buffer* the number of unencrypted data bytes specified by *BytesToRead*. The actual number of bytes placed in

 $[\]ensuremath{\textcircled{O}}$ Copyright Merge Healthcare Solutions Inc. 2023

the *Buffer* is returned at **BytesRead*. SS_NORMAL_COMPLETION must be returned if the read request was satisfied.

If the **SS_Read** function cannot retrieve *BytesToRead* bytes in *Timeout* seconds, it must return SS_TIMEOUT. Please note that the socket connection passed by Merge DICOM Toolkit to the **SS_Session_Start** function is non-blocking. (Note that if SS_TIMEOUT is returned, an outstanding MC_Read_Message call will return MC_TIMEOUT.)

If it is determined that the secure socket connection has closed, or that the underlying transport has closed, SS_SESSION_CLOSED must be returned. This should only occur during a DICOM association if the remote host aborted the association. If a fatal error occurs while processing the read request, SS_ERROR must be returned. (Note that if SS_SESSION_CLOSED or SS_ERROR is returned, an outstanding MC_Read_Message call will return MC_NETWORK_SHUT_DOWN.)

SS_Write

Merge DICOM Toolkit calls the **SS_Write** function whenever it needs to send association data over the secure connection. The *ApplicationContext* parameter is that presented by the *SS_context* parameter of the **MC_Open_Secure_Association** call, and the *SScontext* parameter is that previously returned by the **SS_Session_Start** function. It is the responsibility of the **SS_Write** function to send *BytesToWrite* bytes of the data in *Buffer* over the secure connection, returning the number of bytes actually written at **BytesWritten*. SS_NORMAL_COMPLETION must be returned if the write request was satisfied.

If the **SS_Write** function cannot send *BytesToWrite* bytes in *Timeout* seconds, it must return SS_TIMEOUT. Please note that the socket connection passed by Merge DICOM Toolkit to the **SS_Session_Start** function is non-blocking.

If a fatal error occurs while processing the write request, SS_ERROR must be returned. (Note that if SS_ERROR is returned, an outstanding MC_Send_Request_Message, MC_Send_Request, MC_Send_Response_Message or MC_Send_Response call will return MC_SYSTEM_ERROR.)

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally. <i>AssociationID</i> has been set to the association's identification number.
MC_ASSOCIATION_REJECTED	The remote system rejected the open association request.
MC_NEGOTIATION_ABORTED	The association was aborted during negotiation
MC_CONNECTION_FAILED	The remote system could not be connected to at the TCP/IP level. Check that the remote host name and port number have been configured properly.
MC_NULL_POINTER_PARM	<i>AssociationID</i> , <i>SS_functions</i> or <i>RemoteApplicationTitle</i> was NULL.

MC_NULL_VALUE	One or more of the function parameters within <i>SS_functions</i> was NULL.
MC_INVALID_APPLICATION_TITLE	<i>RemoteApplicationTitle</i> was not 1-16 bytes long.
MC_INVALID_APPLICATION_ID	<i>ApplicationID</i> does not identify a valid Merge DICOM Toolkit application.
MC_INVALID_PORT_NUMBER	<i>RemoteHostPortNumber</i> was not NULL and pointed at a negative number.
MC_INVALID_SERVICE_LIST_NAME	ServiceList was not NULL and pointed at an empty string.
MC_INVALID_HOST_NAME	<i>RemoteHostTCPIPName</i> was not NULL and it was not 1- 39 bytes long.
MC_TIMEOUT	Timeout attempting the association.
MC_SYSTEM_CALL_INTERRUPTED	The operating system interrupted the network call. Retry the connection.
MC_NO_APPLICATIONS_REGISTERE	D An application title for this application has not yet be registered.
MC_LIBRARY_NOT_INITIALIZED	This call was made prior to the initialization of the library
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Open_Association MC_Wait_For_Association MC_Set_Negotiation_Info MC_Clear_Negotiation_Info	MC_Register_Application MC_Wait_For_Secure_Association MC_Get_Negotiation_Info MC_Close_Association
MC_Abort_Association	MC_Wait_For_Connection

MC_Wait_For_Connection_On_Port MC_Process_Association_Request

MC_Process_Secure_Association_Request

MC_Process_Association_Request MC_Process_Secure_Association_Request

Process a standard association request or secure association request that was received by the MC_Wait_For_Connection or MC_Wait_For_Connection_on_Port functions. It also can be used to process an association received by the internet services daemon (inetd) on UNIX systems.

Synopsis

MC STATUS MC Brock	cess Secure Association Request (
MC_SIAIOS MC_FIOC	
—	*ServiceListName,
int *Applic	
int *Associ	•
	EFunctions *SS functions,
	oplication context
)	
,	
Socket	Socket returned by MC_Wait_For_Connection or
	MC_Wait_For_Connection_On_Port.
ServiceListName	Name of a service list in the Merge DICOM Toolkit
DerviceListivarrie	_
	configuration file.
ApplicationID	The identification number of a previously registered
	application is returned here.
1 <i></i>	
AssociationID	The identification number of an association object is
	returned here.
SS functions	Pointer to a structure containing functions that will be
	called by Merge DICOM Toolkit while processing
	network I/O over the secure connection.
	The SecureSocketFunctions structure is defined in
	"mergecom.h" (see MC_Open_Secure_Association)
SS context	An optional pointer to application-specific data that
00_00/110/1	
	Merge DICOM Toolkit passes to the functions declared
	in <i>SS_functions</i> .

Remarks

MC_Process_Association_Request can be used in conjunction with the MC_Wait_For_Connection and MC_Wait_For_Connection_On_Port routines to handle incoming connections.

MC_Process_Secure_Association_Request can also be used in conjunction with MC_Wait_For_Connection and MC_Wait_For_Connection_On_Port to process secure DICOM association requests. See the description to MC_Wait_For_Secure_Association for details on how to use secure callback functions.

MC_Process_Association_Request and **MC_Process_Secure_Association_Request** can also be used on UNIX toolkit platforms in conjunction with a UNIX OS's internet services daemon (inetd). In these situations, inetd is configured to handle the TCP/IP connection request on a specific listen port. Once a connection request is received, inetd performs a fork/exec on a pre-configured user program to handle the connection. inetd supplies the socket to the process by setting stdin to the socket's value. In this scenario, stdin must be used as the *Socket* parameter to **MC_Process_Association_Request**.

Once the user's program is invoked, this function is used in the place of **MC_Wait_For_Association** to complete the association request with one of the previously registered applications.

inetd must be properly configured in order correctly perform the listen on the correct TCP/IP socket and to invoke the user's application. There are some differences between UNIX versions, so the

[©] Copyright Merge Healthcare Solutions Inc. 2023

following is only an example configuration. Please consult the appropriate UNIX man pages or documentation for more information.

The server must be specified in the systems services file (typically in /etc/inet/ or in /etc directory). E.g.

dicom_echo 104/tcp

The server must also be specified in the inetd configuration file (typically /etc/inet/inetd.conf or /etc/inetd.conf). E.g.:

dicom_echo stream tcp nowait root /usr/bin/inetd_echo_scp inetd_echo_scp /usr/bin/merge.ini

Where:

"dicom_echo" must be defined in services as servicing the DICOM port (e.g. 104)

"stream" - always

"tcp" - always

"nowait" - always

"root" - from /etc/passwd: typically root

"/usr/bin/inetd_echo_scp" - full path name to this runtime application

"inetd_echo_scp" - argv[0]

"/usr/bin/merge.ini" - argv[1] - The application's arguments, if any.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_TIMEOUT	The <i>Timeout</i> period expired.	
MC_NEGOTIATION_ABORTED	An incoming association was aborted during negotiation. Normally this situation is handled by retrying the MC_Process_Association_Request call.	
MC_SYSTEM_CALL_INTERRUPTED		
	The system call used to wait for an association request was interrupted by a signal. Normally this situation is handled by retrying the MC_Process_Association_Request call.	
MC_NULL_POINTER_PARM	<i>AssociationID, ApplicationID</i> or <i>ServiceList</i> parameter was NULL.	
MC_NO_APPLICATIONS_REGISTERED		
	No applications have been registered using MC_Register_Application.	

MC_INVALID_SERVICE_LIST_NAME

	ServiceList points at a null string.
MC_SYSTEM_ERROR	An unexpected, potentially serious problem was detected in the operating environment. A message describing the error had been written to the Merge DICOM Toolkit log file.
MC_ASSOCIATION_REJECTED	The association was rejected as invalid.
MC_ASSOCIATION_CLOSED	The association was closed after handling an echo request.
See Also	
MC_Wait_For_Association MC_Wait_For_Connection	MC_Wait_For_Secure_Association MC_Wait_For_Connection_On_Port

MC_Read_Message

Reads the next message sent by the remote application.

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Read Message (
       int AssociationID,
       int Timeout,
       int *MessageID,
       char **ServiceName,
       MC COMMAND *Command
)
 AssociationID
                 The association object's identification number
 Timeout
                 The maximum time (in seconds) to wait for a message to
                 start arriving. A value of zero (0) means "check one time"
                 and a value of minus one (-1) means "wait forever".
 MessageID
                 The ID of a message object containing the attributes of the
```

```
ServiceNameThe name of the service associated with MessageID is<br/>returned here. Note well that this memory is freed when the<br/>MC_Free_Message() function is called for MessageID.CommandThe command which is associated with the message is<br/>returned here. The MC_COMMAND enumerated values are<br/>defined in "mc3msg.h".
```

Remarks

MC_Read_Message allows the caller to wait for the arrival of a request or response message from the remote application. If a message is not received on the association identified by *AssociationID* within *Timeout* seconds, the function returns a status of MC_TIMEOUT.

NOTE: If *Timeout* is set to less than zero, **MC_Read_Message** will wait forever for a message to arrive. If it is set to zero, it will check one time if a message has arrived and then return.

Once a message begins arriving, the **INACTIVITY_TIMEOUT** configuration value is used. While receiving the message, the association will be aborted if there is no activity on the network for the length of time specified by this configuration value.

If a message arrives, **MessageID* is set to the ID of a message object which has been opened to receive the attributes of the message. The message's attribute values have been set by the DICOM message stream received from the remote application. If **MC_Register_Callback_Function** was called for any message attributes, those values were passed on to the registered callback function by the time this function (**MC_Read_Message**) returns.

ServiceName is set to the name of the service associated with *MessageID* and *Command* is set to the command associated with *MessageID*.

If MC_Read_Message returns either MC_ASSOCIATION_CLOSED, MC_ASSOCIATION_ABORTED, MC_NETWORK_SHUT_DOWN, MC_INACTIVITY_TIMEOUT, MC_CONFIG_INFO_ERROR or MC_INVALID_MESSAGE_RECEIVED no further calls may be made for the association and no message ID is returned at *MessageID.

The following commands are supported by the Merge DICOM Toolkit:

Command	Description
C_STORE_RQ	DICOM Composite Store Service Request
C_STORE_RSP	DICOM Composite Store Service Response
C_ECHO_RQ	DICOM Verification Service Request
C_ECHO_RSP	DICOM Verification Service Response
C_FIND_RQ	DICOM Composite Find Service Request
C_FIND_RSP	DICOM Composite Find Service Response
C_CANCEL_FIND_RQ	Cancel DICOM Composite Find Service Request
C_GET_RQ	DICOM Composite Get Service Request
C_GET_RSP	DICOM Composite Get Service Response
C_CANCEL_GET_RQ	Cancel DICOM Composite Get Service Request
C_MOVE_RQ	DICOM Composite Move Service Request
C_MOVE_RSP	DICOM Composite Move Service Response

© Copyright Merge Healthcare Solutions Inc. 2023

NOTE: It is the responsibility of the caller to release the message object. This is done using the MC_Free_Message call.

C_CANCEL_MOVE_RQ Cancel DICOM Composite Move Service Request

N_EVENT_REPORT_RQ	DICOM Normalized Report Service Request
N_EVENT_REPORT_RSP	DICOM Normalized Report Service Response
N_GET_RQ	DICOM Normalized Get Service Request
N_GET_RSP	DICOM Normalized Get Service Request
N_SET_RQ	DICOM Normalized Set Service Request
N_SET_RSP	DICOM Normalized Set Service Response
N_ACTION_RQ	DICOM Normalized Action Service Request
N_ACTION_RSP	DICOM Normalized Action Service Response
N_CREATE_RQ	DICOM Normalized Create Service Request
N_CREATE_RSP	DICOM Normalized Create Service Response
N_DELETE_RQ	DICOM Normalized Delete Service Request
N_DELETE_RSP	DICOM Normalized Delete Service Response

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these commands.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_TIMEOUT	The <i>Timeout</i> period expired.
MC_MUST_CONTINUE_BEFORE_RE	ADING A previous message was not read completely before making this call.
MC_NULL_POINTER_PARM	The <i>MessageID</i> , <i>ServiceName</i> or <i>Command</i> parameter was NULL.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of th	e following are returned:
MC_NETWORK_SHUT_DOWN	The network connect unexpectedly dropped.
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_ASSOCIATION_CLOSED	The association has been closed.
MC_INACTIVITY_TIMEOUT	A timeout occurred in the middle of receiving a message.
MC_CONFIG_INFO_ERROR	The message information file describing the message's service/command pair could not be accessed.

MC_INVALID_MESSAGE_RECEIVED An improperly formatted DICOM message was received. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Read_MessageMC_Free_MessageMC_Register_Callback_FunctionMC_Continue_Read_MessageMC_Continue_Read_Message_To_Tag

MC_Read_Message_To_Tag

Reads the next message sent by the remote applicate up until a specified tag.

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Read Message To Tag (
       int AssociationID,
       int Timeout,
       unsigned long StopTag,
       int *MessageID,
       char **ServiceName,
       MC COMMAND *Command
)
 AssociationID
                 The association object's identification number
 Timeout
                 The maximum time (in seconds) to wait for a message to
                  start arriving. A value of zero (0) means "check one time"
                  and a value of minus one (-1) means "wait forever".
 StopTag
                  Read up to and include this tag when reading from the
                  remote application. Must be outside the command section (
                 greater than or equal to (0001,0000))
                 The ID of a message object containing the attributes of the
 MessageID
                  message is returned here.
 ServiceName
                 The name of the service associated with MessageID is
                  returned here. Note well that this memory is freed when the
                  MC_Free_Message() function is called for MessageID.
 Command
                 The command which is associated with the message is
                  returned here. The MC_COMMAND enumerated values are
                  defined in "mc3msg.h".
```

Remarks

MC_Read_Message_To_Tag allows the caller to wait for the arrival of a request or response message from the remote application. If a message is not received on the association identified by *Association/D* within *Timeout* seconds, the function returns a status of MC_TIMEOUT. Note that if *Timeout* is set to less than zero, **MC_Read_Message_To_Tag** will wait forever for a message to arrive. If it is set to zero, it will check one time if a message has arrived and then return.

Once a message begins arriving, the **INACTIVITY_TIMEOUT** configuration value is used. While receiving the message, the association will be aborted if there is no activity on the network for the length of time specified by this configuration value.

If a message arrives, **MessageID* is set to the ID of a message object which has been opened to receive the attributes of the message, up to, and including, the specified *StopTag*. The message's attribute values have been set by the DICOM message stream received from the remote application. If **MC_Register_Callback_Function** was called for any message attributes, those values were passed on to the registered callback function by the time this function (**MC_Read_Message_To_Tag**) returns.

The message prior to this call must have been read with **MC_Read_Message**, or read up to the last tag (FFFF,FFFF) either explicitly with **MC_Continue_Read_Message_To_Tag**, or by using **MC_Continue_Read_Message**.

The message returned must be read until the last tag (FFF,FFF) either explicitly with **MC_Continue_Read_Message_To_Tag**, or by calling **MC_Continue_Read_Message** before using **MC_Read_Message** or **MC_Read_Message_To_Tag** again.

ServiceName is set to the name of the service associated with *MessageID* and *Command* is set to the command associated with *MessageID*.

NOTE: It is the responsibility of the caller to release the message object. This is done using the MC_Free_Message call.

If MC_Read_Message_To_Tag returns either MC_ASSOCIATION_CLOSED, MC_ASSOCIATION_ABORTED, MC_NETWORK_SHUT_DOWN, MC_INACTIVITY_TIMEOUT, MC_CONFIG_INFO_ERROR or MC_INVALID_MESSAGE_RECEIVED no further calls may be made for the association and no message ID is returned at **MessageID*.

The following commands are supported by the Merge DICOM Toolkit:

Command	Description
C_STORE_RQ	DICOM Composite Store Service Request
C_STORE_RSP	DICOM Composite Store Service Response
C_ECHO_RQ C_ECHO_RSP	DICOM Verification Service Request DICOM Verification Service Response
C_FIND_RQ	DICOM Composite Find Service Request
C_FIND_RSP	DICOM Composite Find Service Response
C_CANCEL_FIND_RQ	Cancel DICOM Composite Find Service Request
C_GET_RQ C_GET_RSP C_CANCEL_GET_RQ	DICOM Composite Get Service Request DICOM Composite Get Service Response Cancel DICOM Composite Get Service Request

[©] Copyright Merge Healthcare Solutions Inc. 2023

C_MOVE_RQ C_MOVE_RSP C_CANCEL_MOVE_RQ	DICOM Composite Move Service Request DICOM Composite Move Service Response Cancel DICOM Composite Move Service Request
N_EVENT_REPORT_RQ	DICOM Normalized Report Service Request
N_EVENT_REPORT_RSP	DICOM Normalized Report Service Response
N_GET_RQ	DICOM Normalized Get Service Request
N_GET_RSP	DICOM Normalized Get Service Request
N_SET_RQ	DICOM Normalized Set Service Request
N_SET_RSP	DICOM Normalized Set Service Response
N_ACTION_RQ	DICOM Normalized Action Service Request
N_ACTION_RSP	DICOM Normalized Action Service Response
N_CREATE_RQ	DICOM Normalized Create Service Request
N_CREATE_RSP	DICOM Normalized Create Service Response
N_DELETE_RQ	DICOM Normalized Delete Service Request
N_DELETE_RSP	DICOM Normalized Delete Service Response

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these commands.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_MUST_CONTINUE_BEFORE_RE	ADING A previous message was not read completely before making this call.	
MC_TIMEOUT	The <i>Timeout</i> period expired.	
MC_NULL_POINTER_PARM	The <i>MessageID</i> , <i>ServiceName</i> or <i>Command</i> parameter was NULL.	
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.	
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.	
MC_INVALID_TAG	<i>StopTag</i> was within the command set (less then (0001,0000)).	
The association is dropped if any of the following are returned:		
MC_NETWORK_SHUT_DOWN	The network connect unexpectedly dropped.	
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.	

© Copyright Merge Healthcare Solutions Inc. 2023

MC_Read_Message MC_Continue_Read_Message MC_Continue_Read_Message_To_Tag	MC_Register_Callback_Function MC_Free_Message «
See Also	
MC_INVALID_MESSAGE_RECEIVED	An improperly formatted DICOM message was received. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CONFIG_INFO_ERROR	The message information file describing the message's service/command pair could not be accessed.
MC_INACTIVITY_TIMEOUT	A timeout occurred in the middle of receiving a message.
MC_ASSOCIATION_CLOSED	The association has been closed.

MC_Read_To_Stream

Reads the message sent by the remote application directly into a streaming function. Especially useful for the transfer of large data which we don't want to load into application memory.

Synopsis

```
#include ``mergecom.h"
```

```
MC_STATUS MC_Read_To_Stream (
    int AssociationID,
    int Timeout,
    char **ServiceName,
    MC_COMMAND *Command
    char *FileName,
    void *UserInfo,
    MC_STATUS (*YourReceiveStreamFunction)()
```

```
)
```

AssociationID	The association object's identification number
Timeout	The maximum time (in seconds) to wait for a message to start arriving. A value of zero (0) means "check one time" and a value of minus one (-1) means "wait forever".
ServiceName	The name of the service is returned here.
Command	The command which is associated with the message is returned here. The MC_COMMAND enumerated values are defined in "mc3msg.h".
Filename	The name of file associated with the receiving message. This may be NULL.
UserInfo	Address of data which will be passed on to <i>YourReceiveStreamFunction</i> each time it is called. This may be NULL.

YourReceiveStre	eamFunctio Name of a function which will be called
п	repeatedly to receive blocks of streamed
	DICOM message data.
The function must	be prototyped as follows:
MC_STATUS Your	rReceiveStreamFunction (
	essageID,
	BuserInfo,
int CBda	
	BdataBuffer,
int CBis int CBis	,
)	suast
/	
CBmessagelD	The internal identifier assigned to the receiving message object. It becomes invalid once the message is received.
CBuserInfo	Address of data which is being passed from the
	MC_Message_To_Stream function. This may be NULL.
CBdataSize	The number of bytes of stream data being provided to you in <i>CBdataBuffer</i> .
CBdataBuffer	The buffer containing stream data from the message object.
CBisFirst	Is TRUE (not zero) when Merge DICOM Toolkit is providing the first block of stream data.
CBisLast	Is TRUE (not zero) when Merge DICOM Toolkit is providing the last block of stream data.

Remarks

MC_Read_To_Stream allows the caller to wait for the arrival of a request or response message from the remote application. If a message is not received on the association identified by *AssociationID* within *Timeout* seconds, the function returns a status of MC_TIMEOUT.

NOTE: If *Timeout* is set to less than zero, **MC_Read_To_Stream** will wait forever for a message to arrive. If it is set to zero, it will check one time if a message has arrived and then return.

Once a message begins arriving, the **INACTIVITY_TIMEOUT** configuration value is used. While receiving the message, the association will be aborted if there is no activity on the network for the length of time specified by this configuration value.

If a message arrives, *YourReceiveStreamFunction* is called repeatedly providing the blocks of data of a message object. The message's attribute values have been set by the DICOM message stream received from the remote application.

ServiceName is set to the name of the service associated with receiving message and *Command* is set to the command associated with *MessageID*.

If MC_Read_To_Stream returns either MC_ASSOCIATION_CLOSED, MC_ASSOCIATION_ABORTED, MC_NETWORK_SHUT_DOWN, MC_INACTIVITY_TIMEOUT,

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_CONFIG_INFO_ERROR or **MC_INVALID_MESSAGE_RECEIVED** no further calls may be made for the association and no message ID is returned at **MessageID*.

The following commands are supported by the Merge DICOM Toolkit:

Command	Description
C_STORE_RQ	DICOM Composite Store Service Request
C_STORE_RSP	DICOM Composite Store Service Response
C_ECHO_RQ	DICOM Verification Service Request
C_ECHO_RSP	DICOM Verification Service Response
C_FIND_RQ	DICOM Composite Find Service Request
C_FIND_RSP	DICOM Composite Find Service Response
C_CANCEL_FIND_RQ	Cancel DICOM Composite Find Service Request
C_GET_RQ	DICOM Composite Get Service Request
C_GET_RSP	DICOM Composite Get Service Response
C_CANCEL_GET_RQ	Cancel DICOM Composite Get Service Request
C_MOVE_RQ	DICOM Composite Move Service Request
C_MOVE_RSP	DICOM Composite Move Service Response
C_CANCEL_MOVE_RQ	Cancel DICOM Composite Move Service Request
N_EVENT_REPORT_RQ	DICOM Normalized Report Service Request
N_EVENT_REPORT_RSP	DICOM Normalized Report Service Response
N_GET_RQ	DICOM Normalized Get Service Request
N_GET_RSP	DICOM Normalized Get Service Request
N_SET_RQ	DICOM Normalized Set Service Request
N_SET_RSP	DICOM Normalized Set Service Response
N_ACTION_RQ	DICOM Normalized Action Service Request
N_ACTION_RSP	DICOM Normalized Action Service Response
N_CREATE_RQ	DICOM Normalized Create Service Request
N_CREATE_RSP	DICOM Normalized Create Service Response
N_DELETE_RQ	DICOM Normalized Delete Service Request
N_DELETE_RSP	DICOM Normalized Delete Service Response

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these commands.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_TIMEOUT	The <i>Timeout</i> period expired.	
MC_MUST_CONTINUE_BEFORE_RE	ADING A previous message was not read completely before making this call.	
MC_NULL_POINTER_PARM	The <i>MessageID</i> , <i>ServiceName</i> or <i>Command</i> parameter was NULL.	
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.	
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.	
The association is dropped if any of the following are returned:		
MC_NETWORK_SHUT_DOWN	The network connect unexpectedly dropped.	
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.	
MC_ASSOCIATION_CLOSED	The association has been closed.	
MC_INACTIVITY_TIMEOUT	A timeout occurred in the middle of receiving a message.	
MC_CONFIG_INFO_ERROR	The message information file describing the message's service/command pair could not be accessed.	
MC_INVALID_MESSAGE_RECEIVED	An improperly formatted DICOM message was received. A message describing the error has been written to the Merge DICOM Toolkit log file.	
See Also		
MC_Read_Message MC_Continue_Read_Message_To_Ta	MC_Continue_Read_Message g	

MC_Register_Application

Registers a Merge DICOM Toolkit application.

Synopsis #include "mergecom.h" MC_STATUS MC_Register_Application (int *ApplicationID, char *AEtitle) ApplicationID A unique application identification number will be returned here

AEtitle The DICOM application title used by this application.

Remarks

MC_Register_Application registers an application with Merge DICOM Toolkit. This function provides Merge DICOM Toolkit with the information it needs to identify this DICOM application and to maintain information specific to this application.

AEtitle is the DICOM name of the application. *AEtitle* must be made known to any other DICOM application which wishes to communicate with this one.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	ApplicationID or AEtitle was NULL.
MC_INVALID_LENGTH_FOR_TITLE	AEtitle must be 1 to 16 bytes long.
MC_ALREADY_REGISTERED	An application title with this name has already been registered.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
	C_Register_Callback_Function IC_Set_Negotiation_Info

MC_Register_Callback_Function MC_Register_pCallback_Function

Registers a callback function to provide or retrieve a standard or private attribute's values as they are transmitted or received on a DICOM association or read or written to a DICOM file

Synopsis

```
#include "mc3msg.h"
MC_STATUS MC_Register_Callback_Function (
            int ApplicationID,
            unsigned long Tag,
            void *UserInfo,
            MC_STATUS (*CallbackFunction)()
)
MC_STATUS MC_Register_pCallback_Function (
            int ApplicationID,
            char *PrivateCode,
            unsigned short Group,
            unsigned char ElementByte,
            void *UserInfo,
            MC_STATUS (*pCallbackFunction)
)
```

ApplicationID	The identifier assigned to this application by the MC_Register_Application function.
Tag	DICOM tag which identifies the attribute to which this callback function applies. <i>Tag</i> must specify an attribute with a value representation of OB, OW, OD, or OF.
PrivateCode	The code string which dictates which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group containing the private attribute. It must be an odd number.
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .
UserInfo	Address of data which will be passed on to <i>CallbackFunction</i> each time it is called. This may be NULL.
<i>CallbackFunction</i> or <i>pCallbackFunction</i>	Name of a function which will be called repeatedly to provide blocks of data or to request blocks of data for the attribute's value.

The functions must be prototyped as follows:

```
MC_STATUS CallbackFunction (
```

```
int CBmessageID,
unsigned long CBtag,
void *CBuserInfo,
CALLBACK_TYPE CBtype,
unsigned long *CBdataSizePtr,
void **CBdataBufferPtr,
int CBisFirst,
int *CBisLastPtr
```

```
MC_STATUS pCallbackFunction (
```

```
int CBmessageID,
```

```
char *CBprivateCode,
unsigned short CBgroup,
unsigned char CBelementByte,
void *CBuserInfo,
CALLBACK_TYPE CBtype,
unsigned long *CBdataSizePtr,
void **CBdataBufferPtr,
int CBisFirst,
int *CBisLastPtr
```

```
)
```

)

```
CBmessageID
```

```
CBtagMC_Open_Message function.CBtagMerge DICOM Toolkit sets this to the DICOM tag<br/>which identifies the attribute.
```

Merge DICOM Toolkit sets this to the identifier assigned to the message, item or file object by the

CBprivateCode	The code string which dictates which block in the private <i>CBgroup</i> "owns" the attribute.
Cbgroup	The number identifying the private group containing the private attribute. It must be an odd number.
CBelementByte	The number identifying the private attribute within the private <i>CBgroup</i> for this <i>CBprivateCode</i> .
CBuserInfo	Merge DICOM Toolkit sets this to the address of data which is being passed from the MC_Register_Callback_Function function. This may be NULL.
CBtype	When the <i>CallbackFunction</i> is called, this is set to one of the following enumerated CALLBACK_TYPE codes as defined in "mergecom.h": REQUEST_FOR_DATA REQUEST_FOR_DATA_LENGTH PROVIDING_DATA PROVIDING_DATA_LENGTH PROVIDING_MEDIA_DATA_LENGTH PROVIDING_OFFSET_TABLE FREE_DATA REQUEST_FOR_DATA_WITH_OFFSET

CbdataSizePtr	If <i>CBtype</i> is REQUEST_FOR_DATA, then you must
	set <i>*CbdataSizePtr</i> to the number of bytes of data
	<u>you</u> are providing at <i>*CBdataBufferPtr.</i> <u>This must be</u>
	<u>an even number and its value must be no larger than</u>
	INT_MAX.
	If CBtype is REQUEST_FOR_DATA_WITH_OFFSET,
	then <i>*CbdataSizePtr</i> is the pointer to long which
	contains the offset of required data in the input
	seekable stream. The stream offset has to be used
	only when <i>CBisFirst</i> to TRUE (not zero), i.e. when you are to provide the first block of data. After the data
	reading <i>*CbdataSizePtr</i> needs to be set to the
	number of bytes of provided data at
	*CBdataBufferPtr. This must be an even number and
	its value must be no larger than INT_MAX.
	If <i>CBtype</i> is PROVIDING_DATA, then Merge DICOM
	Toolkit has set <i>*CBdataSizePtr</i> to the number of
	bytes of data it is providing at <i>*CBdataBufferPtr.</i>
	If <i>CBtype</i> is REQUEST_FOR_DATA_LENGTH then
	you must set <i>*CBdataSizePtr</i> to the total size (in
	bytes) of the attribute's value. <u>This must be an even</u>
	number.
	If <i>CBtype</i> is PROVIDING_DATA_LENGTH then Merge
	DICOM Toolkit has set <i>*CBdataSizePtr</i> to the total
	size (in bytes) of the attribute's value. Merge DICOM
	Toolkit will call the function with <i>CBtype</i> of
	PROVIDING_DATA_LENGTH before it calls the
	function with <i>CBtype</i> of PROVIDING_DATA.
	If <i>CBtype</i> is PROVIDING_MEDIA_DATA_LENGTH
	then Merge DICOM Toolkit has set <i>*CBdataSizePtr</i> to
	the total size (in bytes) of the attribute's value.
	If <i>CBtype</i> is PROVIDING_OFFSET_TABLE then
	Merge DICOM Toolkit has set <i>*CBdataSizePtr</i> to the
	number of bytes of data it is providing at
	*CbdataBufferPtr.
	If <i>CBtype</i> is FREE_DATA then Merge DICOM Toolkit
	has set <i>CBdataSizePtr</i> to NULL.

CBdataBufferPtr	If <i>CBtype</i> is REQUEST_FOR_DATA, REQUEST_FOR_DATA_WITH_OFFSET or PROVIDING_DATA this points at the address of the buffer containing a portion of the attribute's value. The value for this field must be even because of alignment problems on some platforms. If <i>CBtype</i> is REQUEST_FOR_DATA_LENGTH or <i>CBtype</i> is PROVIDING_DATA_LENGTH this parameter is not used.
	If <i>CBtype</i> is PROVIDING_MEDIA_DATA_LENGTH, <i>*CbdataBufferPtr</i> is a pointer to long that contains the offset of the attribute's value from the beginning of the DICOM file.
	If <i>CBtype</i> is FREE_DATA then Merge DICOM Toolkit has set <i>CBdataBufferPtr</i> to NULL.
CBisFirst	If <i>CBtype</i> is REQUEST_FOR_DATA, then Merge DICOM Toolkit will set <i>CBisFirst</i> to TRUE (not zero) the first time it is requesting data for this attribute's value (i.e. when you are to provide the first block of data).
	If <i>CBtype</i> is PROVIDING_DATA , then Merge DICOM Toolkit will set <i>CBisFirst</i> to TRUE (not zero) the first time it is providing data for this attribute's value (i.e. when it is providing the first block of data).
	If <i>CBtype</i> is PROVIDING_OFFSET_TABLE, then Merge DICOM Toolkit will set <i>CBisFirst</i> to TRUE (not zero) the first time it is providing data for the basic offset table.
	If <i>CBtype</i> is REQUEST_FOR_DATA_LENGTH, <i>CBtype</i> is PROVIDING_DATA_LENGTH, or <i>CBtype</i> is PROVIDING_MEDIA_DATA_LENGTH this parameter is not used.
CBisLastPtr	If <i>CBtype</i> is REQUEST_FOR_DATA or REQUEST_FOR_DATA_WITH_OFFSET, then you must set <i>*CBisLastPtr</i> to TRUE (not zero) the last time you are providing data for this attribute's value (i.e. when you are providing the last block of data).
	If <i>CBtype</i> is PROVIDING_DATA, then Merge DICOM Toolkit has set <i>*CBisLastPtr</i> to TRUE (not zero) because this is the last time it will be providing data for this attribute's value (i.e. when it is providing the last block of data).

If *CBtype* is REQUEST_FOR_DATA_LENGTH, *CBtype* is PROVIDING_DATA_LENGTH or *CBtype* is PROVIDING_MEDIA_DATA_LENGTH this parameter is not used.

Remarks

MC_Register_Callback_Function registers *CallbackFunction* to provide or retrieve an attribute's values as they are transmitted or received on a DICOM association or read or written to a DICOM file. The callback will also be called if the user calls **MC_Get_Value_To_Function** or

MC_Set_Value_From_Function although this should not happen because the user already is storing the OB, OW, or OF data. **MC_Register_Callback_Function** may only be used for attributes with a value representation of OB, OW, or OF.

MC_Register_pCallback_Function registers *pCallbackFunction* to provide or retrieve a private attribute's values as they are transmitted or received on a DICOM association or read or written to a DICOM file. The callback will also be called if the user calls MC_Get_pValue_To_Function or MC_Set_pValue_From_Function although this should not happen because the user already is storing the OB, OW, or OF data. MC_Register_Callback_Function may only be used for attributes with a value representation of OB, OW, or OF. Subsequent descriptions in this section are applicable for both standard and private registered callback functions.

- **NOTE:** If a value has already been set for *Tag*, *CallbackFunction* will not be called to set or get its value.
- **NOTE:** The **CALLBACK_MIN_DATA_SIZE** configuration value specifies the minimum sized value for which a callback function is used. This configuration option can be used so that only messages or files with very large attribute values use a callback function.
- **NOTE:** Some message definitions contain OB, OW, or OF data but a value does not have to be set for these attributes (such as C-FIND-RQ and C-FIND-RSP messages). To have the registered callback function ignored, the attribute should be deleted from the message with the **MC_Delete_Attribute** function.

If a callback function was already registered, *CallbackFunction* replaces the function previously registered. A callback function may be "de-registered" by calling **MC_Release_Callback_Function** for the same attribute.

The **MC_Register_Callback_Function** caller may provide the *CallbackFunction* with a pointer to agreed upon data in its *UserInfo* parameter. This is optional and *UserInfo* may be NULL.

Network Communications:

When the Merge DICOM Toolkit communications software receives a message on a DICOM association, it uses **MC_Open_Message** to acquire a message compatible with the message being

[©] Copyright Merge Healthcare Solutions Inc. 2023

received. It then populates the message with the values it receives over the association. When it receives attributes of type OB, OW, or OF (or any other large values) the message library stores the values in a configurable location (normally in temporary files). Then, the receiving application could use any of the "Get" functions (including **MC_Get_Value_To_Function**) to retrieve attribute values from the message object.

Some applications may wish to bypass the storing of these large values in the message object. To do so, the **MC_Register_Callback_Function** is used to register a function which will be called as the attribute's value arrives on the association; blocks of the value are then passed to this *"CallbackFunction"* instead of storing the value in the message.

Similarly, a client application may use MC_Register_Callback_Function to register a callback function to provide Merge DICOM Toolkit with an attribute's value as Merge DICOM Toolkit needs the value for transmission on the network. "*CallbackFunction*" will be called when *Tag* is encountered in any message being transferred between this *ApplicationID* and a remote application.

Media Applications:

When the Merge DICOM Toolkit media software reads a file with **MC_Open_File**, it populates the file object with the values it receives from media. When it receives attributes of type OB, OW, or OF (or any other large values) the file library stores the values in a configurable location (normally in temporary files). Then, the reading application could use any of the "Get" functions (including **MC_Get_Value_To_Function**) to retrieve attribute values from the file object.

Some applications may wish to bypass the storing of these large values in the file object. To do so, the **MC_Register_Callback_Function** is used to register a function which will be called as the attribute's value is read from media. The user has two options when the attribute's value is read from media. If the file is being read with the function **MC_Open_File** or **MC_Open_File_Upto_Tag** it behaves in the same manner as with associations. The OB, OW, or OF data is passed to *CallbackFunction* instead of storing it in the file object. However, when **MC_Open_File_Bypass_OBOW or** MC_Open_File_Upto_Tag_Bypass_Value are used, the OB, OW, or OF data is left on media. The location and length of the data within the file is passed to *CallbackFunction*. The actual data is not passed. Whenever *Tag* is encountered in any file being read within *ApplicationID*, *CallbackFunction* will be called.

In some cases, *CallbackFunction* must also supply Merge DICOM Toolkit with the OB, OW, or OF data. If the file object is converted to a message object with **MC_File_To_Message** and sent over the network, Merge DICOM Toolkit will request the attribute's value from *CallbackFunction* when it needs the value for transmission. Similarly, if the file object is re-written to media, *CallbackFunction* will also be requested for the attribute's value.

The CallbackFunction:

Merge DICOM Toolkit passes the *CallbackFunction* both *CBmessageID* and *CBtag* to identify which attribute is involved. **MC_Get_Message_Service** may be called to identify which service and command this *CBmessageID* relates to. The *CBuserInfo* parameter is optional data from the **MC_Register_Callback_Function** caller.

CallbackFunction must return a status of **MC_NORMAL_COMPLETION** if it could accommodate the callback request, or **MC_CANNOT_COMPLY**, if not. These status codes are defined in "mcstatus.h."

[©] Copyright Merge Healthcare Solutions Inc. 2023

<u>CBtype is set to REQUEST_FOR_DATA_LENGTH</u> if Merge DICOM Toolkit is requesting the size (in bytes) of the attribute's value. In this case *CallbackFunction* must set **CBdataSizePtr* to the attribute's value size. Merge DICOM Toolkit will call the *CallbackFunction* with *CBtype* set to REQUEST_FOR_DATA_LENGTH <u>before</u> it begins calling *CallbackFunction* with *CBtype* set to REQUEST_FOR_DATA.

<u>CBtype is set to PROVIDING_DATA_LENGTH</u> if Merge DICOM Toolkit is providing *CallbackFunction* with the total size of the attribute's value. The value size (in bytes) is placed at **CBdataSizePtr*. Merge DICOM Toolkit will call the *CallbackFunction* with *CBtype* set to PROVIDING_DATA_LENGTH <u>before</u> it begins calling *CallbackFunction* with *CBtype* set to PROVIDING_DATA.

<u>CBtype is set to REQUEST_FOR_DATA</u> if Merge DICOM Toolkit is requesting data from the *CallbackFunction*. In this case *CallbackFunction* sets **CBdataBufferPtr* to the address of a block of the attribute's value, and **CBdataSizePtr* to the number of bytes at **CBdataBufferPtr*. If this is the first time Merge DICOM Toolkit is calling *CallbackFunction* for this attribute's data *CBisFirst* will be TRUE (non-zero). If *CallbackFunction* is returning the last block of data to be provided for this attribute, it must set **CBisLastPtr* to TRUE (non-zero). Merge DICOM Toolkit will repeatedly call *CallbackFunction* until **CBisLastPtr* is set to TRUE or *CallbackFunction* returns with a status other than **MC_NORMAL_COMPLETION**.

<u>CBtype is set to PROVIDING_DATA</u> if Merge DICOM Toolkit is providing data to the CallbackFunction. In this case Merge DICOM Toolkit has set **CBdataBufferPtr* to the address of a block of the attribute's value, and **CBdataSizePtr* to the number of bytes at **CBdataBufferPtr*. If this is the first time Merge DICOM Toolkit is calling CallbackFunction for this attribute's data CBisFirst will be TRUE (non-zero). If this is the last block of data to be provided for this attribute, Merge DICOM Toolkit has set **CBisLastPtr* to TRUE (non-zero). Merge DICOM Toolkit will repeatedly call CallbackFunction until all of the value has been provided. At that time, it sets **CBisLastPtr* to TRUE.

<u>CBtype is set to PROVIDING_OFFSET_TABLE</u> if Merge DICOM Toolkit is providing basic offset table to the *CallbackFunction*. In this case Merge DICOM Toolkit has set **CBdataBufferPtr* to the address of the offset table, and **CBdataSizePtr* to the number of bytes at **CBdataBufferPtr*. If this is the first time Merge DICOM Toolkit is calling *CallbackFunction* for this offset table's data *CBisFirst* will be TRUE (non-zero).

- **NOTE:** **CBdataBufferPtr* may be NULL when **CBisLastPtr* is returned TRUE.
- **NOTE:** The *CBtype* PROVIDING_OFFSET_TABLE is given to the callback function when the pixel data is encapsulated.

NOTE: CallbackFunction is not called with this value when a file is opened with **MC_Open_File_Bypass_OBOW** because the OB, OW, or OF data is left on media.

<u>CBtype is set to PROVIDING_MEDIA_DATA_LENGTH</u> if Merge DICOM Toolkit is reading a file from media by use of the **MC_Open_File_Bypass_OBOW** function. *CallbackFunction* is provided with the total size of the attribute's value and the byte offset from the beginning of the media file of the

[©] Copyright Merge Healthcare Solutions Inc. 2023

attribute's value. The value size (in bytes) is placed at **CBdataSizePtr. **CBdataBufferPtr* is a long int that contains the byte offset of the attribute's value from the beginning of the media file. Merge DICOM Toolkit will not call the *CallbackFunction* with *CBtype* set to PROVIDING_DATA when it calls *CallbackFunction* with *CBtype* set to PROVIDING_MEDIA_DATA_LENGTH.

<u>CBtype is set to FREE_DATA</u> if the memory associated with the enclosing message, item, or file is being freed by Merge DICOM Toolkit, and the USE_FREE_DATA_CALLBACK configuration option has been set to Yes. When this configuration option is set to No, registered callback functions will not be called with the FREE_DATA type.

If *CallbackFunction* returns with a status other than MC_NORMAL_COMPLETION, Merge DICOM Toolkit will stop calling it.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>CallbackFunction</i> , <i>pCallbackFunction</i> or <i>PrivateCode</i> were NULL.
MC_INVALID_APPLICATION_ID	ApplicationID does not identify a valid application object.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Set_Value_From_Function MC_Set_pValue_From_Function MC_Open_File MC_Open_File_Bypass_OBOW MC_Open_File_Upto_Tag_Bypass_V MC_Release_Callback_Function MC_Set_Message_Callbacks	MC_Get_Value_To_Function MC_Get_pValue_To_Function MC_Get_Message_Service MC_Open_File_Upto_Tag /alue MC_Register_Application

MC_Register_Compression_Callbacks

Registers compression/decompression callback functions to compress or decompress a standard or private attribute's values. The callbacks are automatically used when **MC_Set_Encapsulated_Value**, **MC_Set_Next_Encapsulated_Value**, **MC_Get_Encapsulated_Value**,

MC_Get_Next_Encapsulated_Value, MC_Get_Frame_To_Function and MC_Duplicate_Message are called.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Register_Compression_Callbacks (
    int MsgFileItemID,
    MC_STATUS (*CompressionCallbackFunction)()
```

 $[\]ensuremath{\textcircled{O}}$ Copyright Merge Healthcare Solutions Inc. 2023

MC_STATUS (*DecompressionCallbackFunction)()
MsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
CompressionCallb ackFunction	Name of a function which will be called repeatedly with blocks of uncompressed data, and will return compressed data.
DecompressionCal IbackFunction	Name of a function which will be called repeatedly with compressed blocks of data, and will return decompressed data

The functions must be prototyped as follows:

```
MC STATUS YourDestMsg(De)CompressionCallback (
       int CBmsqID,
       void **CBcontext,
       unsigned long CBdataLength,
       void *CBdataValue,
       unsigned long *CBoutdataLength,
       void **CBoutdataValue,
       int CBisFirst,
       int CBisLast
       int CBrelease
)
 CBmsgID
                  The identifier assigned to this message object by the
                   MC_Open_Message, MC_Open_Empty_Message
                  functions.
 CBcontext
                  A data structure that contains user data and has to be
                   preserved between compression calls. The first call to the
                   callback function should initialize this structure.
                  The length of the incoming data pointed to by
 CBdataLength
                   CBdataValue.
 CBdataValue
                  Pointer to incoming data.
```

```
CBoutdataLengtThe length of the outgoing data pointed to by<br/>CBoutdataValue.CBoutdataValuePointer to the outgoing data.CBisFirstThis is TRUE (not zero) the first time<br/>YourDestMsg(De)CompressionCallback is being called.CBisLastThis is TRUE (not zero) the last time<br/>YourDestMsg(De)CompressionCallback is being called.CBreleaseThis is TRUE (not zero) if the callback should release all<br/>the context memory and return.
```

Remarks

MC_Register_Compression_Callbacks may only be used for attributes with a value representation of OB or OW. The built-in callbacks can be used, or the user may implement their own (Please see Users Manual, "Using Compression/Decompression Callback Functions")

The toolkit has a built in compressor/decompressor, represented by functions **MC_Standard_Compressor** and **MC_Standard_Decompressor**, which are capable of handling the following types of pixel data:

Grayscale: JPEG_BASELINE (8 bit) JPEG_EXTENDED_2_4 (8,10,12 bit) JPEG_LOSSLESS_HIER_14 (any from 2 to 16 bit) JPEG_2000 (8,10,12,16 bit) JPEG_2000_LOSSLESS_ONLY (8,10,12,16 bit)

Other possible inputs/outputs: 8 bits/pixel, 3 samples/pixel RGB 8 bits/pixel, 3 samples/pixel YBR_FULL_422

If a callback function was already registered, the passed in callback replaces the function previously registered. A callback function may be "de-registered" by calling **MC_Register_Compression_Callbacks** with NULL as the parameter.

The **MC_Register_Callback_Function** caller may provide the *CallbackFunction* with a pointer to agreed upon data in its *UserInfo* parameter. This is optional and *UserInfo* may be NULL.

NOTE: **CBdataBufferPtr* may be NULL when **CBisLastPtr* is returned TRUE.

NOTE: FOR MC_Standard_(De)compressor:

JPEG_2000/JPEG_2000_LOSSLESS_ONLY will cause an irreversible, respective reversible color transformation when compressing RGB data. The Photometric Interpretation MUST be changed from RGB to:

1. YBR_ICT if JPEG_2000 is used with COMPRESSION_WHEN_J2K_USE_LOSSY = Yes (lossy color transform for lossy compression)

2. YBR_RCT if JPEG_2000_LOSSLESS_ONLY, or JPEG_2000 with COMPRESSION_WHEN_J2K_USE_LOSSY = No (lossless color transform for lossless compression).

Similarly, on the decompression end, the Photometric Interpretation should be changed back to RGB, but the Lossy Image Compression attribute should indicate if it had been compressed lossy.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TRANSFER_SYNTAX	MsgFileItemID's transfer syntax is not one of the encapsulated transfer syntaxes.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Alee	

See Also

MC_Set_Encapsulated_Value_From_Function MC_Set_Next_Encapsulated_Value_From_Function MC_Get_Encapsulated_Value_To_Function MC_Get_Next_Encapsulated_Value_To_Function MC_Get_Frame_To_Function MC_Get_Offset_Table_To_Function MC_Duplicate_Message

MC_Register_Enhanced_MemoryLog_Function

Registers a callback function that will be notified as each log message is written

Synopsis

{

#include ``mc3msg.h"

MemLogFunctionThe name of a function that will be called each timeMerge DICOM Toolkit emits a log message.

MemLogFunction must be prototyped as follows:

```
void (*MemLogFunction) (LogInfo *logInfo),
```

logInfo A structure containing information about the message that is being logged.

The **LogInfo** structure is defined as follows:

```
typedef struct LogInfo_struct
```

MsgLogType	typeCode;
LogTime	timeValues;
char	*type;
char	*prefix;

[©] Copyright Merge Healthcare Solutions Inc. 2023

char char	*processID; *timeStamp;
char	<pre>*timeStamp; *function;</pre>
char	*message;
<pre>} LogInfo;</pre>	
typeCode	The type of message being logged. It will be one of the following constants defined by the MsgLogType enumeration:
	MC_ERROR_MESSAGE
	MC_WARNING_MESSAGE
	MC_INFO_MESSAGE
	MC_T1_MESSAGE
	MC_T2_MESSAGE
	MC_T3_MESSAGE
	MC_T4_MESSAGE
	MC_T5_MESSAGE
	MC_T6_MESSAGE
	MC_T7_MESSAGE
	MC_T8_MESSAGE
	MC_T9_MESSAGE
timeValues	A LogTime structure containing the exact time the message was being logged. (See structure definition below.)
type	The string representation of the message type. For example, "E", "W", "I", "T1" etc.
prefix	An option prefix string. Normally this will be an empty string. The MC_Set_Log_Prefix function is used to provide a string that will be prefixed to all log messages.
processID	A string that represents the ID of the process that generated the message. In some operating environments this may be an empty string.
timeStamp	A string representation of the time the message was logged. This string may or may not be included in the actual log message.
function	A string representation of the Merge DICOM Toolkit function that generated the message. This may be an empty string.
message	The log message. This is the string normally logged to screen or to file.

The **LogTime** structure that is included in the returned *logInfo* is defined as follows:

© Copyright Merge Healthcare Solutions Inc. 2023

typedef struct	LogTime_struct {
<pre>int hour;</pre>	
<pre>int min;</pre>	
<pre>int sec;</pre>	
<pre>int hun;</pre>	
<pre>int day;</pre>	
<pre>int mon;</pre>	
<pre>int year;</pre>	
<pre>} LogTime;</pre>	
hour	hours since midnight - [0,23]
min	minutes after the hour - [0,59]
sec	seconds after the minute - [0,59]
hun	hundredths of a sec after the sec $[0,99]$
day	day of the month - [1,31]
mon	months since January - [0,11]
year	years since 1900

Remarks

This function registers an enhanced message log handler to process system log messages. (See **MC_Register_MemoryLog_Function** for information about registering a standard log handler.)

Once registered, all system log messages are passed to the *MemLogFunction*. Information about the logged message is passed by the *logInfo* parameter. Refer to the above description of the **LogInfo** structure for detailed information about the data that is made available.

Only one enhanced message log handler may exist at any one time. If this function is called again, any existing enhanced log handler will no longer be called. The *MemLogFunction* parameter may be null to de-register the last enhanced handler.

Log messages can be directed to one or more of four destinations by settings in the merge.ini file, or by setting the destinations at runtime using the **MC_Set_Log_Destination** function. Different types of log messages may be directed to the screen, to a file, to memory, or to the "bit bucket". Messages directed to the "bit bucket" are those that have no destination set for them.

Note that all messages with a screen or file destination are also passed on to the *MemLogFunction*. If you wish to capture log messages without sending messages to a file or screen destination, set the memory destination in the merge.ini file.

See Also

MC_Set_Log_Destination MC_Set MC_Register_MemoryLog_Function

MC_Set_Log_Prefix

MC_Register_MemoryLog_Function

Registers a callback function that will be notified as each log message is written

Synopsis

#include ``mc3msg.h"

[©] Copyright Merge Healthcare Solutions Inc. 2023

MemLogFunction The name of a function that will be called each time Merge DICOM Toolkit emits a log message.

MemLogFunction must be prototyped as follows:

```
      void (*MemLogFunction) (

      MsgLogType
      MsgType,

      char
      *Msg

      )
      The type of message being logged. It will be one of the following constants defined by the MsgLogType enumeration:

      MC_ERROR_MESSAGE
      MC_WARNING_MESSAGE

      MC_INFO_MESSAGE
      MC_INFO_MESSAGE

      MC_TRACE_MESSAGE
      MC_TRACE_MESSAGE

      Msg
      The message string that is being logged
```

Remarks

This function registers a standard message log handler to process system log messages. (See **MC_Register_Enhanced_MemoryLog_Function** for information about registering an enhanced log handler.)

Once registered, all system log messages are passed to the *MemLogFunction*. The type of message being logged is provided by the *MsgType* parameter and the message string itself is passed by the *Msg* parameter.

Only one standard message log handler may exist at any one time. If this function is called again, any existing standard log handler will no longer be called. The *MemLogFunction* parameter may be null to de-register the last standard handler.

Log messages can be directed to one or more of four destinations by settings in the merge.ini file, or by setting the destinations at runtime using the **MC_Set_Log_Destination** function. Different types of log messages may be directed to the screen, to a file, to memory, or to the "bit bucket". Messages directed to the "bit bucket" are those that have no destination set for them.

Note that all messages with a screen or file destination are also passed on to the *MemLogFunction*. If you wish to capture log messages without sending messages to a file or screen destination, set the memory destination in the merge.ini file.

See Also

MC_Set_Log_Destination MC_Register_Enhanced_MemoryLog_Function

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Register_Network_Capture_Callbacks (deprecated)

Registers a set of callback functions that will be used to capture network data, overriding the default network capture handler.

NOTE: Use of this call is deprecated.

Synopsis

```
#include ``mergecom.h"
```

AcallbackInfo Pointer to an MC_NetworkCaptureInfo structure that provides an optional application context, plus various callback functions. If NULL, the default Merge DICOM Toolkit handler functions will be used.

AcallbackInfo is a pointer to an MC_NetworkCaptureInfo type prototyped as follows:

```
typedef struct MC NetworkCaptureInfo Struct {
    void *CAP applicationContext;
    int (NOEXP FUNC *CAP initialization) (
        void *AppContext,
        long MaxFileSize,
        char *FileName,
        int NumFiles,
        int RewriteFiles
);
    void (NOEXP FUNC *CAP shutdown) (
        void *AppContext
    );
    void *(NOEXP_FUNC *CAP_getConnectionContext) (
        void *AppContext,
        struct sockaddr *SourceIPaddr,
        struct sockaddr *DestIPaddr
    );
    void (NOEXP FUNC *CAP freeConnectionContext) (
        void *AppContext,
        void *ConnContext
    );
    int (NOEXP FUNC *CAP conectionOpened) (
        void *AppContext,
        void *ConnContext,
        struct sockaddr *SourceIPaddr,
        struct sockaddr *DestIPaddr
    );
    int (NOEXP_FUNC *CAP_conectionClosed) (
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
void *AppContext,
void *ConnContext,
struct sockaddr *SourceIPaddr,
struct sockaddr *DestIPaddr
);
int (NOEXP_FUNC *CAP_dataSent) (
void *AppContext,
void *ConnContext,
struct sockaddr *SourceIPaddr,
struct sockaddr *DestIPaddr,
char *Data,
long DataLength
);
} MC NetworkCaptureInfo;
```

The functions provided in the MC_NetworkCaptureInfo structure <u>will override the default network</u> <u>capture handler</u> provided by Merge DICOM Toolkit. (Please refer to the *Merge DICOM Toolkit Users Manual* for more information about the default handler.)

CAP_applicationContext - An optional pointer to application context data. Merge DICOM Toolkit passes this pointer when it calls the callback functions. It may be NULL.

CAP_initialization - Merge DICOM Toolkit calls this function at startup time (if the NETWORK_CAPTURE configuration parameter is Yes), whenever a network capture configuration value is changed at runtime, and whenever new capture functions are registered. Five configuration parameters may be set in the [TRANSPORT_PARMS] section of the mergecom.pro file:

```
NETWORK_CAPTURE = Yes/No - Determines whether or not Merge DICOM
Toolkit should call the capture handler functions.
```

- CAPTURE_FILE_SIZE = n Sets the maximum size in kilobytes of capture
 files, with zero meaning unlimited size.
- CAPTURE_FILE = <filename> Provides the base name to use for capture
 files.

Each of the above values may also be set at runtime using the appropriate MC_Set_XXX_Config_Value function.

CAP_initialization is called with these parameters:

AppContext	The application context provided by
	CAP_applicationContext.
<i>MaxFileSize</i>	the current value of the CAPTURE_FILE_SIZE
	configuration parameter
FileName	the current value of the CAPTURE_FILE configuration
	parameter
NumFiles	the current value of the NUMBER_OF_CAP_FILES
	configuration parameter

RewriteFiles the current value of the **REWRITE_CAPTURE_FILES** configuration parameter

CAP_shutdown - Merge DICOM Toolkit calls this function when the library is released, or when NETWORK_CAPTURE is turned off after it was on. The function is called with one parameter:

AppContext The application context provided by CAP applicationContext.

CAP_getConnectionContext - Merge DICOM Toolkit calls this function when it has opened a new association connection. The function is expected to return a pointer to a context area that will be used to process the connection's capture information. Merge DICOM Toolkit will pass this pointer (as the *ConnContext* parameter) to other callbacks (see below). Merge DICOM Toolkit provides the TCP/IP addresses used on the connection. The IP addresses are presented in network order (i.e. big endian).

AppContext	The application context provided by
	CAP_applicationContext.
SourceIPaddr	The first Internet Protocol address used on the connection.
DestIPaddr	The second Internet Protocol address used on the connection.

CAP_freeConnectionContext - Merge DICOM Toolkit calls this function when it closes a network connection. Merge DICOM Toolkit will never use the *ConnContext* pointer again.

AppContext	The application context provided by
	CAP_applicationContext.
ConnContext	The connection context provided by the
	CAP_getConnectionContext callback.

CAP_conectionOpened – Merge DICOM Toolkit calls this function when it has opened a new association connection after calling CAP_getConnectionContext. The source and destination addresses are provided. The function is called with these parameters:

AppContext	The application context provided by
	CAP_applicationContext.
ConnContext	The connection context provided by the
	CAP getConnectionContext callback.
SourceIPaddr	The Internet Protocol address of the node that opened
	the connection.
DestIPaddr	The Internet Protocol address of the node that accepted
	the connection.

CAP_conectionClosed - Merge DICOM Toolkit calls this function

AppContext	The application context provided by
	CAP_applicationContext.
ConnContext	The connection context provided by the
	CAP_getConnectionContext callback.
SourceIPaddr	The Internet Protocol address of the node that opened
	the connection.
DestIPaddr	The Internet Protocol address of the node that accepted
	the connection.

CAP_dataSent - Merge DICOM Toolkit calls this function

AppContext	The application context provided by
	CAP applicationContext.
ConnContext	The connection context provided by the
	CAP getConnectionContext callback.
SourceIPaddr	The Internet Protocol address of the node that sent the
	data.
DestIPaddr	The Internet Protocol address of the node that received
	the data.

Merge DICOM Toolkit provides a facility to capture raw data that is transmitted over a network connection. The data is captured in files that can be read by the MergeDPM utility to analyze the network activity. The network capture facility is controlled by the configuration parameters discussed in the above section.

In rare cases you may wish to replace the default Merge DICOM Toolkit network capture facility with one of your own. The **MC_Register_Network_Capture_Callbacks** function is used to register the functions that you will provide to capture network data, overriding the default handler functions.

Refer to the above description of the individual callback functions for details of the network capture interface.

Note that received network data is provided by Merge DICOM Toolkit after it has been optionally decrypted by any secure socket functions (see **MC_Open_Secure_Association** and **MC_Wait_For_-Secure_Association**). Likewise, the data is provided to the callbacks before it sends the data to secure socket functions for possible encryption.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the functions provided in the MC_NetworkCaptureInfo structure was NULL.
See Also	
MC_Open_Secure_Association	MC_Wait_For_Secure_Association

MC_Reject_Association MC_Reject_Association_With_Reason_Codes

Reject a remote application's request for a DICOM association

MC STATUS MC R	Reject Association With Reason Codes (
	pciationID,
unsigned	l short ResultCode,
unsigned	l short SourceCode,
unsigned	l short ReasonCode
)	
AssociationID	The association object's identification number
Reason	The reason the DICOM association is being rejected. Use one of
	the enumerated REJECT_REASON types defined in
	"mergecom.h":
	PERMANENT_NO_REASON_GIVEN,
	PERMANENT_NO_REASON_GIVEN_SERV_USER (same
	as PERMANENT_NO_REASON_GIVEN),
	PERMANENT_NO_REASON_GIVEN_SERV_PROV_ACSE,
	TRANSIENT_NO_REASON_GIVEN,
	TRANSIENT_NO_REASON_GIVEN_SERV_USER (same
	as TRANSIENT_NO_REASON_GIVEN),
	TRANSIENT_NO_REASON_GIVEN_SERV_PROV_ACSE,
	PERMANENT_APPLICATION_CONTEXT_NAME_NOT_
	SUPPORTED,
	PERMANENT_CALLING_AE_TITLE_NOT_RECOGNIZED,
	PERMANENT_CALLED_AE_TITLE_NOT_RECOGNIZED,
	PERMANENT_ACSE_PROTOCOL_VERSION_NOT_
	SUPPORTED, TRANSIENT_TEMPORARY_CONGESTION,
	TRANSIENT_LOCAL_LIMIT_EXCEEDED
ReasonCode	The reason field (byte 10 in A-ASSOCIATE-RJ PDU as per DICOM
	PS3.8, Section 9.3.4)
ResultCode	The result field (byte 8 in A-ASSOCIATE-RJ PDU as per DICOM
nesuncoue	
	PS3.8, Section 9.3.4)
SourceCode	The source field (byte 9 in A-ASSOCIATE-RJ PDU as per DICOM
	PS3.8, Section 9.3.4)

If a **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** function completes normally, one of two functions must be called. **MC_Accept_Association** informs the remote application that the association can proceed. Use **MC_Reject_Association** to reject the association request.

Upon successful return from the **MC_Reject_Association** call, no further calls may be made for the association.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NO_REQUEST_PENDING	There is no pending association request for this AssociationID.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_INVALID_ASSOC_ID

AssociationID is not a valid association object ID.

See Also

MC_Wait_For_Association MC_Accept_Association

MC_Wait_For_Secure_Association

MC_Release_Application

De-register an application.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Release_Application (
    int *ApplicationID
```

)

ApplicationID Address of the application object's identification number

Remarks

MC_Release_Application releases the system resources used by the application. If the application has one or more associations open, <u>they are aborted</u>. Applications should always close or abort associations before calling **MC_Release_Application**.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	The ApplicationID parameter was NULL.
MC_INVALID_APPLICATION_ID	*ApplicationID is not a valid application ID.
See Also	

MC_Register_Application

MC_Release_Callback_Function MC_Release_pCallback_Function

De-register a callback function for a given attribute.

Synopsis #include "mergecom.h" MC_STATUS MC_Release_Callback_Function (int ApplicationID, unsigned long Tag) MC_STATUS MC_Release_pCallback_Function (int ApplicationID, char *PrivateCode,

[©] Copyright Merge Healthcare Solutions Inc. 2023

<pre>unsigned short Group, unsigned char ElementByte,</pre>	
)	
ApplicationID	The application object's identification number
Tag	The identifying tag of an attribute which had a callback function registered for it.
PrivateCode	The code string which dictates which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group containing the private attribute. It must be an odd number.
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .

MC_Release_Callback_Function releases the callback function which was registered for the attribute identified by *Tag.* The callback function will no longer be called when the attribute's value is being received or when the attribute's value is required.

MC_pRelease_Callback_Function releases the callback function which was registered for the private attribute. The callback function will no longer be called when the private attribute's value is being received or when the private attribute's value is required.

The callback function will, however, still be used for messages that were opened before this function call was made.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_APPLICATION_ID	<i>ApplicationID</i> does not identify a valid Merge DICOM Toolkit application.
MC_INVALID_TAG	There was no callback function registered for <i>Tag</i> .
See Also	
MC_Register_Callback_Function	MC_Register_pCallback_Function

MC_Release_Library_Exception_Handler

Disables the internal system exception handler and releases the user-defined exception handler.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value

Meaning

MC_NORMAL_COMPLETION

The function completed normally.

See Also

MC_Set_Library_Exception_Handler

MC_Release_Parent_Association (UNIX Only)

Releases parent's copy of an association object.

Synopsis

#include "mergecom.h"
MC_STATUS MC_Release_Parent_Association (
 int AssociationID
)
AssociationID The ID of an association object returned by a
 MC_Wait_For_Association or
 MC_Wait_For_Secure_Association function.

Remarks

MC_Release_Parent_Association is used by a parent process to free its copy of an association object after spawning a child process (and passing the association object to the child). This is often done in UNIX environments in which a server process loops, calling the **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** function. As each association request arrives, the process spawns a child process to handle the association, then uses this function to free its copy of the association object. IT IS VERY IMPORTANT that **MC_Release_Parent_Association** be called in this situation to avoid memory leaks.

Note that this routine is only needed in the parent process when fork() is called to create a child process to handle the association. It is not required when a thread is created to handle the connection.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	<i>AssociationID</i> does not identify a valid Merge DICOM Toolkit association.

See Also

MC_Wait_For_Association

MC_Wait_For_Secure_Association

MC_Release_Parent_Connection (UNIX Only)

Releases parent's copy of a socket for a connection.

Synopsis #include "mergecom.h" void MC_Release_Parent_Connection (MC_SOCKET Socket) Socket A socket returned by a MC_Wait_For_Connection or MC_Wait_For_Connection_On_Port function.

Remarks

MC_Release_Parent_Connection is used by a parent process to free its copy of a socket after spawning a child process (and passing the socket to the child). This is often done in UNIX environments in which a server process loops, calling the MC_Wait_For_Connection or MC_Wait_For_Connection_On_Port function. As each connection arrives, the process spawns a child process to handle the connection, then uses this function to free its copy of the connection. IT IS VERY IMPORTANT that MC_Release_Parent_Connection be called in this situation to avoid memory leaks.

Note that this routine is only needed in the parent process when fork() is called to create a child process to handle the connection. It is not required when a thread is created to handle the connection.

Return Value

There is no return value.

See Also

MC_Wait_For_Connection MC_Wait_For_Connection_On_Port MC_Process_Association_Request MC_Process_Secure_Association_Request

MC_Report_Memory

Reports information on the memory used by the toolkit's internal memory management.

[©] Copyright Merge Healthcare Solutions Inc. 2023

bytesAllocated	The number of bytes that the toolkit has allocated. This parameter is type size_t on 64-bit Windows and unsigned long on other platforms.
bytesInUse	The number of bytes of non-freeable memory. This parameter is type size_t on 64-bit Windows and unsigned long on other platforms.

This function steps through the toolkits internal memory structures to determine how much memory has been allocated, and how much of that is actually in use. This function is helpful in determining if a MC_Cleanup_Memory call should be made.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_NULL_POINTER_PARM	One or both of the parameters were non-initialized pointers.
See Also	

MC_Cleanup_Memory

MC_Reset_Filename

Resets the filename associated with a file object

Synopsis	Synopsis		
#include "mc3m	nedia.h″		
int File	Reset_Filename (PID, ewFilename		
)			
FileID	The identifier assigned to this object by the MC_Create_Empty_File or MC_Create_File function.		
NewFilename	A pointer to a string containing the filename to be associated with the new file object.		
Remarks			

The **MC_Reset_Filename** function changes the filename associated with a file object.

The filename is passed to the user's callback function when the MC_Write_File function is called.

This function is useful when reusing file objects or saving a copy of a file.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
MC_NULL_POINTER_PARM	<i>NewFilename</i> was NULL.
See Also	
MC_Write_File	MC_Empty_File

MC_Reset_Message_Transfer_Syntax

Resets the transfer syntax over which a message will be sent on the network.

Synopsis

```
#include ``mergecom.h"
MC_STATUS MC_Reset_Message_Transfer_Syntax (
    int MessageID
)
MessageID
The identifier assigned to this object by the
    MC_Read_Message, MC_Open_Message or
    MC_Open_Empty_Message function.
```

Remarks

MC_Reset_Message_Transfer_Syntax resets the DICOM transfer syntax associated with a message. When a message is read off the network, **MC_Read_Message** stores the transfer syntax in which the message was received over the network, similarly to how **MC_Set_Message_Transfer_Syntax** can be used to set the transfer syntax for a message. If the same message is sent again over the network, it must be transferred with the transfer syntax that it was received in.

If the transfer syntax of the message is reset with **MC_Reset_Message_Transfer_Syntax**, Merge DICOM Toolkit will pick any uncompressed transfer syntax to send over the network.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MessageID</i> is not a valid message object ID.
See Also	
MC_Send_Request_Message MC_Set_Message_Transfer_Syntax	MC_Get_Message_Transfer_Syntax

© Copyright Merge Healthcare Solutions Inc. 2023

MC_Send_Request_Message

Sends a request message to the remote application.

Synopsis

<pre>#include ``mergecom.h"</pre>		
MC_STATUS MC_S	end_Request_Message (
<pre>int AssociationID,</pre>		
int MessageID		
)		
AssociationID	The association object's identification number	
MessagelD	A message object's identification number	

Remarks

MC_Send_Request_Message sends a request message to the remote application. The message is identified by *MessageID* and must have been created using the **MC_Open_Message** function. The values of the message attributes must have been set with the "Set..." message functions and/or the **MC_Stream_To_Message** function. The message is sent across an open DICOM association identified by *AssociationID*.

NOTE: If the message being sent was opened with MC_Open_Empty_Message and the pixel data is being supplied to the message with a registered callback function, the pixel data attribute should be explicitly added to the message with the function MC_Add_Standard_Attribute before MC_Send_Request_Message is called.

If **MC_Register_Callback_Function** was called to register a callback function for any attributes in the message, **MC_Send_Request_Message** calls that function to retrieve the attribute values as it sends the message.

- **NOTE:** Some DICOM services require that values for certain "command level" message attributes (i.e. group 0 attributes) be set. (Merge DICOM Toolkit automatically adds command level attributes to the message when **MC_Open_Message** is called. With the exceptions listed below, the caller must set any command level attribute <u>values</u> before sending the message. **MC_Send_Request_Message** does, however, set the following attribute values for you **if you have not set them**:
 - The group length attribute (0000,0000) value is always set by Merge DICOM Toolkit.
 - If the message command requires it, the Affected SOP Class UID attribute (0000,0002) value is set to the service's abstract syntax UID.
 - The command attribute (0000,0100) value is always set by Merge DICOM Toolkit.

- The Message ID attribute (0000,0110) value is set to a unique number for this association.
- If the message command requires it, the Priority attribute (0000,0700) value is set "medium" priority (i.e. zero).
- The Data Set Type attribute (0000,0800) value is always set by Merge DICOM Toolkit.

If **MC_Send_Request_Message** returns a status code of **MC_ASSOCIATION_ABORTED**, no further calls may be made for the association.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_MAX_OPERATIONS_EXCEEDED	D Sending of the message would cause the Maximum Number of Operations Invoked that was negotiated to be exceeded. A response message must be read before this message can be sent.
MC_REQUIRED_ATTRIBUTE_MISSI	
	Some DICOM services require certain command level attributes (group 0) to have values set. This status indicates that a required attribute value for the service related to <i>MessageID</i> has not been set.
MC_UNACCEPTABLE_SERVICE	The service related to <i>MessageID</i> has not been successfully negotiated for this association.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of th	e following are returned:
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Register_Callback_Function MC_Read_Message MC_Send_Response_Message MC_Send_Request_Message_For_Se MC_Send_Request_For_Service	MC_Open_Message MC_Read_To_Stream MC_Send_Response ervice

MC_Send_Request_Message_For_Service

Sends a request message to the remote application.

Synopsis

```
#include ``mergecom.h"
    MC_STATUS MC_Send_Request_Message_For_Service (
    int AssociationID,
    int MessageID,
    char *AServiceName
)
Association/D The association object's identification number
    MessageID A message object's identification number
    AserviceName Name of the service negotiated to utilize for the presentation
    context.
```

Remarks

MC_Send_Request_Message_For_Service sends a request message to the remote application. The presentation context used in the transfer is looked up based on the service specified in the call. This call is used to select the proper presentation context when the object is common to multiple services negotiated. An example is the FilmBox when both BASIC_GRAYSCALE_PRINT_MANAGEMENT and BASIC_COLOR_PRINT_MANAGEMENT are negotiated. The message is identified by *MessageID* and must have been created using the **MC_Open_Message** function. The values of the message attributes must have been set with the "Set..." message functions and/or the **MC_Stream_To_Message** function. The message is sent across an open DICOM association identified by *AssociationID*.

NOTE: If the message being sent was opened with
 MC_Open_Empty_Message and the pixel data is being supplied to the message with a registered callback function, the pixel data attribute should be explicitly added to the message with the function
 MC_Add_Standard_Attribute before
 MC_Send_Request_Message_For_Service is called.

If **MC_Register_Callback_Function** was called to register a callback function for any attributes in the message, **MC_Send_Request_Message_For_Service** calls that function to retrieve the attribute values as it sends the message.

NOTE: Some DICOM services require that values for certain "command level" message attributes (i.e. group 0 attributes) be set. (Merge DICOM Toolkit automatically adds command level attributes to the message when **MC_Open_Message** is called. With the exceptions listed below, the caller must set any command level attribute <u>values</u> before sending the message. **MC_Send_Request_Message_For_Service** does, however, set the following attribute values for you - **if you have not set them**:

- The group length attribute (0000,0000) value is always set by Merge DICOM Toolkit.
- If the message command requires it, the Affected SOP Class UID attribute (0000,0002) value is set to the service's abstract syntax UID.
- The command attribute (0000,0100) value is always set by Merge DICOM Toolkit.
- The Message ID attribute (0000,0110) value is set to a unique number for this association.
- If the message command requires it, the Priority attribute (0000,0700) value is set "medium" priority (i.e. zero).
- The Data Set Type attribute (0000,0800) value is always set by Merge DICOM Toolkit.

If **MC_Send_Request_Message_For_Service** returns a status code of **MC_ASSOCIATION_ABORTED**, no further calls may be made for the association.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_MAX_OPERATIONS_EXCEEDED	D Sending of the message would cause the Maximum Number of Operations Invoked that was negotiated to be exceeded. A response message must be read before this message can be sent.
MC_REQUIRED_ATTRIBUTE_MISSI	NG
	Some DICOM services require certain command level attributes (group 0) to have values set. This status indicates that a required attribute value for the service related to <i>MessageID</i> has not been set.
MC_UNACCEPTABLE_SERVICE	The service related to <i>MessageID</i> has not been successfully negotiated for this association.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of the following are returned:	
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Register_Callback_Function MC_Send_Request_Message MC_Send_Response_Message MC_Read_Message MC_Get_Meta_ServiceName MC_Open_Message MC_Send_Request MC_Send_Response MC_Read_To_Stream

MC_Send_Request

Sends a request message to the remote application directly from a callback function. Especially useful for large data transfers which we don't want to load into application memory

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Send Request (
       int AssociationID,
       char *Filename,
       void *UserInfo,
       MC STATUS (*CallbackFunction) ()
)
                    The association object's identification number
 AssociationID
 Filename
                    This parameter is unused (legacy). It may be anything,
                    including NULL.
 UserInfo
                    Address of user data which will be passed on to
                    CallbackFunction each time it is called. This may be
                    NULL.
 CallbackFunction
                    Name of a function which will be called repeatedly to
                    provide blocks of request data.
```

The function must be prototyped as follows:

```
MC_STATUS CallbackFunction (
    int CBAssociationID,
    unsigned long CBtag,
    void* CBuserInfo,
    CALLBACK_TYPE CBtype,
    unsigned long* CBdataSizePtr,
    void** CBdataBufferPtr,
    int CBisFirst,
    int* CBisLastPtr
)
```

```
where:
```

CBtag

CBAssociationID	The association object's identification number
-----------------	--

Not used and has undefined value (0xFFFFFFFF)

and all other *CallbackFunction* parameters as described in *MC_Register_Callback_Function* section.

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Send_Request sends a request to the remote application using user's *CallbackFunction* as the message data provider. The request message might be associated with DICOM file identified by *Filename* parameter.

The using of *CallbackFunction* mechanism facilitates the sending of large messages, decreases the application memory footprint and in many cases increases the overall performance. The data blocks of the request message will be repeatedly requested through *CallbackFunction* and sent to the remote application across an open DICOM association identified by *AssociationID*.

NOTE: Some DICOM services require that values for certain "command level" message attributes (i.e. group 0 attributes) be present. Merge DICOM Toolkit automatically adds command level attributes to the sent message when **MC_Send_Request** is called (see details in **MC_Send_Request_Message** section).

If **MC_Send_Request** returns a status code of **MC_ASSOCIATION_ABORTED**, no further calls may be made for the association.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_MAX_OPERATIONS_EXCEEDED	Sending of the message would cause the Maximum Number of Operations Invoked that was negotiated to be exceeded. A response message must be read before this message can be sent.
MC_REQUIRED_ATTRIBUTE_MISSING	
	Some DICOM services require certain command level attributes (group 0) to have values set. This status indicates that a required attribute value for the service related to <i>MessageID</i> has not been set.
MC_UNACCEPTABLE_SERVICE	The service related to <i>MessageID</i> has not been successfully negotiated for this association.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of the following are returned:	
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Register_Callback_FunctionMC_Open_MessageMC_Read_MessageMC_Read_To_StreamMC_Send_Response_MessageMC_Send_ResponseMC_Send_Request_Message_For_ServiceMC_Send_Request_For_Service

MC_Send_Request_For_Service

Sends a request message to the remote application directly from a callback function. Especially useful for large data transfers which we don't want to load into application memory.

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Send Request For Service (
       int AssociationID,
       char *AServiceName
       char *Filename,
       void *UserInfo,
       MC STATUS (*CallbackFunction)()
)
 AssociationID
                    The association object's identification number
 AserviceName
                    Name of the service negotiated to utilize for the
                    presentation context.
                    This parameter is unused (legacy). It may be anything,
 Filename
                    including NULL.
 UserInfo
                    Address of user data which will be passed on to
                    CallbackFunction each time it is called. This may be
                    NULL.
 CallbackFunction
                    Name of a function which will be called repeatedly to
                    provide blocks of request data.
```

The function must be prototyped as follows:

```
MC_STATUS CallbackFunction (
    int CBAssociationID,
    unsigned long CBtag,
    void* CBuserInfo,
    CALLBACK_TYPE CBtype,
    unsigned long* CBdataSizePtr,
    void** CBdataBufferPtr,
    int CBisFirst,
    int* CBisLastPtr
)
```

)

where:

CBAssociationID The association object's identification number

[©] Copyright Merge Healthcare Solutions Inc. 2023

CBtag

Not used and has undefined value (0xFFFFFFF)

and all other *CallbackFunction* parameters **as** described in *MC_Register_Callback_Function* section.

Remarks

MC_Send_Request_For_Service sends a request message to the remote application using user's *CallbackFunction* as the message data provider. The request message might be associated with DICOM file identified by *Filename* parameter.

The presentation context used in the transfer is looked up based on the service specified in the call. This call is used to select the proper presentation context when the object is common to multiple services negotiated. An example is the FilmBox when both

BASIC_GRAYSCALE_PRINT_MANAGEMENT and BASIC_COLOR_PRINT_MANAGEMENT are negotiated.

The data blocks of the message will be repeatedly requested through *CallbackFunction* and sent to the remote application across an open DICOM association identified by *AssociationID*

NOTE: Some DICOM services require that values for certain "command level" message attributes (i.e. group 0 attributes) be present. Merge DICOM Toolkit automatically adds command level attributes to the sent message when **MC_Send_Request_For_Service** is called (see details in **MC_Send_Request_Message** section).

If **MC_Send_Request_For_Service** returns a status code of **MC_ASSOCIATION_ABORTED**, no further calls may be made for the association.

Return Value

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.	
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.	
MC_MAX_OPERATIONS_EXCEEDED	Sending of the message would cause the Maximum Number of Operations Invoked that was negotiated to be exceeded. A response message must be read before this message can be sent.	
MC_REQUIRED_ATTRIBUTE_MISSING		
	Some DICOM services require certain command level attributes (group 0) to have values set. This status indicates that a required attribute value for the service related to <i>MessageID</i> has not been set.	

MC_UNACCEPTABLE_SERVICE	The service related to <i>MessagelD</i> has not been successfully negotiated for this association.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
The association is dropped if any of the	e following are returned:
MC_ASSOCIATION_ABORTED	The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Register_Callback_Function MC_Send_Request_Message MC_Send_Response_Message MC_Read_Message MC_Get_Meta_ServiceName	MC_Open_Message MC_Send_Request MC_Send_Response MC_Read_To_Stream

MC_Send_Response_Message

Sends a response to a message received from the remote application.

Synopsis

```
#include "mergecom.h"
MC_STATUS MC_Send_Response_Message (
    int AssociationID,
    RESP_STATUS ResponseStatus,
    int ResponseMessageID
)
AssociationID The association object's identification number
ResponseStatus This is the status code to be returned. It must be one
    of the enumerated RESP_STATUS codes defined in
    "mergecom.h".
ResponseMessageID This is the ID of a message object containing the
```

response message.

Remarks

MC_Send_Response_Message allows the caller to respond to a message received from the remote application. **MC_Send_Response_Message** is called after receiving a request message using **MC_Read_Message or MC_Read_To_Stream**.

ResponseStatus must be set to a valid response code for the service involved. Response codes for standard DICOM services are defined in the "mergecom.h" file.

ResponseMessageID must be set to an identifier returned by **MC_Open_Message**. Many DICOM services do not require a response of more than just a status. Others, however, (e.g. C_FIND_RSP) require the setting of several message attributes.

[©] Copyright Merge Healthcare Solutions Inc. 2023

NOTE: Some DICOM services require that values for certain "command level" message attributes (i.e. group 0 attributes) be set. (Merge DICOM Toolkit automatically adds command level attributes to the message when **MC_Open_Message** is called. With the exceptions listed below, the caller must set any command level attribute <u>values</u> before sending the message. **MC_Send_Response_Message** does, however, set the following attribute values for you - **if you have not set them**:

- The group length attribute (0000,0000) value is always set by Merge DICOM Toolkit.
- If the message command requires it, the Affected SOP Class UID attribute (0000,0002) value is set to the service's abstract syntax UID.
- The command attribute (0000,0100) value is always set by Merge DICOM Toolkit.
- The Response Message ID attribute (0000,0120) value is set to the Message ID value of the last received Request message.
- The Data Set Type attribute (0000,0800) value is always set by Merge DICOM Toolkit.

If **MC_Send_Response_Message** returns **MC_ASSOCIATION_ABORTED**, no further calls may be made for that association.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.
MC_REQUIRED_ATTRIBUTE_MISSING	
	Some DICOM services require certain command level attributes (group 0) to have values set. This status indicates that a required attribute value for the service related to <i>MessageID</i> has not been set.
MC_UNACCEPTABLE_SERVICE	The service related to <i>MessageID</i> has not been successfully negotiated for this association.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

The association is dropped if any of the following are returned:

The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

```
MC_Send_Request_Message
```

```
MC_Read_Message
```

MC_Send_Response

Sends a response to a message received from the remote application. The response is sent directly from a callback function. Especially useful for large data transfers which we don't want to load into application memory.

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Send Response (
       int AssociationID,
       RESP STATUS ResponseStatus,
       char *Filename,
       void *UserInfo,
       MC STATUS (*CallbackFunction) ()
)
 AssociationID
                    The association object's identification number
                    This is the status code to be returned. It must be one of
 ResponseStatus
                    the enumerated RESP_STATUS codes defined in
                    "mergecom.h".
 Filename
                    This parameter is unused (legacy). It may be anything,
                    including NULL.
                    Address of user data which will be passed on to
 UserInfo
                    CallbackFunction each time it is called. This may be
                    NULL.
 CallbackFunction
                    Name of a function which will be called repeatedly to
                    provide blocks of the response message data.
```

The function must be prototyped as follows:

```
MC_STATUS CallbackFunction (
    int CBAssociationID,
    unsigned long CBtag,
    void *CBuserInfo,
    CALLBACK_TYPE CBtype,
    unsigned long *CBdataSizePtr,
    void **CBdataBufferPtr,
    int CBisFirst,
    int *CBisLastPtr
)
where:
```

CBAssociationID	The association object's identification number
CBtag	Not used and has undefined value (0xFFFFFFF)

and all other *CallbackFunction* parameters as described in *MC_Register_Callback_Function* section.

Remarks

MC_Send_Response allows the caller to respond to a message received from the remote application using the callback mechanism. The response data should be provided in the form of a DICOM object. The data blocks of the response message will be repeatedly requested through *CallbackFunction* and sent to the remote application across an open DICOM association identified by *AssociationID*

MC_Send_Response is called after receiving a request message using MC_Read_Message or MC_Read_To_Stream.

ResponseStatus must be set to a valid response code for the service involved. Response codes for standard DICOM services are defined in the "mergecom.h" file.

Many DICOM services do not require a response of more than just a status. Others, however, (e.g. C_FIND_RSP) require the setting of several message attributes.

NOTE: See **MC_Send_Response_Message** section for the details regarding the handling of the required message attributes for DICOM services.

If **MC_Send_Response** returns **MC_ASSOCIATION_ABORTED**, no further calls may be made for that association.

Return Value

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_INVALID_ASSOC_ID	AssociationID is not a valid association object ID.	
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.	
MC_REQUIRED_ATTRIBUTE_MISSING		
	Some DICOM services require certain command level attributes (group 0) to have values set. This status indicates that a required attribute value for the service related to <i>MessageID</i> has not been set.	
MC_UNACCEPTABLE_SERVICE	The service related to <i>MessageID</i> has not been successfully negotiated for this association.	
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.	

The association is dropped if any of the following are returned:

MC_ASSOCIATION_ABORTED	
------------------------	--

The association has been aborted. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Send_Request_Message	MC_S
MC_Read_Message	MC_R

MC_Send_Request MC_Read_To_Stream

MC_Set_Bool_Config_Value

Used to set the value of a boolean toolkit configuration parameter at runtime.

_	C_Set_Bool_Config_Value (arm Aparm,
, Aparm	An enumerated constant identifying the boolean configuration parameter to be (re)set. Aparm can have any of the following values: ACCEPT_ANY_APPLICATION_TITLE ACCEPT_ANY_CONTEXT_NAME ACCEPT_ANY_CONTEXT_NAME ACCEPT_ANY_PRESENTATION_CONTEXT ACCEPT_DIFFERENT_IC_UID ACCEPT_DIFFERENT_VERSION ACCEPT_MULTIPLE_PRES_CONTEXTS ACCEPT_STORAGE_SERVICE_CONTEXTS ALLOW_COMMA_IN_DS_FL_FD_STRINGS ALLOW_EMPTY_PDV_LENGTH ALLOW_INVALID_PRIVATE_ATTRIBUTES ALLOW_INVALID_PRIVATE_CREATOR_CODES ALLOW_OUT_OF_RANGE_BITS_JPEG_LOSSLESS ATT_00081190_USE_UT_VR ATT_00287FE0_USE_UT_VR ATT_0074100A_USE_ST_VR AUTO_ECHO_SUPPORT BLANK_FILL_LOG_FILE CALCULATE_DEFINED_LENGTH_FOR_CB COMPRESSION_ALLOW_FRAGS COMPRESSION_J2K_LOSSY_USE_QUALITY COMPRESSION_WHEN_J2K_USE_LOSSY CREATE_OFFSET_TABLE DEFLATE_ALLOW_FLUSH DICOMDIR_STREAM_STORAGE

DUPLICATE_ENCAPSULATED_ICON ELIMINATE ITEM REFERENCES EMPTY PRIVATE CREATOR CODES EXPLICIT_VR_TO_UN_FOR_LENGTH_GT_64K EXPORT EMPTY PRIVATE CREATOR CODES EXPORT_GROUP_LENGTHS_TO_MEDIA EXPORT GROUP LENGTHS TO NETWORK EXPORT PRIVATE ATTRIBUTES TO MEDIA EXPORT_PRIVATE_ATTRIBUTES_TO_NETWORK EXPORT UN VR TO MEDIA EXPORT_UN_VR_TO_NETWORK EXPORT_UNDEFINED_LENGTH_SQ EXPORT UNDEFINED LENGTH SQ IN DICOMDIR FORCE OPEN EMPTY ITEM HARD_CLOSE_TCP_IP_CONNECTION INSURE EVEN UID LENGTH LIST UN ATTRIBUTES LOG FILE BACKUP MSG FILE ITEM OBJ TRACE NETWORK CAPTURE PRIVATE SYNTAX 1 ENCAPSULATED PRIVATE_SYNTAX_1_EXPLICIT_VR PRIVATE_SYNTAX_1_LITTLE_ENDIAN PRIVATE_SYNTAX_2_ENCAPSULATED PRIVATE SYNTAX 2 EXPLICIT VR PRIVATE_SYNTAX_2_LITTLE ENDIAN REJECT_INVALID_VR RELEASE SQ ITEMS REMOVE PADDING CHARS REMOVE SINGLE TRAILING SPACE RETURN COMMA_IN_DS_FL_FD_STRINGS REWRITE CAPTURE FILES SEND ECHO PRIORITY SEND LENGTH TO END SEND MSG ID RESPONSE SEND RECOGNITION CODE SEND RESPONSE PRIORITY SEND_SOP_CLASS_UID SEND SOP INSTANCE UID TCPIP_DISABLE NAGLE TOLERATE INVALID IN DEFAULT CHARSET UPDATE GROUP 0028 ON DUPLICATE USE FREE DATA CALLBACK These names are the same as those given to the parameters in the toolkit configuration files.

Avalue The boolean value to which Aparm is to be set.

The Merge DICOM Library accesses several configuration files at startup. This call allows your application to (re)set boolean configurable parameters specified in these files at runtime. This call should be made immediately after calling **MC_Library_Initialization** to avoid using these parameters before they are set. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Get_Int_Config_Value MC_Get_Log_Destination MC_Get_Long_Config_Value MC_Get_String_Config_Value	MC_Set_Int_Config_Value MC_Set_Log_Destination MC_Set_Long_Config_Value MC_Set_String_Config_Value

MC_Set_Encapsulated_Value_From_Function

Encapsulates a single frame and stores the encapsulated value into the message object for the given attribute. The value representation of the attribute must be OB or OW. If a compressor is registered, the data will be compressed in the message.

<pre>#include ``mc3msg.h"</pre>		
<pre>MC_STATUS MC_Set_Encapsulated_Value_From_Function (int MsgFileItemID, unsigned long Tag, void *UserInfo, MC_STATUS (*YourSetFunction)())</pre>		
MsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.	
Tag	DICOM tag which identifies the attribute.	
UserInfo	Address of user data which will be passed on to <i>YourSetFunction</i> each time it is called. This may be NULL.	

YourSetFunction Name of a function which will be called repeatedly for blocks of data for the attribute's value.

The function must be prototyped as follows:

<pre>MC_STATUS YourSetFunction (int CBMsgFileItemID, unsigned long CBtag, int CBisFirst, void *CBuserInfo, int *CBdataSizePtr, void **CBdataBufferPtr, int *CBisLastPtr)</pre>		
CBMsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.	
CBtag	DICOM tag which identifies the attribute.	
CBisFirst	This is TRUE (non-zero) the first time Merge DICOM Toolkit calls <i>YourSetFunction</i> to request data blocks.	
CBuserInfo	Address of data which is being passed from the MC_Set_Value_From_Function function. This may be NULL.	
CBdataSizePtr	Set <i>*CBdataSizePtr</i> to the number of bytes you are providing.	
CBdataBufferPtr	Set <i>* CBdataBufferPtr</i> to the address of the data you are providing.	
CBisLastPtr	Set <i>*CBisLastPtr</i> to TRUE (not zero) when you are returning with the last block of OBOW data.	

Remarks

The MC_Set_Encapsulated_Value_From_Function function is used to encapsulate, and then set the value of an attribute which has a value representation of OB or OW. If a compressor is registered with MC_Register_Compression_Callbacks, the data will be compressed using the registered compressor. Such attributes tend to have values of great length. To accommodate this, one uses the MC_Set_Encapsulated_Value_From_Function function to specify the name of a function (*YourSetFunction*) which Merge DICOM Toolkit, in turn, calls repeatedly requesting blocks of the attribute's data value.

MC_Close_Encapsulated_Value must be called when the user is through encapsulating data within this message. If more frames will be encapsulated, use

MC_Set_Next_Encapsulated_Value_From_Function until all frames have been encapsulated, followed by **MC_Close_Encapsulated_Value**.

© Copyright Merge Healthcare Solutions Inc. 2023

An optional *UserInfo* parameter may be used to pass information between the **MC_Set_Encapsulated_Value_From_Function** caller and *YourSetFunction* which receives the data in its *CBuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

YourSetFunction

YourSetFunction must set **CBdataBufferPtr* to the address of a block of data which is a portion of the attribute's value. It also must set **CBdataSizePtr* to the number of bytes in the block.

Merge DICOM Toolkit sets *CBisFirst* to TRUE (non-zero) if it is requesting the first block of the attribute's value. *YourSetFunction* must set **CBisLastPtr* to TRUE (non-zero) if it is supplying the last block of the attribute's value.

YourSetFunction must return **MC_NORMAL_COMPLETION** if all went well. If not, it must return **MC_CANNOT_COMPLY**. Both of these are defined in "mc3msg.h".

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TRANSFER_SYNTAX	The message's transfer syntax is non-encapsulated.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag.</i> The attribute can be added with
	MC_Add_Standard_Attribute.
	Note: This status will not be returned if setting the value of a message object which was opened using
	MC_Open_Empty_Message or MC_Create_Empty_File . In that case, the attribute is automatically added to the object before setting the value.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INCOMPATIBLE_VR	The attribute's value representation was not OB or OW.
MC_CALLBACK_DATA_SIZE_UNEVEN	
	The value set by <i>YourSetFunction</i> in its * <i>CBdataSizePtr</i> parameter was not an even number.
MC_CALLBACK_PARM_ERROR	The value set by <i>YourSetFunction</i> in its * <i>CBdataSizePtr</i> parameter was an odd number
MC_CALLBACK_CANNOT_COMPLY	YourSetFunction returned with MC_CANNOT_COMPLY,
See Also	
MC_Set_Next_Encapsulated_Value_I MC_Close_Encapsulated_Value	From_Function

Return Value

One of these enumerated MC_STATUS codes defined in "mcstatus.h":

MC_Register_Compression_Callbacks

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Set_File_Preamble

Sets the preamble for a file object

Synopsis

<pre>#include ``mc3media.h"</pre>	
MC_STATUS MC_S int Fil char *Pr)	,
FileID	The identifier assigned to this object by the MC_Create_Empty_File or MC_Create_File function.
Preamble	A pointer to the 128 byte preamble to be associated with the file object <i>FileID</i> .

Remarks

MC_Set_File_Preamble changes the preamble associated with a file object. The function copies the 128 byte DICOM file preamble pointed to by *Preamble* into the file object *FileID*. This preamble is written to media when the **MC_Write_File** function is called.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
MC_NULL_POINTER_PARM	Preamble was NULL.
See Also	
MC_Get_File_Preamble	

MC_Set_Int_Config_Value

Used to set the value of a integer toolkit configuration parameter at runtime.

```
#include ``mc3msg.h"
MC_STATUS MC_Set_Int_Config_Value (
    IntParm Aparm,
    int Avalue
)
Aparm An enumerated constant identifying the integer
    configuration parameter to be (re)set. Aparm can have any
    of the following values:
    ARTIM_TIMEOUT
    ASSOC_REPLY_TIMEOUT
```

COMPRESSION_CHROM_FACTOR COMPRESSION J2K LOSSY QUALITY COMPRESSION J2K LOSSY RATIO COMPRESSION_LUM_FACTOR CONNECT_TIMEOUT DEFLATE_COMPRESSION_LEVEL DESIRED LAST PDU SIZE FLATE GROW OUTPUT BUF SIZE IGNORE JPEG BAD SUFFIX INACTIVITY TIMEOUT LARGE DATA SIZE LIST_SQ_DEPTH_LIMIT LIST VALUE LIMIT LOG FILE LINE LENGTH LOG_FILE_SIZE LOG MEMORY SIZE MAX PENDING CONNECTIONS NUM HISTORICAL LOG FILES NUMBER OF CAP FILES **OBOW BUFFER SIZE** PEGASUS NUMBER OF THREADS RELEASE_TIMEOUT TCPIP_KEEP_ALIVE_INTERVAL TCPIP_KEEP_ALIVE_TIME TCPIP LISTEN PORT TCPIP RECEIVE BUFFER SIZE TCPIP SEND BUFFER SIZE WORK BUFFER SIZE WRITE TIMEOUT

These names are the same as those given to the parameters in the toolkit configuration files. A description of the options can be found in Appendix B.

Avalue The integer value to which Aparm is to be set.

Remarks

The Merge DICOM Toolkit Library accesses several configuration files at startup. This call allows your application to (re)set integer configurable parameters specified in these files at runtime. This call should be made immediately after calling **MC_Library_Initialization** to avoid using these parameters before they are set. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_MUST_BE_POSITIVE	Parameter value cannot be negative.

MC_LIBRARY_NOT_INITIALIZED The library has not been properly initialized.

See Also

MC_Get_Int_Config_Value	
MC_Get_Log_Destination	MC_Set_Log_Destination
MC_Get_Long_Config_Value	MC_Set_Long_Config_Value
MC_Get_String_Config_Value	MC_Set_String_Config_Value

MC_Set_Library_Exception_Handler

Initializes the internal system exception handler and registers the user-defined exception handler which will be called if System exception or System signal are raised.

Synopsis

The handler must be prototyped as follows:

```
void UserExceptionHandler (
    int ExceptionId
)
```

Remarks

This call allows you to initialize the internal exception handler and to register a user-defined exception handler callback, which will be called in case System exception or System signal are raised. This allows for the (limited) recovery of the user program by at least giving it a chance to cleanup and exit gracefully and avoid a straight crash.

Note

To enable the system exception handlers, ALLOW_LIBRARY_EXCEPTION_HANDLER configuration parameter should be set to 'Yes'.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Meaning

MC_NORMAL_COMPLETION The function completed normally.

See Also

MC_Release_Library_Exception_Handler

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Set_Log_Destination

Determines where log messages will be written.

Synopsis

<pre>#include ``mc3msg.h"</pre>		
MC_STATUS MC_S LogParm int Aval		
Aparm	An enumerated constant identifying the class of logging that is to be redirected. <i>Aparm</i> can have any of the following values:	
	ERROR_DESTINATIONS, WARNING_DESTINATIONS, INFO_DESTINATIONS, T1_DESTINATIONS, T2_DESTINATIONS, T3_DESTINATIONS, T4_DESTINATIONS, T5_DESTINATIONS, T6_DESTINATIONS, T7_DESTINATIONS, T8_DESTINATIONS, T9_DESTINATIONS	
Avalue	A defined term identifying where the logging is to be directed. <i>Avalue</i> can have any of the following values: File_Destination, Memory_Destination, Screen_Destination, Bitbucket_Destination These values can also be OR'ed together to indicate multiple destinations.	

Remarks

This call allows you to redirect the logging of error, warning, and info messages at runtime. The [DEFAULT_LIBRARY] section of the Merge DICOM Toolkit initialization file contains the setting used at startup. This call should be made immediately after calling **MC_Library_Initialization** to avoid using these parameters before they are set. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

ValueMeaningMC_NORMAL_COMPLETIONThe function completed normally.

O Copyright Merge Healthcare Solutions Inc. 2023

MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Get_Int_Config_Value MC_Get_Log_Destination	MC_Set_Int_Config_Value MC_Get_Long_Config_Value MC_Set_Long_Config_Value MC_Get_String_Config_Value MC_Set_String_Config_Value

MC_Set_Log_Prefix

Sets the string to be placed at the beginning of log statements.

Synopsis

Prefix

A prefix to place in front of all logs

Remarks

This call allows you to place a string in front of all log statements generated by the toolkit.

Return Value

None

See Also

MC_Get_Log_Destination

MC_Set_Log_Destination

MC_Set_Long_Config_Value

Sets the value of a long integer toolkit configuration parameter at runtime.

```
#include "mc3msg.h"
MC_STATUS MC_Set_Long_Config_Value (
    LongParm Aparm,
    long int Avalue
)
Aparm An enumerated constant identifying the long integer
    configuration parameter to be (re)set. Aparm can have only
    the following values:
        CAPTURE_FILE_SIZE,
        PDU_MAXIMUM_LENGTH,
        CALLBACK_MIN_DATA_SIZE,
        PIXEL_BUFFER_SIZE
```

This name is the same as those given to the parameters in
the toolkit configuration files.

Avalue The long integer value to which Aparm is to be set.

Remarks

The Merge DICOM Toolkit Library accesses several configuration files at startup. This call allows your application to (re)set long integer configurable parameters specified in these files at runtime. This call should be made immediately after calling **MC_Library_Initialization** to avoid using this parameter before is set. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_MUST_BE_POSITIVE	Parameter value cannot be negative.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Get_Int_Config_Value	MC_Set_Int_Config_Value
MC_Get_Long_Config_Value	MC_Set_Long_Config_Value
MC_Get_String_Config_Value MC_Get_Log_Destination	MC_Set_String_Config_Value

MC_Set_MergelNI MC_Set_MergelNI_Unicode

Set the path of the merge.ini file so that the MERGE_INI environment variable is not used.

Synopsis

Filename An absolute or relative path containing the location of the merge.ini file. For **MC_Set_MergeINI_Unicode**, it must be specified as a null terminated Unicode string: UTF-16 (wchar_t) for Windows platforms and UTF-8 for non Windows.

The Merge DICOM Toolkit Library accesses several configuration files at startup. The MERGE_INI environment variable is used to point to the main configuration file (merge.ini). **MC_Set_MergeINI** and **MC_Set_Merge_INI_Unicode** can be used to override the use of this environment variable. *Filename* contains an absolute or relative filename where the merge.ini file is located. This call should be made immediately before calling **MC_Library_Initialization**.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_LIBRARY_ALREADY_INITIALIZED The library has already been initialized.

See Also	
MC_Library_Initialization MC_Library_Reset	MC_Library_Release

MC_Set_Message_Callbacks

Associates registered callback functions with a message or file

```
      Synopsis

      #include ``mc3msg.h"

      MC_STATUS MC_Set_Message_Callbacks (
int ApplicationID,
int MessageFileID,

      )

      ApplicationID

      The identifier assigned to this application by the
MC_Register_Application function.

      MessageFileID

      The identifier assigned to this object by the
MC_Open_Message or MC_Create_File functions.
```

Remarks

MC_Set_Message_Callbacks associates the callback functions registered for an application with a message or file object. When this function is not used, the callback function is first associated with a message or file when it is transmitted or received on a DICOM association or read or written to a DICOM file.

When **MC_Set_Message_Callbacks** is used, subsequent calls to **MC_Get_Value_To_Function** or **MC_Set_Value_From_Function** will also call the callback function to store an attribute's values.

[©] Copyright Merge Healthcare Solutions Inc. 2023

See MC_Register_Callback_Function for further details on using callback functions to manage OB, OW, or OF data.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_APPLICATION_ID	ApplicationID does not identify a valid application object.
MC_INVALID_MESSAGE_ID	<i>MessageFileID</i> is not a valid message or file object ID.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Set_Value_From_Function MC_Register_Callback_Function	MC_Get_Value_To_Function MC_Release_Callback_Function

MC_Set_Message_Transfer_Syntax

Sets the transfer syntax over which a message will be sent through the network.

Synopsis

<pre>#include ``mergecom.h"</pre>	
MC_STATUS MC_Set_Message_Transfer_Syntax (int MessageID, TRANSFER SYNTAX TransferSyntax	
)	_
MessagelD	The identifier assigned to this object by the MC_Open_Message or MC_Open_Empty_Message function.
TransferSyntax	The name of the transfer syntax to associate with the message. Use one of the enumerated TRANSFER_SYNTAX types defined in " mc3msg.h " (see MC_Get_Transfer_Syntax_From_Enum).

Remarks

MC_Set_Message_Transfer_Syntax associates a DICOM transfer syntax with a message. This function is only of use for SCU applications that negotiate more than one transfer syntax for a service. This call is used in conjunction with **MC_Send_Request_Message** to specify the transfer syntax to use when sending the message.

Reference the *Merge DICOM Toolkit User's Manual* for a discussion on negotiating multiple transfer syntaxes for a service.

MC_Set_Message_Transfer_Syntax can also be used to set the transfer syntax of a DICOM file object. When used with a DICOM file object, the tag (0002,0010) Transfer Syntax UID will be set with the proper UID for the transfer syntax specified.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MessagelD</i> is not a valid message object ID.
See Also	
MO O and Damas t Massa to	NO Oct Manager Transfer Ormitan

MC_Send_Request_Message MC_Get_Message_Transfer_Syntax MC_Reset_Message_Transfer_Syntax

MC_Set_Negotiation_Info

Registers extended negotiation information.

MC_Set_Negotiation_Info_For_Association and setting of negotiation information through service lists should be used in place of this call and **MC_Clear_Negotiation_Info**.

NOTE: Use of this call is deprecated.

Synopsis

#include ``mergecom.h"

```
MC_STATUS MC_Set_Negotiation_Info (
    int ApplicationID,
    char *ServiceName,
    void *ExtInfoBuffer,
    int ExtInfoLength
```

```
)
```

ApplicationID	The identification number for the registered application.
ServiceName	The name given to a valid DICOM service.
ExtInfoBuffer	A pointer to the buffer containing extended negotiation information.
ExtInfoLength	The number of bytes contained in the <i>ExtInfoBuffer</i> . This must be an even number.

Remarks

The DICOM standard allows application entities to exchange "extended negotiation information" when establishing an association. The contents of the negotiation information must be known to both the association requestor application and the association acceptor application. Such extended negotiation is not often used for DICOM services, but some services may require it.

MC_Set_Negotiation_Info allows the caller to supply Merge DICOM Toolkit with extended negotiation information which it will use when establishing associations.

When a **MC_Open_Association** or **MC_Open_Secure_Association** call is made, Merge DICOM Toolkit sends any registered extended negotiation information to the association acceptor. The

[©] Copyright Merge Healthcare Solutions Inc. 2023

acceptor normally returns an updated version of the negotiation information which can be accessed using the **MC_Get_Negotiation_Info** call.

When a **MC_Wait_For_Association** or **MC_Wait_For_Secure_Association** call is made, Merge DICOM Toolkit stores any received extended negotiation information. This information may be accessed using the **MC_Get_Negotiation_Info** call. **MC_Set_Negotiation_Info** may be used to "update" the extended negotiation information before calling **MC_Accept_Association** to accept the association. Merge DICOM Toolkit will return any registered negotiation information to the remote application.

Use MC_Clear_Negotiation_Info to remove extended information registered for a service.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_APPLICATION_ID	ApplicationID is not a valid application identifier.
MC_NULL_POINTER_PARM	Either ServiceName or ExtInfoBuffer was NULL.
MC_UNKNOWN_SERVICE	<i>ServiceName</i> was not registered in the Merge DICOM Toolkit configuration files.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	

MC_Get_Negotiation_Info MC_Clear_Negotiation_Info MC_Set_Negotiation_Info_For_Association

MC_Set_Negotiation_Info_For_Association

Sets extended negotiation information before acceptance of an association.

ServiceName	The name given to a valid DICOM service.
ExtInfoBuffer	A pointer to the buffer containing extended negotiation information.
ExtInfoLength	The number of bytes contained in the <i>ExtInfoBuffer</i> .

The DICOM standard allows application entities to exchange "extended negotiation information" when establishing an association. The contents of the negotiation information must be known to both the association requestor application and the association acceptor application. Such extended negotiation is not often used for DICOM services, but some services may require it.

MC_Set_Negotiation_Info_For_Association allows the caller to supply Merge DICOM Toolkit with extended negotiation information which it will use when establishing associations. When a MC_Wait_For_Association or MC_Wait_For_Secure_Association call is made, Merge DICOM Toolkit stores any received extended negotiation information. This information may be accessed using the MC_Get_Negotiation_Info call. MC_Set_Negotiation_Info_For_Association may then be used to set extended negotiation information in the association response. After setting the extended negotiation information would be called to accept the association. If MC_Set_Negotiation_Info_For_Association information will be returned by MC_Accept_Association is not used, no extended negotiation information will be returned by MC_Accept_Association in the response.

Note that if no negotiation information was included in the association request, **MC_Set_Negotiation_Info_For_Association** will not set the negotiation information in the association response.

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_ASSOC_ID	AssociationID is not a valid association identifier.
MC_NO_REQUEST_PENDING	<i>AssociationID</i> references to an association that has already been accepted.
MC_NULL_POINTER_PARM	Either ServiceName or ExtInfoBuffer was NULL.
MC_UNKNOWN_SERVICE	<i>ServiceName</i> was not registered in the Merge DICOM Toolkit configuration files or was not negotiated for the association.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Get_Negotiation_Info	MC_Clear_Negotiation_Info

Return Value

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

MC_Set_Next_Encapsulated_Value_From_Function

Encapsulates the next frame and stores the encapsulated value into the message object for the given attribute. The value representation of the attribute must be OB or OW. If a compressor is registered, the data will be compressed in the message.

Synopsis

```
#include ``mc3msg.h"
```

```
sgFileItemID The identifier assigned to this object by the
MC_Open_Message, MC_Open_Empty_Message,
MC_Create_File, MC_Create_Empty_File, or
MC_Open_Item functions.
```

Tag DICOM tag which identifies the attribute.

UserInfo	Address	s of ι	use	r data	whi	ich	wil	l be p	bas	sed or	n to		
	YourSe	tFun	nctio	on eac	ch ti	me	it i	s cal	led	. This r	nay	/be	э
	NULL.												

YourSetFunction Name of a function which will be called repeatedly for blocks of data for the attribute's value.

The function must be prototyped as follows:

```
MC STATUS YourSetFunction (
       int CBMsqFileItemID,
       unsigned long CBtag,
       int CBisFirst,
       void *CBuserInfo,
       int *CBdataSizePtr,
       void **CBdataBufferPtr,
       int *CBisLastPtr
)
 CBMsgFileItemID
                    The identifier assigned to this object by the
                    MC_Open_Message, MC_Open_Empty_Message,
                    MC_Create_File, MC_Create_Empty_File, or
                    MC_Open_Item functions.
 CBtag
                    DICOM tag which identifies the attribute.
 CBisFirst
                    This is TRUE (non-zero) the first time Merge DICOM
                    Toolkit calls YourSetFunction to request data blocks.
 CBuserInfo
                    Address of data which is being passed from the
                    MC_Set_Value_From_Function function. This may be
                    NULL.
                    Set *CBdataSizePtr to the number of bytes you are
 CBdataSizePtr
                    providing.
```

CBdataBufferPtr	Set <i>*CBdataBufferPtr</i> to the address of the data you are providing.
CBisLastPtr	Set <i>*CBisLastPtr</i> to TRUE (not zero) when you are returning with the last block of OBOW data.

Remarks

The **MC_Set_Next_Encapsulated_Value_From_Function** function is used to encapsulate the next frame, and then set the value of an attribute which has a value representation of OB or OW. If a compressor is registered with **MC_Register_Compression_Callbacks**, the data will be compressed using the registered compressor. Such attributes tend to have values of great length. To accommodate this, one uses the **MC_Set_Next_Encapsulated_Value_From_Function** function to specify the name of a function (*YourSetFunction*) which Merge DICOM Toolkit, in turn, calls repeatedly requesting blocks of the attribute's data value.

MC_Close_Encapsulated_Value must be called when the user is through encapsulating data within this message. If more frames will be encapsulated, use

MC_Set_Next_Encapsulated_Value_From_Function until all frames have been encapsulated, followed by **MC_Close_Encapsulated_Value**.

An optional *UserInfo* parameter may be used to pass information between the **MC_Set_Next_Encapsulated_Value_From_Function** caller and *YourSetFunction* which receives the data in its *CBuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

YourSetFunction

YourSetFunction must set **CBdataBufferPtr* to the address of a block of data which is a portion of the attribute's value. It also must set **CBdataSizePtr* to the number of bytes in the block.

Merge DICOM Toolkit sets *CBisFirst* to TRUE (non-zero) if it is requesting the first block of the attribute's value. *YourSetFunction* must set **CBisLastPtr* to TRUE (non-zero) if it is supplying the last block of the attribute's value.

YourSetFunction must return **MC_NORMAL_COMPLETION** if all went well. If not, it must return **MC_CANNOT_COMPLY**. Both of these are defined in "mc3msg.h".

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TRANSFER_SYNTAX	The message's transfer syntax is non-encapsulated.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag.</i> The attribute can be added with MC_Add_Standard_Attribute . This status will not be returned if setting the value of a message object which was opened

	using MC_Open_Empty_Message or MC_Create_Empty_File . In that case, the attribute is automatically added to the object before setting the value.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INCOMPATIBLE_VR	The attribute's value representation was not OB or OW.
MC_CALLBACK_DATA_SIZE_UNEVE	N
	The value set by <i>YourSetFunction</i> in its * <i>CBdataSizePtr</i> parameter was not an even number.
MC_CALLBACK_PARM_ERROR	The value set by <i>YourSetFunction</i> in its * <i>CBdataSizePtr</i> parameter was an odd number
MC_CALLBACK_CANNOT_COMPLY	YourSetFunction returned with MC_CANNOT_COMPLY.
See Also	

MC_Set_Encapsulated_Value_From_Function MC_Close_Encapsulated_Value MC_Register_Compression_Callbacks

MC_Set_Next_pValue... Functions

Appends another value to a private attribute in a message object.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Set_Next_pValue (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC DT DataType, void* Value
)
MC STATUS MC Set Next pValue From Float (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      float Value
)
MC_STATUS MC_Set_Next_pValue_From_Double (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      double Value
MC_STATUS MC_Set_Next_pValue_From_ShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
```

© Copyright Merge Healthcare Solutions Inc. 2023

```
unsigned char ElementByte,
      short int Value
MC STATUS MC Set Next pValue From UShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned short Value
MC STATUS MC Set Next pValue From Int (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      int Value
)
MC STATUS MC Set Next pValue From UInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned int Value
)
MC_STATUS MC_Set_Next_pValue_From_LongInt (
      int MsgFileItemID,
      char*PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      long int Value
MC STATUS MC Set Next pValue From ULongInt (
      int MsqFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long Value
MC_STATUS MC_Set_Next_pValue_From_LongLong (
      int MsgFileItemID,
      char*PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      long long Value
MC STATUS MC Set Next pValue From ULongLong (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long long Value
)
MC_STATUS MC_Set_Next_pValue_From_String (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
unsigned char ElementByte,
       char *Value
MC STATUS MC Set Next pValue From UnicodeString (
       int MsgFileItemID,
       char *PrivateCode,
       unsigned short Group,
       unsigned char ElementByte,
       int ValueLength,
       const MC Uhar *Value
)
 MsgFileItemID
                 The identifier assigned to this object by the
                 MC_Open_Message, MC_Open_Empty_Message,
                 MC_Create_File, MC_Create_Empty_File, or
                 MC_Open_Item functions.
 PrivateCode
                 The code string which identifies which block in the private
                 Group "owns" the attribute.
 Group
                 The number identifying the private group containing the
                 private attribute. It must be an odd number.
                 The number identifying the private attribute within the
 ElementByte
                 private Group for this PrivateCode.
                 One of the enumerated codes identifying the data type of the
 DataType
                 value in Value. The MC_DT enumerated type is defined in
                 "mc3msg.h". They are:
                 String_Type
                                       Null-terminated character string
                 Int_Type
                                       Binary integer number
                 UInt_Type
                                       Binary unsigned integer number
                                       Binary short integer number
                 ShortInt_Type
                 UShortInt_Type
                                       Binary unsigned short integer
                 number
                 LongInt_Type
                                       Binary long integer number
                 ULongInt_Type
                                       Binary unsigned long integer number
                 LongLong_Type
                                       Binary 64-bit integer number
                 UlongLong_Type
                                       Binary 64-bit unsigned integer
                 number Float_Type
                                              Binary Floating point number
                 Double_Type
                                       Binary Floating point number
                 Buffer_Type
                                       Binary byte value
 Value
                 The attribute should be set to this value.
```

Remarks

These functions add another value for an attribute identified by *ElementByte* for the *PrivateCode* within the given *Group*. If no values exist yet for the attribute, *Value* is stored as the attribute's first value. If **MC_Set_Next_pValue** is used, the data type of the *Value* is specified by the *DataType* parameter. The other function names imply the data type. For example,

MC_Set_Next_pValue_From_Int is the same as calling MC_Set_Next_pValue with DataType

specified as **Int_Type**. Each function will assign the value at *Value*, which must be prototyped as the appropriate type.

NOTE: String values are NULL-terminated character arrays.

For **MC_Set_Next_pValue_From_UnicodeString**, the *ValueLength* parameter is for specifying the number of Unicode characters in the input *Value*. This function requires **MC_Enable_Unicode_Conversion** being called first. See documentation for **MC_Enable_Unicode_Conversion**.

Any reasonable conversion will be made from *Value's* data type to the attribute's value representation. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

NOTE: Attributes with a value representation of **SQ** (sequence of items) *Value* must represent an integer which is the *ItemID* of an item object opened previously using **MC_Open_Item**.

Function	May be used to set values to attributes with these Value Representations		
MC_Set_Next_pValue_From_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_Int	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_UInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_LongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_ULongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_ULongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ		
MC_Set_Next_pValue_From_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ		
MC_Set_Next_pValue_From_UnicodeStri	ng LO, LT, PN, SH, ST, UC, UT		

The same rules apply to the **MC_Set_Next_pValue** function, based on the value used in the *DataType* parameter.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

	Value	Meaning
INFORMATIONAL	MESSAGES: MC_NORMAL_COMPLETION	The function completed normally.
WARNINGS: (VALUE WAS ST	ILL ENCODED)	
	MC_INVALID_CHARS_IN_VALUE	A string-type <i>Value</i> had invalid characters for the value representation of <i>Tag</i> .
	MC_INVALID_VALUE_FOR_VR	<i>Value</i> does not conform to the requirements of its value representation.
Erro	RS:	
	MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
	MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
	MC_INVALID_GROUP	<i>Group</i> was not an odd number.
	MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
	MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
	MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table above.
	MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
	MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
	MC_INVALID_DATA_TYPE	<i>DataType</i> is not valid.
	MC_TOO_MANY_VALUES	This attribute contains the most values that a standard attribute can contain (65535).
	MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
	See Also	
	MC_Set_Value Functions MC_Set_Value_To_NULL MC_Set_pValue Functions MC_Set_pValue_To_NULL MC_Set_Next_Value Functions	MC_Set_Value_From_Function MC_Set_Value_To_Empty MC_Set_pValue_From_Function MC_Set_pValue_To_Empty MC_Stream_To_Message

MC_Set_Next_pValue_To_NULL

Sets the next value of a private attribute in a message object to NULL.

Synopsis

<pre>#include ``mc3msg.h"</pre>				
<pre>MC_STATUS MC_Set_pValue_To_NULL (int MsgFileItemID, char *PrivateCode, unsigned short Group, unsigned char ElementByte)</pre>				
MsgFileItemID	The identifier assigned to this object by the MC_Open_Message function or the MC_Open_Item function.			
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.			
Group	The number identifying the private group containing the private attribute. It must be an odd number.			
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .			

Remarks

MC_Set_Next_pValue_To_NULL adds a NULL to the next value of a private multi-valued text type attribute identified by *PrivateCode, Group* and *ElementByte* in the message identified by *MsgFileItemID*. Any existing values remain in the attribute. Note that the function can only be used on attributes which 1) allow the backslash (\) character as a field delimiter and 2) are of a text type. It can <u>not</u> be used on numerical attributes.

If a callback function was registered for the attribute, it is "de-registered."

NOTE: if the attribute has a value representation of SQ (sequence of items), setting the value to null does NOT free the item object identified by the value.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INCOMPATIBLE_VR	The attribute identified by <i>Tag</i> can not be set to NULL.
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>

MC_INVALID_TAG	The message does not contain an attribute identified by Tag.
	Note: This status will not be returned if setting the value of a
	message object which was opened using
	MC_Open_Empty_Message or MC_Create_Empty_File. In
	that case, the attribute is automatically added to the object
	before setting the value to NULL.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
O Al	

See Also

MC_Set_Next_Value_To_NULL

MC_Set_Next_Value... Functions

Appends a value to an attribute in a message object.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Set_Next_Value (
      int MsgFileItemID,
      unsigned long Tag,
      MC DT DataType,
      void *Value
MC_STATUS MC_Set_Next_Value_From_Float (
      int MsgFileItemID,
      unsigned long Tag,
      float Value
)
MC STATUS MC Set Next Value From Double (
      int MsgFileItemID,
      unsigned long Tag,
      double Value
MC_STATUS MC_Set_Next_Value_From_ShortInt (
      int MsgFileItemID,
      unsigned long Tag,
      short int Value
MC STATUS MC Set_Next_Value_From_UShortInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned short Value
)
MC_STATUS MC_Set_Next_Value_From_Int (
      int MsgFileItemID,
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
unsigned long Tag,
      int Value
MC STATUS MC Set Next Value From UInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned int Value
)
MC STATUS MC Set Next Value From LongInt (
      int MsgFileItemID,
      unsigned long Tag,
      long int Value
MC STATUS MC Set Next Value From ULongInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long Value
MC_STATUS MC_Set_Next_Value_From_LongLong (
      int MsgFileItemID,
      unsigned long Tag,
      long long Value
)
MC STATUS MC Set Next Value From ULongLong (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long long Value
MC STATUS MC Set Next Value From String (
      int MsgFileItemID,
      unsigned long Tag,
      char *Value
MC STATUS MC Set Next Value From UnicodeString (
      int MsqFileItemID,
      unsigned long Tag,
      int ValueLength,
      const MC Uhar *Value)
 MsgFileItemID
               The identifier assigned to this object by the
               MC_Open_Message, MC_Open_Empty_Message,
                MC_Create_File, MC_Create_Empty_File, or
                MC_Open_Item functions.
 Tag
               DICOM tag which identifies the attribute.
 DataType
               One of the enumerated codes identifying the data type of the
               value in Value. The MC_DT enumerated type is defined in
                "mc3msg.h". They are:
```

	String_Type	Null-terminated character string
	Int_Type	Binary integer number
	UInt_Type	Binary unsigned integer number
	ShortInt_Type	Binary short integer number
	UShortInt_Type	Binary unsigned short integer
	number	
	LongInt_Type	Binary long integer number
	ULongInt_Type	Binary unsigned long integer number
	LongLong_Type	Binary 64-bit integer number
	UlongLong_Type	Binary 64-bit unsigned integer
	number Float_Type	Binary Floating point number
	Double_Type	Binary Floating point number
	Buffer_Type	Binary byte value
,	The attribute will be set	t to this value.

Remarks

Value

These functions add another value for the attribute with the given Tag. If no values exist yet for the attribute, Value is stored as the attribute's first value. If MC_Set_Next_Value is used, the data type of the Value is specified by the DataType parameter. The other function names imply the data type. For example, MC_Set_Next_Value_From_Int is the same as calling MC_Set_Next_Value with DataType specified as Int_Type. Each function will set the value at Value, which must be prototyped as the appropriate type. Note that string values are NULL-terminated character arrays.

For MC_Set_Next_Value_From_UnicodeString, the ValueLength parameter is for specifying the number of Unicode characters in the input Value. This function requires MC_Enable_Unicode_Conversion being called first. See documentation for MC_Enable_Unicode_Conversion.

Any reasonable conversion will be made from Value's data type to the attribute's value representation. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

NOTE:	For attributes with a value representation of SQ (sequence of items)
	Value must represent an integer which is the <i>ItemID</i> of an item object
	opened previously using MC_Open_Item.

Functions	May be used to set values to attributes with these Value Representations
MC_Set_Next_Value_From_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_Int	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ

© Copyright Merge Healthcare Solutions Inc. 2023

Merge DICOM Toolkit[™] 5.16.0 C/C++ REFERENCE MANUAL

Functions	May be used to set values to attributes with these Value Representations
MC_Set_Next_Value_From_UInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_LongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_ULongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_ULongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Next_Value_From_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ
MC_Set_Next_Value_From_UnicodeStrip	ng LO, LT, PN, SH, ST, UC, UT

The same rules apply to the **MC_Set_Next_Value** function, based on the value used in the *DataType* parameter.

© Copyright Merge Healthcare Solutions Inc. 2023

Return Value

	Value	Meaning
INFORMATIONAL	MESSAGES: MC_NORMAL_COMPLETION	The function completed normally.
WARNINGS: (VALUE WAS STI	LL ENCODED)	
	MC_INVALID_CHARS_IN_VALUE	A string-type <i>Value</i> had invalid characters for the value representation of <i>Tag</i> .
	MC_INVALID_VALUE_FOR_VR	<i>Value</i> does not conform to the requirements of its value representation.
Erroi	RS:	
	MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
	MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> . The attribute can be added with MC_Add_Standard_Attribute .
		Note: This status will not be returned if setting the value of a message object which was opened using MC_Open_Empty_Message or MC_Create_Empty_File . In that case, the attribute is automatically added to the object before setting the value.
	MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
	MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table below.
	MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
	MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
	MC_INVALID_DATA_TYPE	<i>DataType</i> is not valid.
	MC_TOO_MANY_VALUES	This attribute contains the most values that a standard attribute can contain (65535).
	MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
	MC_CANNOT_COMPLY	The function fails to process the request. Check Merge DICOM Toolkit log for detail.

See Also

MC_Set_Value... Functions MC_Set_Value_To_NULL MC_Set_pValue... Functions MC_Set_pValue_To_NULL MC_Set_Next_pValue... Functions MC_Enable_Unicode_Conversion MC_Set_Value_From_Function MC_Set_Value_To_Empty MC_Set_pValue_From_Function MC_Set_pValue_To_Empty MC_Stream_To_Message

MC_Set_Next_Value_To_NULL

Sets the next value of an attribute in a message object to NULL.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Set_Next_Value_To_NULL (
    int MsgFileItemID,
    unsigned long Tag
)
MsgFileItemID The identifier assigned to this object by the
    MC_Open_Message, MC_Open_Empty_Message,
    MC_Create_File, MC_Create_Empty_File, or
    MC_Open_Item functions.
Tag DICOM tag which identifies the attribute.
```

Remarks

This function adds a NULL to the next value of a multi-valued text type attribute with the given *Tag*. Any existing values remain in the attribute. Note that the function can only be used on attributes which 1) allow the backslash (\) character as a field delimiter and 2) are of a text type. It can <u>not</u> be used on numerical attributes.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INCOMPATIBLE_VR	The attribute identified by <i>Tag</i> can not be set to NULL.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>Tag.</i> The attribute can be added with MC_Add_Standard_Attribute . Note: This status will not be returned if setting the value of a message object which was opened using MC_Open_Empty_Message or MC_Create_Empty_File . In that case, the attribute is automatically added to the object before setting the value to NULL.

MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Set_Next_pValue_To_NULL

MC_Set_pValue... Functions

Sets the value of a private attribute in a message object.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Set pValue (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      MC DT DataType,
      void* Value
MC_STATUS MC_Set_pValue_From_Float (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      float Value
)
MC STATUS MC Set pValue From Double (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      double Value
MC STATUS MC Set pValue From ShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      short int Value
MC_STATUS MC_Set_pValue_From_UShortInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned short Value
```

```
MC_STATUS MC_Set_pValue_From_Int (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      int Value
MC STATUS MC Set pValue From UInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned int Value
MC STATUS MC Set pValue From LongInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      long int Value
MC_STATUS MC_Set_pValue_From_ULongInt (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long Value
MC STATUS MC Set pValue From LongLong (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      long long Value
MC STATUS MC Set pValue From ULongLong (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      unsigned long long Value
MC STATUS MC Set pValue From String (
      int MsqFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      char *Value
)
MC STATUS MC Set pValue From UnicodeString (
      int MsgFileItemID,
      char *PrivateCode,
      unsigned short Group,
      unsigned char ElementByte,
      int ValueLength,
```

```
const MC Uhar* Value
)
MC STATUS MC Set pValue From Buffer (
       int MsgFileItemID,
       char *PrivateCode,
       unsigned short Group,
       unsigned char ElementByte,
       void *Value
       unsigned long ValueLength
)
 MsgFileItemID
                 The identifier assigned to this object by the
                 MC_Open_Message, MC_Open_Empty_Message,
                 MC_Create_File, MC_Create_Empty_File, or
                 MC_Open_Item functions.
 PrivateCode
                 The code string which identifies which block in the private
                  Group "owns" the attribute.
                 The number identifying the private group containing the
 Group
                 private attribute. It must be an odd number.
                 The number identifying the private attribute within the
 ElementByte
                 private Group for this PrivateCode.
                 One of the enumerated codes identifying the data type of the
 DataType
                 value in Value. The MC_DT enumerated type is defined in
                 "mc3msg.h". They are:
                 String_Type
                                       Null-terminated character string
                 Int_Type
                                       Binary integer number
                 UInt_Type
                                       Binary unsigned integer number
                 ShortInt_Type
                                       Binary short integer number
                 UShortInt_Type
                                       Binary unsigned short integer
                 number
                 LongInt_Type
                                       Binary long integer number
                 ULongInt_Type
                                       Binary unsigned long integer number
                                       Binary 64-bit integer number
                 LongLong_Type
                 UlongLong_Type
                                       Binary 64-bit unsigned integer
                 number
                 Float_Type
                                       Binary Floating point number
                 Double_Type
                                       Binary Floating point number
                 Buffer_Type
                                       Binary byte value
 Value
                 The attribute should be set to this value.
 ValueLength
                 Length of provided Value.
```

Remarks

These functions store a new value for a private attribute identified by *ElementByte* for the *PrivateCode* within the given *Group*. If one or more values already exist for the attribute, they are first removed. Use **MC_Set_Next_pValue** to append attribute values. If **MC_Set_pValue** is used, the data type of the *Value* is specified by the *DataType* parameter. The other function names imply the data type. For example, **MC_Set_pValue_From_Int** is the same as calling **MC_Set_pValue** with *DataType* specified

NOTE:

as **Int_Type**. Each function will assign the value at *Value*, which must be prototyped as the appropriate type.

NOTE: String values are NULL-terminated character arrays.

For **MC_Set_pValue_From_UnicodeString**, the *ValueLength* parameter is for specifying the number of Unicode characters in the input *Value*. This function requires **MC_Enable_Unicode_Conversion** being called first. See documentation for **MC_Enable_Unicode_Conversion**.

MC_Set_pValue_From_Buffer is intended mainly for setting the value of an attribute whose value representation is unknown. This may occur if **MC_Message_To_Stream** is to be used and one or more of the stream attributes is unknown. The attribute's value is simply copied (memcpy) from the *Value* buffer according to the length provided in *ValueLength*. The function supports other value representations as well – see table below.

Any reasonable conversion will be made from *Value's* data type to the attribute's value representation. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

Function	May be used to set values to attributes with these Value Representations
MC_Set_pValue_From_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_pValue_From_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_pValue_From_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_pValue_From_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_pValue_From_Int	DS, FD, FL, IS, SL, SS, UL, US, SQ
MC_Set_pValue_From_UInt	DS, FD, FL, IS, SL, SS, UL, US, SQ
MC_Set_pValue_From_LongInt	AT, DS, FD, FL, IS, SL, SS, UL, US, SQ
MC_Set_pValue_From_UlongInt	AT, DS, FD, FL, IS, SL, SS, UL, US, SQ
MC_Set_pValue_From_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_pValue_From_ULongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_pValue_From_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ
MC_Set_pValue_From_UnicodeString	LO, LT, PN, SH, ST, UC, UT

For attributes with a value representation of **SQ** (sequence of items) *Value* must represent an integer which is the *ItemID* of an item object opened previously using **MC_Open_Item**.

Function	May be used to set values to attributes with these Value Representations
MC_Set_pValue_From_Buffer	UNKNOWN_VR , OB, OW, OV, OL, OD, OF, UR, UT
MC_Set_pValue_From_Function	UNKNOWN_VR, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR, UT

The same rules apply to the **MC_Set_pValue** function, based on the value used in the *DataType* parameter. The **MC_Set_pValue_From_Function** call is described in its own section.

Return Value

	Value	Meaning
INFORMATION	AL MESSAGES:	
	MC_NORMAL_COMPLETION	The function completed normally.
(VALUE WAS S	WARNINGS: STILL ENCODED)	
	MC_INVALID_CHARS_IN_VALUE	A string-type <i>Value</i> had invalid characters for the value representation of <i>Tag</i> .
	MC_INVALID_VALUE_FOR_VR	<i>Value</i> does not conform to the requirements of its value representation.
	ERRORS:	
	MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
	MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
		Note: This status will not be returned if setting the value of a message object which was opened using
		MC_Open_Empty_Message . In that case, the attribute is automatically added to the object before setting the value.
	MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
	MC_INVALID_GROUP	<i>Group</i> was not an odd number.
	MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
	MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table above.
	MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g., setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).

MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_INVALID_DATA_TYPE	<i>DataType</i> is not valid.
MC_TOO_MANY_VALUES	This attribute contains the most values that a standard attribute can contain (65535).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Set_Value Functions MC_Set_Value_To_NULL MC_Set_pValue_From_Function MC_Set_pValue_To_Empty MC_Set_Next_pValue Functions	MC_Set_Value_From_Function MC_Set_Value_To_Empty MC_Set_pValue_To_NULL MC_Set_Next_Value Functions MC_Stream_To_Message

MC_Set_pValue_From_Function

Sets the value of a message object private attribute which has the binary value representation of OB, OW, OV, OL, OD or OF, or the numeric value representation of AT, SS, US, SL, UL, SV, UV, FL or FD, or the text value representation of UR or UT.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Set_pValue_From_Function (
       int MsgFileItemID,
       char *PrivateCode,
       unsigned short Group,
       unsigned char ElementByte,
       void *UserInfo,
       MC STATUS (*YourSetFunction)()
)
 MsgFileItemID
                  The identifier assigned to this object by the
                  MC_Open_Message, MC_Open_Empty_Message,
                  MC_Create_File, MC_Create_Empty_File, or
                  MC_Open_Item functions.
 PrivateCode
                  The code string which identifies which block in the private
                  Group "owns" the attribute.
                  The number identifying the private group containing the
 Group
                  private attribute. It must be an odd number.
 ElementByte
                  The number identifying the private attribute within the
```

private Group for this PrivateCode.UserInfoAddress of data which will be passed on to YourSetFunction
each time it is called. This may be NULL.

YourSetFunctio Name of a function which will be called repeatedly to request blocks of data for the attribute's value. п

The function must be prototyped as follows:

```
MC STATUS YourSetFunction (
       int CBMsqFileItemID,
       unsigned long Cbtag,
       int CBisFirst,
       void *CBuserInfo,
       int *CBdataSizePtr,
       void **CBdataBufferPtr,
       int *CBisLastPtr
  CBMsgFileItemID
                     The identifier assigned to this object by the
                     MC_Open_Message, MC_Open_Empty_Message,
                     MC_Create_File, MC_Create_Empty_File, or
                     MC_Open_Item functions.
  Cbtag
                     DICOM tag which identifies the attribute.
  CBisFirst
                     This is TRUE (non-zero) the first time Merge DICOM
                     Toolkit calls YourSetFunction to request data blocks.
  CBuserInfo
                     Address of data which is being passed from the
                     MC_Set_pValue_From_Function function. This may be
                     NULL.
  CBdataSizePtr
                     Set *CBdataSizePtr to the number of bytes you are
                     providing.
  CBdataBufferPtr
                     Set * CBdataBufferPtr to the address of the data you are
                     providing.
  CBisLastPtr
                     Set *CBisLastPtr to TRUE (not zero) when you are
                     returning with the last block of OB, OW, or OF data.
```

Remarks

)

The MC_Set_pValue_From_Function function is used to set the value of a private attribute which has a value representation of UN, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR or UT. Such attributes tend to have values of great length. To accommodate this, one uses MC_Set_pValue_From_Function to specify the name of a function (YourSetFunction) which Merge DICOM Toolkit, in turn, calls. This "callback" function is called repeatedly to request blocks of the attribute's data value.

An optional UserInfo parameter may be used to pass information between the MC_Set_pValue_From_Function caller and YourSetFunction which receives the data in its CBuserInfo parameter. If no such exchange of data is necessary, you may use NULL for UserInfo.

YourSetFunction

YourSetFunction must set *CBdataBufferPtr to the address of a block of data which is a portion of the attribute's value. It also must set **CBdataSizePtr* to the number of bytes in the block.

Merge DICOM Toolkit sets *CBisFirst* to TRUE (non-zero) if it is requesting the first block of the attribute's value. *YourSetFunction* must set **CBisLastPtr* to TRUE (non-zero) if it is supplying the last block of the attribute's value.

YourSetFunction must return **MC_NORMAL_COMPLETION** if all went well. If not, it must return **MC_CANNOT_COMPLY**. Both of these are defined in **"mcstatus.h"**.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	PrivateCode or YourSetFunction was NULL.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INCOMPATIBLE_VR	The attribute's value representation was not one of UN, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR or UT.
MC_CALLBACK_PARM_ERROR	The value set by <i>YourSetFunction</i> in its * <i>CBdataBufferPtr</i> parameter was NULL and * <i>CBisLastPtr</i> was not set to true(non-zero).
MC_CALLBACK_DATA_SIZE_NEGAT	IVE
	The value set by <i>YourSetFunction</i> in its * <i>CBdataSizePtr</i> parameter was a negative number
MC_CALLBACK_DATA_SIZE_UNEVE	N
	The value set by <i>YourSetFunction</i> in its * <i>CBdataSizePtr</i> parameter was an odd number
MC_CALLBACK_CANNOT_COMPLY	
	YourSetFunction returned a status other than MC_NORMAL_COMPLETION.
MC_VALUE_TOO_LARGE	The value set by <i>YourSetFunction</i> int <i>*CBdataSizePtr</i> was too large. It may not be larger than the largest unsigned int on your machine.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Set_pValue... Functions MC_Set_pValue_To_Empty MC_Set_pValue_To_NULL MC_Set_Value... Functions MC_Set_Value_To_NULL MC_Set_Value_To_Empty MC_Set_Next_pValue... Functions

MC_Set_Next_Value... Functions

MC_Set_pValue_Representation

Sets a private attribute's value representation code.

Synopsis

#include ``mc3msg.h" MC STATUS MC Set pValue Representation (int MsgFileItemID, char *PrivateCode, unsigned short Group, unsigned char ElementByte, MC VR ValueRep) The identifier assigned to this object by the MsgFileItemID MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions. PrivateCode The code string which identifies which block in the private Group "owns" the attribute. The number identifying the private group containing the Group private attribute. It must be an odd number. The number identifying the private attribute within the **ElementByte** private Group for this PrivateCode. A valid value representation code as defined by the MC_VR ValueRep type in "mc3msg.h": AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR, UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL, OD, OF, SQ

Remarks

MC_Set_pValue_Representation is used to change a private attribute's value representation. The attribute is identified by *ElementByte* for the *PrivateCode* within the given *Group*. A message attribute may have an unknown value representation if it is obtained from a message stream (**MC_Stream_To_Message**) and the attribute is unknown. This method can be used to facilitate reading of the values for such attributes. This function can also be used to manage the attributes that have more than one possible VR.

The valid VR conversions are:

[©] Copyright Merge Healthcare Solutions Inc. 2023

	(inal R	New VR
US	US	SS
		OW
S	S	US
S	S	OW
0	W	US
0	W	SS
UV	SV	OV
		OV
	V	UV
C	V	SV
0	W	OB
0	В	OW
-	V	OB
0	В	OV
U	N	Any VR

Any other conversion will result in an error return status.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_VR_CODE	<i>ValueRep</i> is invalid.
MC_ATTRIBUTE_HAS_VALUES	When <i>ValueRep</i> is SQ, the attribute may not already have a value.
MC_VR_ALREADY_VALID	The attribute already has a valid Value Representation.
See Also	
MC_Set_Value_Representation	

MC_Set_pValue_To_Empty

Removes all values from a private attribute in a message object.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Synopsis

<pre>#include "mc3m</pre>	sg.h″
int MsgF char *Pr unsigned	et_pValue_To_Empty (TileItemID, TivateCode, A short Group, A char ElementByte
MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.
Group	The number identifying the private group containing the private attribute. It must be an odd number.
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .

Remarks

MC_Set_pValue_To_Empty removes any existing values for the attribute identified by *PrivateCode, Group* and *ElementByte* in the message identified by *MsgFileItemID*. If a callback function was registered for the attribute, it is "de-registered."

NOTE: This is not the same as setting a NULL value. Use **MC_Set_pValue_To_NULL** to give the attribute a value length of zero (i.e. a NULL value).

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
See Also	
MC_Set_Value_To_Empty MC_Set_pValue_To_NULL	MC_Set_Value_To_NULL

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Set_pValue_To_NULL

Sets the value of a private attribute in a message object to NULL.

Synopsis

#include "mc3msg.h"		
<pre>MC_STATUS MC_Set_pValue_To_NULL (int MsgFileItemID, char *PrivateCode, unsigned short Group, unsigned char ElementByte</pre>		
)		
MsgFileItemID	The identifier assigned to this object by the	
	MC_Open_Message, MC_Open_Empty_Message,	
	MC_Create_File, MC_Create_Empty_File, or	
	MC_Open_Item functions.	
PrivateCode	The code string which identifies which block in the private <i>Group</i> "owns" the attribute.	
Group	The number identifying the private group containing the private attribute. It must be an odd number.	
ElementByte	The number identifying the private attribute within the private <i>Group</i> for this <i>PrivateCode</i> .	

Remarks

MC_Set_pValue_To_NULL sets the value of the attribute identified by *PrivateCode, Group* and *ElementByte* in the message identified by *MsgFileItem/D* to NULL. That is, it sets the attribute's value length to zero. NULL is a valid value in DICOM. Use **MC_Set_pValue_To_Empty** if the intent is to remove the attribute's value altogether. If a callback function was registered for the attribute, it is "deregistered."

NOTE: If the attribute has a value representation of SQ (sequence of items), setting the value to null does NOT free the item object identified by the value.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_PRIVATE_CODE	The message does not contain any attributes in <i>Group</i> for <i>PrivateCode.</i>
MC_INVALID_GROUP	<i>Group</i> was not an odd number.
MC_INVALID_TAG	The message does not contain an attribute identified by <i>ElementByte</i> in <i>Group</i> for <i>PrivateCode</i> .

MC_INVALID_MESSAGE_ID

MsgFileItemID is not a valid message object ID, file object ID or item object ID.

See Also

MC_Set_pValue_To_Empty MC_Set_Value_To_NULL MC_Set_Value_To_Empty

MC_Set_Service_Command

Associates a Merge DICOM Toolkit message or file object with a given DICOM service/command pair.

Synopsis

#include ``mc3msg.h"

```
      MC_STATUS MC_Set_Service_Command (

      int MessageID,

      char *ServiceName,

      MC_COMMAND Command

      )

      MessageID

      The identifier assigned to the message object by the

      MC_Open_Empty_Message, MC_Open_Message,

      MC_Create_Empty_File, or MC_Create_File function.

      ServiceName
      String name of a DICOM service to be associated with this message object.

      Command
      The command which is to be associated with this message.
```

The **MC_COMMAND** enumerated values are defined in "mc3msg.h".

Remarks

The **MC_Set_Service_Command** function associates a message object with a given DICOM service/command pair. This is necessary if the message object was opened using **MC_Open_Empty_Message** and the message object is to be sent to a network partner, or is to be validated using the **MC_Validate_Message** function.

The *ServiceName* must be a valid service name defined in the Merge DICOM Toolkit service profile file, and *Command* must be a valid DICOM command.

If the service "DICOMDIR" and command "C_STORE_RQ" is specified and the object identified by *MessageID* is a file object, the object will be converted to a DICOMDIR object. The object can then be manipulated with the **MC_Dir_...** functions. **MC_Set_Service_Command** should only be used on empty file objects, because the **MC_Dir_...** functions will not work correctly if any directory records are already contained in the object.

The following commands are supported by the Merge DICOM Toolkit:

Command	Description
C_STORE_RQ	DICOM Composite Store Service Request

[©] Copyright Merge Healthcare Solutions Inc. 2023

DICOM Composite Store Service Response
DICOM Verification Service Request
DICOM Verification Service Response
DICOM Composite Find Service Request
DICOM Composite Find Service Response
Cancel DICOM Composite Find Service Request
DICOM Composite Get Service Request
DICOM Composite Get Service Response
Cancel DICOM Composite Get Service Request
DICOM Composite Move Service Request
DICOM Composite Move Service Response
Cancel DICOM Composite Move Service Request
DICOM Normalized Report Service Request
DICOM Normalized Report Service Response
DICOM Normalized Get Service Request
DICOM Normalized Get Service Request
DICOM Normalized Set Service Request
DICOM Normalized Set Service Response
DICOM Normalized Action Service Request
DICOM Normalized Action Service Response
DICOM Normalized Create Service Request
DICOM Normalized Create Service Response
DICOM Normalized Delete Service Request
DICOM Normalized Delete Service Response

Reference the "DICOM V3.0 Standard, Final Text - October 29, 1993" for more information about these commands.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	ServiceName was NULL.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_Create_Empty_File	MC_Create_File MC_Open_Empty_Message MC_Open_Message MC_Get_MergeCOM_Service MC_Get_UID_From_MergeCOM_Service
See Also	
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_COMMAND	Command is not a supported command.
MC_INVALID_MESSAGE_ID	<i>MessageID</i> is not a valid message or file object ID.

MC_Set_String_Config_Value

Used to set the value of a character string toolkit configuration parameter at runtime.

Synopsis

```
#include ``mc3msg.h"
```

```
MC_STATUS MC_Set_String_Config_Value (
    StringParm Aparm,
    char *Avalue
```

```
)
```

0	,
Aparm	An enumerated constant identifying the character string configuration parameter to be (re)set. Aparm can have any of the following values: MERGECOM 3_PROFILE MERGECOM 3_SERVICES MERGECOM 3_SERVICES MERGECOM 3_APPLICATIONS CAPTURE_FILE COMPRESSION_RGB_TRANSFORM_FORMAT DECODER_TAGFLICIT_LITTLE_ENDIAN_SYNTAX DICTIONARY_ACCESS DICTIONARY_ACCESS DICTIONARY_FILE ENCAPSULATED_UNCOMPRESSED_ELE_SYNTAX EXPLICIT_BIG_ENDIAN_SYNTAX EXPLICIT_LITTLE_ENDIAN_SYNTAX HEVC_H265_MIOP_LEVEL5_1_SYNTAX HEVC_H265_MIOP_LEVEL5_1_SYNTAX HEVC_H265_MIOP_LEVEL5_1_SYNTAX IMPLEMENTATION_VERSION IMPLICIT_BIG_ENDIAN_SYNTAX IMPLEMENTATION_VERSION IMPLICIT_BIG_ENDIAN_SYNTAX JPEG_2000_LOSSLESS_ONLY_SYNTAX JPEG_2000_MC_LOSSLESS_ONLY_SYNTAX JPEG_2000_MC_SYNTAX JPEG_2000_SYNTAX JPEG_2000_SYNTAX JPEG_EXTENDED_2_4_SYNTAX JPEG_EXTENDED_2_4_SYNTAX JPEG_EXTENDED_HIER_17_19_SYNTAX JPEG_FULL_PROG_HIER_24_26_SYNTAX JPEG_FULL_PROG_HIER_12_12_SYNTAX JPEG_LOSSLESS_INTAX JPEG_LOSSLESS_INTAX JPEG_LOSSLESS_NON_HIER_11_13_SYNTAX JPEG_STENDED_HIER_17_19_SYNTAX JPEG_COUSSLESS_NON_HIER_10_12_SYNTAX JPEG_LOSSLESS_NON_HIER_10_12_SYNTAX JPEG_LOSSLESS_NON_HIER_10_12_SYNTAX JPEG_LOSSLESS_NON_HIER_10_12_SYNTAX JPEG_LOSSLESS_NON_HIER_10_SYNTAX JPEG_LOSSLESS_NON_HIER_10_SYNTAX JPEG_LOSSLESS_NON_HIER_10_SYNTAX JPEG_LOSSLESS_NON_HIER_11_3SYNTAX JPEG_LOSSLESS_NON_HIER_14_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_LOSSLESS_NON_HIER_15_SYNTAX JPEG_SPEC_NON_HIER_6_8_SYNTAX JPEG_SPEC_NON_HIER_7_9_SYNTAX JPEG_SPEC_NON_HIER_7_9_SYNTAX JPEG_SPEC_NON_HIER_7_9_SYNTAX JPEG_SPEC_NON_HIER_7_9_SYNTAX JPEG_SPEC_HIER_21_20_SYNTAX JPEG_SPEC_HIER_21_23_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SYNTAX JPEG_SPEC_HIER_21_SY

MPEG2 MPML SYNTAX MPEG4 AVC H264 BDC HP LEVEL 4 1 SYNTAX MPEG4 AVC H264 HP LEVEL 4 1 SYNTAX MPEG4_AVC_H264_HP_LEVEL_4_2_2D_SYNTAX MPEG4 AVC H264 HP LEVEL 4 2 3D SYNTAX MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2_SYNTAX MSG INFO FILE NULL TYPE3 VALIDATION PEGASUS DISP REG NAME PEGASUS DISP_REGISTRATION PEGASUS OP D2SEPLUS NAME PEGASUS_OP_D2SEPLUS_REGISTRATION PEGASUS OP J2KE NAME PEGASUS OP J2KE REGISTRATION PEGASUS_OP_J2KP_NAME PEGASUS OP J2KP REGISTRATION PEGASUS_OP_JLSE_NAME PEGASUS OP JLSE REGISTRATION PEGASUS OP JLSP NAME PEGASUS OP_JLSP_REGISTRATION PEGASUS OP LIE3PLUS NAME PEGASUS OP LIE3PLUS REGISTRATION PEGASUS_OP_LIP3PLUS_NAME PEGASUS OP LIP3PLUS REGISTRATION PEGASUS OP SE2DPLUS NAME PEGASUS OP SE2DPLUS REGISTRATION PEGASUS OPCODE PATH PRIVATE_SYNTAX_1_SYNTAX PRIVATE SYNTAX 2 SYNTAX RECEIVER NAME RLE SYNTAX SMPTE ST 2110 20 UNCOMPRESSED PROGRESSIVE ACTIVE VIDEO SYNTAX SMPTE_ST_2110_20_UNCOMPRESSED_INTERLACED_A CTIVE VIDEO SYNTAX SMPTE ST 2110 30 PCM DIGITAL AUDIO SYNTAX TEMP FILE DIRECTORY UNKNOWN VR CODE These names are the same as those given to the parameters in the toolkit configuration files. The character string value to which Aparm is to be set.

Remarks

Avalue

The Merge DICOM Toolkit Library accesses several configuration files at startup. This call allows your application to (re)set character string configurable parameters specified in these files at runtime. This call should be made immediately after calling **MC_Library_Initialization** to avoid using these parameters before they are set. Please see the detailed description of toolkit configuration elsewhere in this manual.

Return Value

[©] Copyright Merge Healthcare Solutions Inc. 2023

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_PARAMETER_NAME	Parameter specified is not one of the above listed parameters.
MC_NULL_POINTER_PARM	Avalue was a null pointer.
MC_SYSTEM_ERROR	Out of memory condition occurred.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_Get_Int_Config_Value MC_Get_Log_Destination MC_Get_Long_Config_Value MC_Get_String_Config_Value	MC_Set_Int_Config_Value MC_Set_Log_Destination MC_Set_Long_Config_Value MC_Set_String_Config_Value

MC_Set_Value... Functions

Sets the value of an attribute in a message object.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Set_Value (
      int MsgFileItemID,
      unsigned long Tag,
      MC DT DataType,
      void *Value
MC STATUS MC_Set_Value_From_Float (
      int MsgFileItemID,
      unsigned long Tag,
      float Value
)
MC_STATUS MC_Set_Value_From_Double (
      int MsgFileItemID,
      unsigned long Tag,
      double Value
MC STATUS MC Set Value From ShortInt
      (int MsgFileItemID,
      unsigned long Tag,
      short int Value
MC_STATUS MC_Set_Value_From_UShortInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned short Value
)
MC_STATUS MC_Set_Value_From_Int (
      int MsgFileItemID,
      unsigned long Tag,
      int Value
)
```

```
MC_STATUS MC_Set_Value_From_UInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned int Value
MC STATUS MC Set Value From LongInt (
      int MsqFileItemID,
      unsigned long Tag,
      long int Value
MC STATUS MC Set Value From ULongInt (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long Value
MC STATUS MC Set Value From LongLong (
      int MsgFileItemID,
      unsigned long Tag,
      long long Value
MC_STATUS MC_Set_Value_From_ULongLong (
      int MsgFileItemID,
      unsigned long Tag,
      unsigned long long Value
)
MC STATUS MC Set Value From String (
      int MsgFileItemID,
      unsigned long Taq,
      char* Value
MC STATUS MC Set Value From UnicodeString (
      int MsgFileItemID,
      unsigned long Taq,
      int ValueLength,
      const MC_Uhar* Value
MC STATUS MC Set Value From Buffer (
      int MsqFileItemID,
      unsigned long Tag,
      void *Value,
      unsigned long ValueLength
)
 MsgFileItemID
               The identifier assigned to this object by the
                MC_Open_Message, MC_Open_Empty_Message,
                MC_Create_File, MC_Create_Empty_File, or
                MC_Open_Item functions.
 Tag
               DICOM tag which identifies the attribute.
 DataType
               One of the enumerated codes identifying the data type of the
               value in Value. The MC_DT enumerated type is defined in
                "mc3msg.h". They are:
```

	String_Type	Null-terminated character string
	Int_Type	Binary integer number
	Uint_Type	Binary unsigned integer number
	ShortInt_Type	Binary short integer number
	UshortInt_Type	Binary unsigned short integer
	number	
	LongInt_Type	Binary long integer number
	UlongInt_Type	Binary unsigned long integer number
	LongLong_Type	Binary 64-bit integer number
	UlongLong_Type	Binary 64-bit unsigned integer
	number	
	Float_Type	Binary Floating point number
	Double_Type	Binary Floating point number
	Buffer_Type	Binary byte value
Value	The value to be used a	s the attribute's value.
ValueLength	Length of provided <i>Va</i>	lue.

Remarks

These functions store a new value for the attribute with the given *Tag.* If one or more values already exist for the attribute, they are first removed. Use **MC_Set_Next_Value** to append attribute values. If **MC_Set_Value** is used, the data type of the *Value* is specified by the *DataType* parameter. The other function names imply the data type. For example, **MC_Set_Value_From_Int** is the same as calling **MC_Set_Value** with *DataType* specified as **Int_Type**. Each function will set the value at *Value*, which must be prototyped as the appropriate type. Note that string values are NULL-terminated character arrays.

For **MC_Set_Value_From_UnicodeString**, the *ValueLength* parameter is for specifying the number of Unicode characters in the input *Value*. This function requires **MC_Enable_Unicode_Conversion** being called first. See documentation for **MC_Enable_Unicode_Conversion**.

MC_Set_Value_From_Buffer is intended mainly for setting the value of an attribute whose value representation is unknown. This may occur if **MC_Message_To_Stream** is to be used and one or more of the stream attributes is unknown. The attribute's value is simply copied (memcpy) from the *Value* buffer according to the length provided in *ValueLength*. The function supports other value representations as well – see table below.

Any reasonable conversion will be made from *Value's* data type to the attribute's value representation. However, some conversions are illogical (see the table below). An appropriate error code will be returned if the conversion attempt failed.

> **NOTE:** for attributes with a value representation of **SQ** (sequence of items) *Value* must represent an integer which is the *ltem/D* of an item object opened previously using **MC_Open_Item**.

© Copyright Merge Healthcare Solutions Inc. 2023

Merge DICOM Toolkit[™] 5.16.0 C/C++ REFERENCE MANUAL

_

Function	May be used to set values to attributes with these Value Representations
MC_Set_Value_From_Float	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_Double	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_ShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_UShortInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_Int	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_UInt	DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_LongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_ULongInt	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_LongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_ULongLong	AT, DS, FD, FL, IS, SL, SS, SV, UL, US, UV, SQ
MC_Set_Value_From_String	AE, AS, AT, CS, DA, DS, DT, FD, FL, IS, LO, LT, PN, SH, SL, SS, SV, ST, TM, UC, UI, UL, UR, US, UV, UT, SQ
MC_Set_Value_From_UnicodeString	LO, LT, PN, SH, ST, UC, UT
MC_Set_Value_From_Buffer	UNKNOWN_VR, OB, OW, OV, OL, OD, OF, UR, UT
MC_Set_Value_From_Function	UNKNOWN_VR, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR, UT

The same rules apply to the **MC_Set_Value** function, based on the value used in the *DataType* parameter. The **MC_Set_Value_From_Function** call is described in its own section.

Return Value

Value	Meaning
INFORMATIONAL MESSAGES:	
MC_NORMAL_COMPLETION	The function completed normally.
WARNINGS: (VALUE WAS STILL ENCODED)	
MC_INVALID_CHARS_IN_VALU	JE A string-type <i>Value</i> had invalid characters for the value representation of <i>Tag</i> .

MC_INVALID_VALUE_FOR_VR	<i>Value</i> does not conform to the requirements of its value representation.
DR MESSAGES:	
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> The attribute can be added with MC_Add_Standard_Attribute . Note: This status will not be returned if setting the value of a
	message object which was opened using MC_Open_Empty_Message or MC_Create_Empty_File. In that case, the attribute is automatically added to the object before setting the value.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object II or item object ID.
MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table below.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated I the attribute. (E.g. setting an attribute with an unsigned int V to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a tempora file (if so configured). If a file I/O error occurs, this status is returned.
MC_INVALID_DATA_TYPE	<i>DataType</i> is not valid.
MC_TOO_MANY_VALUES	This attribute contains the most values that a standard attribute can contain (65535).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected i the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CANNOT_COMPLY	The function fails to process the request. Check Merge DICOM Toolkit log for detail.
See Also	
MC_Set_Value_From_Function MC_Set_Value_To_Empty MC_Set_pValue_From_Function MC_Set_pValue_To_Empty Functions	MC_Set_Value_To_NULL MC_Set_pValue Functions MC_Set_pValue_To_NULL MC_Set_Next_Value Functions MC_Set_Next_pValue MC_Enable_Unicode_Conversion

MC_Set_Value_From_Function

Sets the value of a message object attribute which has the binary value representation of OB, OW, OV, OL, OD or OF, or the numeric value representation of AT, SS, US, SL, UL, SV, UV, FL or FD, or the text value representation of UR or UT.

Synopsis

```
#include ``mc3msg.h"
```

```
MC_STATUS MC_Set_Value_From_Function (
    int MsgFileItemID,
    unsigned long Tag,
    void *UserInfo,
    MC_STATUS (*YourSetFunction)()
)
MsgFileItemID The identifier assigned to this object by the
    MC_Open_Message, MC_Open_Empty_Message,
    MC_Create_File, MC_Create_Empty_File, or
```

MC_Open_Item functions.

Tag	DICOM tag which identifies the attribute.
-----	---

	NULL.
	YourSetFunction each time it is called. This may be
UserInfo	Address of data which will be passed on to

YourSetFunction Name of a function which will be called repeatedly to request blocks of data for the attribute's value.

The function must be prototyped as follows:

```
MC_STATUS YourSetFunction (
    int CBMsgFileItemID,
    unsigned long Cbtag,
    int CbisFirst,
    void *CbuserInfo,
    int *CbdataSizePtr,
    void **CbdataBufferPtr,
    int *CbisLastPtr
```

)

CBMsgFileItemID	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
Cbtag	DICOM tag which identifies the attribute.
CbisFirst	This is TRUE (non-zero) the first time Merge DICOM Toolkit calls <i>YourSetFunction</i> to request data blocks.
CbuserInfo	Address of data which is being passed from the MC_Set_Value_From_Function function. This may be NULL.

CbdataSizePtr	Set <i>*CbdataSizePtr</i> to the number of bytes you are providing.
CbdataBufferPtr	Set <i>* CbdataBufferPtr</i> to the address of the data you are providing.
CbisLastPtr	Set <i>*CbisLastPtr</i> to TRUE (not zero) when you are returning with the last block of OB, OW, OV, OD or OF data.

Remarks

The **MC_Set_Value_From_Function** function is used to set the value of an attribute which has a value representation of UN, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR or UT. Such attributes tend to have values of great length. To accommodate this, one uses the **MC_Set_Value_From_Function** function to specify the name of a function (*YourSetFunction*) which Merge DICOM Toolkit, in turn, calls. This "callback" function is called repeatedly to request blocks of the attribute's data value.

An optional *UserInfo* parameter may be used to pass information between the **MC_Set_Value_From_Function** caller and *YourSetFunction* which receives the data in its *CbuserInfo* parameter. If no such exchange of data is necessary, you may use NULL for *UserInfo*.

YourSetFunction

YourSetFunction must set **CbdataBufferPtr* to the address of a block of data which is a portion of the attribute's value. It also must set **CbdataSizePtr* to the number of bytes in the block.

Merge DICOM Toolkit sets *CbisFirst* to TRUE (non-zero) if it is requesting the first block of the attribute's value. *YourSetFunction* must set **CbisLastPtr* to TRUE (non-zero) if it is supplying the last block of the attribute's value.

YourSetFunction must return **MC_NORMAL_COMPLETION** if all went well. If not, it must return **MC_CANNOT_COMPLY**. Both of these are defined in "mc3msg.h".

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag.</i> The attribute can be added with MC_Add_Standard_Attribute . Note: This status will not be returned if setting the value of a message object which was opened using MC_Open_Empty_Message or MC_Create_Empty_File . In that case, the attribute is automatically added to the object before setting the value.

MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INCOMPATIBLE_VR	The attribute's value representation was not one of UN, OB, OW, OV, OL, OD, OF, AT, SS, US, SL, UL, SV, UV, FL, FD, UR or UT.
MC_CALLBACK_DATA_SIZE_UNEVE	N
	The value set by <i>YourSetFunction</i> in its * <i>CbdataSizePtr</i> parameter was not an even number.
MC_CALLBACK_PARM_ERROR	The value set by <i>YourSetFunction</i> in its * <i>CbdataSizePtr</i> parameter was an odd number
MC_CALLBACK_CANNOT_COMPLY	
	YourSetFunction returned with MC_CANNOT_COMPLY.
See Also	
MC_Set_Value Functions	MC_Set_Next_Value Functions

MC_Set_Value_Representation

Sets the value representation code of an attribute.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Set_Value_Representation (
    int MsgFileItemID,
    unsigned long Tag,
    MC VR ValueRep
```

```
)
```

MsgFilelteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.
Tag	A tag identifying the attribute.
ValueRep	A valid value representation code as defined by the MC_VR type in "mc3msg.h":
	AE, AS, CS, DA, DS, DT, IS, LO, LT, PN, SH, ST, TM, UC, UR, UT, UI, SS, US, AT, SL, UL, SV, UV, FL, FD, OB, OW, OV, OL, OD, OF, SQ

Remarks

MC_Set_Value_Representation is used to change an attribute's value representation.

A message attribute may have an unknown value representation if it is obtained from a message stream (**MC_Stream_To_Message**) and the attribute is unknown. This method can be used to facilitate reading of the values for such attributes.

This function can also be used to manage the attributes that have more than one possible VR.

[©] Copyright Merge Healthcare Solutions Inc. 2023

The valid VR conversions are:

Original VR		New VR
US	US	SS
		WO
S	S	US
S	S	WO
0	W	US
OW		SS
UV	SV	OV
		OV
OV		UV
OV		SV
OW		OB
OB		WO
OV		OB
OB		OV
UN		Any VR

Any other conversion will result in an error return status.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INVALID_VR_CODE	<i>ValueRep</i> is invalid.
MC_ATTRIBUTE_HAS_VALUES	When <i>ValueRep</i> is SQ, the attribute may not already have a value.
MC_VR_ALREADY_VALID	The attribute already has a valid Value Representation.
See Also	
MC_Set_pValue_Representation	

_ _ _ .

MC_Set_Value_To_Empty

Removes all values from an attribute in a message object.

Synopsis

#include ``mc3msg.h"

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_STATUS MC_Se	et_Value_To_Empty (
int MsgFi	ileItemID,	
unsigned long Tag		
)		
Msgltemlteml D	The identifier assigned to this object by the MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or MC_Open_Item functions.	
Tag	DICOM tag which identifies the attribute.	

Remarks

MC_Set_Value_To_Empty removes any existing values for the attribute identified by *Tag* in the message identified by *MsgFileItemID*. If a callback function was registered for the attribute, it is "deregistered."

NOTE: This is not the same as setting a NULL value. Use **MC_Set_Value_To_NULL** to give the attribute a value length of zero (i.e. a NULL value).

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag</i> .
MC_INVALID_MESSAGE_ID	<i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
See Also	
MC_Set_pValue_To_Empty MC_Set_pValue_To_NULL	MC_Set_Value_To_NULL

MC_Set_Value_To_NULL

Sets the value of an attribute in a message object to NULL.

Synopsis

<pre>#include ``mc3</pre>	msg.h″	
<pre>MC_STATUS MC_Set_Value_To_NULL (int MsgFileItemID, unsigned long Tag</pre>		
)		
MsgFileItemI	The identifier assigned to this object by the	
D	MC_Open_Message, MC_Open_Empty_Message, MC_Create_File, MC_Create_Empty_File, or	
	MC_Open_Item functions.	
Tar		
Tag	DICOM tag which identifies the attribute.	

Remarks

MC_Set_Value_To_NULL sets the value of the attribute identified by *Tag* in the message identified by *MsgFileItem/D* to NULL. That is, it sets the attribute's value length to zero. NULL is a valid value in DICOM. Use **MC_Set_Value_To_Empty** if the intent is to remove the attribute's value altogether.

If a callback function was registered for the attribute, it is "de-registered."

NOTE:	If the attribute has a value representation of SQ (sequence of items),	
	setting the value to null does NOT free the item object identified by the	
	value.	

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_TAG	The message does not contain an attribute with an ID of <i>Tag.</i> The attribute can be added with MC_Add_Standard_Attribute . Note: This status will not be returned if setting the value of a message object which was opened using MC_Open_Empty_Message or MC_Create_Empty_File . In that case, the attribute is automatically added to the object
MC_INVALID_MESSAGE_ID	before setting the value to NULL. <i>MsgFileItemID</i> is not a valid message object ID, file object ID or item object ID.
See Also	
MC_Set_Value_To_Empty MC_Set_pValue_To_NULL	MC_Set_pValue_To_Empty

Return Value

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_SR_Add_Child

Adds a lower level SR node to a specified SR tree management.

Synopsis

<pre>#include ``mc3media.h"</pre>	
MC_STATUS MC_SR_Add_Child (int SRID, int ItemID	
)	
SRID	The identifier assigned to the given SR object to which the child is to be added.
ItemID	The identifier assigned to the child SR object which is to be added to the parent.

Remarks

MC_SR_Add_Child adds a new lower level SR tree management node referenced by the item object SRID, and places it in the SR tree management object identified by ItemID.

The child SR tree management record node is placed at the end of SRID's SR record list and all internal links to the new entity are adjusted by Merge DICOM Toolkit.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_INVALID_ITEM_ID	The <i>ItemID</i> value is not a valid SR tree management record object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child MC_SR_Get_Next_Child MC_Delete_Child	MC_SR_Get_First_Child MC_SR_Get_Root

MC_SR_Add_Root

Adds the root node to the SR tree management structure.

Synopsis

#include ``mc3media.h"

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_STATUS	MC_SR_Add_Root (
int	MsgID,
int	*SRID
)	
MsgID	The identifier assigned to the message.
SRID	Upon successful completion, the root SR item object identifier is returned here.

Remarks

MC_SR_Add_Root adds an node as the root node of an SR tree management object. A valid identifier of a message object must be passed into this function. This returned item identifier can then be used to set and retrieve SR values.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	The <i>SRID</i> parameter was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child MC_SR_Get_Next_Child MC_SR_Delete_Child	MC_SR_Get_First_Child MC_SR_Get_Root

MC_SR_Delete_Child

Deletes a given SR child node.

Synopsis

SRID The identifier of the SR item which is to be deleted is passed into this function here.

Remarks

MC_SR_Delete_Child removes the specified child SR object and all of its associated child objects from the SR tree management structures. It also frees the system resources used by the specified child object and all of its child SR object(s).

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_SR_Delete_Child updates all of the location values for affected SR objects within the SR tree.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child MC_SR_Get_Next_Child MC_SR_Delete_Child	MC_SR_Get_First_Child MC_SR_Get_Root

MC_SR_Get_Child_Count

Returns the number of children of an SR tree management node.

Synopsis

#include ``mc3media.h"

```
MC_STATUS MC_SR_First_Child (
```

int SRID,
int *Count

)

SRID	The identifier assigned to this SR tree management node.
Count	Upon successful completion, this parameter returns the number of children for <i>SRID</i>

Remarks

MC_SR_Get_Child_Count returns the number of children of the SR tree management item ItemID.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	The <i>Count</i> parameter was NULL.
MC_UNABLE_TO_GET_SR_ID	An internal error was encountered, Merge DICOM Toolkit was unable to retrieve the ID of an SR object.

See Also

MC_SR_Add_Child MC_SR_Get_Next_Child MC_SR_Delete_Child MC_SR_Get_First_Child MC_SR_Get_Root

MC_SR_Get_First_Child

Returns a pointer to the first child node at a given parent node of the SR tree management structure.

Synopsis			
<pre>#include ``mc3</pre>	<pre>#include "mc3media.h"</pre>		
MC_STATUS MC_ int SR int *Is int *Is	temID,		
)			
SRID	The identifier assigned to this SR tree management node.		
ltemID	Upon successful completion, the item object identifier of the first SR tree management child record of the given <i>SRID</i> object is returned here.		
lsLast	Upon successful completion, this parameter is set to true (non-zero) if the first record is also the last record in the SR tree management entity (i.e., it is the only element)		

Remarks

MC_SR_Get_First_Child retrieves the identifier of the first SR tree management child record of the SR tree management item *ItemID*. If *ItemID* is an empty root SR tree management entity, **IsLast* is set to non-zero, and *ItemID* is set to -1.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_INVALID_SR_ITEM_ID	The <i>ItemID</i> value is not a valid SR tree management record object ID.
MC_NULL_POINTER_PARM	The ItemID, or IsLast parameter was NULL.
MC_MSGFILE_ERROR	An error occurred attempting to access the configuration data file for the directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_SR_Add_Child MC_SR_Get_Next_Child MC_SR_Delete_Child MC_SR_Get_First_Child MC_SR_Get_Root

MC_SR_Get_Location

Returns the location within the current SR tree management structure.

Synopsis			
<pre>#include "mc3m</pre>	<pre>#include ``mc3media.h"</pre>		
<pre>MC_STATUS MC_SR_Get_Location (int SRID, int sizeOfLocation, char *Location</pre>			
)			
SRID	The identifier assigned to this SR tree management object.		
sizeOfLocation	The size of the character array that the toolkit is to place the value for <i>Location</i> into.		
Location	String name of the location is returned here upon successful completion.		

Remarks

MC_SR_Get_Location is used to obtain a string value which represents the location of a given item within the SR tree management structure. The string that is returned will contain a dot-separated list of numbers. These numbers represent the numerical "position" of the item within the tree structure. For example, if the number returned by this function is "1.2.1.3", then the item referenced by *SRID* is the third item, under the first item, under the second item, under the root entity.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	The Location parameter was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child MC_SR_Next_Child MC_Delete_Child	MC_SR_First_Child MC_SR_Root

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_SR_Get_Next_Child

Retrieves a pointer to the next child SR tree management node linked to a specified SR tree management record

Synopsis

```
#include ``mc3media.h"
MC STATUS MC SR_Get_Next_Child (
       int SRID,
       int *NextChildID,
       int *IsLast
)
 SRID
                  The identifier assigned to this SR tree management object.
 NextChildID
                  Upon successful completion, the item object identifier of the
                  next SR tree management node in the SR tree management
                  entity NextChildID is returned here. At a given parent SRID a
                  pointer to the next record is maintained.
 lsLast
                  Upon successful completion, this parameter is set to true
                  (non-zero) if the next record in the SR tree management
                  entity NextChildID is also the last record.
```

Remarks

MC_SR_Get_Next_Child returns *NextChildID*, the entity object identifier of the next child SR tree management node linked to *SRID*. The value *IsLast* is set if the given *NextChildID* is the last in the SR tree management structure.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	The NextChildID or IsLast parameter was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child MC_SR_Get_Next_Child MC_Delete_Child	MC_SR_Get_First_Child MC_SR_Get_Root

MC_SR_Get_Root

Returns a pointer to the root node of an SR tree management structure.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Synopsis #include "mc3media.h" MC_STATUS MC_SR_Root (int MsgID, int *SRID) MsgID The identifier assigned to this SR tree management object by the MC_Message_To_SR function. SRID Upon successful completion, the root SR item object

SRID	Upon successful completion, the root SR item object
	identifier is returned here.

Remarks

MC_SR_Get_Root returns an identifier to the root node of an SR tree management object. A valid identifier of a message containing an SR tree object must be passed into this function. This item identifier can then be used to set and retrieve SR values.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>MsgID</i> value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	The <i>SRID</i> parameter was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child MC_SR_Get_Next_Child MC_SR_Delete_Child	MC_SR_Get_First_Child MC_SR_Get_Root

MC_SR_To_Message

Converts a toolkit-managed SR tree object into a toolkit-managed message object.

Remarks

MC_SR_To_Message converts an SR tree management object into a normal toolkit message object by deleting the SR tree management structures from the toolkit's memory

Since the SR API calls are used to alter a toolkit DICOM message, this function is used to remove the SR tree management structures from the toolkit's memory. This call can be used as a last step, after modifications are completed to the normal toolkit message.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_SR_Add_Child MC_SR_Get_Next_Child MC_SR_Delete_Child	MC_SR_Get_First_Child MC_SR_Get_Root

MC_Stream_To_Message MC_Stream_To_Message_With_Offset

Request that the values of a message object be retrieved from a DICOM stream.

Synopsis

```
#include ``mc3msg.h"
MC STATUS MC Stream To Message (
      int MessageID,
      unsigned long StartTag,
      unsigned long StopTag,
      TRANSFER SYNTAX SyntaxType,
      unsigned long *ErrorTagPtr,
      void *UserInfo,
      MC STATUS (*YourStreamToFunction) ()
)
MC STATUS MC Stream To Message With Offset (
      int MessageID,
      unsigned long StartTag,
      unsigned long StopTag,
      TRANSFER SYNTAX SyntaxType,
      unsigned long *ErrorTagPtr,
      unsigned long Offset,
      void *UserInfo,
```

MC_STATUS	(*YourStreamToFunction)()
MessageID	The identifier assigned to this object by the MC_Open_Message , MC_Open_Empty_Message , or the MC_Open_Item functions.
StartTag	The DICOM tag identifying the first attribute which should be added to the message.
StopTag	The DICOM tag identifying the last attribute which should be added to the message.
SyntaxType	Specify which DICOM transfer syntax was used to encode the stream data you will be providing. Use one of the enumerated TRANSFER_SYNTAX types defined in "mc3msg.h":
	IMPLICIT_LITTLE_ENDIAN IMPLICIT_BIG_ENDIAN EXPLICIT_LITTLE_ENDIAN
	EXPLICIT_BIG_ENDIAN DEFLATED_EXPLICIT_LITTLE_ENDIAN ENCAPSULATED_UNCOMPRESSED_ELE RLE
	JPEG_BASELINE JPEG_EXTENDED_2_4 JPEG_EXTENDED_3_5
	JPEG_SPEC_NON_HIER_6_8 JPEG_SPEC_NON_HIER_7_9 JPEG_FULL_PROG_NON_HIER_10_12
	JPEG_FULL_PROG_NON_HIER_11_13 JPEG_LOSSLESS_NON_HIER_14 JPEG_LOSSLESS_NON_HIER_15
	JPEG_EXTENDED_HIER_16_18 JPEG_EXTENDED_HIER_17_19
	JPEG_SPEC_HIER_20_22 JPEG_SPEC_HIER_21_23 JPEG_FULL_PROG_HIER_24_26
	JPEG_FULL_PROG_HIER_25_27 JPEG_LOSSLESS_HIER_28 JPEG_LOSSLESS_HIER_29
	JPEG_LOSSLESS_HIER_14 JPEG_2000_LOSSLESS_ONLY JPEG_2000
	JPEG_LS_LOSSLESS JPEG_LS_LOSSY
	HEVC_H265_M10P_LEVEL_5_1 HEVC_H265_MP_LEVEL_5_1 JPIP_REFERENCED

	JPIP_REFERENCED_DEFLATE
	MPEG2_MPML
	MPEG2_MPHL
	MPEG4_AVC_H264_HP_LEVEL_4_1
	MPEG4_AVC_H264_BDC_HP_LEVEL_4_1
	MPEG4_AVC_H264_HP_LEVEL_4_2_2D
	MPEG4_AVC_H264_HP_LEVEL_4_2_3D
	MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2
	JPEG_2000_MC_LOSSLESS_ONLY
	JPEG_2000_MC
	SMPTE_ST_2110_20_UNCOMPRESSED_PROGRESS
	IVE_ACTIVE_VIDEO
	SMPTE_ST_2110_20_UNCOMPRESSED_INTERLACE
	D_ACTIVE_VIDEO
	SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO
	PRIVATE_SYNTAX_1
	PRIVATE_SYNTAX_2
ErrorTagPtr	If an error occurs, <i>*ErrorTagPtr</i> will be set to the tag which caused the error <i>.</i>
Offset	The offset from the beginning of the stream to the attribute after <i>StopTag</i> .
UserInfo	Address of data which will be passed on to
	<i>YourStreamToFunction</i> each time it is called. This may be NULL.
YourStreamToFunctio	Name of a function which will be called repeatedly to
n	request blocks of streamed DICOM message data.
11	request blocks of streamed bloom message data.

The function must be prototyped as follows:

```
MC STATUS YourStreamToFunction (
       int CbmessageID,
       void *CbuserInfo,
       int CBFirstCall,
       int *CbdataSizePtr,
       void **CbdataBuffer,
       int *CbisLastPtr
)
 CbmessageID The identifier assigned to the message object by the
                MC_Open_Message function.
                Address of data which is being passed from the
 CbuserInfo
                MC_Stream_To_Message function. This may be NULL.
 CBFirstCall
                This is TRUE (non-zero) the first time Merge DICOM Toolkit
                calls YourStreamToFunction to request data blocks.
 CbdataSizePt
                Set *CbdataSizePtr to the number of bytes you are providing.
                This must be an even number.
 r
```

CbdataBuffer	Set <i>*CbdataBuffer</i> to the address of the data you are providing.
CbisLastPtr	Set <i>*CbisLastPtr</i> to TRUE (not zero) when you are returning with the last block of stream data.

Remarks

MC_Stream_To_Message requests that the contents of a "streamed message" (i.e. a message in the form defined by the DICOM standard) be converted and placed in the message object or item object identified by *MessageID*. The streamed message is passed to Merge DICOM Toolkit by *YourStreamToFunction*. Merge DICOM Toolkit repeatedly calls *YourStreamToFunction* until all of the streamed message has been processed.

MC_Stream_To_Message can pass data to *YourStreamToFunction* by specifying the data's address in *UserInfo*. Merge DICOM Toolkit passes the address to *YourStreamToFunction* in *CbuserInfo* each time it is called. *UserInfo* may be NULL.

If an error occurs while processing the stream, Merge DICOM Toolkit will put the tag of the attribute in error in **ErrorTagPtr*.

MC_Stream_To_Message_With_Offset is identical to **MC_Stream_To_Message** except that it returns the byte offset from the beginning of the stream to the next attribute after *StopTag*. If there is not a tag after *StopTag*, the length of the file will be returned.

StartTag and *StopTag* specify which attributes in the stream are to be placed in the message object. Any attributes in the stream with tags less than *StartTag* or greater than *StopTag* will be ignored. Neither *StartTag* nor *StopTag* need be in the stream.

SyntaxType must be set to **one of the values listed above.** The transfer syntax specifies the byte order used in the streamed message, whether or not each attribute's value representation is explicitly encoded in the stream, and how the pixel data is encoded in the message.

If the stream contains any attributes with a value representation of SQ (i.e. the stream contains one or more sequence of items), an item object is automatically opened for each item in the stream. The *Item/D* associated with each opened item object is used as the value for each item in the sequence attribute. Later, the **MC_Get_Value...** functions may be used to retrieve the *Item/D*'s from the SQ attribute. Then, again using the **MC_Get_Value...** functions, the attributes of the *Item/D* object may be retrieved.

NOTE: A runtime configuration parameter determines what happens if the input stream contains an attribute which is not in the message. The default is to ignore such attributes (with a warning message logged). If requested, however, such attributes and their values are added to the message. If the Value Representation of the attribute being added cannot be determined, the attribute is given a pseudo Value Representation of "Unknown_VR". Only the MC_Get_Value_To_Buffer function retrieves the value of such attributes. To change the value of such attributes, MC_Set_Value_Representation or MC_Set_PValue_Representation must first be called to assign a valid Value Representation to the attribute. If MC_Message_To_Stream is used, attributes with unknown VRs are simply copied (memcpy) to the stream with no consideration given to byte ordering.

YourStreamToFunction

It is the responsibility of *YourStreamToFunction* to pass blocks of data back to Merge DICOM Toolkit each time it is called. As a convenience, Merge DICOM Toolkit sets *CBFirstCall* to TRUE (non-zero) the first time it calls *YourStreamToFunction* for this attribute.

**CbdataBufferPtr* must be set to the address of the buffer containing the stream data block, and **CbdataSizePtr* must be set to the number of bytes placed at **CbdataBufferPtr.*

YourStreamToFunction must set **CbisLast* to TRUE (non-zero) when it is providing the last block of the streamed message.

Return Value

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_NULL_POINTER_PARM	YourStreamToFunction or ErrorTagPtr was NULL.	
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message object ID.	
MC_INVALID_TRANSFER_SYNTAX	An invalid code was used for the <i>SyntaxType</i> parameter.	
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.	
MC_CALLBACK_CANNOT_COMPLY	<i>YourStreamToFunction</i> returned with a status other than MC_NORMAL_COMPLETION.	
MC_CALLBACK_DATA_SIZE_UNEVEN		
	The <i>CbdataSizePtr</i> parameter returned by the <i>YourStreamToFunction</i> was an uneven number.	
MC_CALLBACK_PARM_ERROR	A callback function registered by your application returned an empty (NULL) data buffer when the buffers length was specified as being non-zero. See MC_Register_Callback_Function .	

MC_OUT_OF_ORDER_TAG	A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TAG	The message contains an invalid tag. The tag is placed at <i>*ErrorTagPtr.</i>
MC_VALUE_TOO_LARGE	An attribute in the stream message (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation.
MC_UNEXPECTED_EOD	<i>YourStreamToFunction</i> stopped without passing the entire value for an attribute.
MC_INVALID_LENGTH_FOR_VR	The value(s) for one of the stream attributes was not legal for its value representation.
Any of the status codes returned by MC_Set_Value may also be returned.	

See Also

MC_Register_Callback_Function MC_Message_To_Stream

MC_Thread_Release

Releases thread specific resources used by the Merge DICOM Toolkit library.

Synopsis

```
#include ``mc3msg.h"
```

Remarks

MC_Thread_Release releases resources allocated for the calling thread.

MC_Thread_Release should be called before the end of each thread that uses either MC_Standard_Compressor or MC_Standard_Decompressor to avoid memory leaks in Pegasus libraries from Accusoft.

This call has no effect on platforms that do not support threads.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_LIBRARY_NOT_INITIALIZED	The library has not been initialized by a call to MC_Library_Initialization .

See Also

MC_Library_Initialization

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_Unicode_Get_Substitution_Characters

A utility function to retrieve the replacement characters for a DICOM character set encoder.

Synopsis

Remarks

MC_Unicode_Get_Substitution_Characters requires MC_Enable_Unicode_Conversion being called first.

Charset should be set to a value corresponding to one of the DICOM defined terms for the Specific Character Set (0008,0005).

When returns status is MC_NORMAL_COMPLETION, *SubChars* contains the substitution characters as a byte array. *Length* contains how many bytes are in the *SubChars* buffer.

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_CANNOT_COMPLY	Fail to process the input. Check the Merge DICOM Toolkit log file for detail of failure.
MC_BUFFER_TOO_SMALL	The output buffer doesn't have enough space to hold the output.
MC_NULL_POINTER_PARM	Any of the arguments is null.An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Byte_To_Unicode MC_Enable_Unicode_Conversion MC_Get_Value_To_UnicodeString MC_Get_Next_Value_To_UnicodeString MC_Set_Value_From_UnicodeString MC_Set_Next_Value_From_UnicodeString MC_Unicode_To_Byte

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

MC_Unicode_To_Byte

A utility function to convert Unicode to DICOM character set.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Unicode_To_Byte (
       char * * Charsets,
       int NumCharsets,
       const MC UChar *Val,
       int InLen,
       int BufferSize,
       int *OutLen,
       char *OutVal
)
 Charsets
                An array of character strings from the DICOM specific
                character set attribute.
 NumCharsets
               Number of character set strings in the Charsets array
 Val
                Input Unicode array
 InLen
                Input Unicode character count
 BufferSize
                Output buffer size
 OutLen
                Output length (returned by this call)
 OutVal
                Output buffer to hold the byte
```

Remarks

MC_Unicode_To_Byte requires MC_Enable_Unicode_Conversion being called first.

Charsets should be set to the value of (0008,0005) as an array of character strings. If NumCharsets is 0, *Charsets* is NULL or an empty string, the default ISO_IR 6 (ASCII) character set will be used.

InLen can be set to -1 and the toolkit will calculate length if the input Unicode buffer is U+0000 terminated.

When the return status is MC_NORMAL_COMPLETION, *OutVal* contains the output multi-byte characters with a NULL terminator. *OutLen* contains the number of bytes present in the *OutVal* buffer (excluding the terminator).

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_CANNOT_COMPLY	Fail to process the input. Check the Merge DICOM Toolkit log file for details of failure.
MC_BUFFER_TOO_SMALL	The output buffer doesn't have enough space to hold the output.

MC_SYSTEM_ERROR

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

See Also

MC_Byte_To_Unicode MC_Enable_Unicode_Conversion MC_Get_Value_To_UnicodeString MC_Get_Next_Value_To_UnicodeString MC_Set_Value_From_UnicodeString MC_Set_Next_Value_From_UnicodeString

MC_Validate_Attribute

Ensures that an attribute meets DICOM validation criteria.

Synopsis

```
#include ``mc3msg.h"
MC_STATUS MC_Validate_Attribute (
       int MessageFileID,
      unsigned long Tag
      VAL ERR **ErrorInfo,
      VAL LEVEL ErrorLevel
)
 MessageFileID
                  The identifier assigned to this object by the
                   MC_Open_Message, MC_Open_Empty_Message,
                   MC_Create_Empty_File or MC_Create_File function.
 Tag
                  Tag value of the attribute to validate.
 ErrorInfo
                  The address of a validation error block describing the first
                  violation encountered is returned here if a validation
                  violation occurs. Otherwise NULL is returned here. The
                   block is the VAL_ERR type defined in "mc3msg.h":
typedef struct ValErr_struct
      unsigned long Tag; /* Tag of attribute with validation
                              violation */
      int MsgItemID;
                           /* ID of message or item object
                              containing the attribute */
       int ValueNumber;
                          /* Value number involved - zero if no
                              value involved */
      MC STATUS Status; /* Validation violation status code */
                          /* ID of parent of message or item
       int ParentMsgID;
                              object containing the attribute */
       int MsgLevel;
                          /* Nesting level in data set hierarchy
                              of the message or item object
                              containing the attribute; */
```

} VAL_ERR;

 ErrorLevel
 The level of validation checking is specified here. Use one of the enumerated VAL_LEVEL types defined in "mc3msg.h":

 Validation_Level1:
 Report only Errors.

 Validation_Level2:
 Report Errors and Warnings.

 Validation_Level3:
 Report Errors, Warnings and Info messages.

Remarks

The **MC_Validate_Attribute** function validates an attribute contained in the object identified by *MessageFileID* to determine if its values are valid. The attribute is validated to ensure that it meets the requirements of the service and command specified when the object was opened by **MC_Open_Message** or **MC_Create_File**, or when **MC_Set_Service_Command** was called for the object.

Performance Tuning **MC_Validate_Attribute** has a lower overhead than the **MC_Validate_Message** and **MC_Validate_File** functions. When validation is only required for a select group of attributes, this function should be called.

While the validation is quite comprehensive there are limitations in validation:

- If the attribute was specified by DICOM Part 3 as being in either a user optional module, or a conditional module, it will always be treated as a Type 3 attribute (optional attribute) by MC_Validate_Attribute.
- The Ultrasound Image Object used in storage services has mutually exclusive Image, Overlay, and Curve information entities. The attributes defined in these entities are all treated as Type 3 (optional attributes) by MC_Validate_Attribute.

In any case, **MC_Validate_Attribute** is a powerful tool that can catch most all problems with an attribute. The DICOM Standard should always be the final word on whether or not an attribute satisfies the DICOM protocol.

ErrorLevel specifies the type of validation violation which will cause the **MC_Validate_Attribute** function to return a status of **MC_DOES_NOT_VALIDATE**.

NOTE: All validation violations encountered will be logged to the Merge DICOM Toolkit log file, regardless of the level specified in *ErrorLevel*.

A list of validation violations is created and the validation error block describing the first violation is returned at **ErrorInfo.* Subsequent violations may be accessed by using the **MC_Get_Next_Validate_Error** function. If the file validates, the status of **MC_MESSAGE_VALIDATES** is returned and NULL is returned at **ErrorInfo.* Also, the results of the validation are logged into the message log (usually merge.log) at the T5 MESSAGE level.

WARNINGS: Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_MESSAGE_VALIDATES	The attribute contained no validation violations of the type specified in <i>ErrorLevel</i> .
MC_DOES_NOT_VALIDATE	The attribute contained at least one validation violation of the type specified in <i>ErrorLevel</i> .
MC_INVALID_MESSAGE_ID	<i>MessageFileID</i> does not identify a valid message or file object.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_TAG	<i>Tag</i> was not in the message.
MC_MSGFILE_ERROR	An error occurred involving the message info file for the file object. The exact error is logged in the Merge DICOM Toolkit log file.
MC_INVALID_ITEM_ID	The message or file object contains a sequence of items attribute with an invalid item ID value.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

Validation Violations

If a status of **MC_DOES_NOT_VALIDATE** is returned, the status of the first validation violation is returned at *ErrorInfo->Status*. The violation status codes are defined in **"mcstatus.h"**. They are arranged by violation type below:

INFO MESSAGES:

o meddaded.	
MC_NO_CONDITION_FUNCTION	The attribute's DICOM type is "1C" or "2C" and no function was available to check the condition. A user-defined condition function was not correctly linked with the program.
MC_UNABLE_TO_CHECK_CONDITIO	DN
	Not enough information is available to check whether or not a DICOM type "1C" or "2C" attribute is required.
MC_NOT_ONE_OF_DEFINED_TERMS	S A value set for this attribute is not one of the valid defined terms assigned to this attribute.
MC_NON_SERVICE_ATTRIBUTE	The attribute is not one of those defined for the file's service. Note that private attributes will not cause this violation
ERRORS:	
MC_INVALID_VR_CODE	The attribute's value representation is unknown.
MC_REQUIRED_ATTRIBUTE_MISSIN	NG
	An attribute which is required for the message service has been deleted from the file object.
MC_REQUIRED_VALUE_MISSING	The attribute is required to have a value and does not.
MC_VALUE_MAY_NOT_BE_NULL	The attribute is DICOM type "1" or type "1C" and it has been encoded with a NULL value.

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

MC_VALUE_NOT_ALLOWED	This DICOM type "1C" or type "2C" attribute may not have a	
	value under current conditions.	
MC_TOO_FEW_VALUES	The attribute is required to have more values set.	
MC_TOO_MANY_VALUES	The attribute has more values set than are allowed.	
MC_INVALID_ITEM_ID	This sequence of items (SQ) attribute has an invalid value assigned to it.	
MC_NOT_ONE_OF_ENUMERATED_VALUES		
	A value set for this attribute is not one of the valid enumerated values assigned to this attribute.	
MC_INVALID_VALUE_FOR_VR	A value for this attribute does not conform to the requirements of its value representation.	
MC_INVALID_CHARS_IN_VALUE	A value for this attribute does not contain valid characters for its value representation.	
See Also		
MC_Get_Next_Validate_Error	MC_Validate_Message MC_Validate_File	

MC_Validate_File

Ensures that a file meets DICOM validation criteria.

Synopsis

```
#include ``mc3media.h"
```

```
MC_STATUS MC_Validate_File (
int FileID,
VAL_ERR **ErrorInfo,
VAL_LEVEL ErrorLevel
```

```
)
```

```
FileIDThe identifier assigned to this object by theMC_Create_Empty_File or MC_Create_File function.
```

ErrorInfo The address of a validation error block describing the first violation encountered is returned here if a validation violation occurs. Otherwise NULL is returned here. The block is the **VAL_ERR** type defined in "mc3msg.h":

```
typedef struct ValErr_struct
```

```
unsigned long Tag; /* Tag of attribute with validation
                      violation */
                   /* ID of message or item object
int MsgItemID;
                      containing the attribute */
int ValueNumber;
                  /* Value number involved - zero if no
                      value involved */
MC STATUS Status; /* Validation violation status code */
int ParentMsgID;
                  /* ID of parent of message or item
                      object containing the attribute */
                   /* Nesting level in data set hierarchy
int MsgLevel;
                      of the message or item object
                      containing the attribute; */
```

VAL_ERR; ErrorLevel The level of validation checking is specified here. Use one of the enumerated VAL_LEVEL types defined in "mc3msg.h": Validation_Level1: Report only Errors.

Validation_Level2: Report Errors and Warnings. Validation_Level3: Report Errors, Warnings and Info messages.

Remarks

The **MC_Validate_File** function validates the file identified by *FileID* to determine that required attributes are present and that all attribute values are valid. The file is validated to ensure that it meets the requirements of the service and command specified when the message was opened with **MC_Create_File** or when **MC_Set_Service_Command** was called for the file object. If the file was opened with **MC_Create_Empty_File**, the **MC_Set_Service_Command** function must be called to specify the service and command to validate against.

While the validation is quite comprehensive there are limitations in validating files:

- If the attribute was specified by DICOM Part 3 as being in either a user optional module, or a conditional module, it will always be treated as a Type 3 attribute (optional attribute) by **MC_Validate_File**.
- The Ultrasound Image Object used in storage services has mutually exclusive Image, Overlay, and Curve information entities. The attributes defined in these entities are all treated as Type 3 (optional attributes) by **MC_Validate_File**.
- **MC_Validate_File** does not validate properly for application profiles that modify the standard content of a message.

In any case, **MC_Validate_File** is a powerful tool that can catch most all problems with a file. The DICOM Standard should always be the final word on whether or not a file satisfies the DICOM protocol.

ErrorLevel specifies the type of validation violation which will cause the **MC_Validate_File** function to return a status of **MC_DOES_NOT_VALIDATE**.

NOTE: All validation violations encountered will be logged to the Merge DICOM Toolkit log file, regardless of the level specified in *ErrorLevel*.

A list of validation violations is created and the validation error block describing the first violation is returned at **ErrorInfo.* Subsequent violations may be accessed by using the **MC_Get_Next_Validate_Error** function. If the file validates, the status of **MC_MESSAGE_VALIDATES** is returned, and NULL is returned at **ErrorInfo.* Also, the results of the validation are logged into the message log (usually merge.log) at the T5 MESSAGE level.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value

Meaning

[©] Copyright Merge Healthcare Solutions Inc. 2023

The file object contained no validation violations of the type specified in <i>ErrorLevel</i> .
The file object contained at least one validation violation of the type specified in <i>ErrorLevel</i> .
<i>FileID</i> does not identify a valid file object.
One or more of the pointer-type parameters was NULL.
An error occurred involving the message info file for the file object. The exact error is logged in the Merge DICOM Toolkit log file.
The file object contains a sequence of items attribute with an invalid item ID value.
An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

Validation Violations

If a status of **MC_DOES_NOT_VALIDATE** is returned, the status of the first validation violation is returned at *ErrorInfo->Status*. The violation status codes are defined in **"mcstatus.h"**. They are arranged by violation type below:

INFO MESSAGES:	
MC_NO_CONDITION_FUNCTION	The attribute's DICOM type is "1C" or "2C" and no function was available to check the condition. A user-defined condition function was not correctly linked with the program.
MC_UNABLE_TO_CHECK_CONDITION	ON
	Not enough information is available to check whether or not a DICOM type "1C" or "2C" attribute is required.
WARNINGS:	
MC_NOT_ONE_OF_DEFINED_TERM	S
	A value set for this attribute is not one of the valid defined terms assigned to this attribute.
MC_NON_SERVICE_ATTRIBUTE	The attribute is not one of those defined for the file's service. Note that private attributes will not cause this violation
MC_INVALID_VR_CODE	The attribute's value representation is unknown.
MC_REQUIRED_ATTRIBUTE_MISSI	•
	An attribute which is required for the message service has been deleted from the file object.
MC_REQUIRED_VALUE_MISSING	
	The attribute is required to have a value and does not.
MC_VALUE_MAY_NOT_BE_NULL	The attribute is DICOM type "1" or type "1C" and it has been encoded with a NULL value.

MC_VALUE_NOT_ALLOWED	This DICOM type "1C" or type "2C" attribute may not have a value under current conditions.	
MC_TOO_FEW_VALUES	The attribute is required to have more values set.	
MC_TOO_MANY_VALUES	The attribute has more values set than are allowed.	
MC_INVALID_ITEM_ID	This sequence of items (SQ) attribute has an invalid value assigned to it.	
MC_NOT_ONE_OF_ENUMERATED_VALUES		
	A value set for this attribute is not one of the valid enumerated values assigned to this attribute.	
MC_INVALID_VALUE_FOR_VR	A value for this attribute does not conform to the requirements of its value representation.	
MC_INVALID_CHARS_IN_VALUE	A value for this attribute does not contain valid characters for its value representation.	
See Also		
MC_Get_Next_Validate_Error	MC_Validate_Message MC_Validate_Attribute	

MC_Validate_Message

Ensures that a message meets DICOM validation criteria.

Synopsis

#include ``mc3msg.h"

```
MC_STATUS MC_Validate_Message (
    int MessageID,
    VAL_ERR **ErrorInfo,
    VAL_LEVEL ErrorLevel
)
```

MessageID

ErrorInfo

{

The message object's identification number

The address of a validation error block describing the first violation encountered is returned here if a validation violation occurs. Otherwise NULL is returned here. The block is the **VAL_ERR** type defined in "mc3msg.h":

```
typedef struct ValErr_struct
```

```
unsigned long Tag; /* Tag of attribute with validation
violation */
int MsgItemID; /* ID of message or item object
containing the attribute */
int ValueNumber; /* Value number involved - zero if no
value involved */
MC_STATUS Status; /* Validation violation status code */
int ParentMsgID; /* ID of parent of message or item
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

int A	AsgLevel; /*	object containing the attribute */ Nesting level in data set hierarchy of the message or item object containing the attribute; */
} VAL_ERR;		
ErrorLevel		12: Report Errors and Warnings.

Remarks

MC_Validate_Message validates the message identified by *MessageID* to determine that required attributes are present and that all attribute values are valid. The message is validated to ensure that it meets the requirements of the service and command specified when the message was opened by MC_Open_Message. If the message was opened with MC_Open_Empty_Message, the MC_Set_Service_Command function must be called to specify the service and command to validate against.

While the validation is quite comprehensive there are limitations in validating messages opened for composite services:

- If the attribute was specified by DICOM Part 3 as being in either a user optional module, or a conditional module, it will always be treated as a Type 3 attribute (optional attribute) by MC_Validate_Message.
- The Nuclear Medicine and Ultrasound Image Objects used in storage services have mutually exclusive Image, Overlay, and Curve information entities. The attributes defined in these entities are all treated as Type 3 (optional attributes) by MC_Validate_Message.
- For normalized services using the N-EVENT-REPORT command, the actual contents of a message specified in DICOM Part 4 are dependent on the Event Type ID being communicated. All attributes used for all Event Type ID's are treated as Type 3 by **MC_Validate_Message**.
- Unique keys used within C-FIND-RQ messages for each of the Query/Retrieve service classes are
 not validated properly because the differences in what is required for request and response
 messages. The C-FIND-RSP messages are validated properly. An error is also reported when a
 unique key is set to NULL. This is an invalid error for the unique keys at the level that is being
 queried.

In any case, **MC_Validate_Message** is a powerful tool that can catch most all problems with a message. The DICOM Standard should always be the final word on whether or not a message satisfies the DICOM protocol.

ErrorLevel specifies the type of validation violation which will cause the **MC_Validate_Message** function to return a status of **MC_DOES_NOT_VALIDATE**. Note, however, that all validation violations encountered will be logged to the Merge DICOM Toolkit log file, regardless of the level specified in *ErrorLevel*.

[©] Copyright Merge Healthcare Solutions Inc. 2023

A list of validation violations is created and the validation error block describing the first violation is returned at **ErrorInfo*. Subsequent violations may be accessed by using the **MC_Get_Next_Validate_Error** function. If the message validates, the status of **MC_MESSAGE_VALIDATES** is returned and NULL is returned at **ErrorInfo*. Also, the results of the validation are logged into the message log (usually merge.log) at the T5 MESSAGE level.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_MESSAGE_VALIDATES	The message contained no validation violations of the type specified in <i>ErrorLevel</i> .
MC_DOES_NOT_VALIDATE	The message contained at least one validation violation of the type specified in <i>ErrorLevel</i> .
MC_INVALID_MESSAGE_ID	MessageID does not identify a valid message object.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_MSGFILE_ERROR	An error occurred involving the message info file for the message. The exact error is logged in the Merge DICOM Toolkit log file.
MC_INVALID_ITEM_ID	The message contains a sequence of items attribute with an invalid item ID value.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

Validation Violations

If a status of **MC_DOES_NOT_VALIDATE** is returned, the status of the first validation violation is returned at *ErrorInfo->Status*. The violation status codes are defined in **"mcstatus.h"**. They are arranged by violation type below:

INFO MESSAGES:	
MC_NO_CONDITION_FUNCTION	The attribute's DICOM type is "1C" or "2C" and no function was available to check the condition. A user-defined condition function was not correctly linked with the program.
MC_UNABLE_TO_CHECK_CONDIT	ION
	Not enough information is available to check whether or not a DICOM type "1C" or "2C" attribute is required.
WARNINGS:	
MC_NOT_ONE_OF_DEFINED_TERM	/IS
	A value set for this attribute is not one of the valid defined terms assigned to this attribute.
MC_NON_SERVICE_ATTRIBUTE	The attribute is not one of those defined for the message's service. Note that private attributes will not cause this violation
Errors:	
MC_INVALID_VR_CODE	The attribute's value representation is unknown.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_REQUIRED_ATTRIBUTE_MISSIN	NG	
	An attribute which is required for the message service has	
	been deleted from the message object.	
MC_REQUIRED_VALUE_MISSING		
	The attribute is required to have a value and does not.	
MC_VALUE_MAY_NOT_BE_NULL	The attribute is DICOM type "1" or type "1C" and it has been encoded with a NULL value.	
MC_VALUE_NOT_ALLOWED	This DICOM type "1C" or type "2C" attribute may not have a value under current conditions.	
MC_TOO_FEW_VALUES	The attribute is required to have more values set.	
MC_TOO_MANY_VALUES	The attribute has more values set than are allowed.	
MC_INVALID_ITEM_ID	This sequence of items (SQ) attribute has an invalid value assigned to it.	
MC_NOT_ONE_OF_ENUMERATED_VALUES		
	A value set for this attribute is not one of the valid enumerated values assigned to this attribute.	
MC_INVALID_VALUE_FOR_VR	A value for this attribute does not conform to the requirements of its value representation.	
MC_INVALID_CHARS_IN_VALUE	A value for this attribute does not contain valid characters for its value representation.	
See Also		
MC_Get_Next_Validate_Error	MC_Validate_Attribute	

MC_Wait_For_Association MC_Wait_For_Association_On_Port MC_Wait_For_Association_On_Address

Waits for an association request from a remote DICOM application

Synopsis

MC STATUS MC Wait For Association On Address (char *ServiceList, int Timeout, int ApplicationID, int Port, const char *Address, int *AssociationID) ServiceList Name of a service list in the Merge DICOM Toolkit configuration file. Timeout The maximum time (in seconds) to wait for an association request to arrive. A value of zero (0) means "check one time" and a value of minus one (-1) means "wait forever". The identification number of a previously registered **ApplicationID** application is returned here, or the application ID to negotiate is supplied here. The identification number of an association object is AssociationID returned here. Port Port number to listen on. The network name or IP address of the network interface to Address listen on. If this argument is NULL the toolkit will accept connections on all available network interfaces.

Remarks

MC_Wait_For_Association, MC_Wait_For_Association_On_Port and

MC_Wait_For_Association_On_Address are used to wait for a remote DICOM application to make a request for an association with one of the applications previously registered using the
 MC_Register_Application function. MC_Wait_For_Association will negotiate for all applications registered with MC_Register_Application, while MC_Wait_For_Association_On_Port and
 MC_Wait_For_Association_On_Address will only negotiate associations for the specific Application passed to the call.

If *Timeout* is a positive number Merge DICOM Toolkit will wait no longer than *Timeout* seconds for an association request. If *Timeout* is negative, Merge DICOM Toolkit will not return until an association request has arrived or an error occurs. If *Timeout* is zero, Merge DICOM Toolkit will check one time for the arrival of an association request and return. A status of MC_TIMEOUT is returned if no valid association request is received in the specified time period.

ServiceList is the name of a section in the Merge DICOM Toolkit configuration file which specifies which DICOM services the registered applications are willing to provide. If a request for any one of these services is received, this function returns MC_NORMAL_COMPLETION and sets **AssociationID* to the identification number of an association object. *AssociationID* is then used in other association function calls to identify this negotiated connection.

Normally, if an association request is received for an application whose title is not registered or is not specified on a call to **MC_Wait_For_Association_On_Port** or

MC_Wait_For_Association_On_Address, Merge DICOM Toolkit rejects the association request and

continues to wait for a valid association request. Optionally, "ACCEPT_ANY_APPLICATION_TITLE = YES" may be specified in the Merge DICOM Toolkit system profile. In that case, requests for associations with an unregistered application will be given to the first registered application entity title which has been registered with **MC_Register_Application**.

Merge DICOM Toolkit validates each received association request based on the protocol rules of DICOM and on configurable specifications contained in the Merge DICOM Toolkit system profile. An association request is rejected if it does not validate.

Upon successful completion (MC_NORMAL_COMPLETION returned) of **MC_Wait_For_Association**, **ApplicationID* is set to the identification number of a registered application. In all other cases, both **ApplicationID* and **AssociationID* are not valid.

Upon successful completion, the caller may examine any expected extended negotiation information by using the **MC_Get_Negotiation_Info** function. The caller must respond to the association requestor by using either **MC_Accept_Association** or **MC_Reject_Association** before any message exchange can occur over the association.

The listen port can
be changed!The port that MC_Wait_For_Association listens on can be changed while an application is running by
calling MC_Set_Int_Config_Value to set the TCPIP_LISTEN_PORT configuration option. The next call
to MC_Wait_For_Association after this configuration option has changed will stop listening on the
previous listen port and start listening on the newly configured listen port.

Return Value

Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_TIMEOUT	The <i>Timeout</i> period expired.	
MC_NEGOTIATION_ABORTED	An incoming association was aborted during negotiation. Normally this situation is handled by retrying the MC_Wait_For_Association call.	
MC_ASSOCIATION_REJECTED	An incoming association was aborted because no services were acceptable.	
MC_SYSTEM_CALL_INTERRUPTED	The system call used to wait for an association request was interrupted by a signal. Normally this situation is handled by retrying the MC_Wait_For_Association call.	
MC_NULL_POINTER_PARM	<i>AssociationID, ApplicationID</i> or <i>ServiceList</i> parameter was NULL.	
MC_NO_APPLICATIONS_REGISTERED		
	No applications have been registered using MC_Register_Application.	
MC_INVALID_SERVICE_LIST_NAME	ServiceList points at a null string.	
MC_CONFIG_INFO_MISSING	Could not access ServiceList configuration parameters.	
MC_CONFIG_INFO_ERROR	The ServiceList contained too many services.	

MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.	
See Also		
MC_Open_Association	MC_Open_Secure_Association MC_Get_Negotiation_Info MC_Accept_Association	
MC_Reject_Association	MC_Wait_For_Secure_Association	
MC_Close_Listen_Port	MC_Wait_For_Connection	
MC_Close_Listen_Port_On_Address	MC_Wait_For_Connection_On_Port	
MC_Process_Association_Request	MC_Wait_For_Connection_On_Address	
MC_Process_Secure_Association_Request		
MC_Wait_For_Secure_Association_On_Address		

MC_Wait_For_Connection MC_Wait_For_Connection_On_Port MC_Wait_For_Connection_On_Address

Waits for an incoming network connection from a remote DICOM application

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Wait For Connection (
       int Timeout,
      MC SOCKET *Socket
)
MC_STATUS MC_Wait_For_Connection_On_Port (
       int Timeout,
      int Port,
      MC SOCKET *Socket
)
MC_STATUS MC_Wait_For_Connection_On_Address (
       int Timeout,
       int Port,
      const char *Address,
      MC SOCKET *Socket
)
 Timeout
               The maximum time (in seconds) to wait for an association
               request to arrive. A value of zero (0) means "check one time"
               and a value of minus one (-1) means "wait forever".
 Socket
               The socket for which the incoming connection has been
               accepted is returned here.
 Port
               Port number to listen on.
```

Address Address of the network interface to listen on. This can be the IP address or the network name of the interface. If this argument is NULL, the listener will be set up on all available network interfaces.

Remarks

MC_Wait_For_Connection, MC_Wait_For_Connection_On_Port and

MC_Wait_For_Connection_On_Address are used to wait for a remote DICOM application to make a request for an association to this application. These routines work in conjunction with the MC_Process_Association_Request and MC_Process_Secure_Association_Request functions which actually process the incoming association request.

If *Timeout* is a positive number Merge DICOM Toolkit will wait no longer than *Timeout* seconds for an association request. If *Timeout* is negative, Merge DICOM Toolkit will not return until an association request has arrived or an error occurs. If *Timeout* is zero, Merge DICOM Toolkit will check one time for the arrival of an association request and return. A status of MC_TIMEOUT is returned if no valid association request is received in the specified time period.

These functions were introduced to address a problem with the **MC_Wait_For_Association** function where a poorly performing DICOM application connecting to Merge DICOM Toolkit could cause delays in processing DICOM associations. With **MC_Wait_For_Association**, when Merge DICOM Toolkit receives an association, it will not process other new incoming connections until it has read the initial association request from the network. The **MC_Wait_For_Connection**,

MC_Wait_For_Connection_On_Port and **MC_Wait_For_Connection_On_Address** routines allow an application to create a thread or process to handle a new network connection as soon as it is received to eliminate this problem. The routines do not read any data from the network connection after it has been accepted.

After a **MC_Wait_For_Connection** returns, the socket returned must be passed to **MC_Process_Association_Request** or **MC_Process_Secure_Association_Request** for the association to be negotiated.

The listen port can
be changed!The port that MC_Wait_For_Connection listens on can be changed while an application is running by
calling MC_Set_Int_Config_Value to set the TCPIP_LISTEN_PORT configuration option. The next call
to MC_Wait_For_Connection after this configuration option has changed will stop listening on the
previous listen port and start listening on the newly configured listen port.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_TIMEOUT	The <i>Timeout</i> period expired.
MC_SYSTEM_CALL_INTERRUPTED	The system call used to wait for an association request was interrupted by a signal. Normally this situation is handled by retrying the MC_Wait_For_Connection call.

MC_UNKNOWN_HOST_CONNECTED	A host whose remote hostname could be be resolved has attempted to connect and the connection has been droped.
MC_NEGOTIATION_ABORTED	An unexpected network error occurred when attempting to accept the connection.
MC_NULL_POINTER_PARM	The <i>Socket</i> parameter was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
See Also	
MC_Open_Association MC_Wait_For_Secure_Association	MC_Open_Secure_Association MC_Close_Listen_Port

MC_Wait_For_Secure_Association MC_Close_Listen_Port MC_Close_Listen_Port_On_Address MC_Wait_For_Association MC_Wait_For_Association_On_Port MC_Process_Association_Request MC_Wait_For_Association_On_Address MC_Wait_For_Secure_Association_On_Address MC_Process_Secure_Association_Request MC_Release_Parent_Connection

MC_Wait_For_Secure_Association MC_Wait_For_Secure_Association_On_Port MC_Wait_For_Secure_Association_On_Address

Waits for an association request from a remote DICOM application using a secure socket connection.

Synopsis

```
#include ``mergecom.h"
MC STATUS MC Wait For Secure Association (
      char *ServiceList,
      int Timeout,
      int *ApplicationID,
      int *AssociationID,
      SecureSocketFunctions *SS functions,
      void *SS context
)
MC STATUS MC Wait For Secure Association On Port (
      char *ServiceList,
      int Timeout,
      int ApplicationID,
      int Port,
      int *AssociationID,
      SecureSocketFunctions *SS functions,
      void *SS context
)
MC STATUS MC Wait For Secure Association On Address (
      char *ServiceList,
      int Timeout,
      int ApplicationID,
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

Merge DICOM Toolkit™ 5.16.0 C/C++ REFERENCE MANUAL

int Port,		
	har *Address,	
	sociationID,	
	ocketFunctions *SS_functions,	
,	S_context	
)		
ServiceList	Name of a service list in the Merge DICOM Toolkit configuration file.	
Timeout	The maximum time (in seconds) to wait for an association request to arrive. A value of zero (0) means "check one time" and a value of minus one (-1) means "wait forever".	
ApplicationID	The identification number of a previously registered application is returned here, or the application ID to negotiate is supplied here.	
AssociationID	The identification number of an association object is returned here.	
Port	Port number to listen on.	
Address	The network nameor IP address of the network interface to listen on. If this argument is NULL the toolkit will accept connections on all available network interfaces.	
SS_functions	Pointer to a structure containing functions that will be called by Merge DICOM Toolkit while processing network I/O over the secure connection. (see below)	
SS_context	An optional pointer to application-specific data that Merge DICOM Toolkit passes to the functions declared in <i>SS_functions</i> .	

Remarks

Use MC_Wait_For_Secure_Association, MC_Wait_For_Secure_Association_On_Port and MC_Wait_For_Secure_Association_On_Address to wait for a remote DICOM application to make a request for an association with one of the applications previously registered using the MC_Register_Application function. MC_Wait_For_Secure_Association will negotiate for all applications registered with MC_Register_Application, while MC_Wait_For_Secure_Association_On_Port and MC_Wait_For_Secure_Association_On_Address will only negotiate associations for the specific Application passed to the call.

If *Timeout* is a positive number Merge DICOM Toolkit will wait no longer than *Timeout* seconds for an association request to arrive. If *Timeout* is negative, Merge DICOM Toolkit will not return until an association request has arrived or an error occurs. If *Timeout* is zero, Merge DICOM Toolkit will check one time for the arrival of an association request and return. A status of MC_TIMEOUT is returned if no valid association request is received in the specified time period.

ServiceList is the name of a section in the Merge DICOM Toolkit configuration file which specifies which DICOM services the registered applications are willing to provide. If a request for any one of these services is received, this function returns MC_NORMAL_COMPLETION and sets **AssociationID* to the identification number of an association object. *AssociationID* is then used in other association function calls to identify this negotiated connection.

Normally, if an association request is received for an application whose title is not registered or is not specified on a call to **MC_Wait_For_Secure_Association_On_Port** or

MC_Wait_For_Secure_Association_On_Address, Merge DICOM Toolkit rejects the association request and continues to wait for a valid association request. Optionally,

"ACCEPT_ANY_APPLICATION_TITLE = YES" may be specified in the Merge DICOM Toolkit system profile. In that case, requests for associations with an unregistered application will be given to the first registered application entity title which has been registered with **MC_Register_Application**.

Merge DICOM Toolkit validates each received association request based on the protocol rules of DICOM and on configurable specifications contained in the Merge DICOM Toolkit system profile. An association request is rejected if it does not validate.

Upon successful completion (MC_NORMAL_COMPLETION returned) of **MC_Wait_For_Secure_Association**, **ApplicationID* is set to the identification number of a registered application. In all other cases, both **ApplicationID* and **AssociationID* are not valid.

Upon successful completion, the caller may examine any expected extended negotiation information by using the **MC_Get_Negotiation_Info** function. The caller must respond to the association requestor by using either **MC_Accept_Association** or **MC_Reject_Association** before any message exchange can occur over the association.

The listen port can
be changed!The port that MC_Wait_For_Secure_Association listens on can be changed while an application is
running by calling MC_Set_Int_Config_Value to set the TCPIP_LISTEN_PORT configuration option.
The next call to MC_Wait_For_Secure_Association after this configuration option has changed will
stop listening on the previous listen port and start listening on the newly configured listen port.

SecureSocketFunctions

When using the **MC_Wait_For_Secure_Association**, **MC_Wait_For_Secure_Association_On_Port** or **MC_Wait_For_Secure_Association_On_Address** calls, Merge DICOM Toolkit establishes a TCP/IP connection and then calls the functions provided by the *SS_functions* parameter to establish the secure connection and to pass data through the secure connection. The *SS_functions* are responsible for sending and receiving all data through the secure connection, thus allowing them to do so using a secure protocol such as Secure Socket Layer (SSL). Merge DICOM Toolkit closes the underlying TCP/IP connection when all association processing has completed and after it calls the **SS_Session_Shutdown** callback.

The **SecureSocketFunctions** structure is declared in mergecom.h as follows:

typedef struct MC_Secure_Socket_Functions_Struct
{
 SS_STATUS (NOEXP_FUNC *SS_Session_Start) (
 MC_SOCKET SocketToUse,
 CONN_TYPE ConnectionType,
 void *ApplicationContext,

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
**SecurityContext
       void
    );
    SS STATUS (NOEXP FUNC *SS Read) (
                    *SScontext,
       void
                    *ApplicationContext,
       void
                    *Buffer,
       char
       unsigned int BytesToRead,
       unsigned int *BytesRead,
       int
                    Timeout
    );
    SS STATUS (NOEXP FUNC *SS Write) (
                    *SScontext,
       void
                   *ApplicationContext,
       void
                    *Buffer,
       char
       unsigned int BytesToWrite,
       unsigned int *BytesWritten,
       int
                    Timeout
   );
   void (NOEXP FUNC *SS Session Shutdown) (
                    *SScontext,
       void
                    *ApplicationContext
       void
    );
} SecureSocketFunctions;
```

You must provide valid function pointers for each of the four fields in the **SecureSocketFunctions** structure.

SS_Session_Start

Merge DICOM Toolkit calls the **SS_Session_Start** function just after it has accepted the TCP/IP connect request made by the remote host. The *SocketToUse* parameter contains the socket assigned to the connection. <u>Please note that the connection is non-blocking</u>. The *ApplicationContext* parameter is the presented by the *SS_context* parameter of the **MC_Wait_For_Secure_Association** call. *ConnectionType* will be the manifest constant ACCEPTOR_CONNECTION if the **SS_Session_Start** function is called as a result of a **MC_Wait_For_Secure_Association** call. (It will be REQUESTER_CONNECTION if it is called as a result of a **MC_Open_Secure_Association** call.)

The **SS_Session_Start** function is responsible for establishing a secure connection using the socket provided. It is assumed, but not required, that the connection will be a Secure Socket Layer (SSL) or Transport Layer Socket (TLS) connection. A pointer to a context block should be returned at **SecurityContext* if the secure connection is established. Merge DICOM Toolkit will provide this pointer when it calls the other callback functions.

SS_Session_Start must return SS_NORMAL_COMPLETION if the secure connection was successfully established, otherwise SS_ERROR must be returned. If it returns SS_ERROR, the TCP/IP connection will be closed and the **MC_Wait_For_Secure_Association** call will return a status of **MC_NEGOTIATION_ABORTED**.

SS_Session_Shutdown

Merge DICOM Toolkit calls the **SS_Session_Shutdown** function when the association is aborted or closed. It is the responsibility of the **SS_Session_Shutdown** function to gracefully close the secure network connection. Merge DICOM Toolkit closes the TCP/IP socket connection after calling

SS_Session_Shutdown. The *ApplicationContext* parameter is that presented by the *SS_context* parameter of the **MC_Wait_For_Secure_Association** call, and the *Sscontext* parameter is that previously returned by the **SS_Session_Start** function.

SS_Read

Merge DICOM Toolkit calls the **SS_Read** function whenever it needs association data from the secure connection. The *ApplicationContext* parameter is that presented by the *SS_context* parameter of the **MC_Wait_For_Secure_Association** call, and the *Sscontext* parameter is that previously returned by the **SS_Session_Start** function. It is the responsibility of the **SS_Read** function to retrieve into *Buffer* the number of unencrypted data bytes specified by *BytesToRead*. The actual number of bytes placed in the *Buffer* is returned at **BytesRead*. SS_NORMAL_COMPLETION must be returned if the read request was satisfied.

If the **SS_Read** function cannot retrieve *BytesToRead* bytes in *Timeout* seconds, it must return SS_TIMEOUT. Please note that the socket connection passed by Merge DICOM Toolkit to the **SS_Session_Start** function is non-blocking. (Note that if SS_TIMEOUT is returned, an outstanding MC_Read_Message call will return MC_TIMEOUT.)

If it is determined that the secure socket connection has closed, or that the underlying transport has closed, SS_SESSION_CLOSED must be returned. This should only occur during a DICOM association if the remote host aborted the association. If a fatal error occurs while processing the read request, SS_ERROR must be returned. (Note that if SS_SESSION_CLOSED or SS_ERROR is returned, an outstanding MC_Read_Message call will return MC_NETWORK_SHUT_DOWN.)

SS_Write

Merge DICOM Toolkit calls the **SS_Write** function whenever it needs to send association data over the secure connection. The *ApplicationContext* parameter is that presented by the *SS_context* parameter of the **MC_Wait_For_Secure_Association** call, and the *Sscontext* parameter is that previously returned by the **SS_Session_Start** function. It is the responsibility of the **SS_Write** function to send *BytesToWrite* bytes of the data in *Buffer* over the secure connection, returning the number of bytes actually written at **BytesWritten*. SS_NORMAL_COMPLETION must be returned if the write request was satisfied.

If the **SS_Write** function cannot send *BytesToWrite* bytes in *Timeout* seconds, it must return SS_TIMEOUT. Please note that the socket connection passed by Merge DICOM Toolkit to the **SS_Session_Start** function is non-blocking.

If a fatal error occurs while processing the write request, SS_ERROR must be returned.

NOTE: If SS_ERROR is returned, an outstanding MC_Send_Request_Message, MC_Send_Response_Message or MC_Send_Response call will return MC_SYSTEM_ERROR.)

Return Value

Value	Meaning

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_NORMAL_COMPLETION	The function completed normally.	
MC_TIMEOUT	The <i>Timeout</i> period expired.	
MC_NEGOTIATION_ABORTED	An incoming association was aborted during negotiation. Normally this situation is handled by retrying the MC_Wait_For_Secure_Association call. This will also be returned if the SS_Session_Start callback function returned an SS_ERROR status.	
MC_ASSOCIATION_REJECTED	An incoming association was aborted because no services were acceptable.	
MC_SYSTEM_CALL_INTERRUPTED	The system call used to wait for an association request was interrupted by a signal. Normally this situation is handled by retrying the MC_Wait_For_Secure_Association call.	
MC_NULL_POINTER_PARM	<i>AssociationID, ApplicationID, SS_functions</i> or <i>ServiceList</i> parameter was NULL.	
MC_NULL_VALUE	One or more of the function parameters within <i>SS_functions</i> was NULL.	
MC_NO_APPLICATIONS_REGISTERED		
	No applications have been registered using MC_Register_Application.	
MC_INVALID_SERVICE_LIST_NAME	ServiceList points at a null string.	
MC_CONFIG_INFO_MISSING	Could not access ServiceList configuration parameters.	
MC_CONFIG_INFO_ERROR	The ServiceList contained too many services.	
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.	
See Also		
MC_Open_Association	MC_Open_Secure_Association MC_Get_Negotiation_Info MC_Accept_Association	
MC_Reject_Association	MC_Wait_For_Association MC_Close_Listen_Port MC_Close_Listen_Port_On_Address	

MC_Write_File MC_Write_File_By_Callback

Writes a file object to media.

Synopsis

#include ``mc3media.h"

MC_STATUS MC_Write_File (int FileID, int NumBytes, void *UserInfo,

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

```
MC STATUS (*YourToMediaFunction()
)
MC STATUS MC Write File By Callback (
       int ApplicationID,
       int FileID,
       int NumBytes,
       void *UserInfo,
       MC STATUS (*YourToMediaFunction()
)
 ApplicationID
                         The identifier for an application object assigned by
                        the MC_Register_Application function.
                         The identifier assigned to this object by the
 FileID
                         MC_Create_Empty_File or MC_Create_File function.
 NumBytes
                         The attribute (FFFC,FFFC) will be added to the file
                         object and given a length such that the total length of
                         the file being written is a multiple of NumBytes. If
                         NumBytes is set to 0, there will be no padding of the
                         file. NumBytes must be an even number.
 UserInfo
                         Address of data which will be passed on to
                         YourToMediaFunction each time it is called. This may
                         be NULL.
 YourToMediaFunctio
                        Name of a function which will be called repeatedly to
                         provide blocks of DICOM file data.
 n
The function must be prototyped as follows:
```

```
MC_STATUS YourToMediaFunction (
    char *Cbfilename,
    void *CbuserInfo,
    int CbdataSize,
    void *CbdataBuffer,
    int CbisFirst,
    int CbisLast
)
```

```
Cbfilename
                The filename associated with the file object by the
                MC_Create_Empty_File or MC_Create_File function.
CbuserInfo
                Address of data which is being passed from the
                MC_Write_File function. This may be NULL.
                The number of bytes of file data being provided to you in
CbdataSize
                CbdataBuffer.
                The buffer containing file data from the file object.
CbdataBuffer
CbisFirst
                Is TRUE (not zero) when Merge DICOM Toolkit is providing
                the first block of file data.
CbisLast
                Is TRUE (not zero) when Merge DICOM Toolkit is providing
                the last block of file data.
```

Remarks

MC_Write_File and **MC_Write_File_By_Callback** request that the contents of the file object identified by *FileID* be put in a form specified by the DICOM standard for files. The file is passed to the user in blocks by calling *YourToMediaFunction* repeatedly until the entire file has been transferred.

MC_Write_File and MC_Write_File_By_Callback can pass data to *YourToMediaFunction* by specifying the data's address in *UserInfo*. Merge DICOM Toolkit passes the address to *YourToMediaFunction* in *CbuserInfo* each time it is called. *UserInfo* may be NULL.

MC_Write_File_By_Callback is identical to **MC_Write_File** except that it can be used in conjunction with **MC_Register_Callback_Function** to have pixel data supplied to the function as it is being streamed out.

NOTE:	If the file object is a DICOMDIR, MC_Write_File and	
	MC_Write_File_By_Callback will resolve the directory record offsets	
	within the object before passing the user the file data.	
NOTE:	If the file contains "group length" attributes (i.e. attributes with tags of	

- **NOTE:** If the file contains "group length" attributes (i.e. attributes with tags of the form gggg0000: any group, element zero), **MC_Write_File** and **MC_Write_File_By_Callback** will automatically calculate the group length value when supplying it to *YourToMediaFunction*.
- **MC_Write_File** and **MC_Write_File_By_Callback** will format the byte stream passed to *YourToMediaFunction* in the attribute transfer syntax UID (0002, 0010).
- **NOTE: MC_Write_File** and **MC_Write_File_By_Callback** will automatically fill in two group 2 attributes within the file meta information **if you have not set them**:
 - The Implementation Class UID (0002,0012) will be filled in with the value set for the IMPLEMENTATION_CLASS_UID configuration value in the mergecom.pro file.
 - The Implementation Version Name (0002,0013) will be filled in with the value set for the **IMPLEMENTATION_VERSION** configuration value in the mergecom.pro file.

YourToMediaFunction

YourToMediaFunction will be called repeatedly to pass blocks of data to it. Merge DICOM Toolkit sets *CbisFirst* to TRUE (non-zero) the first time it calls *YourToMediaFunction* for this attribute and it sets *CbisLast* to TRUE (non-zero) when it calls *YourToMediaFunction* with the final block of file data.

CbdataBuffer is set to the address of a buffer containing the file data block, and *CbdataSize* is be set to the number of bytes placed at *CbdataBuffer*.

Return Value

[©] Copyright Merge Healthcare Solutions Inc. 2023

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INVALID_APPLICATION_ID	The <i>ApplicationID</i> value is not a valid application object ID.
MC_INVALID_FILE_ID	The <i>FileID</i> value is not a valid file object ID.
MC_INVALID_PAD	The <i>NumBytes</i> parameter contained too large of a number or an odd number.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_MESSAGE_EMPTY	The file has no attributes in it.
MC_CALLBACK_CANNOT_COMPLY	<i>YourToMediaFunction</i> returned a value other than MC_NORMAL_COMPLETION.
MC_INVALID_TRANSFER_SYNTAX	The transfer syntax is improperly specified or is not specified in the DICOM File Meta Information attributes.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
See Also	
MC_Open_File MC_Create_Empty_File	MC_Create_File

MC_XML_To_Message

Reads attribute values from a Merge DICOM Model XML string into a message, file or item object.

```
Synopsis
#include ``mc3msg.h"
MC STATUS MC XML To Message (
       int MessageID,
       void *UserInfo,
       MC STATUS (*YourGetXMLFunction) ()
)
 MessageID
                            The identifier of a message, file or item object.
 UserInfo
                            Address of data which will be passed on to
                            YourGetXMLFunction each time it is called.
                            This may be NULL.
 YourGetXMLFunction
                            Name of a function which will be called
                            repeatedly to get the XML string.
```

The function must be prototyped as follows:

[©] Copyright Merge Healthcare Solutions Inc. 2023

```
void *CBuserInfo,
int *CBdataSize,
void **CBdataBuffer,
int CBisFirst,
int *CBisLast
```

```
)
 CBmessageID
                  The identifier assigned to the message object by the
                  MC_Open_Message function.
 CBuserInfo
                  Address of data which is being passed from the
                  MC_XML_To_Message function. This may be NULL.
 CBdataSize
                  Set *CBdataSize to the number of bytes you are providing.
 CBdataBuffer
                  Set *CBdataBuffer to the address of the data you are
                  providing.
 CBisFirst
                  Set to TRUE (not zero) by the toolkit on the first call.
 CBisLast
                  Set *CBisLast to TRUE (not zero) when you are returning
                  with the last block of XML data.
```

Remarks

MC_XML_To_Message requests that the XML buffer be converted into a DICOM message. The XML data is requested from the user in blocks by calling *YourGetXMLFunction* repeatedly until the entire XML content has been received by the Merge DICOM Toolkit.

YourGetXMLFunction

YourGetXMLFunction will be called repeatedly to get blocks of data from it. Merge DICOM Toolkit sets *CBisFirst* to TRUE(non-zero) the first time it calls **YourGetXMLFunction** for this message and **YourGetXMLFunction** should set **CBisLast* to TRUE(non-zero) when it gives Merge DICOM Toolkit the final block of data to be converted.

***CBdataBuffer* is set to the address of a buffer containing the data block to be converted, and *CBdataSize* is set to the number of bytes to be placed at ***CBdataBuffer.*

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	YourGetXMLFunction parameter was NULL.
MC_INVALID_MESSAGE_ID	The <i>MessagelD</i> value is not a valid message, file or item object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CALLBACK_CANNOT_COMPLY	YourGetXMLFunction returned a value other than MC_NORMAL_COMPLETION.
MC_LIBRARY_NOT_INITIALIZED	This call was made prior to the initialization of the library.

© Copyright Merge Healthcare Solutions Inc. 2023

See Also

MC_Message_To_XML

MC_XML_Native_To_Message

Reads attribute values from a Native DICOM Model XML string into a message, file or item object.

Synopsis

```
#include ``mc3msg.h"
```

```
      MC_STATUS MC_XML_Native_To_Message (

      int MessageID,

      void *UserInfo,

      MC_STATUS (*YourGetXMLFunction) ()

      )

      MessageID

      UserInfo

      Address of data which will be passed on to

      YourGetXMLFunction

      Name of a function which will be called

      repeatedly to get the XML string.
```

The function must be prototyped as follows:

```
MC_STATUS YourGetXMLFunction (
    int CBmessageID,
    void *CBuserInfo,
    int *CBdataSize,
    void **CBdataBuffer,
    int CBisFirst,
    int *CBisLast
```

)

CBmessagelD	The identifier assigned to the message object by the MC_Open_Message function.
CBuserInfo	Address of data which is being passed from the MC_XML_Native_To_Message function. This may be NULL
CBdataSize	Set *CBdataSize to the number of bytes you are providing.
CBdataBuffer	Set *CBdataBuffer to the address of the data you are providing.
CBisFirst	Set to TRUE (not zero) by the toolkit on the first call.
CBisLast	Set *CBisLast to TRUE (not zero) when you are returning with the last block of XML data.

Remarks

MC_XML_Native_To_Message requests that the Native DICOM Model XML buffer be converted into a DICOM message. The XML data is requested from the user in blocks by calling

YourGetXMLFunction repeatedly until the entire XML content has been received by the Merge DICOM Toolkit.

YourGetXMLFunction

YourGetXMLFunction will be called repeatedly to get blocks of data from it. Merge DICOM Toolkit sets *CBisFirst* to TRUE(non-zero) the first time it calls *YourGetXMLFunction* for this message and *YourGetXMLFunction* should set **CBisLast* to TRUE(non-zero) when it gives Merge DICOM Toolkit the final block of data to be converted.

***CBdataBuffer* is set to the address of a buffer containing the data block to be converted, and *CBdataSize* is set to the number of bytes to be placed at ***CBdataBuffer.*

Return Value

One of these enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	YourGetXMLFunction parameter was NULL.
MC_INVALID_MESSAGE_ID	The <i>MessageID</i> value is not a valid message, file or item object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_CALLBACK_CANNOT_COMPLY	YourGetXMLFunction returned a value other than MC_NORMAL_COMPLETION.
MC_LIBRARY_NOT_INITIALIZED	This call was made prior to the initialization of the library.
See Also	

MC_Message_To_XML_Native

High Level API Reference

This section of the Reference Manual includes the function reference for the domain-specific functionality of the Toolkit. Each reference page provides:

- a brief description of the API function
- a synopsis of the API function which contains a list of include files required to use the function, the function prototype and a description of each function parameter
- remarks outlining the use of the function
- a list of status codes returned by the function
- a "See also" cross reference to other functions

MC_DDH_Create

Creates a new DICOMDIR file and its associated toolkit object.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Synopsis

```
#include ``mc3media.h"
MC STATUS MC DDH Create (
        const char *FilePath,
       const char *FileSetID,
       int TemplateFileID,
       int *DirMsgIDPtr
)
  FilePath
                         Path to the DICOMDIR file (including the file name).
  FileSetID
                         The value to be assigned to the File Set ID attribute
                         (0004,1130). If this argument is NULL the toolkit will
                         copy the value of the File Set ID attribute from the
                         template file if specified.
  TemplateFileID
                         The ID of a file object containing attribute values for
                         the new DICODMIR. If this argument is 0 the toolkit
                         creates an empty DICOMDIR with a minimal number
                         of attributes.
                         Pointer to a variable receiving the toolkit identifier of
  DirMsgIDPtr
                         the new directory object.
```

Remarks

This function writes a new **DICOMDIR** file with no records and returns the directory object identifier representing the new directory file. If a file with the same name exists, its content is overwritten.

The directory object must be freed via MC_Free_File when it is no longer needed.

The toolkit copies group 2 and 4 attribute values from the specified template file. If the Media Storage SOP Instance UID attribute is not present in the template file or the template file ID is 0 the toolkit will generate a UID based on the **IMPLEMENTATION_CLASS_UID** configuration item.

See MC_DDH_Open function for details on managing the directory object.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	<i>FilePath</i> or <i>DirMsgIDPtr</i> was NULL.
MC_INVALID_MESSAGE_ID	The <i>TemplateFileID</i> is not a valid toolkit file object identifier.
MC_INVALID_CHARS_IN_VALUE	The <i>FileSetID</i> argument contains invalid characters.
MC_INVALID_VALUE_FOR_VR	The <i>FileSetID</i> argument does not conform to the requirements of the CS value representation.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_CALLBACK_CANNOT_COMPLY	A file read or write operation failed.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
See Also	

MC_DDH_Open

MC_Free_File

MC_DDH_Open

Creates a DICOMDIR toolkit object which represents the content of an exising DICOMDIR file.

Synopsis

```
#include "mc3media.h"
MC_STATUS MC_DDH_Create (
    const char *FilePath,
    int *DirMsgIDPtr
)
FilePath
DirMsgIDPtr
Pointer to a variable receiving the toolkit identifier of
the new directory object.
```

Remarks

This function opens an existing DICOMDIR file for reading and/or updating and returns the directory object identifier representing its content. If the specified file does not exist or is not accessible the toolkit returns an error.

As opposed to opening DICODMIR files using **MC_Open_File**, this function does not read the whole content of the DICOMDIR in the memory but instead reads only the first directory record if it exists. Subsequent directory records are read only on demand, when the application requires access to them, providing fast and efficient access to large DICOMDIR's.

The directory object created by **MC_DDH_Open** allows the application to incrementally add new records or delete existing records using **MC_DDH_Add_Record**, **MC_DDH_Delete_Record** and **MC_DDH_Update** function, whithout rewriting the whole file.

The returned directory object identifier can be used with all **MC_DDH...** functions that require a directory ID or with **MC_Get_Value...** functions to obtain group 2 and 4 attribute values, but it can not be used with any **MC_Dir...** function.

The Directory Record Sequence (0004,1220) attribute and its content is not accessible to the application as it is managed internally by the toolkit.

Although the toolkit allows the application to modify the content of existing directory records or the attributes in the directory object itself, the changes will not be reflected in the DICOMDIR file.

The directory file object must be freed via **MC_Free_File** when it is no longer needed.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_UNEXPECTED_EOD	The specified file does not contain all data necessary to decode an attribute's value.
MC_INVALID_FILE	The specified file does not conform to the requirements of a DICOMDIR file. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TRANSFER_SYNTAX	An invalid transfer syntax code was found within the file's DICOM File Meta Information.
MC_OUT_OF_ORDER_TAG	A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TAG	The file contains an invalid tag.
MC_VALUE_TOO_LARGE	An attribute in the file (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation.
MC_CALLBACK_CANNOT_COMPLY	A file read operation failed.
MC_INVALID_LENGTH_FOR_VR	The value length for one of the file attributes was not legal for its value representation.
MC_INVALID_CHARS_IN_VALUE	A value for an attribute does not contain valid characters for its value representation.
MC_INVALID_VALUE_FOR_VR	The <i>FileSetID</i> argument does not conform to the requirements of the CS value representation.
MC_NULL_POINTER_PARM	<i>FilePath</i> or <i>DirMsgIDPtr</i> was NULL.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
MC_UNABLE_TO_GET_ITEM_ID	The toolkit failed to read the first directory record.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
Any of the status codes returned by MC_Set_Value may also be returned.	

See Also

MC_Free_File

MC_DDH_Update

Commits all pending changes to the DICOMDIR file.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Synopsis

#include ``mc3media.h"

```
MC_STATUS MC_DDH_Update (
    int DirMsgID
)
```

DirMsgID

The identifier of a directory object obtained using **MC_DDH_Create** or **MC_DDH_Open** functions.

Remarks

This function updates the DICOMDIR file to reflect all changes to the dicomdir object requested through **MC_DDH_Add_Record** and **MC_DDH_Delete_Record** functions. The toolkit writes new records to the end of the directory record sequence and bridges over deleted records so that they are no longer accessible.

MC_DDH_Update does not make changes to existing records or the attributes of the directory object, except for the values of the offset attributes required to represent the modified record hierarchy.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	<i>DirMsgID</i> is not a valid toolkit directory object identifier.
MC_INVALID_FILE	A failure occured while writing the file. The DICOMDIR file may be corrupted.
MC_CALLBACK_CANNOT_COMPLY	The DICOMDIR file could not be opened for writing or a file read or write operation failed. The DICOMDIR file is in a consistent state but the directory object can not be reliably used and it should be freed using MC_Free_File .
MC_INVALID_DIRECTORY_RECORE	D_OFFSET
	A directory record offset value is not correct and new records cannot be written, the directory object can still be used in read-only mode.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred. The DICOMDIR file may be in an inconsistent state.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file. The DICOMDIR file is in a consistent state but the directory object can not be reliably used and it should be freed using MC_Free_File .
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.

See Also

MC_DDH_Create MC_DDH_Add_Record

MC_DDH_Open MC_DDH_Delete_Record

MC_DDH_Traverse_Records

Provides easy record traversal functionality.

Synopsis

#include ``mc3media.h"

<pre>MC_STATUS MC_DDH_Traverse_Records (int RootID, void *UserData, MC_TRAVERSAL_STATUS (*YourTraverseCallback)())</pre>		
RootID	The object identifier of the record or directory representing the root for the traversal.	
UserData	The pointer to be passed to the user's callback. The value of this argument is not interpreted by the toollkit and it can be NULL	
YourTraverseCallbac k	The function to be called for each record being visisted	
The function must be prototyped as follows:		
MC_TRAVERSAL_STATUS int CurrentRed void *UserData		

```
)
```

CurrentRecl The identifier of the directory record being traversed.

```
D
```

UserData	The value passed in the UserData argument of the
	MC_DDH_Traverse_Records function.

Remarks

This function traverses all directory records that are below the specified directory record, or all records if the RootID is the identifier of a directory object. The application can controll the traversal through the various status codes returned by the callback for each call:

- MC_TS_CONTINUE The traversal continues normally with the lower level records of the current record or with the next record if the current record does not have lower level records.
- MC_TS_STOP_LEVEL The traversal continues with the next record of the current record's parent, skipping all lower level records and next records of the current record.
- MC_TS_STOP_LOWER The traversal continues with the next record of the current record, skipping the current record's lower level records

[©] Copyright Merge Healthcare Solutions Inc. 2023

- MC_TS_STOP Stop the traversal. **MC_DDH_Traverse_Records** will return MC_NORMAL_COMPLETION.
- MC_TS_ERROR Stop the traversal because an error occured in the callback. MC_DDH_Traverse_Records will return MC_CALLBACK_CANNOT_COMPLY.

MC_DDH_Traverse_Records provides an easy and fast way of searching and counting records.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_DICOMDIR_ID	The <i>RootID</i> argument does not represent a valid record or directory object identifier
MC_NULL_POINTER_PARM	The YourTraverseCallback argument was NULL.
MC_CALLBACK_CANNOT_COMPLY	The callback function returned MC_TS_ERROR or an invalid status or a file read operation failed.
MC_INVALID_DIRECTORY_RECORE	D_OFFSET
	A directory record offset value is not correct.
MC_UNEXPECTED_EOD	Not enough data to decode an attribute's value.
MC_OUT_OF_ORDER_TAG	A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TAG	A record item contains an invalid tag.
MC_VALUE_TOO_LARGE	An attribute in the file (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation.
MC_INVALID_LENGTH_FOR_VR	The value length for one of the file attributes was not legal for its value representation.
MC_INVALID_VALUE_FOR_VR	The value for one of the file attributes was not legal for its value representation.
MC_INVALID_CHARS_IN_VALUE	A value for an attribute does not contain valid characters for its value representation.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.

MC_DDH_Get_Record_Type

Gets a directory record's type.

© Copyright Merge Healthcare Solutions Inc. 2023

Synopsis #include "mc3media.h" MC_STATUS MC_DDH_Get_Record_Type (int RecordID, MC_DIR_RECORD_TYPE *RecType) MC_DIR_RECORD_TYPE *RecType) RecordID The toolkit identifier of the record. RecType Pointer to the variable receiving the record type. The MC_DIR_RECORD_TYPE enumeration is defined in mc3media.h.

Remarks

This function can be used to obtain an integer representing the type of the specified directory record as specified by the value of the Directory Record Type attribute (0004,1430).

If the value of the Directory Record Type attribute does not correspond to any of the values in the MC_DIR_RECORD_TYPE enum the toolkit returns MC_REC_TYPE_UNKNOWN as the record type.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_RECORD_ID	The <i>RecordID</i> argument does not represent a valid record object identifier.
MC_NULL_POINTER_PARM	The <i>RecType</i> argument was NULL.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.

MC_DDH_Get_Parent_Record

Gets the parent directory record of the specified record.

Synopsis

```
#include "mc3media.h"
MC_STATUS MC_DDH_Get_Parent_Record (
    int RecordID,
    int *ParentID
)
RecordID The identifier of the record object for which the parent
    record is requested.
ParentID Pointer to the variable receiving the parent record's
    identifier.
```

Remarks

MC_DDH_Get_Parent_Record provides the parent directory record of a record object. If the specified record is a top level record this function sets the value pointed by the *ParentID* argument to 0 and returns MC_NORMAL_COMPLETION.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_RECORD_ID	The <i>RecordID</i> argument does not represent a valid record object identifier.
MC_NULL_POINTER_PARM	The ParentID argument was NULL.
MC_CALLBACK_CANNOT_COMPLY	A file read operation failed.
MC_INVALID_DIRECTORY_RECORD	D_OFFSET
	A directory record offset value is not correct.
MC_UNEXPECTED_EOD	Not enough data to decode an attribute's value.
MC_OUT_OF_ORDER_TAG	A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TAG	A record item contains an invalid tag.
MC_VALUE_TOO_LARGE	An attribute in the file (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation.
MC_INVALID_LENGTH_FOR_VR	The value length for one of the file attributes was not legal for its value representation.
MC_INVALID_VALUE_FOR_VR	The value for one of the file attributes was not legal for its value representation.
MC_INVALID_CHARS_IN_VALUE	A value for an attribute does not contain valid characters for its value representation.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.

MC_DDH_Get_Next_Record

Gets the next directory record in a directory entity.

© Copyright Merge Healthcare Solutions Inc. 2023

Synopsis		
<pre>#include ``mc3media.h"</pre>		
MC_STATUS MC_DDH_G int RecordID int *NextID)		
RecordID	The identifier of the record object of which the next record is requested.	
ParentID	Pointer to the variable receiving the next record's identifier.	

Remarks

Return Value

c ...

MC_DDH_Get_Next_Record retrieves the directory record following the specified record in the directory record hierarchy. If the specified record is the last record in the list of lower level record of its parent record this function sets the value of the *Next/D* to 0 and returns MC_NORMAL_COMPLETION.

The returned record identifier can be used with the **MC_Get_Value...** functions to access the record's attribute values.

1.0

One of the enumerated MC_STATUS codes defined in "mcstatus.h":		
Value	Meaning	
MC_NORMAL_COMPLETION	The function completed normally.	
MC_INVALID_RECORD_ID	The <i>RecordID</i> argument does not represent a valid record object identifier.	
MC_NULL_POINTER_PARM	The <i>NextID</i> argument was NULL.	
MC_CALLBACK_CANNOT_COMPLY	A file read operation failed.	
MC_INVALID_DIRECTORY_RECORD_OFFSET		
	A directory record offset value is not correct.	
MC_UNEXPECTED_EOD	Not enough data to decode an attribute's value.	
MC_OUT_OF_ORDER_TAG	A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file.	
MC_INVALID_TAG	A record item contains an invalid tag.	
MC_VALUE_TOO_LARGE	An attribute in the file (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation.	
MC_INVALID_LENGTH_FOR_VR	The value length for one of the file attributes was not legal for its value representation.	
MC_INVALID_VALUE_FOR_VR	The value for one of the file attributes was not legal for its value representation.	

MC_INVALID_CHARS_IN_VALUE	A value for an attribute does not contain valid characters for its value representation.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.

MC_DDH_Get_First_Lower_Record

Gets the first lower level directory record of the specified record.

Synopsis

#include ``mc3media.h"

```
      MC_STATUS
      MC_DDH_Get_First_Lower_Record (

      int
      ParentID,

      int
      *LowerID

      )
      The identifier of the record object for which the lower

      level record is requested or the identifier of the

      directory object if the first record in the root entity is

      requested.

      LowerID
      Pointer to the variable receiving the first lower level

      record's identifier.
```

Remarks

MC_DDH_Get_First_Lower_Record retrieves the first lower level directory record of the specified record. If the specified record does not have lower level records this function sets the value of the *LowerID* to 0 and returns MC_NORMAL_COMPLETION.

The returned record identifier can be used with the **MC_Get_Value...** functions to access the record's attribute values.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_RECORD_ID	The <i>ParentID</i> argument does not represent a valid record or DICOMDIR object identifier.
MC_NULL_POINTER_PARM	The <i>LowerID</i> argument was NULL.
MC_CALLBACK_CANNOT_COMPLY	A file read operation failed.

MC_INVALID_DIRECTORY_RECORD_OFFSET A directory record offset value is not correct. MC_UNEXPECTED_EOD Not enough data to decode an attribute's value. MC_OUT_OF_ORDER_TAG A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file. MC_INVALID_TAG A record item contains an invalid tag. MC_VALUE_TOO_LARGE An attribute in the file (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation. MC_INVALID_LENGTH_FOR_VR The value length for one of the file attributes was not legal for its value representation. MC_INVALID_VALUE_FOR_VR The value for one of the file attributes was not legal for its value representation. MC_INVALID_CHARS_IN_VALUE A value for an attribute does not contain valid characters for its value representation. MC_TEMPFILE_ERROR A value for an attribute is stored temporarily on file and an I/O error occurred. MC_SYSTEM_ERROR An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file. MC_LIBRARY_NOT_INITIALIZED The library has not been properly initialized.

MC_DDH_Copy_Values

Copies attribute values from one message or item to another message or item.

Synopsis #include "mc3media.h" MC_STATUS MC_DDH_Copy_Values (int SourceID, int DestID, unsigned long *TagList) SourceID The ID of the source message, file, item or record object containing the values to be copied. DestID The ID of the destination message, file, item or record object which will receive the values.

TagList

A zero ended array of tag numbers of the attributes to be copied. If a tag is a group length tag, the attributes in the whole group are copied. If this parameter is NULL, all attributes are copied.

Remarks

This function can be used to fill in directory record values from messages when adding instances to a file set.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_MESSAGE_ID	The source or destiantion ID value is not a valid toolkit object identifier.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.

MC_DDH_Add_Record

Adds a new directory record to a directory object.

```
Synopsis
#include "mc3media.h"
MC_STATUS MC_DDH_Add_Record (
    int ParentID,
    const char *RecordType
    int *RecordID
)
```

```
ParentID
```

The identifier of a directory object obtained using **MC_DDH_Create** or **MC_DDH_Open** functions or the identifier of a record object obtained using one of the **MC_DDH_...** record navigation functions.

RecordType	The directory record type name for the new record or
	the SOP Class UID of the referenced instance. If this
	argument is NULL the toolkit will set the record type
	based on the type of the parent record: "PATIENT" for
	root level records, "STUDY" if the parent is
	"PATIENT", "SERIES" if the parent is "STUDY" and
	"IMAGE" if the parent is "SERIES". The application
	can change the value of the Directory Record Type
	attribute of the new record item after this function
	returns, before calling MC_DDH_Update.
RecordID	Pointer to the variable receiving the identifier of the new record.

Remarks

MC_DDH_Add_Record creates a new directory record object and appends it to the end of the child record list of the specified parent record or to the end of the root entity if the specified parent is the directory object itself.

If the type of the record is not specified the toolkit will try to determine it based on the type of the parent record. For instance level records, the application can specify the SOP Class UID of the referenced object in the *RecordType* argument and the toolkit will create the corresponding record type.

The returned record identifier can be used with the **MC_Set_Value**... or **MC_DDH_Copy_Values** functions to set the attribute values in the new record.

The new record and any values set for its attributes are written to the DICOMDIR file when the application calls **MC_DDH_Update**. If the directory object is freed without calling **MC_DDH_Update**, all changes to the record hierarchy will be lost.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_RECORD_ID	The <i>ParentID</i> argument does not represent a valid record or DICOMDIR object identifier.
MC_NULL_POINTER_PARM	The <i>RecordID</i> argument was NULL.
MC_INVALID_LOWER_DIR_RECORD	The record type specified in the <i>RecordType</i> argument is not a valid lower level record of the parent record.
MC_BAD_DIR_RECORD_TYPE	The type of the parent record is unknown.
MC_CALLBACK_CANNOT_COMPLY	A file read operation failed.
MC_INVALID_DIRECTORY_RECORD	_OFFSET
	A directory record offset value is not correct.
MC_UNEXPECTED_EOD	Not enough data to decode an attribute's value.

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_OUT_OF_ORDER_TAG	A tag was found in the file that was not in ascending order. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_INVALID_TAG	A record item contains an invalid tag.
MC_VALUE_TOO_LARGE	An attribute in the file (other than those with value representations of OB, OW, or OF) was larger than that acceptable for its value representation.
MC_INVALID_LENGTH_FOR_VR	The value length for one of the file attributes was not legal for its value representation.
MC_INVALID_VALUE_FOR_VR	The value for one of the file attributes was not legal for its value representation.
MC_INVALID_CHARS_IN_VALUE	A value for an attribute does not contain valid characters for its value representation.
MC_TEMPFILE_ERROR	A value for an attribute is stored temporarily on file and an I/O error occurred.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	
MC_DDH_Create MC_DDH_Update	MC_DDH_Open

MC_DDH_Delete_Record

Deletes a directory record.

Synopsis

#include ``mc3media.h"

MC_STATUS MC_DDH_Delete_Record (int RecordID

)

The toolkit identifier of the directory record to delete.

Remarks

RecordID

MC_DDH_Delete_Record removes the specified directory record and any lower level record from the record hierarchy.

The resources allocated for the directory record objects are automatically freed.

The changes made by this function are not reflected in the DICOMDIR file until the application calls **MC_DDH_Update**.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_RECORD_ID	The <i>RecordID</i> argument does not represent a valid record object identifier.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.
MC_LIBRARY_NOT_INITIALIZED	The library has not been properly initialized.
See Also	

MC_DDH_Update

MC_DDH_Release_Record

Releases the content of the record.

Synopsis	
<pre>#include ``mc3media.h"</pre>	
MC_STATUS MC_DDH_ int Record:)	
RecordID	The toolkit identifier of the directory record to release.

Remarks

This function releases the specified record and its lower level records to reduce the toolkit's memory usage. Although the record identifier becomes invalid after this call, the record itself is not deleted and can be accessed again using one of the record navigation functions, **MC_DDH_Get_Next_Record** or **MC_DDH_Get_First_Lower_Level_Record**, in which case the toolkit will read the content of the record from the DICOMDIR file again and assign a new record identifier.

This function does not modify the directory record hierarchy.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_RECORD_ID	The <i>RecordID</i> argument does not represent a valid record object identifier.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_LIBRARY_NOT_INITIALIZED The library has not been properly initialized.

See Also

MC_DDH_Get_Next_Record MC_DDH_Get_First_Lower_Level_Record

MC_SRH_Create_SR

Creates a new SR root node.

Synopsis

```
MC_STATUS MC_SRH_Create_SR (
const char *AserviceName,
const char *AtemplateId,
SR_CONT_CONTINUITY Acontinuity,
const char *AconceptNameValue,
const char *AconceptNameScheme,
const char *AconceptNameMeaning,
int *AsrRootID
```

```
)
```

A

AserviceName	The service (Sop Class) the SR uses.
Atemplateld	DICOM Template Identifier that describes the content of this Content Item and its subsidiary Content Items. Can be NULL.
Acontinuity	Enumerated value specifies whether or not its contained Content Items are logically linked in a continuous textual flow, or are separate items.
AconceptNameValue	Code Value describing the concept represented by the root Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning Code Meaning	
AsrRootID	Pointer that receives a numeric identifer for an SR node.

Remarks

MC_SRH_Create_SR creates SR tree management object. This returned item identifier can then be used to set and retrieve SR values.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_NULL_POINTER_PARM MC_SYSTEM_ERROR

One of the parameters was NULL.

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Free_SR

Releases all memory allocated by SR.

Synopsis

```
MC_STATUS MC_SRH_Free_SR (
int AsrRootID
```

)

AsrRootID The identifier of the SR root node to be released.

Remarks

Underlying message object will be released as well.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrRootID value is not a valid SR tree management object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_Child

Adds a child node to the provided parent SR node.

Synopsis

```
      MC_STATUS MC_SRH_Add_Child (

      int AsrNodeID,

      SR_RELATIONSHIP Arelationship,

      int *AsrChildNodeID

      )

      AsrNodelD

      The identifier assigned to the given SR object to which the child is to be added.

      Arelationship
      Enumerated value representing a relationship.
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

AsrChildNodeID which is to be added to the parent. On successful completion returns a new child SR node identifier.

Remarks

This function shall be used for adding existing items to the SR tree. It is recommended to use specific functions like MC_SRH_Add_XXX_Child to create child nodes from scratch.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_TEXT_Child

Creates a new child TEXT node under the provided parent.

Synopsis

```
MC_STATUS MC_SRH_Add_TEXT_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    const char *AconceptNameValue,
    const char *AconceptNameScheme,
    const char *AconceptNameMeaning,
    const char *Atext,
    int *AsrChildNodeID
)
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning
Atext	A text value to be stored in this node.

[©] Copyright Merge Healthcare Solutions Inc. 2023

AsrChildNodeID

Pointer to the variable receiving an identifier of the new SR node.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_TEXT_Node

Creates a new TEXT node.

```
Synopsis
```

AconceptNameValue Code Value describing the concept represented by this Content Item.

AconceptNameScheme Coding Scheme Designator

AconceptNameMeaning Code Meaning

```
Atext A text value to be stored in this node.
```

AsrNodelD Pointer to the variable receiving an identifier of the new SR node.

Return Value

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_Concept_Name

Sets the Concept Name values for the node.

Synopsis

MC_STATUS MC_SRH_	Set_Concept_Name (
int AsrNodeID,		
const char	*AconceptNameValue,	
const char	*AconceptNameScheme,	
const char	*AconceptNameMeaning	
)		

AsrNodeID	Numeric handle pointing to the SR node.	
AconceptNameValue	Code Value describing the concept represented by this Content Item.	
AconceptNameScheme Coding Scheme Designator		
AconceptNameMeaning Code Meaning		
AsrNodeID	Pointer to the variable receiving an identifier of the new SR node.	

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.

MC_SYSTEM_ERROR

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_CODE_Child

Creates a new child CODE node under the provided parent.

Synopsis

```
MC_STATUS MC_SRH_Add_CODE_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    const char *AconceptNameValue,
    const char *AconceptNameScheme,
    const char *AconceptNameMeaning,
    const char *AconceptCodeValue,
    const char *AconceptCodingScheme,
    const char *AconceptCodeMeaning,
    int *AsrChildNodeID
)
```

```
)
```

AsrNodeID	An Identifier of the parent SR item to which the child is to be added.	
Arelationship	Enumerated value representing a relationship.	
AconceptNameValue	Code Value describing the concept represented by this Content Item.	
AconceptNameScheme	Coding Scheme Designator	
AconceptNameMeaning	Code Meaning	
AconceptCodeValue	Coded expression of the concept.	
AconceptCodingScheme Coding Scheme Designator of the concept.		
AconceptCodeMeaning	Code Meaning of the concept.	
AsrChildNodeID	Pointer to the variable receiving an identifier of the new SR node.	

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_SYSTEM_ERROR

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_CODE_Node

Creates a new CODE node.

Synopsis

MC_STATUS MC_SRH_Create_CODE_Node (
const char	*AconceptNameValue,
const char	*AconceptNameScheme,
const char	*AconceptNameMeaning,
const char	*AconceptCodeValue,
const char	*AconceptCodingScheme,
const char	*AconceptCodeMeaning,
<pre>int *AsrNoc</pre>	deID)

AconceptNameValue	Code Value describing the concept represented by this Content Item.	
AconceptNameScheme	Coding Scheme Designator	
AconceptNameMeaning	Code Meaning	
AconceptCodeValue	Coded expression of the concept.	
AconceptCodingScheme Coding Scheme Designator of the concept.		
AconceptCodeMeaning	Code Meaning of the concept.	
AsrNodeID	Pointer to the variable receiving an identifier of the new SR node.	

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_NUM_Child

Creates a new child NUM node under the provided parent.

Synopsis

)

```
MC_STATUS MC_SRH_Add_NUM_Child (
      int AsrNodeID,
      SR_RELATIONSHIP Arelationship,
      const char *AconceptNameValue,
      const char *AconceptNameScheme,
      const char *AconceptNameMeaning,
      const char *AnumericValue,
      const char *AunitsCodeValue,
      const char *AunitsCodingScheme,
      const char *AunitsCodeMeaning,
      int *AsrChildNodeID
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.	
Arelationship	Enumerated value representing a relationship.	
AconceptNameValue	Code Value describing the concept represented by this Content Item.	
AconceptNameScheme Coding Scheme Designator		
AconceptNameMeaning Code Meaning		
AnumericValue	Numeric Value	
AunitsCodeValue	Coded expression of measurement units.	
AunitsCodingScheme	Coding Scheme Designator of the measurement units.	
AunitsCodeMeaning	Code Meaning of the measurement units.	
AsrChildNodelD	Pointer to the variable receiving an identifier of the new SR node.	

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.

MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_NUM_Node

Creates a new NUM node. Note that the Measured Value Sequence (0040, A300) may be empty to convey the concept of a measurement whose value is unknown or missing, or a measurement or calculation failure.

Synopsis

```
MC_STATUS MC_SRH_Create_NUM_Node (
const char *AconceptNameValue,
const char *AconceptNameScheme,
const char *AconceptNameMeaning,
const char *AnumericValue,
const char *AunitsCodeValue,
const char *AunitsCodingScheme,
const char *AunitsCodeMeaning,
int *AsrNodeID
```

```
)
```

AconceptNameValue Code Value describing the concept represented by this Content Item. *AconceptNameScheme* Coding Scheme Designator

AconceptNameMeaning Code Meaning

Numeric Value. Can be NULL.
Coded expression of measurement units.
Coding Scheme Designator of the measurement units.
Code Meaning of the measurement units.
Pointer to the variable receiving an identifier of the new SR node.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.

MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_NUM_Qualifier

Sets the Qualification of Numeric Value in the Measured Value Sequence, or as a reason for the absence of Measured Value Sequence Item.

Synopsis

```
      MC_STATUS MC_SRH_Set_NUM_Qualifier (

      int AsrNodeID,

      const char *AqualifierCodeValue,

      const char *AqualifierCodingScheme,

      const char *AqualifierCodeMeaning

      )

      AsrNodeID

      An Identifier of the NUM SR item.

      AqualifierCodeValue

      Sets Qualification of Numeric Value.
```

AqualifierCodingScheme Coding Scheme Designator

AqualifierCodeMeaning Code Meaning

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_NUM_Next_Data

Sets the next Numeric Value.

Synopsis		
MC_STATUS MC_SRH_Set_NUM_Next_Data (int AsrNodeID, char *AnumericValue		
)		
AsrNodeID	An Identifier of the NUM SR item.	
AnumericValue	Numeric Value	

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_DATETIME_Child

Creates a new child DATETIME node under the provided parent.

```
MC_STATUS MC_SRH_Add_DATETIME_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    const char *AconceptNameValue,
    const char *AconceptNameScheme,
    const char *AconceptNameMeaning,
    const char *Adatetime,
    int *AsrChildNodeID
)
```

AsrNodeID	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning
Adatetime	A date time string in the DICOM format.
AsrChildNodelD	Pointer to the variable receiving an identifier of the new SR node.

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_DATETIME_Node

Creates a new DATETIME node.

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning
Adatetime	A date time string in the DICOM format.
AsrNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_DATE_Child

Creates a new child DATE node under the provided parent.

<pre>MC_STATUS MC_SRH_Add_DATE_Child (int AsrNodeID, SR_RELATIONSHIP Arelationship, const char *AconceptNameValue, const char *AconceptNameScheme, const char *AconceptNameMeaning, const char *Adate, int *AsrChildNodeID</pre>	
)	
AsrNodeID	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.

 $[\]ensuremath{\textcircled{O}}$ Copyright Merge Healthcare Solutions Inc. 2023

AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning
Adate	A date string in the DICOM format.
AsrChildNodelD	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_DATE_Node

Creates a new DATE node.

```
MC_STATUS MC_SRH_Create_DATE_Node (

const char *AconceptNameValue,

const char *AconceptNameScheme,

const char *AconceptNameMeaning,

const char *Adate,

int *AsrNodeID

)

ArrNodeID

ArrNodeID
```

AsrivodelD	child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.

AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning Code Meaning	
Adate	A date string in the DICOM format.
AsrNodelD	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_TIME_Child

Creates a new child DATETIME node under the provided parent.

```
Synopsis

MC_STATUS MC_SRH_Add_TIME_Child (

int AsrNodeID,

SR_RELATIONSHIP Arelationship,

const char *AconceptNameValue,

const char *AconceptNameScheme,

const char *AconceptNameMeaning,

const char *Atime,

int *AsrChildNodeID

)
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning

Atime	A time string in the DICOM format.
AsrChildNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_TIME_Node

Creates a new TIME node.

```
Synopsis
MC_STATUS MC_SRH_Create_TIME_Node (
       const char *AconceptNameValue,
      const char *AconceptNameScheme,
      const char *AconceptNameMeaning,
      const char *Atime,
      int *AsrNodeID
)
                       An Identifier of the parent SR item to which the
AsrNodeID
                       child is to be added.
Arelationship
                       Enumerated value representing a relationship.
                       Code Value describing the concept represented by
AconceptNameValue
                       this Content Item.
AconceptNameScheme Coding Scheme Designator.
AconceptNameMeaning Code Meaning.
```

Atime A time string in the DICOM format.

[©] Copyright Merge Healthcare Solutions Inc. 2023

AsrNodeID

Pointer to the variable receiving an identifier of the new SR node.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_UIDREF_Child

Creates a new child UIDREF node under the provided parent.

```
MC_STATUS MC_SRH_Add_UIDREF_Child (
int AsrNodeID,
SR_RELATIONSHIP Arelationship,
const char *AconceptNameValue,
const char *AconceptNameScheme,
const char *AconceptNameMeaning,
const char *AuidRef,
int *AsrChildNodeID
```

```
)
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning
AuidRef	Unique Identifier (UID) of the entity identified by the Concept Name.
AsrChildNodelD	Pointer to the variable receiving an identifier of the new SR node.

[©] Copyright Merge Healthcare Solutions Inc. 2023

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_UIDREF_Node

Creates a new UIDREF node.

```
MC STATUS MC_SRH_Create_UIDREF_Node (
      const char *AconceptNameValue,
      const char *AconceptNameScheme,
      const char *AconceptNameMeaning,
      const char *AuidRef,
      int *AsrNodeID
)
                       Code Value describing the concept represented by
AconceptNameValue
                       this Content Item.
AconceptNameScheme Coding Scheme Designator
AconceptNameMeaning Code Meaning
                       Unique Identifier (UID) of the entity identified by
AuidRef
                       the Concept Name.
                       Pointer to the variable receiving an identifier of the
AsrNodeID
                       new SR node.
```

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_PNAME_Child

Creates a new child PNAME node under the provided parent.

Synopsis

)

```
MC STATUS MC_SRH_Add_PNAME_Child (
      int AsrNodeID,
      SR RELATIONSHIP Arelationship,
      const char *AconceptNameValue,
      const char *AconceptNameScheme,
      const char *AconceptNameMeaning,
      const char *ApersonName,
      int *AsrChildNodeID
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning
ApersonName	Person name of the person whose role is described by the Concept Name.
AsrChildNodelD	Pointer to the variable receiving an identifier of the new SR node.

Return Value

[©] Copyright Merge Healthcare Solutions Inc. 2023

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_PNAME_Node

Creates a new PNAME node.

Synopsis

```
      MC_STATUS MC_SRH_Create_PNAME_Node (

      const char *AconceptNameValue,

      const char *AconceptNameScheme,

      const char *AconceptNameMeaning,

      const char *AconceptNameMeaning,

      const char *ApersonName,

      int *AsrNodeID

      )

      AconceptNameValue

      Code Value describing the concept represented by

      this Content Item.

      AconceptNameScheme
      Code Meaning

      AconceptNameMeaning
      Code Meaning

      AconceptNameMeaning
      Person name of the person whose role is described

      by the Concept Name.
      Deinternet the variable meaning of the person whose role is described
```

AsrNodelD Pointer to the variable receiving an identifier of the new SR node.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.

MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_SCOORD_Child

Creates a new child SCOORD node under the provided parent.

```
MC_STATUS MC_SRH_Add_SCOORD_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    float Acolumn,
    float Arow,
    SR_GRAPHIC_TYPE AgraphicType,
    int *AsrChildNodeID
)
```

AsrNodeID	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
Acolumn,Arow	Spatial coordinates of a geometric region of interest in the DICOM image coordinate system. The IMAGE Content Item from which spatial coordinates are selected is denoted by a SELECTED FROM relationship. Data shall be provided as an ordered set of (column, row) pairs.
AgraphicType	This attribute defines the type of geometry of the annotated region of interest. The following Enumerated Values are specified for image spatial coordinate geometries: SR_GRT_POINT = a single pixel denoted by a single (column, row) pair SR_GRT_MULTIPOINT = multiple pixels each denoted by an (column, row) pair SR_GRT_POLYLINE = a series of connected line segments with ordered vertices denoted by (column, row) pairs; if the first and last vertices are the same it is a closed polygon SR_GRT_CIRCLE = a circle defined by two (column, row) pairs. The first point is the central pixel. The second point is a pixel on the perimeter of the circle. SR_GRT_ELLIPSE = an ellipse defined by four pixel (column, row) pairs, the first two points specifying the endpoints of the major axis and the second two points specifying the endpoints of the minor axis of an ellipse.

AsrChildNodeID

Pointer to the variable receiving an identifier of the new SR node.

Remarks

The data can be set by using MC_SRH_Set_SCOORD_First_Data and MC_SRH_Set_SCOORD_Next_Data.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table below.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_SCOORD_Next_Data

Appends the next coordinated pair to the SCOORD node.

Synopsis

```
MC_STATUS MC_SRH_Set_SCOORD_Next_Data (
    int AsrNodeID,
    float Acolumn,
    float Arow
)
```

AsrNodeID An Identifier of the SCOORD SR item.

Acolumn,Spatial coordinates of a geometric region of interest in the DICOM image coordinateAcolumn,system. The IMAGE Content Item from which spatial coordinates are selected is denotedArowby a SELECTED FROM relationship. Data shall be provided as an ordered set of (column, row) pairs.

[©] Copyright Merge Healthcare Solutions Inc. 2023

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table below.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_SCOORD_Node

Creates a new SCOORD node.

Synopsis

```
MC_STATUS MC_SRH_Create_SCOORD_Node (
    float Acolumn,
    float Arow,
    SR_GRAPHIC_TYPE AgraphicType,
    int *AsrNodeID
)
```

Acolumn, Arow Spatial coordinates of a geometric region of interest in the DICOM image coordinate system. The IMAGE Content Item from which spatial coordinates are selected is denoted by a SELECTED FROM relationship. Data shall be provided as an ordered set of (column, row) pairs.

AgraphicType	This attribute defines the type of geometry of the annotated region of interest. The following Enumerated Values are specified for image spatial coordinate geometries: SR_GRT_POINT = a single pixel denoted by a single (column, row) pair SR_GRT_MULTIPOINT = multiple pixels each denoted by an (column, row) pair SR_GRT_POLYLINE = a series of connected line segments with ordered vertices denoted by (column, row) pairs; if the first and last vertices are the same it is a closed polygon SR_GRT_CIRCLE = a circle defined by two (column, row) pairs. The first point is the central pixel. The second point is a pixel on the perimeter of the circle. SR_GRT_ELLIPSE = an ellipse defined by four pixel (column, row) pairs, the first two points specifying the endpoints of the minor axis of an ellipse.
AsrNodeID	Pointer to the variable receiving an identifier of the new SR node.

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_TCOORD_R_Child

Creates a new child TCOORD node under the provided parent where data are provided as Referenced Sample Positions.

Synopsis

```
MC STATUS MC SRH Add TCOORD R Child (
       int AsrNodeID,
       SR RELATIONSHIP Arelationship,
       unsigned long ArefSamplePosition,
       SR TRANGE TYPE AtempRangeType,
       int *AsrChildNodeID
)
                    An Identifier of the parent SR item to which the child is
AsrNodeID
                    to be added.
Arelationship
                    Enumerated value representing a relationship.
                    The first value from the list of samples within a multiplex
ArefSamplePosition group specifying temporal points of the referenced
                    data. Position of first sample is 1.
                    Defines the type of temporal extent of the region of
                    interest. A temporal point (or instant of time) may be
                    defined by a waveform sample offset (for a single
                    waveform multiplex group only), time offset, or absolute
                    time. Following enumerated values shall be used:
                    SR_TRT_POINT a single temporal point
                    SR_TRT_MULTIPOINT multiple temporal points
                    SR_TRT_SEGMENT a range between two temporal
AtempRangeType
                    points
                    SR_TRT_MULTISEGMENT multiple segments, each
                    denoted by two temporal points
                    SR_TRT_BEGIN a range beginning at one temporal
                    point, and extending beyond the end of the acquired
                    data
                    SR_TRT_END a range beginning before the start of the
                    acquired data, and extending to (and including) the
                    identified temporal point
                    Pointer to the variable receiving an identifier of the new
AsrChildNodeID
                    SR node.
Return Value
```

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning

© Copyright Merge Healthcare Solutions Inc. 2023

MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_TCOORD_R_Node

Creates a new child TCOORD node under the provided parent where data are provided as Referenced Sample Positions.

```
MC_STATUS MC_SRH_Create_TCOORD_R_Node (
    unsigned long ArefSamplePosition,
    SR_TRANGE_TYPE AtempRangeType,
    int *AsrNodeID
)
```

ArefSamplePosition	The first value from the list of samples within a multiplex group specifying temporal points of the referenced data. Position of first sample is 1.
	Defines the type of temporal extent of the region of interest. A temporal point (or instant of time) may be defined by a waveform sample offset (for a single waveform multiplex group only), time offset, or absolute time. Following enumerated values shall be used:
AtempRangeType	SR_TRT_POINT a single temporal point
	SR_TRT_MULTIPOINT multiple temporal points
	SR_TRT_SEGMENT a range between two temporal points
	SR_TRT_MULTISEGMENT multiple segments, each denoted by two temporal points

	SR_TRT_BEGIN a range beginning at one temporal point, and extending beyond the end of the acquired data
	SR_TRT_END a range beginning before the start of the acquired data, and extending to (and including) the identified temporal point
AsrNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_TCOORD_R_Next_Data

Adds a value to the existing TCOORD node where data are provided as Referenced Sample Positions.

Synopsis	
int AsrNod	_Set_TCOORD_R_Next_Data(eID, ong ArefSamplePosition
AsrNodeID	An Identifier of the TCOORD SR item.
ArefSamplePosition	Sequential value from the list of samples within a multiplex group specifying temporal points of the referenced data. Position of first sample is 1.

Return Value

[©] Copyright Merge Healthcare Solutions Inc. 2023

MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The <i>SRID</i> value is not a valid SR tree management object ID.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_TCOORD_O_Child

Creates a new child TCOORD node under the provided parent where data are provided as Referenced Time Offsets.

```
MC_STATUS MC_SRH_Add_TCOORD_O_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    const char *AtimeOffset,
    SR_TRANGE_TYPE AtempRangeType,
    int *AsrChildNodeID
)
```

AsrNodeID	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AtimeOffset	The first point from temporal points for reference by number of seconds after start of data Data shall be provided as a DICOM DS string format.
	Defines the type of temporal extent of the region of interest. A temporal point (or instant of time) may be defined by a waveform sample offset (for a single waveform multiplex group only), time offset, or absolute time. Following enumerated values shall be used:
	SR_TRT_POINT a single temporal point
	SR_TRT_MULTIPOINT multiple temporal points
AtempRangeType	SR_TRT_SEGMENT a range between two temporal points
	SR_TRT_MULTISEGMENT multiple segments, each denoted by two temporal points
	SR_TRT_BEGIN a range beginning at one temporal point, and extending beyond the end of the acquired data
	SR_TRT_END a range beginning before the start of the acquired data, and extending to (and including) the identified temporal point
AsrChildNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_TCOORD_O_Node

Creates a new TCOORD node where data are provided as Referenced Time Offsets.

Synopsis MC_STATUS MC_SRH_Create_TCOORD_O_Node (const char *AtimeOffset, SR_TRANGE_TYPE AtempRangeType, int *AsrNodeID)

AtimeOffsetThe first point from temporal points for reference byAtimeOffsetnumber of seconds after start of data Data shall be
provided as a DICOM DS string format.

	Defines the type of temporal extent of the region of interest. A temporal point (or instant of time) may be defined by a waveform sample offset (for a single waveform multiplex group only), time offset, or absolute time. Following enumerated values shall be used:
	SR_TRT_POINT a single temporal point
	SR_TRT_MULTIPOINT multiple temporal points
AtempRangeType	SR_TRT_SEGMENT a range between two temporal points
	SR_TRT_MULTISEGMENT multiple segments, each denoted by two temporal points
	SR_TRT_BEGIN a range beginning at one temporal point, and extending beyond the end of the acquired data
	SR_TRT_END a range beginning before the start of the acquired data, and extending to (and including) the identified temporal point
AsrNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_TCOORD_O_Next_Data

Adds a value to the existing TCOORD node where data are provided as Referenced Time Offsets.

Synopsis

```
MC_STATUS MC_SRH_Set_TCOORD_O_Next_Data (
    int AsrNodeID,
    const char *AtimeOffset
)
```

AsrNodelD An Identifier of the TCOORD SR item.

[©] Copyright Merge Healthcare Solutions Inc. 2023

AtimeOffset Sequential temporal point for reference by number of seconds after start of data Data shall be provided as a DICOM DS string format.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_TCOORD_D_Child

Creates a new child TCOORD node under the provided parent where data are provided as Referenced DateTime.

```
MC_STATUS MC_SRH_Add_TCOORD_D_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    const char *AdateTimeOffset,
    SR_TRANGE_TYPE AtempRangeType,
    int *AsrChildNodeID
)
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AdateTimeOffset	The first point from temporal points list for reference by absolute time. Data shall be provided as DICOM DT string.
AtempRangeType	Defines the type of temporal extent of the region of interest. A temporal point (or instant of time) may be defined by a waveform sample offset (for a single waveform multiplex group only), time offset, or absolute time. Following enumerated values shall be used: SR_TRT_POINT a single temporal point SR_TRT_MULTIPOINT multiple temporal points

	SR_TRT_SEGMENT a range between two temporal points
	SR_TRT_MULTISEGMENT multiple segments, each denoted by two temporal points
	SR_TRT_BEGIN a range beginning at one temporal point, and extending beyond the end of the acquired data
	SR_TRT_END a range beginning before the start of the acquired data, and extending to (and including) the identified temporal point
AsrChildNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_TCOORD_D_Node

Creates a new TCOORD node where data are provided as Referenced DateTime.

```
      MC_STATUS MC_SRH_Create_TCOORD_D_Node (

      const char *AdateTimeOffset,

      SR_TRANGE_TYPE AtempRangeType,

      int *AsrNodeID

      AsrNodeID

      An Identifier of the parent SR item to which the child is to be added.

      Arelationship
      Enumerated value representing a relationship.

      The first point from temporal points list for reference by absolute time. Data shall be provided as DICOM DT string.
```

	Defines the type of temporal extent of the region of
	interest. A temporal point (or instant of time) may be
	defined by a waveform sample offset (for a single
	waveform multiplex group only), time offset, or absolute
	time. Following enumerated values shall be used:
	SR_TRT_POINT a single temporal point
	SR_TRT_MULTIPOINT multiple temporal points
AtempRangeType	SR_TRT_SEGMENT a range between two temporal points
	SR_TRT_MULTISEGMENT multiple segments, each
	denoted by two temporal points
	SR_TRT_BEGIN a range beginning at one temporal point, and extending beyond the end of the acquired data
	SR_TRT_END a range beginning before the start of the acquired data, and extending to (and including) the identified temporal point
AsrChildNodeID	Pointer to the variable receiving an identifier of the new SR node.

Remarks

None.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_TCOORD_D_Next_Data

Adds a value to the existing TCOORD node where data are provided as Referenced DateTime.

Synopsis

```
MC_STATUS MC_SRH_Set_TCOORD_D_Next_Data (
    int AsrNodeID,
    const char *AdateTimeOffset
)
```

AsrNodelD An Identifier of the TCOORD SR item.

[©] Copyright Merge Healthcare Solutions Inc. 2023

AdateTimeOffsetSequential point from temporal points list for referenceAdateTimeOffsetby absolute time. Data shall be provided as DICOM DT
string.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_COMPOSITE_Child

Creates a new child COMPOSITE node under the provided parent.

Synopsis

```
MC_STATUS MC_SRH_Add_COMPOSITE_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    const char *AsopClassUid,
    const char *AsopInstanceUid,
    int *AsrChildNodeID
)
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AsopClassUid	A SOP Class UID of a Composite object which is not an Image or Waveform.
AsopInstanceUid	A SOP Instance UID of a Composite object which is not an Image or Waveform.
AsrChildNodelD	Pointer to the variable receiving an identifier of the new SR node.

Return Value

[©] Copyright Merge Healthcare Solutions Inc. 2023

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_COMPOSITE_Node

Creates a new COMPOSITE node.

Synopsis MC_STATUS MC_SRH_Create_COMPOSITE_Node (const char *AsopClassUid, const char *AsopInstanceUid, int *AsrNodeID) AsopClassUid ASOP Class UID of a Composite object which is not an Image or Waveform. AsopInstanceUid ASOP Instance UID of a Composite object which is not an Image or Waveform. AsrNodeID Pointer to the variable receiving an identifier of the new SR node.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_IMAGE_Child

Creates a new child IMAGE node under the provided parent.

Synopsis

```
MC_STATUS MC_SRH_Add_IMAGE_Child (
    int AsrNodeID,
    SR_RELATIONSHIP Arelationship,
    const char *AsopClassUid,
    const char *AsopInstanceUid,
    int *AsrChildNodeID
)
```

An Identifier of the parent SR item to which the child is to be added.
Enumerated value representing a relationship.
A SOP Class UID of an Image object.
A SOP Instance UID of an Image object.
Pointer to the variable receiving an identifier of the new SR node.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_IMAGE_Node

Creates a new IMAGE node.

Synopsis MC_STATUS MC_SRH_Create_IMAGE_Node (const char *AsopClassUid, const char *AsopInstanceUid, int *AsrNodeID

```
)
```

© Copyright Merge Healthcare Solutions Inc. 2023

AsopClassUid	An SOP Class UID of an Image object.
AsopInstanceUid	An SOP Instance UID of an Image object.
AsrChildNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_IMAGE_Frames

Sets frame numbers within Referenced SOP Instance to which the reference applies.

Synopsis

```
MC_STATUS MC_SRH_Set_IMAGE_Frames (
    int AsrNodeID,
    long AreferencedFrames[],
```

```
      long AframesArraySize

      )

      AsrNodelD

      An Identifier of the IMAGE SR item.

      Identifies the frame numbers within the Referenced

      AreferencedFrames

      SOP Instance to which the reference applies. The first frame shall be denoted as frame number 1.
```

AframesArraySize Size of the AreferencedFrames array.

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.

MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_WAVEFORM_Child

Creates a new child WAVEFORM node under the provided parent.

Synopsis

```
      MC_STATUS MC_SRH_Add_WAVEFORM_Child (

      int AsrNodeID,

      SR_RELATIONSHIP Arelationship,

      const char *AsopClassUid,

      const char *AsopInstanceUid,

      int *AsrChildNodeID

      )

      AsrNodeID

      An Identifier of the parent SR item to which the child is to be added.
```

SR node.

Arelationship	Enumerated value representing a relationship.
AsopClassUid	An SOP Class UID of a Waveform object.
AsopInstanceUid	An SOP Instance UID of a Waveform object.
AsrChildNodeID	Pointer to the variable receiving an identifier of the new

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.

MC_SYSTEM_ERROR

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_WAVEFORM_Node

Creates A new WAVEFORM node.

Synopsis

```
MC_STATUS MC_SRH_Create_WAVEFORM_Node (
    const char *AsopClassUid,
    const char *AsopInstanceUid,
    int *AsrNodeID
)
AsopClassUid An SOP Class UID of a Waveform object.
AsopInstanceUid An SOP Instance UID of a Waveform object.
```

AsrNodelD Pointer to the variable receiving an identifier of the new SR node.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_WAVEFORM_Channels

Sets channels referenced within the Referenced SOP Instance to which the reference applies.

```
Synopsis

MC_STATUS MC_SRH_Set_WAVEFORM_Channels (

int AsrNodeID,

unsigned short ArefChannels[],

long ArefChannelsSize

)

AsrNodeID An Identifier of the SR WAVEFORM node.
```

ArefChannels	A multi-value attribute which lists the channels referenced. Each channel is specified as a pair of values (M,C), where the first value is the sequence item number of the Waveform Sequence (5400,0100) attribute in the referenced object (i.e. the Multiplex Group Number), and the second value is the sequence item number of the Channel Definition Sequence (003A,0200) attribute (i.e., the Channel Number) within the multiplex group.
--------------	--

ArefChannelsSize Size of the ArefChannels array.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The ArefChannelsSize value is < 1.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_CONTAINER_Child

Creates a new child CONTAINER node under the provided parent.

Synopsis

```
      MC_STATUS MC_SRH_Add_CONTAINER_Child (

      int AsrNodeID,

      SR_RELATIONSHIP Arelationship,

      SR_CONT_CONTINUITY Acontinuity,

      const char *AtemplateId,

      int *AsrChildNodeID

      )

      AsrNodeID

      An Identifier of the parent SR item to which the child is to be added.
```

Arelationship	Enumerated value representing a relationship.
---------------	---

AcontinuityEnumerated value specifies whether or not its containedAcontinuityContent Items are logically linked in a continuous textual
flow, or are separate items.

AtemplateId Template Identifer that describes the content of this Content Item and its subsidiary Content Items.

AsrChildNodeID Pointer to the variable receiving an identifier of the new SR node.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_CONTAINER_Node

Creates a new CONTAINER node.

Synopsis

```
      MC_STATUS MC_SRH_Create_CONTAINER_Node (
SR_CONT_CONTINUITY Acontinuity,
const char *AtemplateId,
int *AsrNodeID

      )
      Enumerated value specifies whether or not its contained
Content Items are logically linked in a continuous textual flow,
or are separate items.

      AtemplateId
      DICOM Template Identifer that describes the content of this
Content Item and its subsidiary Content Items. Can be NULL.

      AsrNodeID
      Pointer to the variable receiving an identifier of the new SR
node.
```

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value

Meaning

© Copyright Merge Healthcare Solutions Inc. 2023

MC_NORMAL_COMPLETION MC_NULL_POINTER_PARM MC_SYSTEM_ERROR

The function completed normally.

One or more of the pointer-type parameters was NULL.

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_TABLE_Child

Create a new child TABLE node under the provided parent.

```
Synopsis
```

```
MC_STATUS MC_SRH_Add_TABLE_Child (
        SR_RELATIONSHIP Arelationship,
        const char *AconceptNameValue,
        const char *AconceptNameScheme,
        const char *AconceptNameMeaning,
        int Arows,
        int Acolumns,
        TABLE_ROW_DEF *AtableRowDefs,
        int AtableRowDefsCount,
        TABLE_COL_DEF *AtableColDefs,
        int AtableColDefsCount,
        TABLE_CELL_VALUES *AtableCellValues,
        int AtableCellValuesCount,
        int AtableCellValuesCount,
        int *AsrChildNodeID
)
```

AsrNodeID	An Identifier of the parent SR item to which the child is to be added.
Arelationship	Enumerated value representing a relationship.
AconceptNameValue	Code Value describing the concept represented by this Content Item.
AconceptNameScheme	Coding Scheme Designator
AconceptNameMeaning	Code Meaning
Arows	The number of rows in this TABLE item
Acolumns	The number of columns in this TABLE item
AtableRowDefs	The concepts that define the meaning of the rows of the table
AtableRowDefsCount	The number of items in the Table Row Definition Sequence
AtableColDefs	The concepts that define the meaning of the columns of the table

AtableColDefsCount	The number of items in the Table Column Definition Sequence
AtableCellValues	The values of each populated cell in the table, identified by row and column
AtableCellValuesCount	The number of items in the Cell Values Sequence
AsrChildNodeID	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Create_TABLE_Node

Creates a new TABLE node.

```
Synopsis
MC_STATUS MC_SRH_Create_TABLE_Node (
      const char *AconceptNameValue,
      const char *AconceptNameScheme,
      const char *AconceptNameMeaning,
      int Arows,
      int Acolumns,
      TABLE_ROW_DEF *AtableRowDefs,
      int AtableRowDefsCount,
      TABLE COL DEF *AtableColDefs,
      int AtableColDefsCount,
      TABLE CELL VALUES *AtableCellValues,
      int AtableCellValuesCount,
      int *AsrNodeID
)
                      Code Value describing the concept represented
AconceptNameValue
                      by this Content Item.
AconceptNameScheme
                      Coding Scheme Designator
```

AconceptNameMeaning	Code Meaning
Arows	The number of rows in this TABLE item
Acolumns	The number of columns in this TABLE item
AtableRowDefs	The concepts that define the meaning of the rows of the table
AtableRowDefsCount	The number of items in the Table Row Definition Sequence
AtableColDefs	The concepts that define the meaning of the columns of the table
AtableColDefsCount	The number of items in the Table Column Definition Sequence
AtableCellValues	The values of each populated cell in the table, identified by row and column
AtableCellValuesCount	The number of items in the Cell Values Sequence
AsrNodelD	Pointer to the variable receiving an identifier of the new SR node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_TABLE_Next_Row_Definition

Add a new item to the Table Row Definition Sequence in the TABLE Content Item.

Synopsis

MC	STATUS	MC	SRH	Set	TABLE	Next	Row	Definition	(
	int	Asi	rNode	eID,				-	
	TAB	LE_E	ROW_I)EF 🤊	*Atable	eRowDe	ef		
)			_						

AsrNodeID An Identifier of the TABLE SR item.

AtableRowDefPointer to the structure representing the Table RowDefinition Sequence item to be added to the TABLE
Content Item

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_TABLE_Next_Column_Definition

Add a new item to the Table Column Definition Sequence in the TABLE Content Item.

Synopsis

```
      MC_STATUS MC_SRH_Set_TABLE_Next_Column_Definition (

      int AsrNodeID,

      TABLE_COL_DEF *AtableColDef

      )

      AsrNodelD

      An Identifier of the TABLE SR item.

      Pointer to the structure representing the Table Column

      AtableColDef

      Definition Sequence item to be added to the TABLE

      Content Item
```

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Set_TABLE_Next_Cell_Values

Creates a new CONTAINER node.

Synopsis		
<pre>MC_STATUS MC_SRH_Set_TABLE_Next_Cell_Values (int AsrNodeID,</pre>		
TABLE_CELL_VALUES *AtableCellValues		
)		
AsrNodelD	An Identifier of the TABLE SR item.	
AtableCellValues	Pointer to the structure representing the Cell Values Sequence item to be added to the TABLE	

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Add_Reference

Adds a Referenced Content Item.

Synopsis MC_STATUS MC_SRH_Add_Reference (int AsrNodeID, SR_RELATIONSHIP Arelationship, int AsrRefNodeID) AsrNodelD An Identifier of the parent SR item to which the child is to be added. Arelationship Enumerated value representing a relationship. AsrRefNodelD SR node that will be referenced from the current node.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_VALUE_NOT_ALLOWED	The Arelationship value is not valid.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_Reference

Returns a Referenced Content Item.

Synopsis

```
MC_STATUS MC_SRH_Get_Reference (
    int AsrNodeID,
    int *AsrRefNodeID
)
```

AsrNodelD	An Identifier of the parent SR item to which the child is to be
ASINUUEID	added.

AsrRefNodelD SR node that is referenced from the current node.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_NodeType

Returns node type for the specified SR node.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Synopsis	Synopsis		
MC_STATUS MC_SRH_Get_NodeType (int AsrNodeID, SR_CONTENT_TYPE *AnodeType)			
AsrNodeID	An Identifier of the SR node.		
AnodeType	Pointer to the variable that receives an enumerated node type from the list below		

Enumerated values for the Node Type

	•
SR_NODE_TEXT	Free text, narrative description of unlimited length. May also be used to provide a label or identifier value.
SR_NODE_NUM	Numeric value fully qualified by coded representation of the measurement name and unit of measurement.
SR_NODE_CODE	Categorical coded value. Representation of nominal or non-numeric ordinal values.
SR_NODE_DATETIME	Date and time of occurrence of the type of event denoted by the Concept Name.
SR_NODE_DATE	Date of occurrence of the type of event denoted by the Concept Name.
SR_NODE_TIME	Time of occurrence of the type of event denoted by the Concept Name.
SR_NODE_UIDREF	Unique Identifier (UID) of the entity identified by the Concept Name.
SR_NODE_PNAME	Person name of the person whose role is described by the Concept Name.
SR_NODE_COMPOSITE	A reference to one Composite SOP Instance which is not an Image or Waveform.
SR_NODE_IMAGE	A reference to one Image. IMAGE Content Item may convey a reference to a Softcopy Presentation State associated with the Image.
SR_NODE_WAVEFORM	A reference to one Waveform.
SR_NODE_SCOORD	Spatial coordinates of a geometric region of interest in the DICOM image coordinate system. The IMAGE Content Item from which spatial coordinates are selected is denoted by a SELECTED FROM relationship.
SR_NODE_TCOORD_D	Temporal Coordinates (i.e. time or event based coordinates) of a region of interest in the DICOM waveform coordinate system. The WAVEFORM or IMAGE or SCOORD Content Item from which Temporal Coordinates are selected is denoted by a SELECTED

FROM relationship. The value is stored as a Referenced DateTime here.

- SR_NODE_TCOORD_O Temporal Coordinates (i.e. time or event based coordinates) of a region of interest in the DICOM waveform coordinate system. The WAVEFORM or IMAGE or SCOORD Content Item from which Temporal Coordinates are selected is denoted by a SELECTED FROM relationship. The value is stored as Referenced Time Offsets here.
- SR_NODE_TCOORD_R
 Temporal Coordinates (i.e. time or event based coordinates) of a region of interest in the DICOM waveform coordinate system. The WAVEFORM or IMAGE or SCOORD Content Item from which Temporal Coordinates are selected is denoted by a SELECTED FROM relationship. The value is stored as Referenced Sample Positions here.

 SR_NODE_CONTAINER
 CONTAINER groups Content Items and defines the baseding or category of observation that applies to that
- SR_REFERENCE
 CONTAINER gloups content terms and domines the heading or category of observation that applies to that content. The heading describes the content of the CONTAINER Content Item and may map to a document section heading in a printed or displayed document.

 SR_REFERENCE
 This item is not a child SR content node. It is a reference

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

to another node.

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_First_Child

Returns the ID of the first child in the list of children that the parent node maintains.

```
MC_STATUS MC_SRH_Get_First_Child (
    int AsrNodeID,
    int *AsrChildNodeID,
```

[©] Copyright Merge Healthcare Solutions Inc. 2023

<pre>SR_RELATIONSHIP *Arelationship, SR_CONTENT_TYPE *AnodeType, int *AisLast)</pre>			
AsrNodeID	The identifier assigned to this SR tree item.		
AsrChildNodelD	Upon successful completion, the item object identifier of the first SR tree management node in the SR tree management entity AsrChildNodeID is returned here. At a given parent <i>SRID</i> a pointer to the next record is maintained.		
Arelationship	Enumerated value representing a relationship between parent and child nodes SR_REL_CONTAINS, SR_REL_HAS_OBS_CONTEXT, SR_REL_HAS_ACQ_CONTEXT, SR_REL_HAS_CONCEPT_MOD, SR_REL_HAS_PROPERTIES, SR_REL_INFERRED_FROM, SR_REL_SELECTED_FROM		
AnodeType	Returns an enumerated node type. See MC_SRH_Get_NodeType function.		
AisLast	Upon successful completion, this parameter is set to true (non-zero) if the first record in the SR tree management entity AsrChildNodeID is also the last record (i.e. it is the only element)		

Remarks

This function is an extension of the MC_SR_Get_First_Child function and it is using it internally.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_Next_Child

Returns the ID of the next child in the list of children that the parent node maintains.

Synopsis

MC_STATUS MC_SRH_Get_Next_Child (

[©] Copyright Merge Healthcare Solutions Inc. 2023

Merge DICOM Toolkit™ 5.16.0 C/C++ REFERENCE MANUAL

int AsrNodeID,		
	ChildNodeID,	
_	'IONSHIP *Arelationship, NT TYPE *AnodeType,	
int *Ais		
)		
1		
AsrNodeID	The identifier assigned to this SR tree item.	
AsrChildNodeID	Upon successful completion, the item object identifier of the next SR tree management node in the SR tree management entity AsrChildNodeID is returned here. At a given parent <i>SRID</i> a pointer to the next record is maintained.	
Arelationship	Enumerated value representing a relationship between parent and child nodes SR_REL_CONTAINS, SR_REL_HAS_OBS_CONTEXT, SR_REL_HAS_ACQ_CONTEXT, SR_REL_HAS_CONCEPT_MOD, SR_REL_HAS_PROPERTIES, SR_REL_INFERRED_FROM, SR_REL_SELECTED_FROM	
AnodeType	Returns an enumerated node type. See MC_SRH_Get_NodeType function.	
AisLast	Upon successful completion, this parameter is set to true (non-zero) if the next record in the SR tree management entity AsrChildNodeID is also the last record.	

Remarks

This function is an extension of the MC_SR_Get_Next_Child function and it is using it internally.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_Concept_Name

Returns the Concept Name values for the node.

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

Synopsis

```
MC_STATUS MC_SRH_Get_Concept_Name (
    int AsrNodeID,
    char AconceptNameValue[],
    int AconceptNameValueSize,
    char AconceptNameScheme[],
    int AconceptNameSchemeSize,
    char AconceptNameMeaning[],
    int AconceptNameMeaningSize
)
```

AsrNodeID	An Identifier of the SR item.	
AconceptNameValue	Returns Code Value describing the concept represented by this Content Item.	
AconceptNameValueSize	Code Value buffer size	
AconceptNameScheme	Returns Coding Scheme Designator	
AconceptNameSchemeSize Coding Scheme Designator buffer size		
AconceptNameMeaning	Returns Code Meaning	
AconceptNameMeaningSize	Code Meaning buffer size	

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TEXT_Data

Returns a TEXT node data.

```
MC_STATUS MC_SRH_Get_TEXT_Data (
    int AsrNodeID,
    char Atext[],
    int AtextSize
)
```

AsrNodelD An Identifier of the TEXT SR item.

Atext Returns A text value stored in this node

AtextSize Text buffer size

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_CODE_Data

Returns the Code Name values for the node.

```
MC_STATUS MC_SRH_Get_CODE_Data (
    int AsrNodeID,
    char AconceptCodeValue[],
    int AconceptCodeValueSize,
    char AconceptCodingScheme[],
    int AconceptCodingSchemeSize,
    char AconceptCodeMeaning[],
    int AconceptCodeMeaningSize
)
```

```
AsrNodelDAn Identifier of the CODE SR item.AconceptCodeValueReturns Code ValueAconceptCodeValueSizeCode Value buffer sizeAconceptCodingSchemeReturns Coding Scheme DesignatorAconceptCodeMeaningSizeCode MeaningAconceptCodeMeaningSizeCode Meaning buffer size
```

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_NUM_Data

Reads value from the NUM node.

```
MC_STATUS MC_SRH_Get_NUM_Data (
    int AsrNodeID,
    char AnumericValue[],
    int AnumericValueSize,
    int *AvalueCount,
    char AunitsCodeValue[],
    int AunitsCodeValueSize,
    char AunitsCodingScheme[],
    int AunitsCodeMeaning[],
    int AunitsCodeMeaningSize
```

```
)
```

AsrNodeID	An Identifier of the NUM SR item.
AnumericValue	Numeric Value
AnumericValueSize	Numeric Value buffer size
AvalueCount	Number of values stored
AunitsCodeValue	Coded expression of measurement units
AunitsCodeValueSize	AunitsCodeValue buffer size
AunitsCodingScheme	Coding Scheme Designator of the measurement units.
AunitsCodingSchemeSize	AunitsCodingScheme buffer size
AunitsCodeMeaning	Code Meaning of the measurement units.

AunitsCodeMeaningSize AunitsCodeMeaning buffer size

Remarks

The Measured Value Sequence (0040, A300) may be empty to convey the concept of a measurement whose value is unknown or missing, or a measurement or calculation failure in that case function will return the MC_EMPTY_VALUE code.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_NUM_Next_Data

Reads next value from the NUM node.

```
Synopsis

MC_STATUS MC_SRH_Get_NUM_Next_Data (

int AsrNodeID,

char AnumericValue[],

int AnumericValueSize

)

AsrNodeID An Identifier of the NUM SR item.

AnumericValue Numeric Value

AnumericValueSize Numeric Value buffer size
```

Remarks

The Measured Value Sequence (0040,A300) may be empty to convey the concept of a measurement whose value is unknown or missing, or a measurement or calculation failure in that case function will return the MC_EMPTY_VALUE code.

[©] Copyright Merge Healthcare Solutions Inc. 2023

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_NUM_Qualifier

Gets Qualification of Numeric Value in Measured Value Sequence, or reason for absence of Measured Value Sequence Item.

```
MC_STATUS MC_SRH_Get_NUM_Qualifier (
    int AsrNodeID,
    char AqualifierCodeValue[],
    int AqualifierCodeValueSize,
    char AqualifierCodingScheme[],
    int AqualifierCodingSchemeSize,
    char AqualifierCodeMeaning[],
    int AqualifierCodeMeaningSize
```

```
)
```

AsrNodeID	An Identifier of the NUM SR item.
AqualifierCodeValue	Sets Qualification of Numeric Value
AqualifierCodeValueSize	Size of AqualifierCodeValue buffer
AqualifierCodingScheme	Coding Scheme Designator
AqualifierCodingSchemeSize	Size of AqualifierCodingScheme buffer
AqualifierCodeMeaning	Code Meaning
AqualifierCodeMeaningSize	Size of AqualifierCodeMeaning buffer

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_DATETIME_Data

Returns a DATETIME node data.

Synopsis

```
MC_STATUS MC_SRH_Get_DATETIME_Data (
    int AsrNodeID,
    char AdateTime[],
    int AdateTimeSize
)
AsrNodeID An Identifier of the DATETIME SR item.
```

AdateTime Returns a text value stored in this node

AdateTimeSize Datetime buffer size

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_DATE_Data

Returns a DATE node data.

Synopsis

```
MC_STATUS MC_SRH_Get_DATE_Data (
    int AsrNodeID,
    char Adate[],
    int AdateSize
)
```

AsrNodelD An Identifier of the DATE SR item.

Adate Returns A text value stored in this node.

AdateSize Date buffer size

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TIME_Data

Returns a TIME node data.

Synopsis MC_STATUS MC_SRH_Get_TIME_Data (int AsrNodeID, char Atime[], int AtimeSize)

AsrNodelD An Identifier of the TIME SR item.

Atime Returns a text value stored in this node.

AtimeSize Time buffer size

[©] Copyright Merge Healthcare Solutions Inc. 2023

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_UIDREF_Data

Returns a UIDREF node data.

Synopsis

```
MC_STATUS MC_SRH_Get_UIDREF_Data (
    int AsrNodeID,
    char AuidRef[],
    int AuidRefSize
)
```

AsrNodelD An Identifier of the UIDREF SR item.

AuidRef Returns a uid value stored in this node.

AuidRefSize uidRef buffer size

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_PNAME_Data

Returns a PNAME node data.

```
Synopsis
MC STATUS MC SRH Get PNAME Data (
      int AsrNodeID,
      char ApersonName[],
      int ApersonNameSize
)
AsrNodeID
                   An Identifier of the PNAME SR item.
ApersonName
                   Returns a person name value stored in this node.
ApersonNameSize ApersonName buffer size
```

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_SCOORD_First_Data

Returns an SCOORD node data.

```
Synopsis
MC_STATUS MC_SRH_Get_SCOORD_First_Data (
      int AsrNodeID,
      float *Acolumn,
      float *Arow,
      int *AvalueCount,
      SR GRAPHIC TYPE *AgraphicType
)
```

AsrNodeID An Identifier of the SCOORD SR item.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Acolumn, Arow	Spatial coordinates of a geometric region of interest in the DICOM image coordinate system. The IMAGE Content Item from which spatial coordinates are selected is denoted by a SELECTED FROM relationship. Data shall be provided as an ordered set of (column, row) pairs.
AvalueCount	Returns number of value pairs stored.
AgraphicType	This attribute defines the type of geometry of the annotated region of interest. The following Enumerated Values are specified for image spatial coordinate geometries: SR_GRT_POINT = a single pixel denoted by a single (column, row) pair SR_GRT_MULTIPOINT = multiple pixels each denoted by an (column, row) pair SR_GRT_POLYLINE = a series of connected line segments with ordered vertices denoted by (column, row) pairs; if the first and last vertices are the same it is a closed polygon SR_GRT_CIRCLE = a circle defined by two (column, row) pairs. The first point is the central pixel. The second point is a pixel on the perimeter of the circle. SR_GRT_ELLIPSE = an ellipse defined by four pixel (column, row) pairs, the first two points specifying the endpoints of the major axis and the second two points specifying the endpoints of the minor axis of an ellipse.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_SCOORD_Next_Data

Returns an SCOORD node data.

Synopsis

MC_STATUS MC_SRH_Get_SCOORD_Next_Data (
 int AsrNodeID,

 $[\]ensuremath{\textcircled{O}}$ Copyright Merge Healthcare Solutions Inc. 2023

```
float *Acolumn,<br/>float *Arow)AsrNodelDAn Identifier of the SCOORD SR item.Spatial coordinates of a geometric region of interest in the<br/>DICOM image coordinate system. The IMAGE Content Item<br/>from which spatial coordinates are selected is denoted by a<br/>SELECTED FROM relationship. Data shall be provided as an
```

ordered set of (column, row) pairs.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table below.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TCOORD_D_First_Data

Returns a TCOORD node's data that were provided as a Referenced DateTime.

```
MC_STATUS MC_SRH_Get_TCOORD_D_First_Data (
    int AsrNodeID,
    char AdateTimeOffset[],
    int AdateTimeOffsetSize,
    int *AvalueCount,
    SR_TRANGE_TYPE *AtempRangeType
)
```

AsrNodeID	An Identifier of the TCOORD SR item.
AdateTimeOffset	Specifies temporal point for reference by absolute time. Data is formated as DICOM DT string.
AdateTimeOffsetSize	Specifies temporal point buffer size.
AvalueCount	Returns number of values stored.
	Defines the type of temporal extent of the region of interest. A temporal point (or instant of time) may be defined by a waveform sample offset (for a single waveform multiplex group only), time offset, or absolute time. Following enumerated values shall be used:
	SR_TRT_POINT a single temporal point
	SR_TRT_MULTIPOINT multiple temporal points
AtempRangeType	SR_TRT_SEGMENT a range between two temporal points
	SR_TRT_MULTISEGMENT multiple segments, each denoted by two temporal points
	SR_TRT_BEGIN a range beginning at one temporal point, and extending beyond the end of the acquired data
	SR_TRT_END a range beginning before the start of the acquired data, and extending to (and including) the identified temporal point

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TCOORD_D_Next_Data

Returns a TCOORD node's data that were provided as a Referenced DateTime.

Synopsis

```
MC STATUS MC SRH Get TCOORD D Next Data (
       int AsrNodeID,
       char AdateTimeOffset[],
       int AdateTimeOffsetSize
)
AsrNodeID
                        An Identifier of the TCOORD SR item.
                        Specifies temporal points for reference by absolute
AdateTimeOffset
                        time. Data is formatted as DICOM DT string.
                        Specifies temporal points for buffer size. Note that it is
AdateTimeOffsetsSize
                        a multi-value string that can long.
AvalueCount
                        Returns number of values stored.
                        Defines the type of temporal extent of the region of
                        interest. A temporal point (or instant of time) may be
                        defined by a waveform sample offset (for a single
                        waveform multiplex group only), time offset, or absolute
                        time. Following enumerated values shall be used:
                        SR_TRT_POINT a single temporal point
                        SR_TRT_MULTIPOINT multiple temporal points
                        SR_TRT_SEGMENT a range between two temporal
AtempRangeType
                        points
                        SR_TRT_MULTISEGMENT multiple segments, each
                        denoted by two temporal points
                        SR_TRT_BEGIN a range beginning at one temporal
                        point, and extending beyond the end of the acquired
                        data
                        SR_TRT_END a range beginning before the start of the
                        acquired data, and extending to (and including) the
                        identified temporal point
```

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.

MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error
	has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TCOORD_O_First_Data

Returns a TCOORD node's data that were provided as Referenced Time Offsets.

```
MC_STATUS MC_SRH_Get_TCOORD_O_First_Data (
       int AsrNodeID,
       char AtimeOffset[],
       int AtimeOffsetSize,
       int *AvalueCount,
       SR TRANGE TYPE *AtempRangeType
)
AsrNodeID
                      An Identifier of the TCOORD SR item.
                      Specifies temporal points for reference by number of
AtimeOffset
                      seconds after start of data.
AdateTimeOffsetSize Specifies temporal point buffer size.
AvalueCount
                      Returns number of values stored.
                     Defines the type of temporal extent of the region of
                     interest. A temporal point (or instant of time) may be
                      defined by a waveform sample offset (for a single
                     waveform multiplex group only), time offset, or
                     absolute time. Following enumerated values shall be
                      used:
                     SR_TRT_POINT a single temporal point
                     SR_TRT_MULTIPOINT multiple temporal points
                     SR_TRT_SEGMENT a range between two temporal
AtempRangeType
                      points
                     SR_TRT_MULTISEGMENT multiple segments, each
                     denoted by two temporal points
                     SR_TRT_BEGIN a range beginning at one temporal
                      point, and extending beyond the end of the acquired
                      data
                     SR_TRT_END a range beginning before the start of the
                     acquired data, and extending to (and including) the
                      identified temporal point
```

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TCOORD_O_Next_Data

Returns a TCOORD node's data that were provided as Referenced Time Offsets.

Synopsis

```
      MC_STATUS MC_SRH_Get_TCOORD_O_Next_Data (

      int AsrNodeID,

      char AtimeOffset[],

      int AtimeOffsetSize

      )

      AsrNodeID

      An Identifier of the TCOORD SR item.

      Specifies temporal points for reference by number of
```

AdateTimeOffsetSize Specifies temporal point buffer size.

Return Value

AtimeOffset

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

seconds after start of data.

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table below.

MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by
	the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TCOORD_R_First_Data

Returns a TCOORD node's data that were provided as Referenced Sample Positions.

Synopsis MC STATUS MC SBH Get TCOORD B First Data (
MC STATUS MC SRH Get TCOORD R First Data (
<pre>MC_STATUS MC_SRH_Get_TCOORD_R_First_Data (int AsrNodeID, unsigned long *ArefSamplePosition, int *AvalueCount, SR_TRANGE_TYPE *AtempRangeType)</pre>		
AsrNodelD An Identifier of the TCOORD SR item.		
List of samples within a multiplex group specifying <i>ArefSamplePosition</i> temporal points of the referenced data. Position of fin sample is 1. Only the first value returned here.	st	
AvalueCount Returns number of values stored.		
AtempRangeTypeDefines the type of temporal extent of the region of interest. A temporal point (or instant of time) may be defined by a waveform sample offset (for a single waveform multiplex group only), time offset, or absol time. Following enumerated values shall be used: SR_TRT_POINT a single temporal point SR_TRT_MULTIPOINT multiple temporal points SR_TRT_SEGMENT a range between two temporal pointsSR_TRT_MULTISEGMENT multiple segments, each denoted by two temporal pointsSR_TRT_BEGIN a range beginning at one temporal 	ute	

SR_TRT_END a range beginning before the start of the acquired data, and extending to (and including) the identified temporal point

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TCOORD_R_Next_Data

Returns a TCOORD node's data that were provided as Referenced Sample Positions.

Synopsis

<pre>MC_STATUS MC_SRH_Get_TCOORD_R_Next_Data (int AsrNodeID, unsigned long *ArefSamplePosition)</pre>		
AsrNodeID	An Identifier of the TCOORD SR item.	
ArefSamplePosition	List of samples within a multiplex group specifying temporal points of the referenced data. Position of the first sample is 1. Only the first value is returned here.	

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).

MC_SYSTEM_ERROR

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_COMPOSITE_Data

Returns a COMPOSITE node data.

Synopsis

```
MC_STATUS MC_SRH_Get_COMPOSITE_Data (
    int AsrNodeID,
    char AsopClassUid[],
    int AsopClassUidSize,
    char AsopInstanceUid[],
    int AsopInstanceUidSize
)
```

AsrNodeID	An Identifier of the COMPOSITE SR item.
AsopClassUid	A SOP Class UID of a Composite object which is not an Image or Waveform.
AsopClassUidSize	The size of the SOP Class UID buffer.
AsopInstanceUid	A SOP Instance UID of a Composite object which is not an Image or Waveform.
AsopInstanceUidSize	The size of the SOP Instance UID buffer.

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_IMAGE_Data

Returns an IMAGE node data.

Synopsis

MC_STATUS MC_SRH_Get_IMAGE_Data (
<pre>int AsrNodeID,</pre>		
char AsopCl		
int AsopCla		
-	stanceUid[],	
int *ArefFr	tanceUidSize,	
· · · · · · · · · · · · · · · · · · ·	amescount	
)		
AsrNodeID	An Identifier of the IMAGE SR item.	
AsopClassUid	An SOP Class UID of a Composite object which is an Image.	
AsopClassUidSize	The size of the SOP Class UID buffer.	
AsopInstanceUid	An SOP Instance UID of a Composite object which is an Image.	
AsopInstanceUidSize	The size of the SOP Instance UID buffer.	
ArefFramesCount	Returns a count of references frames within the Referenced SOP Instance to which the reference applies. Frames array can be retrieved via MC_SRH_Get_IMAGE_Frames.	

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_IMAGE_Frames

Gets frame numbers within a Referenced SOP Instance to which the reference applies.

Synopsis

MC_STATUS MC_SRH_Get_IMAGE_Frames (
 int AsrNodeID,
 long AreferencedFrames[],

 $[\]textcircled{O}$ Copyright Merge Healthcare Solutions Inc. 2023

long Afram	esArraySize
AsrNodeID	An Identifier of the IMAGE SR item.
AreferencedFrames	Identifies the frame numbers within the Referenced SOP Instance to which the reference applies. The first frame shall be denoted as frame number 1.
AframesArraySize	Size of the AreferencedFrames array.

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INCOMPATIBLE_VR	The attribute's value representation cannot be derived from <i>Value</i> . See the table below.
MC_VALUE_OUT_OF_RANGE	A numeric <i>Value</i> was larger than could be accommodated by the attribute. (E.g. setting an attribute with an unsigned int VR to the value 123.45, or setting it to 128000, or setting it to -1).
MC_TEMP_FILE_ERROR	If the attribute's value is large, it may be stored in a temporary file (if so configured). If a file I/O error occurs, this status is returned.
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_WAVEFORM_Data

Returns an IMAGE node data.

Synopsis

```
MC_STATUS MC_SRH_Get_WAVEFORM_Data (
    int AsrNodeID,
    char AsopClassUid[],
    int AsopClassUidSize,
    char AsopInstanceUid[],
    int AsopInstanceUidSize,
    int *ArefChannelsCount
)
```

AsrNodelD An Identifier of the WAVEFORM SR item.

[©] Copyright Merge Healthcare Solutions Inc. 2023

AsopClassUid	An SOP Class UID of a Composite object which is a Waveform.
AsopClassUidSize	Size of the SOP Class UID buffer.
AsopInstanceUid	An SOP Instance UID of a Composite object which is a Waveform.
AsopInstanceUidSize	Size of the SOP Instance UID buffer.
ArefChannelsCount	A value count for the channels referenced. Channels array can be retrieved via MC_SRH_Get_WAVEFORM_Channels.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_WAVEFORM_Channels

Gets channels referenced within the Referenced SOP Instance to which the reference applies.

```
Synopsis
MC_STATUS MC_SRH_Get_WAVEFORM_Channels (
    int AsrNodeID,
    short ArefChannels[],
    long ArefChannelsSize
)
```

AsrNodelD An Identifier of the WAVEFORM SR item.

ArefChannels	A multi-value attribute which lists the channels referenced. Each channel is specified as a pair of values (M,C), where the first value is the sequence item number of the Waveform Sequence (5400,0100) attribute in the referenced object (i.e. the Multiplex Group Number), and the second value is the sequence item number of the Channel Definition Sequence (003A, 0200) attribute (i.e., the Channel Number) within the multiplex group.
ArefChannelsSize	Size of the ArefChannels array that shall be enough to fit all data.

One of the enumerated MC_STATUS codes defined in "mcstatus.h"

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TABLE_First_Cell_Values_Counts

Retrieve the value counts of the internal attributes of the first item in the Cell Values Sequence in the TABLE Content Item. These value counts will be necessary for allocating the memory space for the subsequent retrieval of those attributes.

<pre>MC_STATUS MC_SRH_Get_TAB int AsrNodeID, int *AvalueCount, int *ArefContentIt int *AselectorAttr unsigned long *Ase int *AcellValueSID)</pre>	ValueCount, lectorAttrTag,
AsrNodeID	An Identifier of the TABLE content item.
AvalueCount	Returning the number of items in the Cell Values Sequence

ArefContentItemIdCount	Returning the number of values in the Referenced Content Item Identifier.
AselectorAttrValueCount	Returning the number of values in the Selector XX Value data element denoted by the Selector Attribute VR.
AselectorAttrTag	Returning the Selector XX Value data element denoted by the Selector Attribute VR.
AcellValuesID	Returning the numeric handle of the first item in the Cell Value Sequence.

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TABLE_First_Column_Definition

Retrieve the data of the first item of the Table Column Definition Sequence in the TABLE Content Item.

```
MC_STATUS MC_SRH_Get_TABLE_First_Column_Definition (
    int AsrNodeID,
    TABLE_ROW_DEF *AtableColDef,
    int *AvalueCount
)
AsrNodelD An Identifier of the TIME SR item.
```

AtableColDef	Pointer to a supplied structure in which the next Table
AlableColDel	Column Definition Sequence item is to be retrieved.

[©] Copyright Merge Healthcare Solutions Inc. 2023

AvalueCount Returns the number of items stored in the Table Column Definition Sequence

Return Value

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TABLE_First_Row_Definition

Retrieve the data of the first item of the Table Row Definition Sequence in the TABLE Content Item.

Synopsis

<pre>MC_STATUS MC_SRH_Get_TABLE_First_Row_Definition (int AsrNodeID, TABLE_ROW_DEF *AtableRowDef, int *AvalueCount)</pre>		
AsrNodeID	An Identifier of the TIME SR item.	
AtableRowDef	Pointer to a supplied structure in which the next Table Row Definition Sequence item is to be retrieved.	
AvalueCount	Returns the number of items stored in the Table Row Definition Sequence	

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.

MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in
	the operating environment. A message describing the error
	has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TABLE_Next_Cell_Values_Counts

Retrieve the value counts of the internal attributes of the next Cell Values Sequence item in the TABLE Content Item. These value counts will be necessary for allocating the memory space for the subsequent retrieval of those attributes.

Synopsis

```
MC_STATUS MC_SRH_Get_TABLE_Next_Cell_Values_Counts (
    int AsrNodeID,
    int *ArefContentItemIdCount,
    int *AselectorAttrValueCount,
    unsigned long *AselectorAttrTag,
    int *AcellValueSID
)
```

AsrNodeID	An Identifier of the TABLE content item.
ArefContentItemIdCount	Returning the number of values in the Referenced Content Item Identifier.
AselectorAttrValueCount	Returning the number of values in the Selector XX Value data element denoted by the Selector Attribute VR.
AselectorAttrTag	Returning the Selector XX Value data element denoted by the Selector Attribute VR.
AcellValuesID	Returning the numeric handle of the first item in the Cell Value Sequence.

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.

© Copyright Merge Healthcare Solutions Inc. 2023

MC_NULL_VALUE MC_SYSTEM_ERROR

The attribute's value is NULL (i.e. its length is zero).

An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TABLE_Next_Column_Definition

Retrieve the data in the next item of the Table Column Definition Sequence in the TABLE Content Item.

Synopsis MC_STATUS MC_SRH_Get_TABLE_Next_Column_Definition (int AsrNodeID, TABLE_COL_DEF *AtableColDef AsrNodeID An Identifier of the TABLE content item. AsrNodeID An Identifier of the TABLE content item. AtableColDef Pointer to a supplied structure in which the next Table Column Definition Sequence item is to be retrieved. AtimeSize Time buffer size

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TABLE_Next_Row_Definition

Retrieve the data in the next item of the Table Row Definition Sequence in the TABLE Content Item.

Synopsis		
MC_STATUS MC_SRH_Get_TABLE_Next_Row_Definition (int AsrNodeID, TABLE_ROW_DEF *AtableRowDef		
)		
AsrNodelD	An Identifier of the TABLE content item.	
AtableRowDef	Pointer to a supplied structure in which the next Table Row Definition Sequence item is to be retrieved.	

Return Value

One of the enumerated MC_STATUS codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_TABLE_Cell_Values_Item

Retrieve the data in a Cell Value Sequence item in a TABLE Content item.

MC_STATUS MC_SRH_Get_TABLE_Cell_Values_Item (int AcellValuesID,		
TABLE CELL VALUES* AtableCellValues		
) _	_	
AcellValuesID	Numeric handle pointing to the Cell Value Sequence Item to be retrieved.	
AtableCellValues	Pointer to a supplied structure in which the attributes of the Cell Value Sequence item are to be retrieved.	

One of the enumerated **MC_STATUS** codes defined in "mcstatus.h":

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.
MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

MC_SRH_Get_CONTAINER_Data

Returns a CONTAINER node data.

Synopsis

```
      MC_STATUS MC_SRH_Get_CONTAINER_Data (

      int AsrNodeID,

      SR_CONT_CONTINUITY *Acontinuity,

      char AtemplateId[],

      int AtemplateIdSize

      )

      AsrNodeID

      An Identifier of the TIME SR item.

      Enumerated return value that specifies whether or not its contained Content Items are logically linked in a continuous textual flow, or are separate items.

      DICOM Template Identifer that describes the content of this Content Item and its subsidiary
```

```
Content Items. Can be NULL.AtemplateIdSizeThe size in bytes of the template Identifier array.
```

Return Value

Value	Meaning
MC_NORMAL_COMPLETION	The function completed normally.
MC_INVALID_SR_ID	The AsrNodeID value is not a valid SR tree management object ID.

MC_NULL_POINTER_PARM	One or more of the pointer-type parameters was NULL.
MC_INCOMPATIBLE_VALUE	The <i>Acontinuity</i> value is not valid.
MC_EMPTY_VALUE	The attribute has no value assigned to it.
MC_NULL_VALUE	The attribute's value is NULL (i.e. its length is zero).
MC_SYSTEM_ERROR	An unexpected, potentially serious, problem was detected in the operating environment. A message describing the error has been written to the Merge DICOM Toolkit log file.

Appendix A: Writing a DICOM Conformance Statement

Detailed below is a guideline for writing a conformance statement for your application. Since the Toolkit is *not* an application, this section only gives an outline of the DICOM services it supports. Responsibility for full DICOM conformance to particular SOP classes rests with the application developer, since many of the requirements for such conformance lie outside the realm of the Toolkit. For example, the high level behavior of Query/Retrieve service class SCUs and SCPs as defined in Part 4 of the DICOM standard, is implemented by the application developer in conjunction with the toolkit functionality.

Conformance Statement Sections

Implementation Model

The Implementation model consists of three sections:

- the Application Data Flow Diagram which specifyies the relationship between the Application Entities and the "external world" or Real-World activities
- a functional description of each Application Entity
- the sequencing constraints among them.

Application Data Flow

As part of the Implementation model, an Application Data Flow Diagram shall be included. This diagram represents all of the Application Entities present in an implementation, and graphically depicts the relationship of the AE's use of DICOM to Real-World Activities as well as any applicable User interaction.

The Merge DICOM Toolkit provides the core functionality required to facilitate data flow between SCUs and SCPs.

Application conformance statements will include a data flow diagram. An example is shown below for a simple Storage Service Class SCP.




Figure A-1: MERGE_STORE_SCP application data flow diagram

Functional Definition of Application Entities (AE)

This part shall contain a functional definition for each individual, local Application Entity. It will describe in general terms the functions to be performed by the AE, and the DICOM services used to accomplish these functions. In this sense, "DICOM services" refers not only to DICOM Service Classes, but also to lower level DICOM services, such as Association Services.

Application conformance statements will fill in this section with a general specification of functions to be performed by SCU or SCP.

Sequencing of Real World Activities

If applicable, this section shall contain a description of sequencing as well as potential constraints on real-world activities. These include any applicable user interaction as performed by all the AEs. A UML sequence diagram that depicts the real-world activities as vertical bars, and shows events exchanged between them as arrows, is strongly recommended.

Application conformance statements fill this section in with any associated sequence of real-world activities. For example, a Storage Service Class SCP might perform the following real-world activities: store an image, modify it in some defined manner and act as a Storage Service Class SCU and forward the modified image somewhere.

AE Specifications

The next section in the DICOM Conformance Statement is a set of Application Entity specifications. There shall be one specification for AE. Each individual AE specification has a subsection. There are as many of these subsections as there are different AE's in the implementation. That is, if there are two distinct AEs, then there will be two subsections. The Merge DICOM Toolkit is using the mergecom.app configuration file to read configuration parameters for each AE. The following subsections are filled in for each AE:

- Application Entity
 - SOP Classes
 - Association Policies
 - o General
 - o Number of Associations
 - o Asynchronous Nature
 - o Implementation Identifying Information
 - Association Initiation Policy
 - o Activity

[©] Copyright Merge Healthcare Solutions Inc. 2023

- Description and Sequencing of Activities
- Proposed Presentation Contexts
- SOP Specific Conformance for SOP Class(es)
- Association Acceptance Policy
 - o Activity
 - Description and sequencing of Activities
 - Accepted Presentation Contexts
 - SOP Specific Conformance for SOP Class(es)

SOP Classes

Application conformance statements specify the DICOM SOPs which are supported by each Application Entity. For SCP Entities, the initiation of associations. Please see the "System Profile" section in ANNEX B: CONFIGURATION PARAMETERS and the "MC_Wait_For_Association" or "MC_Wait_For_Secure_Association" definition in this Reference Manual. For SCU Entities, the list of supported SOP classes will correspond to the services specified in "mergecom.app" for any SCPs to which the SCU wishes to connect.

Number of Associations

The Merge DICOM Toolkit does not impose any limit on the number of simultaneous associations that can be requested or accepted. The only limitation on the number of simultaneous associations is imposed by the operating system and available resources. However, if your application enforces that limit, it is defined here.

The MAX_PENDING_CONNECTIONS setting in the "mergecom.pro" file refers to the maximum number of outstanding connection requests per listener socket. It does not limit the maximum number of simultaneous associations.

Asynchronous Nature

Merge DICOM Toolkit does not currently support multiple outstanding transactions over a single association.

Implementation Identifying Information

Application conformance statements specify the Implementation Class Unique Identifier (UID) for the application, as well as the Implementation version name. These identifiers are taken from the mergecom.pro configuration file under the following keys:

IMPLEMENTATION_CLASS_UID

IMPLEMENTATION_VERSION

This UID must follow the syntax rules specified in Part 5 of the DICOM standard.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Proposed or Accepted Presentation Contexts

Application conformance statements specify all presentation contexts that are used for association negotiation. A presentation context consists of:

- an Abstract Syntax which is a DICOM service class name and unique identifier(UID)
- a transfer syntax name and UID. A transfer syntax represents a set of data encoding rules that are specified in the "mergecom.pro" file. Please see the "System Profile" section in Appendix B: Configuration Parameters.
- the role that the application will perform within the service class. The roles associated with a particular service class are discussed in Part 4 of the DICOM standard.
- any extended negotiation information used when creating associations. See the "MC_Get_Negotiation_Info" function in thi Reference Manual.
- any rules that govern the acceptance of presentation contexts for the AE. This includes rules for which combinations of Abstract/Transfer Syntaxes are acceptable, and rules for prioritization of presentation contexts. Rules that govern selection of transfer syntax within a presentation context shall be stated here. Please see the "Application Profile" section in the Appendix B: Configuration Parameters. Also, see the "MC_Get_Association_Info" function in this Reference Manual to learn about the presentation contexts that are queryable by an application program.

Refer to Table A-1 for an example.

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Extended
Name	UID	Name List	UID List		Negotiation
Computed Radiography Image Storage	1.2.840.10008.5.1.4.1.1.1	DICOM Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		DICOM Explicit VR Little Endian	1.2.840.10008.1.2.1		
		DICOM Explicit VR Big Endian	1.2.840.10008.1.2.2		

Table A-1: Example Presentation Context

Merge DICOM Toolkit uses mergecom.app configuration settings to specify presentation contexts shown above.

SOP Specific Conformance

This section includes the SOP specific behavior, i.e. error codes, error and exception handling and time-outs, etc. The information is described in the SOP specific Conformance Statement section of PS 3.4 (or relevant private SOP definition).

Transfer Syntax Selection Policies

Merge DICOM Toolkit uses the following policy when selecting transfer syntax:

- An SCU offers any transfer syntaxes which are defined in its mergecom.pro file.
- The SCP prefers its native byte ordering, and will prefer explicit over implicit VR.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Network Interfaces

Physical Network Interface

Merge DICOM Toolkit runs over the TCP/IP protocol stack on any physical interconnection media supporting the TCP/IP stack.

IPv4 and IPv6 Support

Merge DICOM Toolkit supports both IPv4 and IPV6 protocols and is configurable in the system profile.

Configuration

Refer to the Appendix B: Configuration Parameters for complete configuration information.

Applications reference four (4) configuration files. The first, merge.ini, is found through the MERGE_INI environment variable. They are as follows:

- merge.ini Specifies the names of the other three (3) configuration files and also contains message logging parameters.
- mergecom.pro Specifies run-time parameters for the application
- mergecom.app Defines service lists and applications on other network nodes to which connections are possible.
- mergecom.srv Service and sequence definitions.

AE Title/Presentation Address Mapping

Presentation address mapping is configured in the mergecom.app file. The Presentation Address of an SCU/SCP application is specified by configuring the Listen Port in the mergecom.pro file, and specifying the AE title for the SCU/SCP within the application itself.

Configurable Parameters

The mergecom.pro configuration file can be used to set or modify other lower-level communication parameters. This includes time-outs and other parameters. Some information about supported SOP classes is also stored here. **Most parameters in this file should NEVER be changed. Doing so may compromise DICOM conformance.** Before modifying any parameters, such as time-out, be sure to have a backup of the originally supplied mergecom.pro file. Also, before modifying other parameters, you should consider contacting Merge Healthcare for advice.

PDU size

The maximum PDU size is configurable with a minimum of 4,096 bytes.

Application conformance statements specify the chosen PDU (Protocol Data Units) size and any general rules governing the initiation of associations. Please see the "System Profile" section of this Reference Manual for further information about configuring the PDU size.

Extensions/Specializations/Privatizations

Standard Extended/Specialized/Private SOPs

Application conformance statements list extended, specialized, or private SOPs that are supported.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Private Transfer Syntaxes

This section describes private transfer syntaxes that are listed in the Transfer Syntax Tables. See the System Profile section in Appendix B: Configuration Parameters for details.

Appendix B: Configuration Parameters

This appendix describes each configuration parameter in detail.

Merge DICOM Toolkit is a highly configurable toolkit, and understanding its configuration is critical to using the library effectively.

Related parameters are grouped into sections in a configuration file as follows:

```
[SECTION_1]

PARAMETER_1 = value1

PARAMETER_2 = value2

[SECTION_2]

PARAMETER 3 = value3
```

Information contained in these tables are parameter names including descriptions and sections where they are contained. The parameters are listed alphabetically and organized by the initialization file where they are used.

The structure of initialization and configuration files are detailed in the "Configuration" section in this Reference Manual.

Initialization File

The Merge DICOM Toolkit Initialization file (usually called merge.ini) provides the DICOM Toolkit with its top-level configuration. It specifies the location of the other three configuration files, along with message and error logging characteristics.

There are two ways of accessing the merge.ini file in your runtime environment. The functions MC_Set_MergeINI() or MC_Set_MergeINI_Unicode() can be used to assign the path where the merge.ini file is located. You can also set the MERGE_INI environment variable to point to the Merge Initialization File. This variable can be set within a command shell. For example:

```
In Unix C-shell:
setenv MERGE_INI /users/mc3adv/merge.ini
```

In Unix Bourne, Korn, or Bash shell: MERGE INI=/users/mc3adv/merge.ini; export MERGE INI

```
In DOS command shell:
set MERGE_INI=\mc3adv\merge.ini
```

See the Merge DICOM Toolkit Platform Notes for your platform if none of these methods apply.

The initialization file contains one [MergeCOM3] section that points to the other three Merge DICOM Toolkit initialization files. It specifies characteristics of the message/error log kept by the DICOM Toolkit library, turns particular types of logging on and off, and specifies where the messages are logged (file, screen, both or none). In most cases the INFO, WARNING and ERROR messages are sufficient. The Tn_MESSAGE settings (where *n* is an integer between 1 and 9) turns on lower-level protocol tracing capabilities. These capabilities can prove useful when running into

difficulties communicating with other implementations of DICOM over a network and can be used by Merge Healthcare service engineers in diagnosing lower-level network problems.

The following parameters are recognized by Merge DICOM Toolkit in the initialization file.

Name	Section	Description
ALLOW_LIBRARY_EXCEPTION_ HANDLER	MergeCOM3	This parameter, if set to YES, directs the toolkit to call the internal (and the user defined, if present) exception handler in the case System exception or System signal is raised. This allows for the (limited) recovery of the user program by at least giving it a chance to cleanup and exit gracefully and avoid a straight crash. DEFAULT: NO
BLANK_FILL_LOG_FILE	MergeCOM3	This parameter informs the toolkit whether or not to expand the log file to its maximum size on initialization. Setting this value to "NO" will decrease the time spent in MC_Library_Initialization() but increase the time spent doing actual logging while the application is running. DEFAULT: YES
ERROR_MESSAGE	MergeCOM3	This parameter instructs the toolkit to which destination (File, Screen and/or Memory) to log error messages.
INFO_MESSAGE	MergeCOM3	This parameter instructs the toolkit to which destination (File, Screen and/or Memory or none) to log information messages.
LOG_FILE	MergeCOM3	This is the name of the Merge DICOM Toolkit message log. The file will be [re-]created by Merge DICOM Toolkit. This parameter is ignored by embedded toolkits.
		The path to the file can be specified using environment variables (including the pseudo environment variable MC3INIDIR which does not need to be set as the toolkit will resolve it internally to the directory where the merge.ini file resides).
		Unicode paths can now be specified through the environment variables. Both Windows style (%) and Unix style (\$) notations for the environment variables are accepted.
		DEFAULT: ./merge.log
LOG_FILE_BACKUP	MergeCOM3	This parameter tells the Merge DICOM Toolkit to create a backup of the log file before starting a new log. If "ON", any existing log file is renamed with a file extension of .L <i>nn</i> where <i>nn</i> is an integer number between 01 and 99. DEFAULT: OFF.

Name	Section	Description
LOG_FILE_LINE_LENGTH	MergeCOM3	This option specifies the number of characters that occur on a line within the merge.log file. DEFAULT: 78 MINIMUM: 16 MAXIMUM: 254
LOG_FILE_SIZE	MergeCOM3	This is the number of lines which will be created for the log file. If BLANK_FILL_LOG_FILE is set to YES, the file is initialized to all binary zeros before the first message is logged. DEFAULT: 1000 MINIMUM: 100 MAXIMUM: 30720 (30 * 1024)
LOG_MEMORY_SIZE	MergeCOM3	This is the number of lines of length equal to LOG_FILE_LINE_LENGTH which will be created for the memory log. Used when logging is set to "Memory" in merge.ini. DEFAULT: 1
MERGECOM_3_APPLICATIONS	MergeCOM3	 File containing the Merge DICOM Toolkit application configurations. The path to the file can be specified using environment variables (including the pseudo environment variable MC3INIDIR which does not need to be set as the toolkit will resolve it internally to the directory where the merge.ini file resides). Unicode paths can now be specified through the environment variables. Both Windows style (%) and Unix style (\$) notations for the environment variables are accepted.
MERGECOM_3_PROFILE	MergeCOM3	File containing the Merge DICOM Toolkit system profile parameters. The path to the file can be specified using environment variables (including the pseudo environment variable MC3INIDIR which does not need to be set as the toolkit will resolve it internally to the directory where the merge.ini file resides). Unicode paths can now be specified through the environment variables. Both Windows style (%) and Unix style (\$) notations for the environment variables are accepted.

Name	Section	Description
MERGECOM_3_SERVICES	MergeCOM3	File containing the Merge DICOM Toolkit system service and message definitions.
		The path to the file can be specified using environment variables (including the pseudo environment variable MC3INIDIR which does not need to be set as the toolkit will resolve it internally to the directory where the merge.ini file resides).
		Unicode paths can now be specified through the environment variables. Both Windows style (%) and Unix style (\$) notations for the environment variables are accepted.
NUM_HISTORICAL_LOG_FILES	MergeCOM3	This parameter informs the toolkit of the number of historical log files to keep. The valid range of number for this parameter is 1 – 99. The historical log files are named <i>basename</i> .L01 to <i>basename</i> .LXX where <i>basename</i> .LXX is the latest log file. The <i>basename</i> is determined by the LOG_FILE parameter. When the maximum number of historical log files is met, the oldest log file is deleted and the log files are renamed. Note that a new log file is created each time the library is initialized. This parameter is only used when LOG_FILE_BACKUP is set to YES.
T1_MESSAGE	MergeCOM3	Not used (internal tracing).
T2_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log the entire contents of messages sent or received over the network. The format is similar to MC_List_Message's output.
T3_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log messages relating to association negotiation.
T4_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log messages when incoming associations are automatically rejected.
T5_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log messages relating to regular and extended validation.
T6_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log messages relating to configuration.
T7_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log messages relating to logging of command level attributes in messages sent or received.
T8_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log messages relating to the streaming in and out of messages and file objects.

Name	Section	Description
T9_MESSAGE	MergeCOM3	This logging level parameter instructs the toolkit to log messages relating to PDU's sent and received. Note: Receipt and transmission of P-DATA PDU's are logged; not the actual PDU itself.
WARNING_MESSAGE	MergeCOM3	This parameter instructs the toolkit to which destination (File, Screen and/or Memory or none) to log warning messages.

NOTE: Note: The destination for the logging from the T2_MESSAGE to T9_MESSAGE trace levels can be one or more of File, Screen and Memory or none.

Application Profile

The Merge DICOM Toolkit application profile (usually called mergecom.app) specifies the characteristics of your application entity and the Application Entities (AE) your application connects with over a network. The name and location of this file is specified in the [MergeCOM3] section of the Merge DICOM Toolkit initialization file.

When your application acts as a client (SCU), the application profile must specify the network address of the server (SCP) and the application entities you wish to connect (open an association) to. Your client refers to the Application Entity using a DICOM AE title, which is the same way it is referred to in the application profile. The AE title consists of a string of characters containing no spaces and having a length of 16 characters or less. A section of the profile exists for each Server AE that you wish to connect to.

For example, if your application is an image source and also performs query and retrieval of images from two separate DICOM AE's, it might contain sections like the following:

= 104
= acme_sun1
= Storage_Service_List
= 104
= acme_hp2
= Query_Service_List

Acme_Store_SCP and Acme_QR_SCP are the AE titles for the applications you wish to connect to. The storage server runs on a SUN Solaris computer with the host name acme_sun1, while the query/retrieve server runs on an HP workstation with the host name acme_hp2. Both servers listen on port 104 (the standard DICOM listen port). The host name and port combined make up the TCP/IP network address for a listening server application. See Figure 9.

Besides entering a hostname for the HOST_NAME parameter, it is also possible to simply enter an IP address. Both IPv4 addresses and IPv6 addresses are allowed in this field.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Sections

Table B-2: Application profile section headings

The application profile contains the following sections:		
Section	Description	
<remote_application_title></remote_application_title>	Section describing a remote DICOM Application Entity title(s). The remote Application Entity titles listed here must be 1 to 16 bytes in length with no embedded spaces. Simply, this section is where you list the DICOM applications you want to communicate with.	
<service_list_name></service_list_name>	List(s) of DICOM services that will be provided by the Application Entities listed in the [<remote_application_title>] sections. The service names listed here must be 1 to 33 bytes in length with no embedded spaces. Simply, this section is where you list the services that are provided by the remote DICOM applications.</remote_application_title>	
<syntax_list_name></syntax_list_name>	List(s) of DICOM transfer syntaxes that will be supported by the services listed in the [<service_list_name>] sections. The transfer syntaxes must be one of those listed in Table 5.</service_list_name>	

The application profile contains the following sections:

Parameters

Table B-3: Application profile section headers.

The application profile contains the following parameters:

Parameter	Section	Description
PORT_NUMBER	<remote_application_title></remote_application_title>	This parameter is the TCP/IP port on which the remote DICOM system <u>listens</u> for connections. The commonly used port number is 104. This default value may be overridden by the MC_Open_Association or MC_Open_Secure_Association function call.
HOST_NAME	<remote_application_title></remote_application_title>	This parameter is the name of the remote host as it is known to your TCP/IP system. This default value may be overridden by the MC_Open_Association or MC_Open_Secure_Association function call. The parameters value must be 1 to 19 bytes in length with no embedded spaces. Note: A numeric internet address may be used: e.g. 192.204.32.1

SERVICE_LIST	<remote_application_title></remote_application_title>	This parameter is the name of a section in the application profile which provides a list of services for which local applications will negotiate when attempting to establish an association. This is a default list; another list may be specified in the MC_Open_Association or MC_Open_Secure_Association function call. The parameters value names must be 1 to 33 bytes in length with no embedded spaces.
--------------	---	--

Service List

The SERVICE_LIST section of the application profile is used to describe the DICOM services that will be negotiated by the listed Application Entity.

For example, these sections might look like the following:

```
[Storage_Service_List]
SERVICES_SUPPORTED = 11  # Number of Services in list
SERVICE_1  = STANDARD_MR
SERVICE_2  = STANDARD_CR
SERVICE_3  = STANDARD_CT
SERVICE_4  = STANDARD_CURVE
SERVICE_5  = STANDARD_MODALITY_LUT
SERVICE_6  = STANDARD_OVERLAY
SERVICE_7  = STANDARD_OVERLAY
SERVICE_8  = STANDARD_US
SERVICE_9  = STANDARD_US
SERVICE_10  = STANDARD_US_MF
SERVICE_11  = STANDARD_VOI_LUT
SERVICE_11  = STANDARD_NM
[Query_Service_List]
SERVICE_2  = STUDY_ROOT_FIND
SERVICE_2  = STUDY_ROOT_MOVE
```

[Storage_Service_List] lists the storage services that are requested, while [Query_Service_List] lists the type of query/retrieve that are requested. These service names are the strings used in Merge DICOM Toolkit to identify standard DICOM services, and are discussed further in specific application guides. Any services listed must be defined in the service profile, discussed below.




Figure B-1: A sample configuration of DICOM applications

A service list also needs to be defined for each of your own server AE's. Even though you do not need a section for your server AE title (since it is running on your local machine), you do need to specify a service list that your application supports as an SCP. If your application also acts as a storage server, it

Might use [Storage_Service_List]. You also need to specify a listen port for your server AE in the system profile, which is discussed below.

The parameter values are text strings recognizable by the Merge DICOM Toolkit. These strings are defined in mergecom.srv in the toolkit distribution. They are also made available to the application writer at compile time in the header file mc3services.h and source file mc3services.c as symbolic constants. The following is a list of currently supported services:

Merge DICOM Toolkit Service Parameter	DICOM Service Class
ACQUISITION_CONTEXT_SR	Storage
ADVANCED_BLENDING_PRESENTATION_STATE	Storage
ARTERIAL_PULSE_WAVEFORM	Storage
AUDIO_WAVEFORM_REAL_TIME_COMMUNICATION	Storage
AUTOREFRACTION_MEASUREMENTS	Storage
BASIC_ANNOTATION_BOX	Print Management
BASIC_COLOR_IMAGE_BOX	Print Management
BASIC_FILM_BOX	Print Management

[©] Copyright Merge Healthcare Solutions Inc. 2023

Merge DICOM Toolkit Service Parameter	DICOM Service Class
BASIC_FILM_SESSION	Print Management
BASIC_GRAYSCALE_IMAGE_BOX	Print Management
BASIC_PRINT_IMAGE_OVERLAY_BOX	Print Management
BASIC_STRUCTURED_DISPLAY	Storage
BODY_POSITION_WAVEFORM	Storage
BREAST_IMAGING_RPI_QUERY	Relevant Patient Information Query
BREAST_PROJ_PRESENT	Storage
BREAST_PROJ_PROCESS	Storage
BREAST_TOMO_IMAGE_STORAGE	Storage
C_ARM_PHOTON_ELECTRON_RADIATION	Storage
C_ARM_PHOTON_ELECTRON_RADIATION_RECORD	Storage
CARDIAC_RPI_QUERY	Relevant Patient Information Query
CHEST_CAD_SR	Storage
COLON_CAD_SR	Storage
COLOR_PALETTE_FIND	Query/Retrieve
COLOR_PALETTE_GET	Query/Retrieve
COLOR_PALETTE_MOVE	Query/Retrieve
COLOR_PALETTE_STORAGE	Storage
COMPOSITE_INST_RET_NO_BULK_GET	Query/Retrieve
COMPOSITE_INSTANCE_ROOT_RET_GET	Query/Retrieve
COMPOSITE_INSTANCE_ROOT_RET_MOVE	Query/Retrieve
COMPOSITING_PLANAR_MPR_VOLUMETRIC_PS	Storage
COMPREHENSIVE_3D_SR	Storage
CONTENT_ASSESSMENT_RESULTS	Storage

Merge DICOM Toolkit Service Parameter	DICOM Service Class
CORNEAL_TOPOGRAPHY_MAP	Storage
CT_DEFINED_PROCEDURE_PROTOCOL	Storage
CT_PERFORMED_PROCEDURE_PROTOCOL	Storage
DEFINED_PROCEDURE_PROTOCOL_FIND	Query/Retrieve
DEFINED_PROCEDURE_PROTOCOL_GET	Query/Retrieve
DEFINED_PROCEDURE_PROTOCOL_MOVE	Query/Retrieve
DEFORMABLE_SPATIAL_REGISTRATION	Storage
DERMOSCOPIC_PHOTOGRAPHY_IMAGE	Storage
DETACHED_INTERP_MANAGEMENT	Results Management
DETACHED_PATIENT_MANAGEMENT	Patient Management
DETACHED_RESULTS_MANAGEMENT	Results Management
DETACHED_STUDY_MANAGEMENT	Study Management
DETACHED_VISIT_MANAGEMENT	Patient Management
DICOMDIR	Media Storage
DISPLAY_SYSTEM	Display System Management
ELECTROMYOGRAM_WAVEFORM	Storage
ELECTROOCULOGRAM_WAVEFORM	Storage
ENCAPSULATED_CDA	Storage
ENCAPSULATED_MTL	Storage
ENCAPSULATED_OBJ	Storage
ENCAPSULATED_STL	Storage
ENHANCED_CT_IMAGE	Storage
ENHANCED_MR_COLOR_IMAGE	Storage
ENHANCED_MR_IMAGE	Storage
ENHANCED_PET_IMAGE	Storage

Merge DICOM Toolkit Service Parameter	DICOM Service Class
ENHANCED_US_VOLUME	Storage
ENHANCED_XA_IMAGE	Storage
ENHANCED_XRAY_RADIATION_DOSE_SR	Storage
ENHANCED_XRF_IMAGE	Storage
EXTENSIBLE_SR	Storage
G_P_PERFORMED_PROCEDURE_STEP_RETIRED	Study Management
G_P_SCHEDULED_PROCEDURE_STEP_RETIRED	Study Management
G_P_WORKLIST_RETIRED	Basic Worklist Management
GENERAL_AUDIO_WAVEFORM	Storage
GENERAL_RPI_QUERY	Relevant Patient Information Query
GENERIC_IMPLANT_TEMPLATE	Storage
GENERIC_IMPLANT_TEMPLATE_FIND	Query/Retrieve
GENERIC_IMPLANT_TEMPLATE_GET	Query/Retrieve
GENERIC_IMPLANT_TEMPLATE_MOVE	Query/Retrieve
GRAYSCALE_PLANAR_MPR_VOLUMETRIC_PS	Storage
HANGING_PROTOCOL	Hanging Protocol Storage
HANGING_PROTOCOL_FIND	Hanging Protocol Query/Retrieve
HANGING_PROTOCOL_GET	Hanging Protocol Query/Retrieve
HANGING_PROTOCOL_MOVE	Hanging Protocol Query/Retrieve
IMAGE_OVERLAY_BOX_RETIRED	Print Management
IMPLANT_ASSEMBLY_TEMPLATE	Storage
IMPLANT_ASSEMBLY_TEMPLATE_FIND	Query/Retrieve
IMPLANT_ASSEMBLY_TEMPLATE_GET	Query/Retrieve

Merge DICOM Toolkit Service Parameter	DICOM Service Class
IMPLANT_ASSEMBLY_TEMPLATE_MOVE	Query/Retrieve
IMPLANT_TEMPLATE_GROUP	Storage
IMPLANT_TEMPLATE_GROUP_FIND	Query/Retrieve
IMPLANT_TEMPLATE_GROUP_GET	Query/Retrieve
IMPLANT_TEMPLATE_GROUP_MOVE	Query/Retrieve
IMPLANTATION_PLAN_SR_DOCUMENT	Storage
INSTANCE_AVAIL_NOTIFICATION	Instance Availability Notification
INTRAOCULAR_LENS_CALCULATIONS	Storage
INVENTORY	Storage
INVENTORY_CREATION	Storage Management
INVENTORY_FIND	Query/Retrieve
INVENTORY_GET	Query/Retrieve
INVENTORY_MOVE	Query/Retrieve
KERATOMETRY_MEASUREMENTS	Storage
KEY_OBJECT_SELECTION_DOC	Storage
LEGACY_CONVERTED_ENHANCED_CT_IMAGE	Storage
LEGACY_CONVERTED_ENHANCED_MR_IMAGE	Storage
LEGACY_CONVERTED_ENHANCED_PET_IMAGE	Storage
LENSOMETRY_MEASUREMENTS	Storage
MACULAR_GRID_THIICKNESS_VOLUME	Storage
MAMMOGRAPHY_CAD_SR	Storage
MEDIA_CREATION_MANAGEMENT	Media Creation Management
MICROSCOPY_BULK_SIMPLE_ANNOTATIONS	Storage
MODALITY_WORKLIST_FIND	Modality Work list
MR_SPECTROSCOPY	Storage

Merge DICOM Toolkit Service Parameter	DICOM Service Class
MULTI_CHANNEL_RESPIRATORY_WAVEFORM	Storage
MULTIPLE_VOLUME_RENDERING_VOLUMETRIC_PRESENTATION_ST ATE	Storage
OPHT_VIS_FIELD_STATIC_PERIM_MEAS	Storage
OPHTHALMIC_AXIAL_MEASUREMENTS	Storage
OPHTHALMIC_OCT_BSCAN_VOLUME_ANALYSIS	Storage
OPHTHALMIC_OCT_EN_FACE_IMAGE	Storage
OPHTHALMIC_TOMOGRAPHY_IMAGE	Storage
OPM_THICKNESS_MAP	Storage
PARAMETRIC_MAP	Storage
PATIENT_RADIATION_DOSE_SR	Storage
PATIENT_ROOT_QR_FIND	Query/Retrieve
PATIENT_ROOT_QR_GET	Query/Retrieve
PATIENT_ROOT_QR_MOVE	Query/Retrieve
PATIENT_STUDY_ONLY_QR_FIND_RETIRED	Query/Retrieve
PATIENT_STUDY_ONLY_QR_GET_RETIRED	Query/Retrieve
PATIENT_STUDY_ONLY_QR_MOVE_RETIRED	Query/Retrieve
PERFORMED_IMAGING_AGENT_ADMINISTRATION_SR	Storage
PERFORMED_PROCEDURE_STEP	Study Management
PERFORMED_PROCEDURE_STEP_NOTIFY	Study Management
PERFORMED_PROCEDURE_STEP_RETRIEVE	Study Management
PLANNED_IMAGING_AGENT_ADMINISTRATION_SR	Storage
PRESENTATION_LUT	Print Management
PRINT_JOB	Print Management
PRINT_QUEUE_MANAGEMENT	Print Management
PRINTER	Print Management

Merge DICOM Toolkit Service Parameter	DICOM Service Class
PRINTER_CONFIGURATION	Print Management
PROCEDURAL_EVENT_LOGGING	Application Event Logging
PROCEDURE_LOG	Storage
PRODUCT_CHARACTERISTICS_QUERY	Query/Retrieve
PROTOCOL_APPROVAL	Storage
PROTOCOL_APPROVAL_FIND	Query/Retrieve
PROTOCOL_APPROVAL_GET	Query/Retrieve
PROTOCOL_APPROVAL_MOVE	Query/Retrieve
PULL_PRINT_REQUEST	Print Management
RADIOPHARMACEUTICAL_RADIATION_DOSE_SR	Storage
RAW_DATA	Storage
REAL_WORLD_VALUE_MAPPING	Storage
REFERENCED_IMAGE_BOX	Print Management
RENDITION_SELECTION_DOCUMENT_REAL_TIME_COMMUNICATION	Storage
REPOSITORY_QUERY	Query/Retrieve
RESPIRATORY_WAVEFORM	Storage
ROBOTIC_ARM_RADIATION	Storage
ROBOTIC_ARM_RADIATION_RECORD	Storage
ROUTINE_SCALP_ELECTROENCEPHALOGRAM_WAVEFORM	Storage
RT_BEAMS_DELIVERY_INSTRUCTION	Storage
RT_BRACHY_APP_SETUP_DELIVERY_INSTR	Storage
RT_CONVENTIONAL_MACHINE_VERIFICATION	Verification
RT_ION_MACHINE_VERIFICATION	Verification
RT_PHYSICIAN_INTENT	Storage
RT_RADIATION_RECORD_SET	Storage

	-
Merge DICOM Toolkit Service Parameter	DICOM Service Class
RT_RADIATION_SALVAGE_RECORD	Storage
RT_RADIATION_SET	Storage
RT_RADIATION_SET_DELIVERY_INSTRUCTION	Storage
RT_SEGMENT_ANNOTATION	Storage
RT_TREATMENT_PREPARATION	Storage
SC_MULTIFRAME_GRAYSCALE_BYTE	Storage
SC_MULTIFRAME_GRAYSCALE_WORD	Storage
SC_MULTIFRAME_SINGLE_BIT	Storage
SC_MULTIFRAME_TRUE_COLOR	Storage
SEGMENTATION	Storage
SEGMENTED_VOLUME_RENDERING_VOLUMETRIC_PRESENTATION_ STATE	Storage
SIMPLIFIED_ADULT_ECHO_SR	Storage
SLEEP_ELECTROENCEPHALOGRAM_WAVEFORM	Storage
SPATIAL_FIDUCIALS	Storage
SPATIAL_REGISTRATION	Storage
SPECTACLE_PRESCRIPTION_REPORT	Storage
STANDARD_BASIC_TEXT_SR	Storage
STANDARD_BLENDING_SOFTCOPY_PS	Storage
STANDARD_COLOR_SOFTCOPY_PS	Storage
STANDARD_COMPREHENSIVE_SR	Storage
STANDARD_CR	Storage
STANDARD_CT	Storage
STANDARD_CURVE	Storage
STANDARD_DX_PRESENT	Storage
STANDARD_DX_PROCESS	Storage

Merge DICOM Toolkit Service Parameter	DICOM Service Class
STANDARD_ECHO	Verification
STANDARD_ENCAPSULATED_PDF	Storage
STANDARD_ENHANCED_SR	Storage
STANDARD_GRAYSCALE_SOFTCOPY_PS	Storage
STANDARD_HARDCOPY_COLOR	Storage
STANDARD_HARDCOPY_GRAYSCALE	Storage
STANDARD_IO_PRESENT	Storage
STANDARD_IO_PROCESS	Storage
STANDARD_IVOCT_PRESENT	Storage
STANDARD_IVOCT_PROCESS	Storage
STANDARD_MG_PRESENT	Storage
STANDARD_MG_PROCESS	Storage
STANDARD_MODALITY_LUT	Storage
STANDARD_MR	Storage
STANDARD_NM	Storage
STANDARD_NM_RETIRED	Storage
STANDARD_OPHTHALMIC_16_BIT	Storage
STANDARD_OPHTHALMIC_8_BIT	Storage
STANDARD_OVERLAY	Storage
STANDARD_PET	Storage
STANDARD_PET_CURVE	Storage
STANDARD_PRINT_STORAGE	Storage
STANDARD_PSEUDOCOLOR_SOFTCOPY_PS	Storage
STANDARD_RT_BEAMS_TREAT	Storage
STANDARD_RT_BRACHY_TREAT	Storage

Merge DICOM Toolkit Service Parameter	DICOM Service Class
STANDARD_RT_DOSE	Storage
STANDARD_RT_IMAGE	Storage
STANDARD_RT_ION_BEAMS_TREAT	Storage
STANDARD_RT_ION_PLAN	Storage
STANDARD_RT_PLAN	Storage
STANDARD_RT_STRUCTURE_SET	Storage
STANDARD_RT_TREAT_SUM	Storage
STANDARD_SEC_CAPTURE	Storage
STANDARD_US	Storage
STANDARD_US_MF	Storage
STANDARD_US_MF_RETIRED	Storage
STANDARD_US_RETIRED	Storage
STANDARD_VIDEO_ENDOSCOPIC	Storage
STANDARD_VIDEO_MICROSCOPIC	Storage
STANDARD_VIDEO_PHOTOGRAPHIC	Storage
STANDARD_VL_ENDOSCOPIC	Storage
STANDARD_VL_MICROSCOPIC	Storage
STANDARD_VL_PHOTOGRAPHIC	Storage
STANDARD_VL_SLIDE_MICROSCOPIC	Storage
STANDARD_VOI_LUT	Storage
STANDARD_WAVEFORM_12_LEAD_ECG	Storage
STANDARD_WAVEFORM_AMBULATORY_ECG	Storage
STANDARD_WAVEFORM_BASIC_VOICE_AUDIO	Storage
STANDARD_WAVEFORM_CARDIAC_EP	Storage
STANDARD_WAVEFORM_GENERAL_ECG	Storage

Merge DICOM Toolkit Service Parameter	DICOM Service Class
STANDARD_WAVEFORM_HEMODYNAMIC	Storage
STANDARD_XRAY_ANGIO	Storage
STANDARD_XRAY_ANGIO_BIPLANE	Storage
STANDARD_XRAY_RF	Storage
STEREOMETRIC_RELATIONSHIP	Storage
STORAGE_COMMITMENT_PULL	Storage Commitment
STORAGE_COMMITMENT_PUSH	Storage Commitment
STUDY_COMPONENT_MANAGEMENT	Study Management
STUDY_CONTENT_NOTIFICATION	Study Content Notification
STUDY_ROOT_QR_FIND	Query/Retrieve
STUDY_ROOT_QR_GET	Query/Retrieve
STUDY_ROOT_QR_MOVE	Query/Retrieve
SUBJ_REFRACTION_MEASUREMENTS	Storage
SUBSTANCE_ADMIN_LOGGING	Storage
SUBSTANCE_APPROVAL_QUERY	Storage
SURFACE_SCAN_MESH	Storage
SURFACE_SCAN_POINT_CLOUD	Storage
SURFACE_SEGMENTATION	Storage
TOMOTHERAPEUTIC_RADIATION	Storage
TOMOTHERAPEUTIC_RADIATION_RECORD	Storage
TRACTOGRAPHY_RESULTS	Storage
UPS_EVENT_SOP	Unified Procedure Step Management
UPS_EVENT_SOP_TRIAL_RETIRED	Unified Procedure Step Management
UPS_PULL_SOP	Unified Procedure Step Management

Merge DICOM Toolkit Service Parameter	DICOM Service Class
UPS_PULL_SOP_TRIAL_RETIRED	Unified Procedure Step Management
UPS_PUSH_SOP	Unified Procedure Step Management
UPS_PUSH_SOP_TRIAL_RETIRED	Unified Procedure Step Management
UPS_QUERY_SOP	Unified Procedure Step Management
UPS_WATCH_SOP	Unified Procedure Step Management
UPS_WATCH_SOP_TRIAL_RETIRED	Unified Procedure Step Management
VIDEO_ENDOSCOPIC_IMAGE_REAL_TIME_COMMUNICATION	Storage
VIDEO_PHOTOGRAPHIC_IMAGE_REAL_TIME_COMMUNICATION	Storage
VISUAL_ACUITY_MEASUREMENTS	Storage
VL_WHOLE_SLIDE_MICROSCOPY_IMAGE	Storage
VOI_LUT_BOX	Print Management
VOLUME_RENDERING_VOLUMETRIC_PRESENTATION_STATE	Storage
WIDE_FIELD_OPHTHALMIC_PHOTO_3D_COORDINATES	Storage
WIDE_FIELD_OPHTHALMIC_PHOTO_STEREOGRAPHIC_PROJ	Storage
XA_DEFINED_PROCEDURE_PROTOCOL	Storage
XA_PERFORMED_PROCEDURE_PROTOCOL	Storage
XA_XRF_GRAYSCALE_SOFTCOPY_PS	Storage
XRAY_3D_ANGIO_IMAGE	Storage
XRAY_3D_CRANIO_IMAGE	Storage
XRAY_RADIATION_DOSE_SR	Storage
BASIC_COLOR_PRINT_MANAGEMENT (META_SOP)	Print Management
BASIC_GRAYSCALE_PRINT_MANAGEMENT (META_SOP)	Print Management

Merge DICOM Toolkit Service Parameter	DICOM Service Class
DETACHED_PATIENT_MANAGEMENT_META (META_SOP)	Patient Management
DETACHED_RESULTS_MANAGEMENT_META (META_SOP)	Results Management
G_P_WORKLIST_MANAGEMENT_META_RETIRED (META_SOP)	Basic Worklist Management
PULL_STORED_PRINT_MANAGEMENT (META_SOP)	Print Management
REF_COLOR_PRINT_MANAGEMENT (META_SOP)	Print Management
REF_GRAYSCALE_PRINT_MANAGEMENT (META_SOP)	Print Management
STUDY_MANAGEMENT (META_SOP)	Study Management

Transfer Syntax Lists

For DICOM Toolkit users, Merge DICOM Toolkit allows you to define transfer syntaxes supported for each service in a service list. This functionality is implemented through the use of transfer syntax lists. The basic service lists discussed above can be modified to include these transfer syntax lists. The following is an example service list that has transfer syntaxes specified for each service:

[Storage_Service_List]	
SERVICES_SUPPORTED	= 3 # Number of Services
SERVICE_1	= STANDARD_MR
SYNTAX_LIST_1	= MR_Syntax_List
SERVICE_2	= STANDARD_US
SYNTAX_LIST_2	= US_Syntax_List
SERVICE_3	= STANDARD_CT
SYNTAX_LIST_3	= CT_Syntax_List
[MR_Syntax_List] SYNTAXES_SUPPORTED	
SYNTAX_1	= JPEG_BASELINE
SYNTAX_2	= EXPLICIT_BIG_ENDIAN
SYNTAX_3	= EXPLICIT_LITTLE_ENDIAN
SYNTAX_4	= IMPLICIT_LITTLE_ENDIAN
[US_Syntax_List]	
_	= 2 # Number of Syntaxes
SYNTAX_1	= RLE
SYNTAX_2	= IMPLICIT_LITTLE_ENDIAN
[CT_Syntax_List]	
—	= 2 # Number of Syntaxes
SYNTAX_1	= EXPLICIT_LITTLE_ENDIAN
SYNTAX_2	= IMPLICIT_LITTLE_ENDIAN

[Storage_Service_List] lists various standard storage service class services used by Merge DICOM Toolkit. The SYNTAX_LIST_N parameter has been added to this example to specify a transfer syntax list for each service. This optional parameter is set to the name of another section in the application profile which lists DICOM transfer syntaxes to be negotiated. If this parameter is not set, the default non-compressed transfers syntaxes (implicit VR little endian, explicit VR little endian, and explicit VR big endian) are negotiated.

The [MR_Syntax_List], [US_Syntax_List] and [CT_Syntax_List] sections each define a separate transfer syntax list for the MR, US, and CT services, respectively. Merge DICOM Toolkit currently supports all transfer syntaxes specified in the DICOM standard. The names used for these transfer syntaxes are defined below in this Reference Manual.

For server (SCP) applications, the order in which transfer syntaxes are specified in a transfer syntax list dictates the priority Merge DICOM Toolkit places on them during association negotiation. For example, using the [US_Syntax_List] specified above, if a client (SCU) proposes the Ultrasound storage service with the RLE compressed transfer syntax and the implicit VR little endian transfer syntax, Merge DICOM Toolkit selects the RLE transfer syntax because it is listed first in the transfer syntax list.

When a transfer syntax list is not specified in a service list, the priority Merge DICOM Toolkit places on transfer syntaxes during association negotiation is dependent on the hardware platform. On little endian machines (Intel and DEC Alpha based systems) the priority order is: Explicit VR Little Endian, Implicit VR Little Endian, and Explicit VR Big Endian. On big endian machines the priority order is: Explicit VR Big Endian, Explicit VR Little Endian, and Implicit VR Little Endian.

The following is a list of the currently supported transfer syntaxes:

Merge DICOM Toolkit Transfer Syntax Description Parameter	
IMPLICIT_LITTLE_ENDIAN	Implicit VR Little Endian: Default Transfer Syntax for DICOM
IMPLICIT_BIG_ENDIAN	Implicit VR Big Endian
EXPLICIT_LITTLE_ENDIAN	Explicit VR Little Endian
EXPLICIT_BIG_ENDIAN	Explicit VR Big Endian
DEFLATED_EXPLICIT_LITTLE_ENDIAN	Deflated Explicit VR Little Endian
ENCAPSULATED_UNCOMPRESSED_ELE	Encapsulated Explicit VR Little Endian
RLE	Run length Encoding
HEVC_H265_M10P_LEVEL_5_1	HEVC/H.265 Main 10 Profile / Level 5.1
HEVC_H265_MP_LEVEL_5_1	HEVC/H.265 Main Profile / Level 5.1

Table B-5: Transfer Syntax List Parameters

[©] Copyright Merge Healthcare Solutions Inc. 2023

(Process 6 & 8)JPEG_SPEC_NON_HIER_7_9JPEG Spectral Selection, Non-Hierarchical (Process 7 & 9)JPEG_FULL_PROG_NON_HIER_10_12JPEG Full Progression, Non-Hierarchical (Process 10 & 12)JPEG_FULL_PROG_NON_HIER_11_13JPEG Full Progression, Non-Hierarchical (Process 11 & 13)JPEG_LOSSLESS_NON_HIER_14JPEG Lossless, Non-Hierarchical (Proces 14)JPEG_LOSSLESS_NON_HIER_15JPEG Lossless, Non-Hierarchical (Proces 15)JPEG_EXTENDED_HIER_16_18JPEG Extended, Hierarchical (Process 16 & 18)		
Transfer Syntax for Lossy JPEG 8 Bit Image Compression JPEG_EXTENDED_2_4 JPEG_EXTENDED_3_5 JPEG_SPEC_NON_HIER_6_8 JPEG_SPEC_NON_HIER_7_9 JPEG_FULL_PROG_NON_HIER_10_12 JPEG_FULL_PROG_NON_HIER_11_13 JPEG_LOSSLESS_NON_HIER_14 JPEG_LOSSLESS_NON_HIER_15 JPEG_EXTENDED_HIER_16_18 JPEG_EXTENDED_HIER_16_18 JPEG_SPEC_NON_HIER_16_18 JPEG_SPES_NON_HIER_16_18 JPEG_SPES_HIER_20_22 JPEG_SPES_HIER_20_22		Description
Transfer Syntax for Lossy JPEG 12 Bit Image Compression (Process 4 only)JPEG_EXTENDED_3_5JPEG Extended (Process 3 & 5)JPEG_SPEC_NON_HIER_6_8JPEG Spectral Selection, Non-Hierarchic (Process 6 & 8)JPEG_SPEC_NON_HIER_7_9JPEG Spectral Selection, Non-Hierarchic (Process 7 & 9)JPEG_FULL_PROG_NON_HIER_10_12JPEG Full Progression, Non-Hierarchical 	JPEG_BASELINE	Transfer Syntax for Lossy JPEG 8 Bit
JPEG_SPEC_NON_HIER_6_8JPEG Spectral Selection, Non-Hierarchical (Process 6 & 8)JPEG_SPEC_NON_HIER_7_9JPEG Spectral Selection, Non-Hierarchical (Process 7 & 9)JPEG_FULL_PROG_NON_HIER_10_12JPEG Full Progression, Non-Hierarchical (Process 10 & 12)JPEG_FULL_PROG_NON_HIER_11_13JPEG Full Progression, Non-Hierarchical (Process 11 & 13)JPEG_LOSSLESS_NON_HIER_14JPEG Lossless, Non-Hierarchical (Proces 14)JPEG_LOSSLESS_NON_HIER_15JPEG Lossless, Non-Hierarchical (Proces 15)JPEG_EXTENDED_HIER_16_18JPEG Extended, Hierarchical (Process 17 & 19)JPEG_SPEC_HIER_20_22JPEG Spectral Selection, Hierarchical	JPEG_EXTENDED_2_4	Transfer Syntax for Lossy JPEG 12 Bit
(Process 6 & 8)JPEG_SPEC_NON_HIER_7_9JPEG Spectral Selection, Non-Hierarchical (Process 7 & 9)JPEG_FULL_PROG_NON_HIER_10_12JPEG Full Progression, Non-Hierarchical (Process 10 & 12)JPEG_FULL_PROG_NON_HIER_11_13JPEG Full Progression, Non-Hierarchical (Process 11 & 13)JPEG_LOSSLESS_NON_HIER_14JPEG Lossless, Non-Hierarchical (Process 14)JPEG_LOSSLESS_NON_HIER_15JPEG Lossless, Non-Hierarchical (Process 15)JPEG_EXTENDED_HIER_16_18JPEG Extended, Hierarchical (Process 16 & 18)JPEG_EXTENDED_HIER_17_19JPEG Extended, Hierarchical (Process 17 & 19)JPEG_SPEC_HIER_20_22JPEG Spectral Selection, Hierarchical	JPEG_EXTENDED_3_5	JPEG Extended (Process 3 & 5)
JPEG_FULL_PROG_NON_HIER_10_12JPEG Full Progression, Non-Hierarchical (Process 10 & 12)JPEG_FULL_PROG_NON_HIER_11_13JPEG Full Progression, Non-Hierarchical (Process 11 & 13)JPEG_LOSSLESS_NON_HIER_14JPEG Lossless, Non-Hierarchical (Process 14)JPEG_LOSSLESS_NON_HIER_15JPEG Lossless, Non-Hierarchical (Process 15)JPEG_EXTENDED_HIER_16_18JPEG Extended, Hierarchical (Process 16 & 18)JPEG_SPEC_HIER_20_22JPEG Spectral Selection, Hierarchical	JPEG_SPEC_NON_HIER_6_8	JPEG Spectral Selection, Non-Hierarchical (Process 6 & 8)
JPEG_FULL_PROG_NON_HIER_11_13JPEG Full Progression, Non-Hierarchical (Process 11 & 13)JPEG_LOSSLESS_NON_HIER_14JPEG Lossless, Non-Hierarchical (Process 14)JPEG_LOSSLESS_NON_HIER_15JPEG Lossless, Non-Hierarchical (Process 15)JPEG_EXTENDED_HIER_16_18JPEG Extended, Hierarchical (Process 16 & 18)JPEG_EXTENDED_HIER_17_19JPEG Extended, Hierarchical (Process 17 & 19)JPEG_SPEC_HIER_20_22JPEG Spectral Selection, Hierarchical	JPEG_SPEC_NON_HIER_7_9	JPEG Spectral Selection, Non-Hierarchical (Process 7 & 9)
Image: Control (Process 11 & 13)JPEG_LOSSLESS_NON_HIER_14JPEG Lossless, Non-Hierarchical (Process 14)JPEG_LOSSLESS_NON_HIER_15JPEG Lossless, Non-Hierarchical (Process 15)JPEG_EXTENDED_HIER_16_18JPEG Extended, Hierarchical (Process 16 & 18)JPEG_EXTENDED_HIER_17_19JPEG Extended, Hierarchical (Process 17 & 19)JPEG_SPEC_HIER_20_22JPEG Spectral Selection, Hierarchical	JPEG_FULL_PROG_NON_HIER_10_12	
14)JPEG_LOSSLESS_NON_HIER_15JPEG_EXTENDED_HIER_16_18JPEG_EXTENDED_HIER_16_18JPEG_EXTENDED_HIER_17_19JPEG_EXTENDED_HIER_17_22JPEG_SPEC_HIER_20_22JPEG Spectral Selection, Hierarchical	JPEG_FULL_PROG_NON_HIER_11_13	
15) JPEG_EXTENDED_HIER_16_18 JPEG_EXTENDED_HIER_17_19 JPEG_EXTENDED_HIER_17_19 JPEG_SPEC_HIER_20_22 JPEG Spectral Selection, Hierarchical	JPEG_LOSSLESS_NON_HIER_14	JPEG Lossless, Non-Hierarchical (Process 14)
& 18) JPEG_EXTENDED_HIER_17_19 JPEG_SPEC_HIER_20_22 JPEG Spectral Selection, Hierarchical	JPEG_LOSSLESS_NON_HIER_15	JPEG Lossless, Non-Hierarchical (Process 15)
& 19) JPEG_SPEC_HIER_20_22 JPEG Spectral Selection, Hierarchical	JPEG_EXTENDED_HIER_16_18	JPEG Extended, Hierarchical (Process 16 & 18)
	JPEG_EXTENDED_HIER_17_19	JPEG Extended, Hierarchical (Process 17 & 19)
	JPEG_SPEC_HIER_20_22	
JPEG_SPEC_HIER_21_23 JPEG Spectral Selection, Hierarchical (Process 21 & 23)	JPEG_SPEC_HIER_21_23	
JPEG_FULL_PROG_HIER_24_26 JPEG Full Progression, Hierarchical (Process 24 & 26)	JPEG_FULL_PROG_HIER_24_26	
JPEG_FULL_PROG_HIER_25_27 JPEG Full Progression, Hierarchical (Process 25 & 27)	JPEG_FULL_PROG_HIER_25_27	
JPEG_LOSSLESS_HIER_28 JPEG Lossless, Hierarchical (Process 28	JPEG_LOSSLESS_HIER_28	JPEG Lossless, Hierarchical (Process 28)
JPEG_LOSSLESS_HIER_29 JPEG Lossless, Hierarchical (Process 29	JPEG_LOSSLESS_HIER_29	JPEG Lossless, Hierarchical (Process 29)

Merge DICOM Toolkit Transfer Syntax Parameter	Description
JPEG_LOSSLESS_HIER_14	JPEG Lossless, Hierarchical, First-Order Prediction (Process 14 [Selection Value 1]): Default Transfer Syntax for Lossless JPEG Image Compression
JPEG_2000_LOSSLESS_ONLY	JPEG 2000, Lossless
JPEG_2000	JPEG 2000, Lossless or Lossy
JPEG_2000_MC_LOSSLESS_ONLY	JPEG 2000 Part 2 Multi-component Image Compression (Lossless Only)
JPEG_2000_MC	JPEG 2000 Part 2 Multi-component Image Compression
JPEG_LS_LOSSLESS	JPEG LS Lossless
JPEG_LS_LOSSY	JPEG LS Lossy (Near-Lossless)
JPIP_REFERENCED	JPIP Referenced
JPIP_REFERENCED_DEFLATE	JPIP Referenced Deflate
MPEG2_MPHL	MPEG2 Main Profile / High Level
MPEG2_MPML	MPEG2 Main Profile @ Main Level
MPEG4_AVC_H264_BDC_HP_LEVEL_4_1	MPEG-4 AVC/H.264 BD-compatible High Profile / Level 4.1
MPEG4_AVC_H264_HP_LEVEL_4_1	MPEG-4 AVC/H.264 High Profile / Level 4.1
MPEG4_AVC_H264_HP_LEVEL_4_2_2D	MPEG-4 AVC/H.264 High Profile / Level 4.2 For 2D Video
MPEG4_AVC_H264_HP_LEVEL_4_2_3D	MPEG-4 AVC/H.264 High Profile / Level 4.2 For 3D Video
MPEG4_AVC_H264_STEREO_HP_LEVEL_4_2	MPEG-4 AVC/H.264 Stereo High Profile / Level 4.2
SMPTE_ST_2110_20_UNCOMPRESSED_INTE RLACED_ACTIVE_VIDEO	SMPTE ST 2110-20 Uncompressed Interlaced Active Video
SMPTE_ST_2110_20_UNCOMPRESSED_PROG RESSIVE_ACTIVE_VIDEO	SMPTE ST 2110-20 Uncompressed Progressive Active Video
SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO	SMPTE ST 2110-30 PCM Digital Audio

Merge DICOM Toolkit Transfer Syntax Parameter	Description
PRIVATE_SYNTAX_1	Private transfer syntax 1 with the characteristics specified by the PRIVATE_SYNTAX_1_LITTLE_ENDIAN, PRIVATE_SYNTAX_1_EXPLICIT_VR, and PRIVATE_SYNTAX_1_ENCAPSULATED configuration options.
PRIVATE_SYNTAX_2	Private transfer syntax 2 with the characteristics specified by the PRIVATE_SYNTAX_2_LITTLE_ENDIAN, PRIVATE_SYNTAX_2_EXPLICIT_VR, and PRIVATE_SYNTAX_2_ENCAPSULATED configuration options.

Role Negotiation

Merge DICOM Toolkit also supports DICOM role negotiation through its service lists. Whereas in previous examples, the same service list could be used for both client (SCU) and server (SCP), these service lists are specific to the role to be negotiated for each service.

[SCU_Service_List] SERVICES_SUPPORTED SERVICE_1 ROLE 1	<pre>= 1 # Number of Services = STORAGE_COMMITMENT_PUSH = SCU</pre>
SCP_Service_List] SERVICES_SUPPORTED SERVICE_1 ROLE_1	<pre>= 1 # Number of Services = STORAGE_COMMITMENT_PUSH = SCP</pre>

In this case, the [SCU_Service_List] supports the Storage Commitment Push SOP class as an SCU, and the [SCP_Service_List] supports the Storage Commitment Push SOP class as an SCP. Merge DICOM Toolkit negotiates the association based on the settings for these roles.

The role for a service can be defined as SCU, SCP, BOTH or remain undefined. Table 9 contains a complete list of configurable roles for both requestors and acceptors along with the resultant negotiated roles.

NOTE: In some cases, a service will be rejected because the roles being negotiated do not match.

Requestor's Configured Role	Acceptor's Configured Role	Requestor's Negotiated Role	Acceptor's Negotiated Role
SCU	SCP	SCU	SCP
	SCU	Rejected	Rejected
	вотн	SCU	SCP
	NOT DEFINED	SCU	SCP
SCP	SCP	Rejected	Rejected
	SCU	SCP	SCU
	вотн	SCP	SCU
	NOT DEFINED	Rejected	Rejected
вотн	SCP	SCU	SCP
	SCU	SCP	SCU
	вотн	вотн	BOTH
	NOT DEFINED	SCP	SCP
NOT DEFINED	SCP	SCU	SCP
	SCU	Rejected	Rejected
	вотн	SCU	SCP
	NOT DEFINED	SCU	SCP

Table B-6: Negotiated Roles

DICOM Asynchronous Communication

Merge DICOM Toolkit optionally supports DICOM Asynchronous Operations Window Negotiation through service lists. The service list in this case can be used for both the client (SCU) and server (SCP). The following is a sample service list that configures DICOM asynchronous communication negotiation:

```
[SCU_Or_SCP_Service_List]
SERVICES_SUPPORTED = 1 # Number of Services
MAX_OPERATIONS_INVOKED = 10
MAX_OPERATIONS_PERFORMED = 10
SERVICE_1 = STANDARD_MR
```

In this case, the [SCU Or SCP Service List] supports the Standard MR SOP Class. For all services, it supports 10 maximum operations invoked and 10 maximum operations performed. When MAX OPERATIONS INVOKED and MAX OPERATIONS PERFORMED are not included in the service list, asynchronous communications are not negotiated. See a subsequent section for details on implementing DICOM asynchronous communications with Merge DICOM Toolkit.

Extended Negotiation

Merge DICOM Toolkit optionally supports configuration of DICOM extended negotiation information in service lists. Currently, the DICOM standard allows extended negotiation information for Storage and Query/Retrieve Service classes as defined in PS3.4 of the standard. The extended negotiation information can be set for only the client (SCU). Server applications utilizing extended negotiation must set this information with a call to

```
MC Set Negotiation Info For Association.
[SCU Service List]
```

-	SERVICES SUPPORTED	=	2 # Number of Services
	SERVICE_1	=	STUDY_ROOT_QR_FIND
	EXT_NEG_INFO_1	=	0x01
	SERVICE_2	=	STUDY_ROOT_QR_MOVE
	EXT_NEG_INFO_2	=	0x01

In this case, the [SCU Service List] supports the Study Root Q/R Find and Move services. Both services have set a single byte of extended negotiation information set to hexadecimal 0x01 (this implies the Client supports relational Queries and Moves).

Multiple hexadecimal bytes can be set in the service list by listing each byte in the format "0x00 0x010x02".

Related General SOP Classes and Service Classes

DICOM Supplement 90 defines a mechanism in association negotiation for identifying when an SOP Class is a customization of a generalized SOP Class. It also defines a method to identify the service class of an SOP Class that is proposed by an SCU. This allows flexibility in an SCP to support service classes for which it supports the generalized version of an SOP class, but does not explicitly support the customized SOP class. It also allows a mechanism to easily make an SCP that supports all storage service class SOP classes that are proposed to it.

Related General SOP classes can be supported in the application profile by defining a service list containing the related general SOP classes for a given SOP class, then assigning the service list to the SOP class. The following example demonstrates how this is done:

[SCU_SR_General] SERVICES_SUPPORTED SERVICE_1 SERVICE_2	<pre>= 2 = STANDARD_ENHANCED_SR = STANDARD_COMPREHENSIVE_SR</pre>
[SCU_DX_General] SERVICES_SUPPORTED SERVICE_1	= 1 = STANDARD_DX_PRESENT

[SCU_Service_List]	
SERVICES_SUPPORTED	= 3
SERVICE 1	= STANDARD_BASIC_TEXT_SR
REL_GENERAL_1	= SCU_SR_General
SERVICE_CLASS_1	= 1.2.840.10008.4.2
SERVICE 2	= STANDARD_IO_PRESENT
REL_GENERAL_2	= SCU_DX_General
SERVICE_CLASS_2	= 1.2.840.10008.4.2
SERVICE 3	= STANDARD CT
SERVICE_CLASS_3	= 1.2.840.10008.4.2

In this case, the SCU_SR_General service list contains the related general SOP Classes for the STANDARD_BASIC_TEXT_SR service. The REL_GENERAL_1 option points to the service list to use as the related general services for SERVICE_1.

The example above also shows how the service class can be defined for each SOP Class within a service list. For instance, <code>SERVICE_CLASS_3</code> in the example specifies the service class for <code>SERVICE_3</code>. The UID for the Storage Service Class as defined in Supplement 90 is used.

The service lists above are only utilized by SCU applications. For SCP applications, there are several configuration options that define how Merge DICOM Toolkit negotiates an association when related general SOP Classes are included, or the Service Class is included for an SOP Class.

When the ACCEPT_STORAGE_SERVICE_CONTEXTS configuration option is set to Yes, Merge DICOM Toolkit accepts any proposed SOP Class that is defined as supporting the Storage Service Class.

When the ACCEPT_RELATED_GENERAL_SERVICES configuration option is set to Yes, Merge DICOM Toolkit accepts any SOP Class proposed if the SCP lists any of the related general SOP Classes defined for the SOP Class proposed.

System Profile

The System Profile is used to define system-wide parameters. These parameters apply across all associations with other DICOM application entities. The location of this file is provided by the MERGECOM 3 PROFILE parameter of the [MergeCOM3] section of the MERGE.INI file.

The following are a few notes to keep in mind concerning the System Profile:

- You must specify your own unique DICOM Implementation Class UID and place it in this file along with an optional Implementation Version. These need to be documented in your DICOM conformance statement.
- There are several exception options specified at both the association and DIMSE levels of DICOM communication. You should not have to modify these options in normal circumstances and doing so could make your application non DICOM conformant.
- The DICOM Upper Layer section network time-outs can be modified. This is useful on slower or less-predictable networks (e.g. WAN's).
- The section of the System Profile dealing with transport parameters is important. This is where you specify the default **TCP/IP listen port** for a DICOM server (SCP) application.

[©] Copyright Merge Healthcare Solutions Inc. 2023

Most importantly, you must place the license number you received when you purchased the Merge DICOM Toolkit in the [ASSOC_PARMS] section of the system profile. If the license you received with your Toolkit was DA53-31D34, you would set it in the [ASSOC_PARMS] section as follows:

```
[ASSOC_PARMS]
LICENSE = DA53-31D34
IMPLEMENTATION_CLASS_UID = 2.16.840.1.113669.2.1.2
IMPLEMENTATION_VERSION = MergeCOM3_340
ACCEPT_MULTIPLE_PRES_CONTEXTS = Yes
```

Both Toolkit sample applications and your own applications which use the DICOM Toolkit Library will not work without a valid license number.

The example of the [ASSOC_PARMS] section of the system profile also contains a sample implementation class UID and implementation version configuration values. The implementation class UID is intended by the DICOM standard to be unique for major revisions of an Application Entity. The implementation version is intended to be unique for minor revisions of an Application Entity. These configuration values are used during association negotiation by Merge DICOM Toolkit, and are intended to aid in tracking versions of applications in the field.

The ACCEPT_MULTIPLE_PRES_CONTEXTS configuration value is used by server (SCP) applications. This value determines whether multiple presentation contexts can be negotiated for a single DICOM Service.

As mentioned earlier, a listen port must be identified for your server AE. Port 104 is the standard DICOM listen port. This, along with the number of simultaneous TCP connection requests that can be queued up for acceptance (pending) for Merge DICOM Toolkit, is specified in the [TRANSPORT PARMS] section.

The MAX_PENDING_CONNECTIONS setting in the "mergecom.pro" file refers to the maximum number of outstanding connection requests per listener socket. Note that this is not the "Maximum number of simmultaneous associations". The value of this configuration is passed by the toolkit to the listen() call on the socket as the backlog parameter and it specifies how many pending connections can be queued at any given time.

The MAX_PENDING_CONNECTIONS configuration option affects the accepting of associations but not the requesting of associations and it affects the behavior at the TCP level. In the default case, if more than 5 association requests arrive at once then only the first 5 will be accepted by TCP and passed to Merge DICOM Toolkit, the others would be refused at the TCP level.

[TRANSPORT_PARMS]	
TCPIP_LISTEN_PORT	= 104
<pre># Max number of open listen channels</pre>	
MAX_PENDING_CONNECTIONS	= 5

An important section of the System Profile is the [MESSAGE PARMS] section:

[MESSAGE_PARMS]	
LARGE_DATA_STORE	= FILE # MEM Default = FILE
LARGE_DATA_SIZE	= 200
OBOW_BUFFER_SIZE	= 4096
DICTIONARY_ACCESS	= MEM # FILE Default = FILE

[©] Copyright Merge Healthcare Solutions Inc. 2023

DICTIONARY FILE	<pre>= /users/mc3adv/mrgcom3.dct</pre>
TEMP FILE DIRECTORY	= /users/mc3adv/tmp files/
MSG INFO FILE	<pre>= /users/mc3adv/mrgcom3.msg</pre>

The LARGE_DATA_STORE parameter informs the toolkit where it should store large data: either in memory, or in temporary files on disk. Large data is defined as a value for an attribute larger than LARGE_DATA_SIZE bytes. Pixel data associated with a medical image would most certainly be considered large data. If you are running your process on a resource rich system that supplies plenty of physical and virtual memory, you should select LARGE_DATA_STORE = MEM to improve your performance. If your process is not so fortunate or you are dealing with messages with very large data values, you will want to use LARGE_DATA_STORE = FILE. In this case, the DICOM Toolkit will manage the large data in temporary files located in the TEMP_FILE_DIRECTORY you specify.

Large data that is of value representation OB (Other Byte), OW (Other Word), or OF (Other Float) is treated specially by the toolkit. Pixel Data, Curves, and Overlays are composed of this type of data. You can let the toolkit manage OB/OW data for you like any other large data, or register your own Callback Function in your applications to deal with such data as it is being received or transmitted over the network. The use of Callbacks will be covered later when we discuss developing DICOM applications with the toolkit.

The OBOW_BUFFER_SIZE is used to tell the toolkit what size 'chunks' in bytes of OB/OW data it should read in before either writing the data to a temporary file or passing it to your Callback Function. Choosing a large number for OBOW_BUFFER_SIZE means less time spent by your application process writing to temporary files or making callbacks, but results in a larger process size.

You must also take into account that there is a 2^16 limit in how many "chunks" can be stored in the temporary files. For example, when the chunk size is 16KB (as set by <code>OBOW_BUFFER_SIZE</code>), the maximum value for pixel data is only 1GB (i.e., 16K * 64K). According to the DICOM standard, the maximum pixel data size can be 4GB. In order to support storing the maximum possible pixel data, the <code>OBOW_BUFFER_SIZE</code> must not be less than 64KB. If you need to use temporary files, you should tune this parameter to maximize performance within the constraints of your runtime environment.

The DICTIONARY_ACCESS parameter specifies whether the DICOM Data Dictionary should be loaded and stored into memory the first time it is accessesed, or be accessed from a file every time it is referenced. The **dictionary file** is a binary file supplied with your toolkit with the default name of mrgcom3.dct. This file is accessed very frequently by the toolkit and you should set this parameter to MEM in all but the most extreme cases of memory limitation. You also specify the location and name of the data dictionary file using the DICTIONARY_FILE parameter.

Another binary file supplied with the Toolkit is the message info file. This file contains binary encoded message objects and is accessed when an application opens a message. Once open, these objects reside in memory, and are "filled in" by your application. They then become a message object instance that can be exchanged over the network. The message info file along with the data dictionary file, make possible the powerful message validation capabilities of the DICOM Toolkit. The message info file is a binary file supplied by your Toolkit with the default name of

[©] Copyright Merge Healthcare Solutions Inc. 2023

mrgcom3.msg. You also specify the location and name of the message info file using the MSG_INFO_FILE parameter.

Tables B-7 through B-12 define how each parameter should be defined within each section of the system profile.

Name	Description
ACCEPT_ANY_APPLICATION_TITLE [†]	If set to YES, the remote system need not specify a correct DICOM application title when requesting an association. If set to NO a correct application title must be used. When this value is set to YES, the toolkit will report the remote application as connecting to the first application registered. DEFAULT: NO
ACCEPT_ANY_CONTEXT_NAME [†]	If set to YES, the remote system need not specify the LOCAL_APPL_CONTEXT_NAME when requesting an association. If set to NO, the correct context name must be used. DEFAULT: NO
ACCEPT_ANY_HOSTNAME	If set to YES, the toolkit will not check if applications connecting to an SCP can have their hostname resolved through the SCP's hostfile or domain name server. If set to NO, the toolkit will automatically reject associations from unknown hosts. DEFAULT: NO
ACCEPT_ANY_PRESENTATION_CONTEXT [†]	If set to YES, the toolkit will not validate that the presentation context ID contained in a message's PDU header information matches the ID of the presentation context negotiated for the type of message contained in the PDU. If set to NO, the toolkit will abort associations when these values do not match. DEFAULT: NO
ACCEPT_DIFFERENT_IC_UID †	If set to NO, the remote system must specify the local IMPLEMENTATION_CLASS_UID when requesting an association. If set to YES, a different implementation class UID may be used. DEFAULT: YES
ACCEPT_DIFFERENT_VERSION [†]	If set to NO, the remote system must specify the local IMPLEMENTATION_VERSION when requesting an association. If set to YES, a different implementation version may be used. DEFAULT: YES

Table B-7 [ASSOC_PARMS] section of system profile parameters

Name	Description
ACCEPT_LIST_OF_APPLICATION_TITLES	List of AE titles which the remote system might use when requesting an association. The parameters line should contain all AE titles separated by one of predefined delimiters: ',' '\' '/' ','. The length of each AE title can not exceed 16 characters. Example: ACCEPT_LIST_OF_APPLICATION_TITLES = MERGE_STORE_SCP/MERGE_STORE_SCU/MERGE_STORE _RQ DEFAULT: <none></none>
ACCEPT_MULTIPLE_PRES_CONTEXTS	If set to YES, SCP applications will allow multiple presentation contexts to be negotiated for a single DICOM service. If set to NO, an SCP will only accept a single presentation context for a DICOM service. DEFAULT:YES
ACCEPT_RELATED_GENERAL_SERVICES	This parameter sets the Merge DICOM Toolkit behavior in regards to support for DICOM Supplement 90. Supplement 90 defines a method for association requestors to specify the generalized version of a SOP Class. When set to YES, Merge DICOM Toolkit will allow association acceptors to accept a presentation context whose generalized SOP Class is supported; however, the customized SOP Class is not specifically supported. DEFAULT: NO
ACCEPT_STORAGE_SERVICE_CONTEXTS	This parameter sets the Merge DICOM Toolkit behavior in regards to support for DICOM Supplement 90. When set to YES, Merge DICOM Toolkit will accept any presentation context which is defined as a Storage Service Class SOP Class. DEFAULT: NO
ALLOW_EMPTY_PDV_LENGTH	The DICOM standard specifies that PDVs shall not be sent without any content in the fragment. The toolkit however can send and accept empty PDVs. To enforce the standard requirement, this setting should be set to No. DEFAULT: YES
AUTO_ECHO_SUPPORT	If set to YES, the toolkit automatically handles C-ECHO requests when the application doesn't explicitly include STANDARD_ECHO in its supported service list. If set to NO, the toolkit rejects C-ECHO requests when the application doesn't explicitly include STANDARD_ECHO in its supported service list. DEFAULT: YES
DEFLATED_EXPLICIT_LITTLE_ENDIAN_SYNTAX	This value defines the UID of the Deflated Explicit VR Little Endian transfer syntax. DEFAULT: 1.2.840.10008.1.2.1.99

Name	Description
ENCAPSULATED_UNCOMPRESSED_ELE_SYNTAX	This value defines the UID of the Encapsulated Uncompressed Explicit VR Little Endian transfer syntax. DEFAULT: 1.2.840.10008.1.2.1.98
EXPLICIT_BIG_ENDIAN_SYNTAX	This value defines the UID of the Explicit VR Big Endian transfer syntax. DEFAULT: 1.2.840.10008.1.2.2
EXPLICIT_LITTLE_ENDIAN_SYNTAX	This value defines the UID of the Explicit VR Little Endian transfer syntax. DEFAULT: 1.2.840.10008.1.2.1
HARD_CLOSE_TCP_IP_CONNECTION	This parameter specifies how TCP/IP connections are closed by the toolkit. When set to YES, TCP/IP connections are instantaneously closed with an RST packet. When set to NO, TCP/IP connections are closed gracefully with a FIN packet. Note, that in the NO case the toolkit must wait for an operating system dependent amount of time for the response to the FIN packet. DEFAULT: YES
IMPLEMENTATION_CLASS_UID	The DICOM Implementation Class UID (as specified in your DICOM conformance statement).
IMPLEMENTATION_VERSION	The Implementation Version Number (as specified in your DICOM conformance statement).
IMPLICIT_BIG_ENDIAN_SYNTAX	The Implicit VR Big Endian transfer syntax is not defined by the DICOM standard. This value is provided to supply compatibility with private implementations. DEFAULT: <none></none>
IMPLICIT_LITTLE_ENDIAN_SYNTAX	The Implicit VR Little Endian transfer syntax is the default network transfer syntax of the DICOM standard. The Implicit VR Little Endian transfer syntax must always be defined. DEFAULT: 1.2.840.10008.1.2
JPEG_2000_LOSSLESS_ONLY_SYNTAX	This value defines the UID for JPEG 2000, Lossless transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.90
JPEG_2000_MC_LOSSLESS_ONLY_SYNTAX	This value defines the UID for JPEG 2000 Part 2 Multi- component Image Compression (Lossless Only) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.92
JPEG_2000_MC_SYNTAX	This value defines the UID for JPEG 2000 Part 2 Multi- component Image Compression transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.93

Name	Description
JPEG_2000_SYNTAX	This value defines the UID for JPEG 2000 transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.91
JPEG_BASELINE_SYNTAX	This value defines the UID for JPEG Baseline (Process 1) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.50
JPEG_EXTENDED_2_4_SYNTAX	This value defines the UID for JPEG Extended (Process 2 & 4) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.51
JPEG_EXTENDED_3_5_SYNTAX	This value defines the UID for JPEG Extended (Process 3 & 5) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.52
JPEG_EXTENDED_HIER_16_18_SYNTAX	This value defines the UID for JPEG Extended, Hierarchical (Process 16 & 18) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.59
JPEG_EXTENDED_HIER_17_19_SYNTAX	This value defines the UID for JPEG Extended, Hierarchical (Process 17 & 19) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.60
JPEG_FULL_PROG_HIER_24_26_SYNTAX	This value defines the UID for JPEG Full Progression, Hierarchical (Process 24 & 26) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.63
JPEG_FULL_PROG_HIER_25_27_SYNTAX	This value defines the UID for JPEG Full Progression, Hierarchical (Process 25 & 27) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.64
JPEG_FULL_PROG_NON_HIER_10_12_SYNTAX	This value defines the UID for JPEG Full Progression, Non- Hierarchical (Process 10 & 12) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.55
JPEG_FULL_PROG_NON_HIER_11_13_SYNTAX	This value defines the UID for JPEG Full Progression, Non- Hierarchical (Process 11 & 13) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.56
JPEG_LOSSLESS_HIER_14_SYNTAX	This value defines the UID for JPEG Lossless, Non-Hierarchical, First-Order Prediction (Process 14, Selection Value 1) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.70
JPEG_LOSSLESS_HIER_28_SYNTAX	This value defines the UID for JPEG Lossless, Hierarchical (Process 28) transfer syntax. DEFAULT: 1.2.840.10008.1.2.4.65

Name	Description
JPEG_LOSSLESS_HIER_29_SYNTAX	This value defines the UID for JPEG Lossless, Hierarchical (Process 29) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.66
JPEG_LOSSLESS_NON_HIER_14_SYNTAX	This value defines the UID for JPEG Lossless, Non-Hierarchical (Process 14) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.57
JPEG_LOSSLESS_NON_HIER_15_SYNTAX	This value defines the UID for JPEG Lossless, Non-Hierarchical (Process 15) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.58
JPEG_LS_LOSSLESS_SYNTAX	This value defines the UID for JPEG LS Lossless transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.80
JPEG_LS_LOSSY_SYNTAX	This value defines the UID for JPEG LS Lossy (Near Lossless) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.81
JPEG_SPEC_HIER_20_22_SYNTAX	This value defines the UID for JPEG Spectral Selection, Hierarchical (Process 20 & 22) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.61
JPEG_SPEC_HIER_21_23_SYNTAX	This value defines the UID for JPEG Spectral Selection, Hierarchical (Process 21 & 23) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.62
JPEG_SPEC_NON_HIER_6_8_SYNTAX	This value defines the UID for JPEG Spectral Selection, Non-Hierarchical (Process 6 & 8) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.53
JPEG_SPEC_NON_HIER_7_9_SYNTAX	This value defines the UID for JPEG Spectral Selection, Non-Hierarchical (Process 7 & 9) transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.54
JPIP_REFERENCED_DEFLATE_SYNTAX	This value defines the UID for JPIP Referenced Deflate transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.95
JPIP_REFERENCED_SYNTAX	This value defines the UID for JPIP Referenced transfer syntax.
	DEFAULT: 1.2.840.10008.1.2.4.94
LICENSE	The Merge DICOM Toolkit license number that was supplied when the toolkit was purchased.

Application Context Name (UID) (as specified in the indard).
.2.840.10008.3.1.1.1
efines the UID for MPEG2 Main Profile @ High Level ax. I.2.840.10008.1.2.4.101
efines the UID for MPEG2 Main Profile @ Main Level ax. I.2.840.10008.1.2.4.100
efines the UID for MPEG-4 AVC/H.264 ble High Profile / Level 4.1 transfer syntax. I.2.840.10008.1.2.4.103
efines the UID for MPEG-4 AVC/H.264 High Profile / nsfer syntax. I.2.840.10008.1.2.4.102
efines the UID for MPEG-4 AVC/H.264 High Profile / 2D Video transfer syntax. .2.840.10008.1.2.4.104
efines the UID for MPEG-4 AVC/H.264 High Profile / 3D Video transfer syntax. .2.840.10008.1.2.4.105
efines the UID for MPEG-4 AVC/H.264 Stereo High I 4.2 transfer syntax. I.2.840.10008.1.2.4.106
m size of Protocol Data Units that can be received e DICOM Toolkit implementation. This value will also on how large PDU values being sent can be. Setting a PDU fits within an even multiple of the default 6 (Maximum Segment Size) of 1460 will optimize formance. Note that 6 bytes for the PDU header must the configured maximum PDU size when calculating the MSS.
see the TCPIP_SEND_BUFFER_SIZE and EIVE_BUFFER_SIZE configuration values for erformance.
460*44)-6 = 64234 PDU Size
34234
K NONE

Name	Description
PRIVATE_SYNTAX_1_ENCAPSULATED	When set to YES, Merge DICOM Toolkit will interpret private transfer syntax 1 as having its pixel data tag (7fe0,0010) being encoded as undefined length in the same manner as the JPEG and RLE transfer syntaxes are encoded.
	DEFAULT: NO
PRIVATE_SYNTAX_1_EXPLICIT_VR	When set to YES, Merge DICOM Toolkit will interpret private transfer syntax 1 as being encoded in explicit VR format. DEFAULT: YES
PRIVATE_SYNTAX_1_LITTLE_ENDIAN	When set to YES, Merge DICOM Toolkit will interpret private transfer syntax 1 as being encoded in little endian format. DEFAULT: YES
PRIVATE_SYNTAX_1_SYNTAX	The unique identifier (UID) Merge DICOM Toolkit will use to identify private transfer syntax 1. When this value is set to " <none>", private transfer syntax support is shut off. DEFAULT: <none></none></none>
PRIVATE_SYNTAX_2_ENCAPSULATED	When set to YES, Merge DICOM Toolkit will interpret private transfer syntax 2 as having its pixel data tag (7fe0,0010) being encoded as undefined length in the same manner as the JPEG and RLE transfer syntaxes are encoded. DEFAULT: NO
PRIVATE_SYNTAX_2_EXPLICIT_VR	When set to YES, Merge DICOM Toolkit will interpret private transfer syntax 2 as being encoded in explicit VR format. DEFAULT: YES
PRIVATE_SYNTAX_2_LITTLE_ENDIAN	When set to YES, Merge DICOM Toolkit will interpret private transfer syntax 2 as being encoded in little endian format. DEFAULT: YES
PRIVATE_SYNTAX_2_SYNTAX	The unique identifier (UID) Merge DICOM Toolkit will use to identify private transfer syntax 2. When this value is set to " <none>", private transfer syntax support is shut off. DEFAULT: <none></none></none>
RLE_SYNTAX	This value defines the UID of the RLE Lossless transfer syntax. DEFAULT: 1.2.840.10008.1.2.5
SMPTE_ST_2110_20_UNCOMPRESSED_INTERLAC ED_ACTIVE_VIDEO_SYNTAX	This value defines the UID for SMPTE ST 2110-20 Uncompressed Interlaced Active Video transfer syntax. DEFAULT: 1.2.840.10008.1.2.7.2
SMPTE_ST_2110_20_UNCOMPRESSED_PROGRESS IVE_ACTIVE_VIDEO_SYNTAX	This value defines the UID for SMPTE ST 2110-20 Uncompressed Progressive Active Video transfer syntax. DEFAULT: 1.2.840.10008.1.2.7.1

Name	Description
SMPTE_ST_2110_30_PCM_DIGITAL_AUDIO_SYN TAX	This value defines the UID for SMPTE ST 2110-30 PCM Digital Audio transfer syntax. DEFAULT: 1.2.840.10008.1.2.7.3

† These options allow for non-standard DICOM operations. Such exceptions, if used, should be noted in your DICOM conformance statement.

* Performance tuning.

Name	Description
INITIATOR_NAME †	The DICOM standard has retired the old ACR/NEMA Initiator Name attribute in command messages. To generate such an attribute in command messages, specify an initiator name. <none> means do not put initiator name in messages. DEFAULT: <none></none></none>
RECEIVER_NAME [†]	The DICOM standard has retired the old ACR/NEMA Receiver Name attribute in command messages. To generate such an attribute in command messages, specify a receiver name. <none> means do not put receiver name in messages. DEFAULT: <none></none></none>
SEND_ECHO_PRIORITY [†]	The DICOM standard has retired the message priority attribute in echo command messages. To generate such an attribute in command messages, specify YES. To NOT use message priority in echo messages, specify NO. DEFAULT: NO
SEND_LENGTH_TO_END [†]	The DICOM standard has retired the old Group-Length-To-End attribute in command messages. To generate such an attribute in command messages, specify YES. If you do not want to generate Group-Length-To-End, specify NO. DEFAULT: NO
SEND_MSG_ID_RESPONSE [†]	The DICOM standard has retired the message ID attribute in response command messages. To generate such an attribute in command messages, specify YES. To NOT use message ID in response messages, specify NO. DEFAULT: NO
SEND_RECOGNITION_CODE [†]	The DICOM standard has retired the old Recognition Code attribute in command messages. To generate such an attribute in command messages, specify YES. If you do not want to generate such an attribute, specify NO. DEFAULT: NO

Name	Description
SEND_RESPONSE_PRIORITY [†]	The DICOM standard has retired the message priority attribute in response messages. To generate such an attribute in response messages, specify YES. To NOT use message priority in response messages, specify NO. DEFAULT: NO
SEND_SOP_CLASS_UID [†]	Certain DICOM service classes demand that the affected SOP class UID be present in the message. To prevent the library from ensuring that this is done, specify NO. To ensure that Affected SOP class UID is present, specify YES. DEFAULT: YES
SEND_SOP_INSTANCE_UID †	Certain DICOM service classes demand that the affected SOP instance UID be present in the message. To prevent the library from ensuring that this is done, specify NO. To ensure that Affected SOP instance UID is present, specify YES. DEFAULT: YES

† These options allow for non-standard DICOM operations. Such exceptions, if used, should be noted in your DICOM conformance statement.

Table B-9: [DUL_PARMS] section of system profile parameters

Name	Description
ARTIM_TIMEOUT	The number of seconds to use as a time out waiting for an association request or waiting for the peer to shut down an association. DEFAULT: 30
ASSOC_REPLY_TIMEOUT	The number of seconds to wait for a reply to an associate request. DEFAULT: 15.
CONNECT_TIMEOUT	The number of seconds to wait for a network connect to be accepted. DEFAULT: 15.
INACTIVITY_TIMEOUT	The number of seconds to wait in between packets of data received over the network after the initial packet of data in a message is received. Used by the MC_Read_Message() and MC_Read_To_Stream functions. DEFAULT: 15.
INSURE_EVEN_UID_LENGTH †	Set to NO, if odd-length UIDs in PDU's should NOT be padded with a NULL to ensure even length unique Ids. Set to YES to ensure even length UIDs in PDUs. DEFAULT: NO

Name	Description
RELEASE_TIMEOUT	The number of seconds to wait for a reply to an associate release. DEFAULT: 15.
WRITE_TIMEOUT	The number of seconds to wait for a network write to be accepted. DEFAULT: 15.

† These options allow for non-standard DICOM operations. Such exceptions, if used, should be noted in your DICOM conformance statement.

Name	Description
DICOMDIR_STREAM_STORAGE	When set to yes, DICOMDIRs read in leave their directory records internally in "stream" format and are not parsed until the directory record is referenced. This can greatly reduce memory usage when reading in large DICOMDIRs when the entire DICOMDIR is not referenced. Default: NO
EXPORT_GROUP_LENGTHS_TO_MEDIA *	When set to NO, do not write group length attributes with <pre>MC_Write_File() and MC_Write_File_By_Callback().</pre> DEFAULT: YES
EXPORT_PRIVATE_ATTRIBUTES_TO_MEDI A	When set to NO, disable the exporting of private attributes in files written with the MC_Write_File() and MC_Write_File_By_Callback() functions. DEFAULT: YES
EXPORT_UN_VR_TO_MEDIA	When set to NO, disable the exporting of attributes with a VR of UN in files written with the MC_Write_File() and MC_Write_File_By_Callback() functions. DEFAULT: YES
EXPORT_UNDEFINED_LENGTH_SQ_IN_DIC OMDIR *	When set to NO, DICOMDIRs written with MC_Write_File() are created with their sequence attributes having defined lengths. Setting this option to Yes will increase performance. DEFAULT: YES

Table B-10: [MEDIA_PARMS] section of system profile parameters

* Performance tuning.

Table B-11: {MESSAGE_PARMS] section of system profile parameters

Name	Description
ALLOW_COMMA_IN_DS_FL_FD_STRINGS	When set to Yes, a comma or a period will be allowed in the value passed to MC_Set_Value_From_String() for attributes with a VR of DS, FL or FD. When set to No, only a period will be acceptable as a decimal separator. Note that

© Copyright Merge Healthcare Solutions Inc. 2023

Name	Description
	the toolkit will always ensure that DS attributes use a period decimal separator when streaming to the network or to a file, regardless of current locale settings. DEFAULT: NO
ALLOW_INVALID_LENGTH_FOR_VR	When set to 'No', attribute data values with incorrect length according to DICOM, i.e., more than the maximum allowed length or not a multiple of the unit size for fixed length VRs (SS, US, AT, SL, UL, SV, UV, FL, FD) will cause an MC_INVALID_LENGTH_FOR_VR error on DICOM data reading. DEFAULT: YES
ALLOW_INVALID_PRIVATE_ATTRIBUTES	When reading messages or file objects, this parameter specifies if private attributes encoded in an invalid format should be ignored or parsed. DEFAULT: NO
ALLOW_INVALID_PRIVATE_CREATOR_CODES	When reading messages or file objects, this parameter specifies if private creator codes encoded with invalid characters should be ignored or parsed. DEFAULT: NO
ALLOW_OUT_OF_RANGE_BITS_JPEG_LOSSLESS	During decompression of JPEG lossless images, the Pegasus decompressor may discover that the original compressor had failed to mask off the out-of-range bits for the image bit depth. However, if all other lossless JPEG computations are correct, the original image, including such incorrect out-of-range bits, can be losslessly recovered. The Pegasus decompressor will return a warning status, along with the fully decoded image.
	If this flag is set by the application, the out-of-range bits in output pixels will not be masked off, but returned in the decoded image. Without this flag, out-of-range bits will be masked off to keep pixel values in range.
	DEFAULT: NO
ATT_00081190_USE_UT_VR	In the 2014b edition of the DICOM Standard, the value representation of attribute (0008,1190) Retrieve URL was changed from UT to the newly introduced UR. For backward compatibility, this parameter specifies that, when reading messages or file objects, the attribute is expected to have the old UT value representation. DEFAULT: NO
ATT_00287FE0_USE_UT_VR	In the 2014b edition of the DICOM Standard, the value representation of attribute (0028,7FE0) Pixel Data Provider URL was changed from UT to the newly introduced UR. For backward compatibility, this parameter specifies that, when reading messages or file objects, the attribute is expected to have the old UT value representation.

News	Description
Name	Description
	DEFAULT: NO
ATT_0040E010_USE_UT_VR	In the 2014b edition of the DICOM Standard, the value representation of attribute (0040,E010) Retrieve URI was changed from UT to the newly introduced UR. For backward compatibility, this parameter specifies that, when reading messages or file objects, the attribute is expected to have the old UT value representation. DEFAULT: NO
ATT_0074100A_USE_ST_VR	In the 2014b edition of the DICOM Standard, the value representation of attribute (0074,100A) Contact URI was changed from ST to the newly introduced UR. For backward compatibility, this parameter specifies that, when reading messages or file objects, the attribute is expected to have the old ST value representation. DEFAULT: NO
CALCULATE_DEFINED_LENGTH_FOR_CB	This parameter is applied when a registered callback function expects the data length to be provided to it and the data length is undefined. If the parameter is set to No, the undefined length will be passed as is to the callback function. If the parameter value is Yes, the toolkit will calculate the actual value of the data length before passing it tot the callback function. DEFAULT: NO
CALLBACK_MIN_DATA_SIZE	When using the MC_Register_Callback_Function() call to store large data such as pixel data, this option specifies the minimum size of value for which the callback function should be used. This option was specifically added so pixel data contained in icons are not managed with a callback function. DEFAULT: 1
COMPRESSION_ALLOW_FRAGS	Configuration Parameter for MC_Standard_Compressor. The Pegasus libraries allow compressed image data to be returned as it continues to compress more image data. This may result in an image frame having one or more fragments. This is perfectly legal, however some viewers may not be able to display the image if they do not support multiple fragments per frame. DEFAULT: YES
COMPRESSION_CHROM_FACTOR	Configuration Parameter for MC_Standard_Compressor. Values 0 through 255. The chrominance compression factor is used to adjust the default chrominance quantization table values. When ChromFactor is 32, the default chrominance quantization table values are used as is. A value of 255 corresponds to high compression, low quality. DEFAULT: 32

Name	Description
COMPRESSION_J2K_LOSSY_QUALITY	Configuration Parameter for MC_Standard_Compressor. When JPEG_2000 with COMPRESSION_WHEN_J2K_USE_LOSSY = Yes, and COMPRESSION_J2K_LOSSY_USE_QUALITY = Yes, a quality can be specified. Valid values are 1 to 10, 1 being highest quality image. DEFAULT: 1
COMPRESSION_J2K_LOSSY_RATIO	Configuration Parameter for MC_Standard_Compressor. When JPEG_2000 with COMPRESSION_WHEN_J2K_USE_LOSSY = Yes, and COMPRESSION_J2K_LOSSY_USE_QUALITY = No, a ratio can be specified. The compressor attempts to reduce the image size to 1/COMPRESSION_J2K_LOSSY_RATIO. DEFAULT: 10
COMPRESSION_J2K_LOSSY_USE_QUALITY	Configuration Parameter for MC_Standard_Compressor. When JPEG_2000 with COMPRESSION_WHEN_J2K_USE_LOSSY = Yes, this indicates which metric should be used for lossy compression, ratio or quality. DEFAULT: YES
COMPRESSION_LUM_FACTOR	Configuration Parameter for MC_Standard_Compressor. Values 0 through 255. 0 is the highest quality, giving a quantization table of all 1's. 32 corresponds to the standard quantization tables. For values between 0 and 128, the standard tables are scaled linearly. For values between 128 and 255, the standard tables are scaled non-linearly and the compression increases (and the quality decreases) by a very large amount. DEFAULT: 32
COMPRESSION_RGB_TRANSFORM_FORMAT	This parameter allows the user to select the output format when doing Lossy JPEG compression of RGB images. The value can be set to YBR_FULL or YBR_FULL_422 to specifiy what photometric interpretion Merge DICOM Toolkit should compress into when compressing RGB images. DEFAULT: YBR_FULL_422
COMPRESSION_USE_HEADER_QUERY	If set to YES, it instructs the toolkit to give precedence to the image parameters (rows, columns, etc.) from the JPEG header, in case disagreement is suspected between the the DICOM header the JPEG header. If set to NO, the DICOM header will be used. DEFAULT: NO
COMPRESSION_WHEN_J2K_USE_LOSSY	Configuration Parameter for MC_Standard_Compressor. When JPEG_2000 is used as a transfer syntax, this could mean either lossy or lossless compression. This parameter specifies the intended syntax.

Nome	Description
Name	Description
	DEFAULT: No
CREATE_OFFSET_TABLE	This parameter specifies if an offset table is created when MC_Duplicate_Message() is used to compress a DICOM message or file. It also specifies if an offset table is created when the MC_Set_Encapsulated_Value_From_Function() and MC_Set_Next_Encapsulated_Value_From_Function() routines are used. DEFAULT: Yes
DECODER_TAG_FILTER	 Specifies the list of tags to be ignored when reading DICOM files or messages. The values are separated by commas and can be specified in different formats: Single tag, e.g.: 00080020 Tag range, e.g.: 00080020-000800FF
	 Single group, e.g.: G0020-G0022 Group range, e.g: G0020-G0022 All private as: PRIVATE
	All ranges are inclusive, meaning that G0020-G0022 will filter groups 20 and 22. DEFAULT: (empty)
DEFLATE_ALLOW_FLUSH	Allows deflate to flush data occasionally to limit buffering. DEFAULT: Yes
DEFLATE_COMPRESSION_LEVEL	Allows the compression level of deflate to be specified when using deflated explicit VR little endian transfer syntax. 0 is no compression, 1 is fastest, and 9 compresses best. DEFAULT: -1
DESIRED_LAST_PDU_SIZE	This parameter allows the user to configure the length of the last PDU sent. This allows for interoperability with other DICOM implementations that may be intolerant with either a zero or two byte final PDU length. The default value used is 8. Note: Starting with release 3.5.1, this configuration option has a limited effect.
DICTIONARY_ACCESS	This parameter specifies whether or not the DICOM dictionary is to be loaded into memory or accessed from the dictionary file. FILE means access information directly from the dictionary file. MEM means load the dictionary into memory and access it there. Note: Starting with the 3.5.1 Merge DICOM Toolkit release, dictionary access is always memory based and can no
	dictionary access is always memory based and can no longer be file based. This option is now ignored. DEFAULT: MEM

Name	Description
DICTIONARY_FILE	This parameter specifies the name (path) of the DICOM dictionary. An absolute or relative path may be specified.
	The path to the file can also be specified using environment variables (including the pseudo environment variable MC3INIDIR which does not need to be set as the toolkit will resolve it internally to the directory where the merge.ini file resides).
	Unicode paths can now be specified through the environment variables. Both Windows style (%) and Unix style (\$) notations for the environment variables are accepted.
	Note: This parameter is ignored if the dictionary has been pre-compiled.
	DEFAULT:/mc3msg/mrgcom3.dct
DUPLICATE_ENCAPSULATED_ICON	When duplicating to an encapsulated transfer syntax, this configuration value specifies whether an ICON IMAGE SEQUENCE should also be encapsulated.
	DEFAULT: NO
ELIMINATE_ITEM_REFERENCES *	This parameter specifies the behavior of the message/item/file handling functions MC_Free_Message(), MC_Empty_Message(), MC_Free_Item(), MC_Empty_Item(), MC_Free_File() and MC_Empty_File(). If this parameter is set to YES, the above functions will search for references in every currently open object to delete when they encounter an item to free within an object. DEFAULT: NO.
EMPTY_PRIVATE_CREATOR_CODES	If set to NO, private creator codes contained in messages are not emptied when the MC_Empty_Message() or MC_Empty_File() function calls are made. DEFAULT: YES
EXPLICIT_VR_TO_UN_FOR_LENGTH_GT_64K	If set to YES, the toolkit will allow encoding in explicit VR of data elements whose VR is none of OB, OW, OD, OF, SQ or UT and whose value length exceeds 65534 bytes by effectively changing the VR to UN (as per CP-1066).
	If set to NO, the attempt to encode such date elements will result in an MC_INVALID_LENGTH_FOR_VR error.
	DEFAULT: NO
EXPORT_EMPTY_PRIVATE_CREATOR_CODES	If set to NO it prevents the toolkit from exporting private creator data elements which don't have any private attributes in the private block. If set to YES, exporting private creator data elements with empty private blocks is allowed. DEFAULT: YES
EXPORT_GROUP_LENGTHS_TO_NETWORK *	When set to NO, do not export group length attributes when using the MC_Send_Request_Message(),

© Copyright Merge Healthcare Solutions Inc. 2023

Name	Description
	<pre>MC_Send_Request(), MC_Send_Response_Message() and MC_Send_Response() functions DEFAULT: YES</pre>
EXPORT_PRIVATE_ATTRIBUTES_TO_NETWORK	When set to NO, disable the exporting of private attributes in messages written to the network with the MC_Send_Request_Message(), MC_Send_Request(), MC_Send_Response_Message() and MC_Send_Response() functions. DEFAULT: YES
EXPORT_UN_VR_TO_NETWORK	<pre>When set to NO, disable the exporting of attributes with a VR of UN in messages written to the network with the MC_Send_Request_Message(), MC_Send_Request(), MC_Send_Response_Message() and MC_Send_Response() functions. DEFAULT: YES</pre>
EXPORT_UNDEFINED_LENGTH_SQ *	If YES, messages transferred over the network or written to disk have their sequence attributes encoded as undefined length. This increases performance of the library. DEFAULT: NO
FLATE_GROW_OUTPUT_BUF_SIZE *	The size that the output buffer of deflate or inflate should grow to when its size is insufficient. An Info message is logged each time the buffer grows. DEFAULT: 1024
FORCE_OPEN_EMPTY_ITEM *	When set to YES, the MC_Open_Item() function will act similar to the MC_Open_Empty_Message() function. The up-front performance cost of the MC_Open_Item() function will be reduced, but the amount of validation done when adding tags to the item is reduced. Setting this value to YES will also improve the performance of the DICOMDIR directory functions. This configuration value does not have any effect on embedded platforms. DEFAULT: NO
IGNORE_JPEG_BAD_SUFFIX	Configuration Parameter for MC_Standard_Decompressor to deal with lossless JPEG images whose suffix have been invalidly written according to the JPEG specification. These images have a 16-zero-bit suffix following a -32768 prefix where the JPEG spec says the suffix is omitted following a -32768 prefix. The following are the valid settings: -1 = Default, fail on these images 0 = Ignore when user detects such images 1 = Let the toolkit detect and ignore automatically

Name	Description
LARGE_DATA_SIZE	Defines "Large Data" to the toolkit. "Large Data" is defined as an attribute value which has a length of LARGE_DATA_SIZE or more. DEFAULT: 200.
LARGE_DATA_STORE	This parameter specifies where "Large Data" values should be stored. FILE means store the values in temporary files. MEM means store the values in memory. Note: Embedded systems should ignore this parameter and always use MEM. DEFAULT: MEM
LIST_SQ_DEPTH_LIMIT	Limit the depth of sequences listing. This parameter should be set to the maximum number of levels any sequence should be listed. DEFAULT: is 0 - means do not limit the listing of sequences
LIST_UN_ATTRIBUTES	If No, attributes with Unknown VR will not be listed by MC_List_Message() and T2 logging option. DEFAULT: Yes
LIST_VALUE_LIMIT	Limit the size of listed values by MC_List_Message() or T2 logging option. This parameter should be set to the maximum number of lines to be printed for any attribute in the list. DEFAULT: 0 - means show the whole value.
MSG_FILE_ITEM_OBJ_TRACE	This parameter allows the tracking of the creation, referencing and freeing of message, file and item objects. This option can be used if the user suspects a memory leak in their application from not freeing one of these object types. The logging is done at the T1 trace level which must be enabled in the merge.ini file. DEFAULT: NO
MSG_INFO_FILE	This parameter specifies the name (path) of the DICOM message information file. An absolute or relative path may be specified. The path to the file can also be specified using environment variables (including the pseudo environment variable MC3INIDIR which does not need to be set as the toolkit will resolve it internally to the directory where the merge.ini file resides). Unicode paths can now be specified through the environment variables. Both Windows style (%) and Unix style (\$) notations for the environment variables are accepted. Note: This parameter is ignored if the message database has been pre-compiled. DEFAULT:/mc3msg/mrgcom3.msg

Name	Description
NULL_TYPE3_VALIDATION	This parameter specifies how the toolkit will validate a single NULL value in a type 3 attribute with VM > 1. Valid values are ERR, WARN and INFO. DEFAULT: ERR
OBOW_BUFFER_SIZE	This parameter specifies the number of bytes of "Large Data" that should be buffered before they are written to disk. This value is only used when the parameter LARGE_DATA_STORE is set to FILE. DEFAULT: 4096
PEGASUS_DISP_REG_NAME	When using your own Pegasus license to remove the 3 frames/second limitation, this should have the company name that was used to generate your Pegasus license.
PEGASUS_DISP_REGISTRATION	When using your own Pegasus license to remove the 3 frames/second limitation, this should have the registration code that goes with the Pegasus dispatcher.
PEGASUS_NUMBER_OF_THREADS	Certain Pegasus opcodes can operate in a multithreaded manner. Use this setting to specify the number of threads to be used by the opcode. DEFAULT: 1
PEGASUS_OP_*_NAME	When using your own Pegasus license to remove the 3 frames/second limitation, this should have the company name that was used to generate your Pegasus license.
PEGASUS_OP_*_REGISTRATION	When using your own Pegasus license to remove the 3 frames/second limitation, this should have the registration code that goes with its respective PEGASUS_OP_*_NAME.
PEGASUS_OPCODE_PATH	This parameter specifies the directory where Pegasus opcode DLLs are to be loaded from. The opcode DLL refers to files like picn6220 and not the dispatcher DLL picn20. If the option is empty, the SSM/DLL is loaded from the same directory as the dispatcher DLL. If these files are not found, opcode SSM/DLL is loaded using the directory order Windows uses when loading DLLs. The SSM/DLL is loaded from the current directory if '.' is specified. DEFAULT: (empty)
RELEASE_SQ_ITEMS	If set to NO, existing item IDs will not be freed when setting a null value or an empty value or a new value to a sequence attribute. Setting it to YES will allow sequence items that have no other references to be freed. DEFAULT: No
REJECT_INVALID_VR	This parameter specifies whether or not to reject invalid VR values in DICOM messages. If set to Yes, the parsing is aborted and the data set is rejected with a status of MC_INVALID_VR. This is useful in some scenarios when

Name	Description
	invalid attribute VR and length can result in runaway read/copy operations which may lead to crashes. DEFAULT: No
REMOVE_PADDING_CHARS	When set to Yes, Merge DICOM Toolkit will remove space padding characters from all text based attributes. This removal will occur when the attribute is encoded with one of the MC_Set_Value functions, or when the attribute is read with one of the streaming or network read functions. DEFAULT: No
REMOVE_SINGLE_TRAILING_SPACE	If set to YES, the toolkit will strip a single trailing padding space character from an attribute value of string type. Otherwise it will not. DEFAULT: YES
RETURN_COMMA_IN_DS_FL_FD_STRINGS	When set to Yes, Merge DICOM Toolkit will return a comma character as a decimal separator in a value when MC_Get_Value_To_String() is called for an attribute with a VR of DS, FL, or FD. When set to No, a period will always be returned for the decimal separator. Note that DS values will always be properly encoded with a period in DICOM message objects. DEFAULT: No
TEMP_FILE_DIRECTORY	This parameter specifies the directory in which temporary files should be created. This parameter is used only if LARGE_DATA_STORE = FILE. An absolute or relative path may be specified. The path to the directory can also be specified using environment variables (including the pseudo environment variable MC3INIDIR which does not need to be set as the toolkit will resolve it internally to the directory where the merge.ini file resides). Unicode paths can now be specified through the environment variables. Both Windows style (%) and Unix style (\$) notations for the environment variables are accepted. DEFAULT: ./
TOLERATE_INVALID_IN_DEFAULT_CHARSET	This parameter specifies if non-ASCII characters are to be tolerated in the default repertoire. When set to Yes, the validation of the attribute/message will not be enforced, but a warning message will still be logged. DEFAULT: Yes
UN_VR_CODE	VR Code to use for attributes with unknown VRs. This may be set to 'OB' if an implementation does not understand 'UN'. DEFAULT: UN VALID VALUES: UN, OB

Name	Description
UPDATE_GROUP_0028_ON_DUPLICATE	When set to Yes, the group 0028 attributes within a message will be updated when duplicating a message or file with MC_Duplicate_Message() and the standard compressor or decompressor. The Photometric Interpretation will be updated as appropriate, and the Lossy Image Compression, Lossy Image Compression Ratio and Lossy Image Compression Method tags will be updated if Lossy Image Compression was applied to the image. DEFAULT: No
USE_FREE_DATA_CALLBACK	When set to Yes, all registered callback functions registered with MC_Register_Callback_Function are called with the FREE_DATA callback type when the memory associated with the callback is to be freed, because the enclosing message, file, or item is being freed. DEFAULT: No
WORK_BUFFER_SIZE *	This parameter specifies the amount of data that is buffered in the toolkit before being stored internally or passed to a user's callback function. This option impacts the MC_Message_To_Stream(), MC_Stream_To_Message(), MC_Send_Request_Message(), MC_Send_Request(), MC_Send_Response_Message(), MC_Send_Response(), MC_Read_Message(), MC_Read_To_Stream(), MC_Open_File(), MC_Open_File_Bypass_OBOW(), MC_Open_File_Upto_Tag(), MC_Write_File() and MC_Write_File_By_Callback() functions. Setting this option to values larger than 28K will in most cases cause the toolkit to use the operating system's memory management scheme instead of the toolkit's internal mechanism. DEFAULT: 28K

* Performance tuning.

Table B-12: [TRANSPORT_PARMS] section	n of system profile parameters
---------------------------------------	--------------------------------

Name	Description
CAPTURE_FILE	This parameter specifies the base name to use for capture files. (Capture files are generated if the NETWORK_CAPTURE value is set to Yes.) If only one capture file is requested (see NUMBER_OF_CAP_FILES), the capture file will have the name specified. If more than one is requested, nnn will be appended to the base file name specified (e.g. merge001.cap) DEFAULT: merge.cap (in the current directory) Note: Use of this parameter is deprecated.
CAPTURE_FILE_SIZE	This parameter specifies the maximum size (in kilobytes) that capture files are allowed to grow (capture files are generated if the NETWORK_CAPTURE value is set to Yes). If more than one capture file is requested (see NUMBER_OF_CAP_FILES), each file generated will have this maximum size. If a value less than 1 is specified only one capture file of unlimited length will be generated. DEFAULT: 0 Note: Use of this parameter is deprecated.
IP_TYPE	This parameter specifies the preferred IP type for network communications. When set to IPV4, Merge DICOM Toolkit will attempt to utilize only IPV4 network connections. When set to IPV6, Merge DICOM Toolkit will attempt to use only IPV6 network connections. When set to AVAILABLE in an SCP, Merge DICOM Toolkit will prefer IPV6 if it is enabled in the operating system over IPV4. If IPV6 is used, the socket is put into dual stack mode, if supported by the operating system, to accept connections from both IPV4 and IPV6. When set to AVAILABLE in an SCU, Merge DICOM Toolkit will use the available type of IP networking. DEFAULT: AVAILABLE VALID VALUES: AVAILABLE, IPV4, IPV6
MAX_PENDING_CONNECTIONS	This parameter specifies the maximum number of open listen channels. Its value is used as the second argument of a TCP listen() call. DEFAULT: 5
NETWORK_CAPTURE	This parameter specifies whether or not network data should be captured in files suitable to be read by the MergeDPM utility. Use these parameters to customize the network capture: CAPTURE_FILE CAPTURE_FILE_SIZE NUMBER_OF_CAP_FILES REWRITE_CAPTURE_FILES DEFAULT: No Note: Use of this parameter is deprecated.

Name	Description
	Description
NUMBER_OF_CAP_FILES	This parameter specifies the number of capture files to generate (capture files are generated if the NETWORK_CAPTURE value is set to Yes). Each capture file generated will have maximum size specified by CAPTURE_FILE_SIZE. If CAPTURE_FILE_SIZE is less than 1 (unlimited size) this parameter's value is ignored. DEFAULT: 1
	Note: Use of this parameter is deprecated.
REWRITE_CAPTURE_FILES	This parameter specifies whether or not the capture files should be rewritten when all files have reached the maximum size specified by CAPTURE_FILE_SIZE (capture files are generated if the NETWORK_CAPTURE value is set to Yes). If Yes is specified, the oldest file will be rewritten. If No is specified and all requested files have been written (see NUMBER_OF_CAP_FILES), no more data will be captured.
	DEFAULT: Yes
	Note: Use of this parameter is deprecated.
TCPIP_DISABLE_NAGLE	This parameter specifies if the Nagle Algorithm should be used when sending packets at the TCP/IP level. Most operating systems enable this by default. It allows small segments of data to delay sending a fixed amount of time to possibly be combined with other small segments and be sent as one larger packet. Disabling this may cause high network traffic.
	DEFAULT: No
TCPIP_LISTEN_PORT	This parameter specifies the TCP/IP port on which server applications are to listen for associate requests. DEFAULT: 104
TCPIP_RECEIVE_BUFFER_SIZE *	This parameter specifies the TCP/IP receive buffer size for each connection. Note that the maximum values for this constant and TCPIP_SEND_BUFFER_SIZE are operating system dependent. If the values of these options are set too high, a message will be logged to the toolkit's log files, although no errors will be returned through the toolkit's API.
	Larger values for these constants will greatly improve network performance on networks with minimal network activity. Note that for optimum performance, these values should be at least slightly larger than the PDU_MAXIMUM_LENGTH configuration value.
	Note also that setting these values to an even multiple of the TCP/IP MSS (Maximum Segment Size) of 1460 bytes can help increase performance.
	Note, also that some operating systems such as Linux have auto-tuning of TCP/IP buffer sizes implemented when an explicit TCP/IP Send and Receive buffer size are not set. These options can be set to zero to disable Merge DICOM Toolkit's setting of each buffer size.
	DEFAULT: 131400
	MAXIMUM: Operating System dependent

Name	Description
TCPIP_SEND_BUFFER_SIZE *	This parameter specifies the TCP/IP send buffer size for each connection. Note that the maximum values for this constant and TCPIP_RECEIVE_BUFFER_SIZE are operating system dependent. If the values of these options are set too high, a message will be logged to the toolkit's log files, although no errors will be returned through the toolkit's API.
	Larger values for these constants will greatly improve network performance on networks with minimal network activity. Note that for optimum performance, these values should be at least slightly larger than the PDU_MAXIMUM_LENGTH configuration value.
	Note also that setting these values to an even multiple of the TCP/IP MSS (Maximum Segment Size) of 1460 bytes can help increase performance.
	Note, also that some operating systems such as Linux have auto-tuning of TCP/IP buffer sizes implemented when an explicit TCP/IP Send and Receive buffer size are not set. These options can be set to zero to disable Merge DICOM Toolkit's setting of each buffer size.
	DEFAULT: 131400
	MAXIMUM: Operating System dependent

* Performance tuning.

Service Profile

The Service Profile (usually called mergecom.srv) contains DICOM standard services and commands and is a useful reference (along with the message.txt file) to find the Merge DICOM Toolkit names for the standard DICOM services and items. It is used by the library to negotiate the proper SOP Class UIDs and to access the binary dictionary and message information files when creating instances of message objects and validating messages.

In most cases, it will not be necessary to modify the Service Profile. However, if you are using an extended toolkit to create your own private services, you will need to add specifications for these private services to the Service Profile. See the *Merge DICOM Toolkit: Database Manual* for further details.

The location of the Service Profile is provided by the MERGECOM_3_SERVICES parameter of the [MergeCOM3] section of the MERGE.INI file.

Remember, the Service Profile is GENERATED by the Merge DICOM Toolkit Profile Database Utilities. Unless you are absolutely confident about changes being made, DO NOT CHANGE THE CONTENTS OF THIS FILE.

The Service Profile contains the following sections:

[©] Copyright Merge Healthcare Solutions Inc. 2023

Table B-13: Service profile parameters.

Name	Description
[SERVICE_TABLE]	List of service names and numbers. This list registers every service available to an Application Entity. The parameters associated with [SERVICE_LIST] are NUMBER_OF_SERVICES_SUPPORTED (the number of service names that will be listed immediately following NUMBER_OF_SERVICES_SUPPORTED) and one entry for each supported service.
[<service_number>]</service_number>	One section number for each of the above services registered in [SERVICE_TABLE]. Each section contains a Service Name, a DICOM SOP Class UID for the Service, a flag that tells whether it is a BASE or META Service (SOP) and a list of commands supported for that service.
[ITEM_TABLE]	One item name and number for each DICOM item that can be encoded in an attribute of Value representation SQ (Sequence of Items).